

Advances and Challenges in Wearable Sensors for Health Monitoring

Laís Canniatti Brazaca,* Alvaro Moreno Lozano, Beatriz Mayol, Desiree Tamara Scheidt, Domenico De Fazio, Enji Kim, Gabriel Maroli, Giulio Rosati, Hatice Ceren Ates, Hui Wang, Hye Jin Kim, Ji-Hwan Ha, Jin Yang, Jisoo Jeon, Jörg Goldhahn, Keval K. Sonigara, La Li, Mingyu Zhou, Seokjoo Cho, Shicheng Fan, Shuang Wu, Vanessa Matos Gonçalves, Vitor H. B. Dei Santi, Weixin Zhou, Wenzheng Heng, Won Gi Chung, Xinming Li, Yang Zou, Yunxia Jin, Zixuan Wu, Amay J. Bandodkar, Arben Merkoçi, Can Dincer, Chwee Teck Lim, Dae-Hyeong Kim, Emanuel Carrilho, Giovanni Antonio Salvatore, Guozhen Shen, Inkyu Park, Itthipon Jeerapan, Jang-Ung Park, Jin Wu, John S. Ho, John Rogers, Joseph Wang, Kevin W. Plaxco, Maria Cristina Ferreira de Oliveira, Martin Pumera, Noé Brasier, Omid Veisheh, Pooi See Lee, Vladimir V. Tsukruk, Wei Gao, Yong Zhu, Zhong Lin Wang, Zhou Li, Juliane Sempionatto,* and Osvaldo N. Oliveira, Jr.*



Cite This: *ACS Appl. Mater. Interfaces* 2026, 18, 39567–39694



Read Online

ACCESS |


Metrics & More

Article Recommendations

ABSTRACT: Analytical tools may revolutionize healthcare by enabling accessible, rapid, and decentralized testing. Wearable (bio)sensors, in particular, provide frequent or continuous patient monitoring through non- to minimally invasive measurements. This approach yields unprecedented amounts of health-related information, leading to more informed clinical decision-making and closer patient follow-up. In this mega-review article, we bring together leading researchers in the field to discuss the state of the art in wearable devices for health monitoring. We begin by providing a broad overview of the field through citation network analysis. We then review the application of chemical (bio)sensors in biofluids (e.g., sweat, saliva, tears, interstitial fluid, and cerebrospinal fluid), highlighting the challenges and advantages associated with each.

Subsequently, we discuss the construction of wearable devices and their main formats (e.g., smart contact lenses, textiles, mouthguards, watches/wristbands, and implantable systems). Physical sensors are addressed in a dedicated section focusing on the assessment of heart rate, blood pressure, and body temperature. The role of soft electronics in wearable devices is also examined, as these technologies are essential for enhancing user comfort and sensor reliability, which demands advances in materials science. Furthermore, we present strategies for signal acquisition and transmission, as well as approaches for on-body energy harvesting and device self-powering. The use of artificial intelligence and machine learning is then discussed as a means of enhancing analytical performance and managing the large volumes of data generated by wearable devices. Finally, business, regulatory, and ethical considerations are examined. We expect that this review will provide an overview of sensing and biosensing technologies for health-related applications, identify promising research directions, and inspire future developments.

KEYWORDS: wearable devices, sensors, biosensors, health monitoring, perspectives



1. INTRODUCTION

Wearable (bio)sensors are commonly designed for the continuous monitoring of an individual's health status. These devices can provide real-time tracking of physical, physiological, and biochemical parameters, offering insights into clinical information.^{1–3} Sensors designed to monitor physical parameters are classified as physical wearable sensors.⁴ They include: (1) inertial sensors, to monitor body motion such as

Received: March 4, 2026

Revised: May 25, 2026

Accepted: May 28, 2026

Published: July 16, 2026






Figure 1. Accessing various body fluids to serve the various needs along the life cycle, ranging from saliva analysis through pacifier to in-ear sweat analysis in advanced age. Reproduced with permission from Brasier et al.,²³ © 2024 Springer Nature.

acceleration and angular velocity; (2) optical sensors, which rely on light's transmission or reflection properties to monitor heart rate and oxygen saturation; and (3) electrical sensors, to detect biopotentials such as the electrical signals generated by the body, allowing for the measurement of cardiac activity (electrocardiogram) or muscle function (electromyography).^{5–7} This type of monitoring using smartphones, smartwatches, and patches outside healthcare institutions is transforming medical care. Representative examples include cardiac diseases, shifting away from bulky, time-limited classical Holter electrocardiograms (ECGs) toward more seamless solutions for detecting^{8–10} and managing¹¹ atrial fibrillation in real-life settings. Additionally, physically assessed parameters can guide managing health conditions, including (i) respiratory diseases such as chronic obstructive pulmonary disease (COPD)¹² and interstitial lung disease,¹³ (ii) neurological disorders like multiple sclerosis,¹⁴ Parkinson disease,¹⁵ and Alzheimer's disease,¹⁶ (iii) autoimmune diseases such as inflammatory arthritis¹⁷ and Raynaud's phenomenon in systemic sclerosis,¹⁸ (iv) dermatological conditions like atopic dermatitis,¹⁹ and (v) mental health disorders such as depression²⁰ and schizophrenia.²¹ While physical measurements have supported the management of health conditions, including COVID-19,²² they mostly still lack disease-specific molecular biomarkers, such as those obtained from blood analysis.

The next generation of wearable devices should enable analysis of body fluids, including sweat, saliva, and interstitial fluid (ISF), among others. This approach represents a paradigm shift in healthcare, enabling continuous, lab-independent collection of molecular health information.²³ There are pros and cons attributable to the use of each body fluid and device designs, which can follow the needs along the life cycle (Figure 1). It is of utmost importance, however, to understand biomarker partitioning mechanisms²⁴ as well as body fluids' (patho-)physiology as, for example, sweat rate is influenced by a variety of internal and external factors.^{25,26} Latest developments in the field have demonstrated sweat's potential to manage hydration,²⁷ heart failure,^{28,29} disorders of substance abuse,³⁰ and biological age.³¹ For breath, smart masks provide straightforward analysis of exhaled breath condensate to manage asthma or COPD³² and detect SARS-CoV-2.³³ In terms of validated clinical use, continuous glucose monitoring through ISF analysis has transformed care for diabetic patients with a positive impact on Hb1Ac in patients with Type II Diabetes³⁴ among others.³⁵ A body fluid sensor, for example, could be selected based on clinical indication to personalize therapeutic drug monitoring for antibiotic treatment.³⁶ Patients with pneumonia might undergo breath analysis, while those with soft tissue infections could use wearable devices for sweat analysis to monitor tissue penetration, drug concentrations, and ensure compliance with narrow therapeutic windows.




Figure 2. Number of papers published per year since 2005 for the search “wearable sensors” in the OpenAlex database in April, 2026.

Wearable sensors drive person-centered health management, having the potential to support overcoming population-based health challenges, such as antimicrobial resistance (AMR), climate change, and aging. For AMR, for example, incentives to develop novel antibiotics remain scarce.³⁷ Therefore, optimizing drug therapy is crucial in fighting the silent pandemic of antimicrobials,³⁸ with therapeutic drug monitoring being able to reduce antibiotic resistance and improve efficiency.³⁹ While blood-based therapeutic drug monitoring is indicated during reserve antibiotic therapy, wearable devices for body fluid analysis may allow for understanding drug tissue penetration.³⁶ Climate change leads to increased weather extremes, including heat waves.⁴⁰ It is predicted that by the end of the 21st century, more than 95% of countries will be exposed to heat stress,^{41,42} which can lead to health issues and a decline in work performance.^{43,44} Wearable devices allow for monitoring hydration status in workers exposed to heat, ensuring safety, and enhancing productivity through haptic feedback,²⁷ facilitating personalized heat strain management. Recent investigations have further demonstrated the potential to monitor metabolic changes related to heat stress, going beyond hydration monitoring.⁴⁵ Aging societies increasingly challenge medical care systems as the elderly population grows, and their needs become more complex, requiring high-quality, equitable care. Wearable sensors have significant potential to support a livable and autonomous life at home and beyond.⁴⁶

One of the most important tasks of medical doctors is to understand the derived health data in its broader context. Thus, efficient wearable sensing platforms can be coupled to artificial intelligence (AI) for automated data processing for interpretation, detection, prediction, and clinical outcome visualization.^{23,47} Once fully validated, these sensors can be integrated into closed-loop management approaches, enabling automated and person-centered therapies. Therefore, sensors and actuators are coupled and coordinated through communication protocols.⁴⁸ Various applications have been identified in areas such as diabetes mellitus,⁴⁹ wound care,⁵⁰ and Parkinson's disease,¹⁵ and in just-in-time adaptive interven-

tions in behavioral health⁵¹ when combined with digital therapeutics.

In this Mega Review, we describe and discuss the state of the art of wearable (bio)sensors, including their diverse applications and formats, the methods for fabricating wearable devices, the basis for energy harvesting and self-powering, soft electronics, the use of machine learning (ML) and AI for data analysis, and the main challenges and prospects of the area. We expect this Review to serve as a reference for both early-career and experienced researchers in wearable devices, providing insightful information for future research.

2. LANDSCAPE OF WEARABLE SENSORS FOR HEALTH MONITORING

2.1. Citation Network

Landscapes of scientific fields can be generated using computational methods based on network analysis and natural language processing (NLP),⁵² particularly with large-language models. For instance, citation networks and NLP have been employed to identify the most relevant areas in journals covering chemistry and materials sciences⁵³ and in providing a historical account of the journal.⁵⁴ This approach provides a bird's eye view of a particular field and estimates its size in terms of the number of publications. We have employed it here by generating citation networks for papers retrieved with searches using various queries to the OpenAlex database.⁵⁵ For that, we applied the following queries: “wearable sensors or health monitoring”; “wearable sensors”; “wearable health monitoring”; “wearable health technology”; “wearable sensors and health monitoring”; “wearable medical devices”; “materials for wearable sensors”; “wearable health monitoring technology”; “wearable sensors for healthcare”; “wearable sensors and human health monitoring”. The searches were conducted with a title and abstract filter (“title and abstract”). The broadest term in scope was the first (“wearable sensors or health monitoring”), which retrieved a considerable number of papers related to the monitoring of the “health” of constructions, bridges, machines, and batteries, not being associated with monitoring the health




Figure 3. Citation network with 35,935 articles obtained from the search “wearable sensors”.

of humans and other animals. Judging from the communities (i.e., clusters representative of topics), the network that best represented the topic of our interest was obtained with the query “wearable sensors”. Though “health monitoring” was not included in the query, most of the papers are related to it, as it will be discussed next. The networks generated with the other queries were smaller and could be considered as being contained in the larger network for “wearable sensors”. The search using “wearable sensors” led to 68,310 papers, most of which were published in the last 20 years. Figure 2 shows updated numbers of papers published until the end of 2025, featuring a trend of increasing attention given to wearable sensors.

The largest connected component of the citation network, obtained with the query “wearable sensors”, contains 35,935 papers (Figure 3). It is worth mentioning that the citation network includes only papers that cited others in the set of papers retrieved. Besides, this citation network may be formed by several subnetworks that are disconnected among themselves (the so-called “connected components”), and the subnetwork considered and shown in the figure corresponds to the largest of these connected components, with the smaller ones being ignored. Nevertheless, owing to the large number of papers, the network should be representative of the field. A community detection algorithm was applied to the network to identify clusters of densely connected papers and the major topics addressed by the papers in each cluster, following the methodology described elsewhere.^{52–54} These clusters are

shown in different colors in the figure, and also their most relevant keywords characterizing the topics. The largest clusters, from A to F, can all be associated with sensors used for human health monitoring.

The largest cluster, A (in blue, with 13,962 articles), focuses on flexible and stretchable pressure and strain sensors. The second largest cluster, B (in orange, with 11,418 articles), relates to sensors used for monitoring human activities, such as gait and physiological conditions. Cluster C (in green, with 2,580 articles) is related to wearable physical sensors monitoring physiological conditions and is located near cluster B.⁴ Cluster D (in red, with 1,945 articles) explores the monitoring of emotional states and physiological stress using mobile sensors and learning algorithms. Cluster E (purple, with 1,871 articles) focuses on IoT infrastructure applied to healthcare, indicating attention to connectivity and intelligent systems for remote health surveillance. Cluster F (in brown, with 1,667 articles) brings together studies of heart rate and sleep monitoring, particularly using photoplethysmography (PPG).

The additional clusters covered more specific applications. Cluster G (in pink, with 967 articles) is dedicated to gesture-based interfaces, especially for *sign language* and *hand gesture recognition*, using glove-based sensors to translate finger and hand movements into digital commands. Cluster I (in light yellow, with 195 articles) addresses social interaction and group behavior. It includes studies that use wearable sensors to track interpersonal dynamics and proximity patterns. The small

cluster J (in light blue, with 159 articles) focuses on assistive navigation for visually impaired individuals. This cluster highlights the use of wearable systems to enhance spatial awareness and autonomy for people with visual impairments. There is a cluster significantly distant from the others: cluster H (in dark gray, with 937 articles), which relates to tool monitoring, including for the oil industry. This cluster is obviously very different as it does not pertain to human or animal health.

Considering the clusters unequivocally related to health monitoring with wearable sensors, we estimate the field's size at approximately 36,000 articles in the citation network, equivalent to about 60,000 articles in total.

Though the other, more specific searches led to networks that could be considered subsets of that of Figure 3, they are useful to reveal additional topics that could have been hidden in the broader search. Indeed, clusters emerged from the other networks that could be associated with machine learning and artificial intelligence methods in health monitoring, wireless communication to transfer data from the sensors, specific monitoring of cardiovascular conditions, and the use of textiles in wearable devices. The main conclusion from the analysis of the network in Figure 3 is that the main topics of the field of "wearable sensors for health monitoring" are flexible, stretchable pressure and strain sensors, followed by various technologies to monitor health using wearable sensors, mostly physical sensors, which include data science methods, machine learning, and artificial intelligence. Physiological conditions monitored include heartbeat, blood pressure, and other vital signs, sleep, stress, gait, while the most monitored disease is diabetes.

2.2. Recent Review Papers

The large growth in the number of papers on wearable sensors for health monitoring over the past few years is reflected by a significant increase in review papers. A search conducted on the Web of Science using the query "*health monitoring AND wearable sensor**" across all fields (i.e., without restricting it to topic or title) retrieved 1,357 review papers. To provide a flavor of what is being covered in review papers, we made a manual selection of comprehensive review articles, with 22 being chosen, published from 2020 to 2026. Overall, the articles emphasize two major dimensions in the development of wearable sensor technology. On one side, there is a strong focus on the advancement of fabrication techniques and novel materials—such as 3D printing of polymer nanocomposites, conductive polymer-based hydrogels, graphene-based composites, and gold nanomaterials—to create sensors that are flexible, stretchable, and biocompatible. These methods aim to overcome the limitations of traditional rigid sensors, improving wearability and enabling continuous, noninvasive health monitoring. On the other side, a significant number of reviews concentrate on the integration of AI, particularly ML and deep learning (DL), to enhance data processing, pattern recognition, and predictive analytics in real time. We will cover the specific reviews on the use of AI for wearable sensors and health monitoring at the end of this subsection.

Most reviews examine physical sensors, such as motion sensors, inertial measurement units, ECG, PPG, and electromyography (EMG) devices along with their applications. These sensors are the most researched, serving today as the foundation of wearable technology. Chemical sensors, which are crucial for detecting biochemical markers in body fluids, are

less explored and present significant challenges related to stability, specificity, and integration into wearable formats. Many of these challenges will be discussed throughout the different sections in this review paper.

Discrepancies arise mainly in the debate over optimal fabrication strategies and material selection. While some studies highlight the promise of novel materials like polymer nanocomposites and hydrogels for their superior flexibility and biocompatibility, others point out issues such as mechanical durability and signal stability. Additionally, although the integration of energy-harvesting solutions is recognized as important, there is less consensus on the best approaches for power management in long-term autonomous sensor operation. Overall, the review papers highlight required improvements in several areas: enhancing sensor durability and reliability; refining energy harvesting and storage solutions to support continuous operation; advancing AI algorithms for more robust and efficient on-device analysis; and increasing research on chemical sensors to achieve multianalyte detection with high specificity. Addressing these challenges is important for transitioning wearable sensors from the research stage to practical and scalable solutions in personalized healthcare.

The use of AI, particularly ML, in conjunction with wearable sensors has been reviewed by many authors from varied perspectives. A systematic review indicated a preference for standard ML over DL methods, primarily due to ML's superior interpretability in classification tasks.⁵⁶ The ability to detect complex patterns in large, high-dimensional data sets was highlighted concerning data from wearable sensors.⁵⁷ Some of the reviews focused on specific areas of wearable sensors. For instance, AI-assisted cardiovascular health monitoring was examined by analyzing health indicators and noninvasive methods.⁵⁸ Meanwhile, the literature on AI and smart devices for tracking the physical activity of people with diabetes examined the relationship between diabetes management and physical activity.⁵⁹ A thorough analysis was conducted on the use of ML and DL for processing data from hydrogel sensors, particularly pressure and strain sensors.⁶⁰ The role of AI in advancing electrochemical sensors for healthcare was discussed, with a focus on developing affordable point-of-care (POC) and point-of-use (POU) devices to improve diagnostic accuracy and health monitoring efficiency.⁶¹ AI and ML methods have primarily been used for data processing and prediction, relying entirely on software-based ML. However, there is growing research in the opposite direction, where wearable sensors are integrated with other devices to enable ML via hardware. Applications in both approaches have been reviewed, covering flexible sensors and the use of artificial synapses.⁶²

As expected, all of these reviews express optimism about the transformative potential of AI in health monitoring. However, they also emphasize the challenges of fully harnessing AI. These challenges include obtaining consistent experimental data for model training, adapting algorithms to the evolving properties of flexible electronics, transferring trained models between devices, and enhancing the computational power for real-time updates. Additionally, difficulties in model generalization and the need for robust, interpretable ML methods remain significant obstacles. There is also concern about the societal implications of the use of AI.




Figure 4. Comparison of molecular weight cutoff, production rates, and lag time (vs blood) for biomarkers in different biofluids.

3. WEARABLE (BIO)SENSORS FOR THE DETECTION OF MOLECULAR BIOMARKERS

Wearable sensors can detect biochemical markers in various biofluids such as sweat, saliva, tears, wound fluid, and urine. These biomarkers typically originate through partitioning from blood plasma due to diffusion or active transport across biological membranes.^{63,64} Blood acts as the central transport system, carrying a range of molecules, such as hormones, proteins, metabolites, drugs, and inflammatory mediators, either in a free form or bound to carrier proteins. For partitioning to occur, these biomarkers must first pass through the vascular endothelium and into the surrounding tissue or fluid compartments. Size, charge, solubility, transport mechanism, metabolism, and degradation are some of the factors that influence the time lag, type, and concentration of biomarkers detectable in each biofluid, with the molecule size being an important limiting factor. Smaller molecules tend to partition easily into sweat, making it suitable primarily for detecting electrolytes, small metabolites, and certain hormones.⁶⁵ Conversely, tears and saliva can contain relatively larger molecules, including hormones and specific proteins, due to the selective permeability of gland membranes.^{64,66} Interstitial and wound fluids usually contain biomarkers of a larger size due to the structure, permeability, and proximity to capillaries from the tissue.^{67,68} Figure 4 summarizes the size, production rates, and time required to find biomarkers in each of the biofluids compared to blood.⁶⁹

The selection of a biofluid for wearable biosensing platforms involves careful consideration of several factors, including fluid accessibility, biomarker concentration, sampling invasiveness, stability, and correlation with physiological or pathological conditions. Sweat and saliva offer minimally invasive or noninvasive collection and continuous accessibility, making them attractive for real-time monitoring. However, biomarker concentrations in these fluids can vary widely due to factors such as hydration, diet, circadian rhythms, and physical activity, posing challenges in establishing consistent baseline levels.⁷⁰ Saliva biomarkers generally show good correlation with systemic conditions, particularly hormones and stress-related markers,⁷¹ while sweat biomarkers often correlate well with electrolyte balance, metabolic activities, and hydration status.⁶⁵ Tears and wound fluid contain specialized but highly

relevant biomarker profiles for some health conditions. However, their sampling requires specialized sensors that are suited to limited fluid volumes and specific anatomical regions. Tears provide biomarkers relevant to ocular diseases, inflammation, and certain metabolic disorders,⁷² but their collection is challenging because of low volumes, rapid evaporation, and concentration instability. Wound fluid, rich in inflammatory markers, proteins, and indicators of healing progress, varies significantly depending on the wound type, infection status, and individual physiological responses, thereby complicating standardized biomarker detection.⁷³ Urine, easily collectible and abundant, facilitates the analysis of metabolic waste biomarkers, although its intermittent availability limits continuous monitoring applications.

Various challenges must be addressed to ensure reliable biomarker detection in wearable systems. First, biofluid composition can vary among individuals depending on their physiological status, hydration levels, diet, and external factors such as environmental temperature or physical activity. This biological variability can complicate the establishment of baseline biomarker concentrations and the diagnostic thresholds. Second, biofouling, the nonspecific adsorption of proteins or biomolecules on sensor surfaces, can reduce sensor performance, demanding antifouling strategies and rigorous calibration protocols.⁷⁴ Despite these limitations, the advantages of wearable biomarker detection are substantial. Real-time continuous monitoring of biofluids enables early detection of disease onset, proactive health management, timely intervention, and personalized therapeutic decisions, enhancing the overall efficiency of healthcare systems by potentially reducing the number of hospital visits and costs.

The following sections will explore each biofluid in detail, highlighting its attributes, challenges, and opportunities in wearable sensor applications.

3.1. Sweat

Sweat is a vital physiological fluid with significant implications for maintaining thermal homeostasis and for monitoring health status.^{75–77} The autonomic nervous system controls the secretion of sweat through activation of the sweat glands, primarily driven by sympathetic cholinergic stimulation.^{78,79} Eccrine glands are distributed throughout the human body, with the highest density in areas such as the palm and

forehead.⁸⁰ Sweat produced by these glands is primarily responsible for thermoregulation at elevated body temperatures, assisting the body to maintain an optimal temperature through evaporative cooling processes.^{76,81,82} Sweat also maintains skin hydration and establishes a slightly acidic environment as a protective barrier against external pathogens.^{83–85} Additionally, sweat helps to regulate electrolyte concentration, and it assists in the excretion of metabolic waste products to maintain fluid and biochemical balance within the body.^{25,86–88}

In terms of its use in health monitoring, sweat functions as a rich and noninvasively accessible biofluid, best understood through three primary categories of application: exposures to exogenous species, loss of substances from the body, and biochemical assessments of physiological health. In the first, sweat provides a means to monitor contact with hazardous substances in the environment, such as heavy metals and toxic volatile organic compounds.^{89–92} As sweat glands reside in the dermis, they carry information due to direct physical exposure to the environment. They also provide indications of ingestion or inhalation, resulting from diffusive transport from the surrounding ISF and blood. This possibility also has implications in monitoring for the use of illicit drugs, either recreationally or for sports performance.^{93,94} In the second category, sweat serves as a channel for the loss of essential substances that require replenishment. The excretion of water and electrolytes such as sodium (Na^+), potassium (K^+), and chloride (Cl^-) during physical exercise or heat exposure represents an obvious example, but other species, such as amino acids, are also important to consider.^{95,96} In the context of tracking parameters related to physiological health, sweat contains biomolecules that mirror internal body processes. Metabolites like glucose, lactate, and urea,^{97–99} along with hormones such as cortisol and estradiol,^{100–102} can pass from blood and ISF into sweat via transcellular and paracellular transport mechanisms. Peptides and amino acids are also present in sweat, as biomarkers of muscle activity and systemic health.^{45,103} Unlike the first two categories of application, for such types of medical purposes, reliable correlations between sweat and blood chemistry are necessary.

Traditional methods for generating eccrine sweat rely on natural mechanisms that follow from elevation in body temperature due to physical exercise or psychological stress, or from exposure to warm, humid environments such as those in saunas or baths/showers. While these natural processes are effective, they require standardization and control over sweat rate and composition. Moreover, these mechanisms for sweat generation are not effective for diseased, pediatric, and elderly subjects. For collection, absorbent pads represent one of the earliest and simplest methods. Weighing the pads before and after a period of sweating yields information on local sweat loss. Scrapping, squeezing, or spinning these pads allows for the removal of sweat for subsequent laboratory analysis. These methods do not, however, support capabilities in real-time data collection, and they require laboratory processing steps performed in specialized facilities by trained personnel. Additionally, the potential for contamination and evaporation during collection can compromise the accuracy of the analysis.^{104–106}

These limitations for the induction and collection of sweat motivate the development of alternative methods. A good example is the approach used as the main clinical diagnostic for cystic fibrosis (CF), which relies on an analysis of the

concentration of chloride in sweat.¹⁰⁷ The system (the Macroduct Sweat Collection System, developed by ELITechGroup Inc.), introduced in the late 20th century, employs transdermal iontophoretic delivery of a pharmacological agent, pilocarpine, to stimulate the action of eccrine sweat glands. The collection of sweat that subsequently appears on the surface of the stimulated region of the skin occurs via a coiled plastic tube in a manner that prevents contamination. Transfer of sweat from this tube into a laboratory instrument, known as a chloridometer, allows for analysis of chloride concentration, which is the critical marker in CF diagnosis.¹⁰⁸ Although this system represents a well-established clinical standard, failure rates due to insufficient sample collection can be high.^{109,110} Furthermore, its use is confined to clinical environments with clinical staff, and real-time monitoring is not possible.

A first attempt to overcome these limitations involved a skin-interfaced, textile-based patch with a pair of light-emitting diodes for optical measurements of pH and sodium using colorimetric indicators for each.^{111,112} Pairing this platform with a separate, external data acquisition system allows for sweat analysis in a system that we refer to as partially wearable. A few years later, a transferable tattoo paper was developed, which was suitable for mounting on the skin and designed for subsequent potentiometric readout with an external unit.^{113,114} The first fully wearable device for on-skin, quantitative analysis of sweat appeared in 2014. This platform used a porous elastomeric slab of PDMS with integrated colorimetric assays and a passive LC oscillator to allow options in optical and radio frequency measurements of sweat rate, loss, and pH, with additional capabilities in determining the concentrations of copper and iron in sweat.¹¹⁵ Later in the same year, an active radio frequency identification (RFID) tag incorporating a sweat sensing module was introduced for measuring electrolytes, mainly sodium, with power and wireless links provided by smartphones for user engagement.¹¹⁶ These early papers catalyzed a broad range of activities in this area, including an impressive, advanced version of this type of device, but with integrated electronics, batteries, and multimodal electrochemical sensors. Also in 2016, Ahyeon Koh et al. introduced a fully integrated microfluidic platform, as a “lab on the skin” technology, with collections of channels, valves, and reservoirs for capture, storage, and biomarker analysis of sweat. Examples included chloride, pH, glucose, and lactate sensing via colorimetric and options in wireless electronics for data communication.¹¹⁷

Microfluidic approaches quickly became an integral aspect of design in wearable sweat sensors (see Section 5.1 for an extensive account of microfluidics). These systems leverage capillary bursting valves and precisely patterned channels to direct small volumes of sweat with high spatial and temporal control, and with straightforward mechanisms for quantifying sweat rate and sweat loss for monitoring and interpreting biomarker concentrations. By engineering valves with varied bursting pressures, these platforms enable chrono-sampling (i.e., time sequential collection of separate small volumes of sweat into different reservoirs without mixing late sweat with early sweat), thereby linking biomarker concentrations to collection time to track dynamic physiological changes. Polydimethylsiloxane (PDMS) is a common material for these microfluidic systems, though its intrinsic hydrophobicity can hinder fluid collection and transport in certain cases. Introducing hydrophilic modifiers or integrating porous, wicking materials within the microchannels leads to enhanced

transport efficiency and improved sensing fidelity.^{118,119} PDMS-based devices can also be susceptible to mechanical deformation and high-water vapor permeability.^{120–122} To mitigate mechanical deformation, researchers developed hard/soft composite structures that maintain device integrity under deformation.^{123,124} These systems employ printing materials that combine flexibility and robustness, allowing them to endure mechanical strain during wear while maintaining reliable sweat uptake, storage, and analytical performance. Reeder et al. addressed water vapor transport concerns by integrating block copolymers into the device design, improving barrier properties and overall performance.¹²⁵ Laser engraving methods offer an alternative to molding approaches for forming the microfluidic constructs. Examples include laser-engraved polyimide (PI)-based microfluidic channels, with electrochemical sensors for monitoring electrolytes and metabolites.^{126,127} Techniques in 3D printing are also of interest due to their ability to support complex microfluidic architectures, enabling the hard/soft structures mentioned above. When implemented with optical-grade polymers, fabrication of microcuvettes for high-precision spectroscopic, colorimetric, and fluorometric assays is possible.^{123,124,128}

Traditional laboratory-based analysis techniques based on spectroscopic methods, enzyme-linked immunosorbent assays (ELISA), and liquid chromatography–mass spectrometry (LC-MS), exhibit high sensitivity and specificity in sweat analysis.^{100,129–131} Requirements in complex workflows, expensive instrumentation, dedicated facilities, and trained personnel are not, however, suitable for continuous monitoring in compact devices for wearable platforms. Optical sensing modalities, including colorimetric and fluorometric methods, suitably calibrated against these traditional laboratory approaches, are attractive for sensors that can be implemented in thin, flexible formats on the skin. This is partly due to their simplicity, their ability to operate without supporting electronics and capacity for wireless readout using analysis of images captured with digital cameras in smartphones. Colorimetric sensors use enzymatic or chemically responsive assay components that produce quantifiable color changes upon reacting with target analytes, thereby allowing accurate measurements based on relating color information extracted from images to analyte concentration. Successful examples for sweat analysis include electrolytes,^{118,132,133} pH,^{134,135} lactate,^{134,136–138} metal ions,^{139,140} ketones,^{141,142} creatinine,¹⁴³ urea,^{144–146} uric acid,^{147,148} glucose,^{149–151} cortisol,^{152,153} ammonia,¹⁵⁴ and ethanol.^{154,155} The sensitivity of colorimetric assays may, however, be insufficient for detecting sweat biomarkers that are present only at low abundance, such as cytokines, which exist at picogram-per-milliliter levels.^{156,157} For these targets and others, fluorometric sensors offer superior sensitivity and specificity, as demonstrated in the detection of ions,¹⁵⁸ amino acids,¹⁵⁹ and other biomolecules in sweat. Lateral flow assays (LFAs), long established in POC diagnostics for their straightforward visual interpretation, represent additional options for use in wearable sweat sensing. A microfluidic platform integrating paper-based LFAs was used for immunoassays of cortisol in sweat.¹⁶⁰ Other examples exploit LFA formats in a similar manner for aptamer-based cortisol sensing¹⁵² and potassium¹⁶¹ monitoring.

Electrochemical sensors are also important, particularly for real-time detection and application to chemical species where no colorimetric or fluorometric assays exist. Potentiometric sensors can be used to measure electrolyte concentrations (e.g.,

Na⁺, K⁺, Cl[−]),^{162–164} while amperometric sensors, often functionalized with enzymes or molecularly imprinted polymers (MIPs), enable detection of metabolites such as glucose, lactate, and ethanol through current responses to redox reactions.^{165–168} Impedimetric sensors allow label-free, highly specific, and sensitive detection of low-concentration biomarkers like cortisol.^{169–172} Conductometric sensors can capture ionic strength and sweat volume via shifts in conductivity.¹⁷³ These electrochemical platforms, when integrated with wireless electronics, offer high specificity, fast response times, and continuous monitoring capabilities. Recent developments in multiplexed sensor arrays allow simultaneous tracking of multiple biomarkers in one wearable device, including pH, electrolytes, metabolites, and inflammation markers like C-reactive protein (CRP), providing a comprehensive view of physiological health.¹⁷⁴

The technologies above can support a broad range of applications, such as the detection of hazardous heavy metals (e.g., Zn, Cd, Pb, Cu, and Hg), using electrochemical square-wave anodic stripping voltammetry (SWASV).^{175–177} Other sensors can monitor exposure to addictive substances, such as caffeine,¹⁷⁸ nicotine,¹⁷⁹ alcohol,¹⁸⁰ and 2-fluoro-methamphetamine (2-FMA).¹⁸¹ As examples in the second area of application, reports describe sensors that can measure loss of water, electrolytes, amino acids, and other metabolites through eccrine sweat.^{27,182,183} The results are important for monitoring hydration status and maintaining homeostasis during physical activities associated with sports and manual labor. In one specific case, a microfluidic wearable band with channels for measuring sweat rate and total loss, containing colorimetric assays for sweat pH and lactate, permitted monitoring of indicators of muscle metabolism and fatigue.¹³⁸

The third application area aims to establish sweat and noninvasive measurements of its chemistry as an alternative to traditional blood and urine assays for health monitoring and disease management. A key challenge in such uses is in establishing quantitative connections between sweat and blood, a task complicated by factors such as secretion dynamics and variable gland responses.^{24,104,184} Early studies were often hindered by inconsistencies in sweat induction and collection protocols, leading to unreliable data across many analytes.^{185–187} Even today, commonly targeted species such as sodium, chloride,¹⁸⁸ glucose,^{189–191} and lactate¹⁹² continue to pose challenges due to their high sensitivity to factors like sweat rate and local metabolic activity. However, recent advances in wearable platforms, particularly those integrating microfluidic sampling with standardized iontophoretic stimulation, have begun to establish robust connections between sweat and blood chemistries. Under such controlled conditions, strong and reproducible correlations have been demonstrated for certain analytes, including potassium¹⁸⁸ and ethanol.^{193,194} Urea¹⁹⁵ and cortisol¹⁹⁶ have also shown promising alignment with serum concentrations, offering potential pathways for noninvasive management of chronic kidney disease and stress. While challenges remain, these advances reaffirm the growing utility of sweat sensors in tracking dynamic physiological trends with increasing precision and clinical relevance.

Integration of capabilities for iontophoretic transdermal delivery of cholinergic agents (e.g., pilocarpine, carbachol) into wearable sweat analysis platforms provides a means for triggering localized sweat gland activation. This scheme builds on well-established methods for capturing samples of sweat for




Figure 5. (Left) Schematic diagram illustrating a skin-adhered sensor fabricated with a screen-printed carbon electrode (SPCE) on a microbial nanocellulose (MNC) substrate. The zoom displays an actual picture of the device adhered to the human skin. The pictorial representation on the right demonstrates that the self-adherent MNC is suitable for a wearable skin sensor capable of detecting biomarkers in sweat. Adapted from Silva et al.¹⁷⁷ © 2020 Elsevier.

assessments of chloride concentration in CF diagnosis. Typically, the stimulation uses hydrogels containing pilocarpine/carbachol with a 0.1–0.25 mA current, followed by microfluidic collection and multiplexed ion, metabolites, and nutrients analysis.^{197,198} These combined stimulation and sensing devices support monitoring of daily health parameters, including diet, supplement intake, and stress levels. Some studies suggest an ability for continuous sweat glucose monitoring in perspiration, as an alternative to blood glucose assessment.¹⁹⁹ Additional published data demonstrate the ability to track the metabolism of nutrients and vitamins in sweat after 24 h of supplement intake.¹⁹⁷ Also, by analyzing hormones such as cortisol,^{200,201} sweat sensors can provide insights into stress response and overall well-being. In clinical contexts, such technologies can help to diagnose and manage various diseases. For example, CF wearable devices can quantify chloride and sodium concentrations using either colorimetric or electrochemical approaches.^{197,202–204} Tracking glucose and uric acid levels in sweat is relevant for diabetes and gout.^{126,205} Additional possibilities are in monitoring for cardiac conditions, where the combined tracking of sweat loss biomolecule concentrations can complement measurements of other vital signs (blood pressure, heart rate, etc.) for a comprehensive view of physiological state.²⁸

It is worth recalling the challenges in using sweat as a matrix for quantitative and real-time detection. First, sweat availability is related to the patient's transpiration, which can be passive or induced. The passive one requires specific environmental conditions (high temperature and humidity), which are not always possible to obtain. The active transpiration typically relies on drugs and direct or reverse iontophoresis, which may be difficult to implement. A second issue is sweat rate normalization, as the rate of sweating of a human changes over time. This influences the concentration of any analyte in sweat, creating an artifact in the measurements. For example, the concentration of a metabolite such as lactic acid, passively generated during physical exercise, is initially extremely high due to the reduced amount of sweat produced at the beginning

of physical activity. A few minutes afterward, the volume of sweat drastically increases; thus, the lactic acid concentration measured results are much lower.²⁰⁶

Wearability is also crucial for sensors applied to sweat. Most skin-conformable sensors are fabricated using synthetic polymers due to their wide range of compositions and ease of functionalization. However, it may be advantageous to explore natural, skin-substitute polymers such as microbial nanocellulose (MNC), which is also used in wound dressings. Advantages of MNC include its strong adhesion to skin without the need for additional adhesives and its potential use in implantable devices, as it resists degradation in the human body while maintaining structural stability. Silva et al.¹⁷⁷ demonstrated the use of MNC as a substrate for screen-printed carbon electrodes (SPCEs). This concept is schematically illustrated in Figure 5, which highlights the suitability of self-adherent MNC for the noninvasive detection of biomarkers in sweat. The zoom of Figure 5 also shows a photograph of a SPCE-based electrochemical sensor deposited on an MNC substrate adhered to the human skin. A significant enhancement in analytical performance for the detection of cadmium (Cd^{2+}), lead (Pb^{2+}), uric acid, and 17β -estradiol was obtained by pretreating the surface of the SPCEs in diluted sulfuric acid, leading to faster electron transfer rates and lower limits of detection (LODs). The reason for such an enhancement was related to the creation of oxygen groups on the carbon surfaces, which increased wettability and hydrophilicity.

Detection of the protein metabolite urea, a biomarker of renal function, is relevant for patients at risk of heart failure. Ideally, urea levels should be monitored continuously, which can be achieved using wearable sensors to detect urea in sweat.²⁰⁷ Urea levels in sweat can be correlated with those in blood, but their electrochemical detection is influenced by sweat pH, which varies among individuals. This limitation was addressed by Ibáñez-Redín et al.,²⁰⁷ who developed wearable potentiometric biosensors on SPCEs capable of quantifying both urea concentration and pH. The biosensor architecture and fabrication procedure are illustrated schematically in




Figure 6. (A) (i) Fabrication procedure for potentiometric sensors that could detect urea and measure pH in sweat. (ii) Change in potential as increasing concentrations of urea were added for the sensor with the PVC membrane. (iii) Same as in (B) with a sensor containing a Nafion membrane. (iv) Calibration curves for urea detection using the two types of sensor (with Nafion and PVC membranes). (B) (i) Potentiometric

Figure 6. continued

response of the sensor at distinct pHs, which led to the calibration plot in (ii). (iii) Sensor response at different urea concentrations, from which the calibration plot was obtained (iv). (v) Photo of the sensor attached to the forehead of a volunteer. (vi) Urea and pH measurements at 3 different days with sweat of the same volunteer. (vii) Photo of the wearable device. (viii) demonstration that the wearable device could be bent multiple times without compromising the potentiometric response. Reproduced with permission from Ibáñez-Redín et al.²⁰⁷ © 2023 Elsevier.




Figure 7. (A) Fabrication process of the screen-printed sensors on a plastic glove. (B) Schematic representation of the 3-electrode sensor where the auxiliary (AE), reference (RE), and working (WE) electrodes are marked. (C) The working electrodes varied depending on the analyte to be detected. (D) Schematic representation of the concept of therapeutic drug monitoring as quantification in sweat may indicate how much of a given drug has been metabolized by the patient. Also shown in this part are scanning electron microscopy images of the working electrodes on the different fingers. (E) Representative signals in DPV for the 4 analytes detected upon touching human sweat on the skin. Reproduced with permission from Raymundo-Pereira et al.²⁰⁸ © 2022 Elsevier.

Figure 6Ai, showing a reference electrode and two working electrodes (one for pH and one for urea). The dual sensor included polyaniline ink, urease bioink, and a polyvinyl chloride (PVC) membrane. The working electrode for urea featured a polyaniline (PANI)-based ink that detected pH increases resulting from urea hydrolysis by urease. The enzyme was immobilized using a chitosan-containing ink, which helped preserve its activity. Data from the potentiometric measurements are presented in Figure 6Aii-iv, where a change in potential occurs as urea concentration increases. The biosensor presented a rapid response, with an incubation time of only 5 min.

The wearable device's ability to detect urea in a real-world scenario was demonstrated by attaching the sensor to a volunteer's forehead. The results and corresponding photos are presented in Figure 6B. The potentiometric response and calibration plots for urea and pH in the volunteer's sweat sample are shown in Figure 6Bi-iv, while a photo of the sensor on the forehead is shown in Figure 6Bv. Having determined the pH of the sweat sample as 7.2, the urea concentration could be precisely determined using the corresponding pH value in the calibration curve. A urea concentration of 10.9

mM, typical of healthy individuals, was obtained and remained stable over 3 days, as shown in Figure 6Bvi. The epidermal device, depicted in Figure 6Bvii, includes an opening to enable direct contact between the electrodes and the skin. It demonstrated flexibility, as it could be bent multiple times without affecting the potentiometric response, as evidenced by Figure 6Bviii.

A plastic glove used to determine therapeutic drugs and biomarkers in human sweat samples was developed by Raymundo-Pereira et al.²⁰⁸ The electrochemical sensor array was screen-printed onto four fingers of the glove, as in the picture of Figure 7A, each of which was designed for a specific analyte. The sensor on the index finger was functionalized with carbon black to quantify uric acid to a LOD of 1.37×10^{-6} mol L⁻¹. The middle and ring finger sensors were coated with Printex Carbon to detect paracetamol and paroxetine with LODs of 2.47×10^{-7} and 4.93×10^{-7} mol L⁻¹, respectively. Detection of the ethinylestradiol hormone was carried out with a pretreated screen-printed carbon electrode on the little finger, with an LOD of 9.35×10^{-7} mol L⁻¹. Besides the choice of sensing materials, measuring conditions in differential pulse voltammetry (DPV) were carefully chosen to yield the

high sensitivity and selectivity achieved with the glove-embedded sensors.

Adequate signal acquisition and transmission are essential for the development of lightweight and compact wearable devices that do not hinder the natural movements and daily activities of the patient (see Section 9 for further details). In sweat-based sensing platforms, wireless communication technologies such as near-field communication (NFC) and Bluetooth are commonly employed to enable data transmission to external readers or smartphones. Integrating wireless communication and signal-conditioning functionalities into compact electronic modules allows a significant reduction in system size and system complexity compared to traditional wired readout architectures. Indeed, a fully printed wearable system was developed for the detection of copper ions in sweat, correlated to the diagnosis of Wilson's disease. The printed microfluidic devices were connected to an ad hoc Bluetooth system for the wireless readout with a customized smartphone app. The readout device was also designed with the possibility of performing reverse iontophoresis to get a sweat sample in adverse environmental conditions. A multi-layer microfluidic passive system (capillary-based) was designed to (i) maximize the area of collection of the sweat, (ii) take the sample to the screen-printed carbon device for the copper measurement by SWASV, (iii) let the sample flow through interdigitated electrodes to measure its conductivity, and (iv) let the sample flow along a channel surrounded by two parallel long electrodes to measure its volume by the change of impedance among them and taking advantage of the measured conductivity. Then, the sweat rate would be defined and used for the copper concentration normalization. The microfluidic circuit, which was built using patterned biadhesive plastic tape, ended in a sponge, so that the drying of the sweat in the channel after the measurements was granted, allowing for new tests and keeping the capillary effect moving the sample onto the electrodes. The system showed a LOD of 396 ppb, a linear range up to 2500 ppb, and a sensitivity of 2.3 nA/ppb in artificial sweat.

With regard to commercially available devices, Eccrine Systems was one of the earliest companies focused on wearable sweat sensors for diagnostic purposes via electrochemical approaches. They established important technical and clinical benchmarks but ultimately ceased operations. Another company, Epicore Biosystems, launched shortly after Eccrine Systems, found commercial success in comparatively simple devices that use microfluidics and colorimetry for monitoring sweat rate, total sweat loss, and electrolyte concentration. The Gx Sweat Patch, launched in the spring of 2021, was the first such product, through a partnership with Gatorade. This thin, flexible microfluidic device pairs with a graphical user interface that operates on a smartphone to capture images that allow for quantitative measurements of local sweat loss and electrolyte concentration, the latter through colorimetric measurements of the concentration of chloride. The result is a personalized recommendation for rehydration following physical exertion in sports training or competition, with specific guidance on replenishment of lost water and electrolytes.²⁰⁹ Advanced versions of this technology support wireless electronics for continuous measurements without the need for visual access to the device, designed for applications in worker safety associated with the oil/gas, construction, manufacturing, mining, and agricultural industries. Specifically, data is passed on to a mobile app and cloud dashboard, allowing safety

managers to intervene before heat stress occurs, thereby improving workforce safety and performance.²⁷ Another key player in this space, Nix Biosensors, reports a reusable hydration biosensor designed primarily for endurance athletes and military applications, providing hydration recommendations with historical and predictive data. Other emerging efforts in commercialization are at the company Persperion, where the focus is on sweat for inferring glucose concentrations in blood by electrochemical detection of glucose levels through sweat collected from the fingertips.²¹⁰ Persperity Health is also advancing wearable sweat-sensing technology focused on women's health. Their platform uses iontophoresis-based on-demand sweat induction to enable real-time assessment of key reproductive hormones like estradiol and progesterone.¹⁰²

As these examples illustrate, the commercial landscape of sweat collection and analysis technologies has evolved rapidly over the last several years, transitioning from simple absorbent pads for collection and separate analysis using conventional techniques to integrated microfluidic systems capable of real-time monitoring in a wearable format. The potential is to transform noninvasive, continuous health monitoring practices and disease diagnosis from wearable devices that track biophysical parameters such as heart rate, respiration rate, physical activity, and temperature to those that include diverse biochemical markers. The result may lead to proactive health management strategies, ultimately improving overall well-being in a cost-efficient manner, applicable across various populations and in resource-constrained areas of the world.

3.2. Saliva

Saliva is a complex biofluid that can provide valuable physiological and biochemical information about both local and systemic health. It is secreted primarily by the parotid, submandibular, and sublingual glands and contains a wide range of biomarkers, including electrolytes (Na^+ , Cl^- , Ca^{2+}), small molecules (glucose, lactate, uric acid, cortisol, and urea), proteins (enzymes, immunoglobulins, mucins), nucleic acids, and microbial metabolites.^{211,212} As saliva originates partly from the filtration, diffusion, or active transport of blood plasma, it carries molecular signatures that reflect not only oral but also systemic physiological states, making it useful for tracking health status in real time.²¹³ Unlike blood, saliva can be collected repeatedly without pain or specialized equipment, to monitor hormone regulation, immune activity, metabolic function, and stress responses.²¹⁴ Salivary biomarkers have been linked to numerous conditions such as diabetes, oral cancer, cardiovascular diseases, stress, and infectious diseases.^{215–219}

Metabolism "begins in the mouth," and saliva plays an active biochemical role in early digestion. Enzymes such as α -amylase, which initiates starch breakdown, and lingual lipase, which begins lipid digestion, transform the oral cavity into a dynamic metabolic environment immediately after food enters the mouth.²¹⁴ The salivary flow rate, electrolyte content, and enzyme activity are influenced by dietary intake, stress, circadian rhythms, and autonomic nervous system activation, factors that make saliva a rich medium for monitoring metabolic physiology. Saliva also provides clinically useful information about oral hygiene and dental health. Its buffering capacity, antimicrobial proteins, and mineral composition directly regulate enamel demineralization and plaque formation. Simple measurements such as salivary pH can offer

rapid insights: low pH is associated with a higher risk of caries and enamel erosion, whereas high pH and increased buffering may indicate calculus formation or altered microbiome composition.²²⁰ Because pH strips and miniaturized pH electrodes are easy to integrate into sensors, salivary pH remains one of the most accessible and informative biomarkers for everyday oral-health monitoring.

Engineering wearable or intraoral biosensors for saliva presents significant challenges. Saliva contains abundant proteases, amylases, and lipases that can degrade analytes or foul sensor surfaces, reducing sensor stability.²²¹ The mouth is also chemically unpredictable because eating, drinking, toothpaste, and mouthwash introduce rapid fluctuations in pH, ionic strength, and viscosity. Moreover, the oral cavity is mechanically harsh, exposing devices to constant moisture, shear forces, and temperature changes. Electronics must be encapsulated in biocompatible, waterproof materials while still enabling wireless communication and stable sensing performance. Additionally, a stable sample acquisition should be ensured, accounting for variations in salivary flow rate and minimizing interference from food residues and microbial contamination.²²² The flow and viscosity of saliva can affect analyte concentration, requiring integrated microfluidic systems for controlled sampling and normalization. Comfort and user acceptance are additional obstacles. Oral devices must not interfere with speech, swallowing, or breathing. Nevertheless, several platforms have shown strong promise. Smart mouthguards, originally developed for athletes, have been adapted to measure analytes such as glucose, lactate, and uric acid during real-world activities.^{223,224} Orthodontic retainers offer discrete, long-term placement and can host microfluidics and electrodes with minimal discomfort.²²¹ Additional information on mouthguards as saliva-based biosensors can be found in Section 5.4. Ultrathin tooth-mounted “tattoo” sensors, including graphene or RF-based enamel patches, demonstrate the potential for invisible, continuous monitoring of oral chemistry and bacteria.²²⁵ For infants and pediatric care, pacifier-based biosensors can continuously track electrolytes and metabolites, providing stress-free sampling.^{226,227} Lollipop-based microfluidic systems (“lab-on-a-lollipop”) encourage natural saliva production and enable easy point-of-care testing, especially in children.²²⁸

The composition and characteristics of saliva depend strongly on the mode of secretion, which can be classified as unstimulated (basal) or stimulated. Both types have distinct biochemical profiles, flow rates, and implications for biomarker detection, which are relevant for the design and interpretation of wearable biosensors.²²⁹ Unstimulated saliva is continuously produced mainly by the submandibular and sublingual glands, with a flow rate of $\sim 0.3\text{--}0.5$ mL/min.²¹² It reflects the body's resting physiological state and contains higher concentrations of proteins, hormones, and small molecules such as cortisol, IL-6, and oxidative stress markers. Because it is less affected by transient stimuli, unstimulated saliva provides a reproducible biofluid for longitudinal monitoring and is preferred for chronic or systemic biomarker detection. However, low flow and high viscosity may complicate sample handling and limit the volume available for analysis, particularly in patients with dry mouth. A collection period of at least 5 min is generally recommended to ensure representative sampling.²²⁹

The analysis of unstimulated saliva can be achieved through intraoral wearable sensors that maintain direct contact with the oral cavity. One example of an intraoral device using basal

saliva is the biosensor reported by García-Carmona et al.²²⁷ The system was designed for infant glucose monitoring, integrating a Prussian Blue-modified carbon electrode functionalized with glucose oxidase (GOx) in a wireless, baby-safe platform. The design uses the infant's natural sucking motion to draw saliva through a rectifying channel toward an external electrochemical cell, isolating all electronic components from the mouth. The unidirectional saliva flow enabled reproducible measurements without backflow contamination, while the external cell configuration avoided direct contact of electrode materials with oral tissue. The biosensor demonstrated a strong correlation between salivary and blood glucose in diabetic adults, confirming saliva's diagnostic relevance and illustrating how unstimulated saliva can be harnessed for continuous, noninvasive biochemical monitoring in sensitive populations such as neonates. Similarly, Ichikawa et al. developed a mouthguard-type optical sensor for monitoring salivary turbidity as an indicator of oral hygiene.²³⁰ The system integrates a 405 nm LED and a phototransistor within a double-layer mouthguard, sealed to protect electronics from moisture and powered by a wireless Bluetooth module. By tracking light transmittance through the thin salivary film formed naturally in the mouth, the device could continuously quantify turbidity values across 1–4000 FTU, showing strong agreement with benchtop spectrophotometric measurements. This design demonstrates the potential of optical readouts for continuous oral monitoring using naturally secreted saliva without any stimulation or external intervention.

Stimulated saliva is secreted upon mechanical or gustatory stimulation (e.g., chewing, citric acid), predominantly from the parotid glands. The flow rate increases up to 1–3 mL/min, while protein content decreases and pH and electrolyte levels (Na^+ , Cl^- , HCO_3^-) rise.²¹² This makes stimulated saliva less viscous and easier to collect, facilitating faster analyses and better reproducibility in subjects with low basal secretion. However, its composition varies with the stimulation method, which can influence biomarker concentration and sensor response. Standardizing the stimulation type and collection time is thus essential for accurate comparison and calibration across studies. Wearable systems that sample stimulated saliva offer additional design flexibility and safety advantages, such as reducing risks associated with prolonged intraoral contact. However, as the analyte concentrations in stimulated saliva are typically more dilute, higher sensor sensitivity and calibration strategies are required. A notable example of this approach is the paper-based microfluidic mouthguard sensor reported by de Castro et al.²³¹ In this semiwearable device, stimulated saliva obtained after a brief mechanical activation was introduced into a microfluidic paper analytical device (μ PAD) integrated into a 3D-printed mouthguard. The μ PAD incorporated colorimetric detection zones for glucose and nitrite, achieving detection limits of 27 and 7 μM , respectively. This work highlighted the practicality of low-cost, disposable colorimetric assays embedded in oral accessories, representing a bridge between conventional sampling and fully integrated wearables.

More recently, saliva-sensing dental floss has been introduced as an on-demand platform for stress assessment. During flossing, mechanical action stimulates saliva secretion and simultaneously guides it through the floss's thread-based microfluidic pathway toward a flexible electrochemical sensor. The device employs a MIP deposited onto porous laser-engraved graphene, enabling highly selective cortisol recog-




Figure 8. Wearable biosensing platforms for salivary analysis. (a) Pacifier-based biosensor for unstimulated salivary glucose monitoring. Top: On-body demonstration and representative amperometric response. Bottom: Schematic of the fluidic pathway, showing passive saliva pumping during sucking and enzymatic glucose detection via GOx/Prussian Blue in an external sensing chamber. Exploded views illustrate the integration of the sensor, chamber, and electronics into a baby-safe pacifier platform. Reproduced from García-Carmona et al.²²⁷ © 2019 American Chemical Society. (b) Wearable electrochemical biosensing based on laser-induced graphene (LIG) electrodes for molecular detection in saliva. Top: motivation for detecting vanillin as a food additive. Bottom: fabrication steps and the hand-held reader for on-body measurement. Reproduced with permission from Ma et al.²²⁸ © 2024 Royal Society of Chemistry. (c) μ PAD-integrated mouthguard device for colorimetric detection in saliva. The assembly consists of a 3D-printed holder, a removable microfluidic paper analytical device (μ PAD), and a top cover. Reproduced with permission from de Castro et al.²³¹ © 2019 Springer Nature. (d) Thread-based microfluidic dental floss sensor for stimulated salivary cortisol detection. Left: fabrication of the porous laser-engraved graphene (PLEG) electrode, electropolymerized molecularly imprinted polymer (eMIP), and template extraction steps forming cortisol-selective binding sites. Right: on-demand cortisol sensing during flossing, with wireless readout to a portable potentiostat and smartphone interface for stress assessment. Reproduced from Sharma et al.²³² © 2025 American Chemical Society.

nitition with ultralow detection limits and strong correlation to ELISA.²³² Some of such saliva (bio)sensor formats are illustrated in Figure 8.

Future efforts are focused on improving signal stability, antifouling performance, and real-time calibration. The oral cavity presents unique challenges, including variable pH and the presence of enzymes and bacteria that may degrade the sensing elements. Strategies such as applying coatings, implementing ratiometric sensing, and monitoring the salivary flow rate *in situ* are being developed to address these issues. The transition from single-analyte enzymatic devices to multiplexed and integrated intraoral systems holds great promise for real-time, personalized health monitoring. As wireless, biocompatible microelectronics, materials, flexible sensing technologies, and microfluidic control continue to evolve, next-generation saliva-based wearable biosensors are poised to play a key role in continuous biochemical monitoring and early disease detection.

3.3. Tears

The eyes can provide important physiological signals, including intraocular pressure and temperature, besides allowing a minimally invasive analysis of diverse biomarkers in tears. These include electrolytes (K^+ , Na^+ , Cl^- , HCO_3^- , Mg^{2+} , Ca^{2+}), proteins, lipids, and small molecules (e.g., glucose, urea,

lactate). These biomarkers carry information not only from the eyes but from the entire body, as tears are created from the filtration of blood plasma. The main challenge, in this case, is to detect biomarkers precisely in tears and correlate their contents with those in blood.²³³ Biomarkers in tears have already been associated with different types of cancer, Parkinson's disease, multiple sclerosis, diabetes, ocular allergies, glaucoma, and others.²³⁴

The analysis of basal tears (i.e., the protective layer of tears that lubricate, nourish, and clean the eyes) is of interest due to the higher analyte content and greater sample reproducibility when compared to stimulated tears. This can be performed, for example, through direct contact of the analytical device and the eyes, including contact lenses, which will be further discussed in Section 5.3. Park's research group²³⁵ developed a wireless, soft, smart contact lens for noninvasive detection of glucose in tears. The wearable sensor comprised functional devices (rectifier, light-emitting diode (LED), glucose sensor) that collected tears and performed a colorimetric reaction. A stretchable, transparent conductor was used for data transmission, interpretation, and display. This platform for personalized eye health management was demonstrated with *in vivo* tests in a rabbit wearing the contact lens, including monitoring of temperature on the rabbit's eye. A smart contact lens was also developed to correlate glucose in tears and blood,

illustrated in Figure 18.²³⁶ The key components of the lens include a glucose biosensor, an NFC chip, and an antenna to transmit the tear glucose data via an app. This work is also relevant because it strictly tracks the effects of wearing the prepared contact lens without interference from other causes, such as pH value. The tears were collected by an external mechanical stimulation to the conjunctiva while wearing the devices. In addition, the designed glucose biosensor ensures long-duration continuous measurements of tear glucose. This enabled the precise identification of the lag time in all of the tested subjects (rabbit, dog, and healthy and diabetic humans) for the in-depth correlation analysis between tear glucose and blood glucose.

Wearable devices that analyze stimulated tears are also an interesting option because of the lower chances of promoting infections or the leakage of metals or other harmful substances into the eye. In this case, however, the use of more sensitive devices is important due to the dilution of the analytes as well as the monitoring of the generated tear volume. Sempionatto et al.,²³⁷ for example, developed modified eyeglasses for the real-time and noninvasive detection of glucose, ethanol, and vitamins in tears (Figure 9). In this case, tears were stimulated using a menthol stick and captured through a superhydrophilic microfluidic device attached to the nose bridge pad of the eyeglasses. Carbon electrodes modified with Prussian blue and containing GOx or alcohol oxidase (AOx) were applied for the detection of glucose and ethanol, respectively. Bare carbon electrodes were used for the simultaneous detection of vitamins B2, C, and B6 through square-wave voltammetry (SWV). The devices were applied for multiple rounds of analyte monitoring, showing great reproducibility and responsiveness throughout the full day.

Physical parameters can also be an important source of health information for the eyes. For example, an elevated intraocular pressure may be an indication of glaucoma, an irreversible ocular disease affecting 80 million people worldwide.^{238,239} Many glaucoma patients miss the optimal time for treatment because of the lack of frequent intraocular pressure monitoring. Continuous monitoring was made possible with a graphene-based smart contact lens.²⁴⁰ The lens had a high sensitivity of $1.0164 \text{ mV mm Hg}^{-1}$ on a silicone eye, and $3.166 \text{ mV mm Hg}^{-1}$ *in vitro* on porcine eyes, displaying remarkable linearity. In a later work, a smart closed-loop system that combines a $\text{Ti}_3\text{C}_2\text{T}_x$ MXene-based soft contact lens sensor, wireless data transmission units, display, and warning components to realize continuous and nondestructive pressure monitoring was developed.²⁴¹ The $\text{Ti}_3\text{C}_2\text{T}_x$ contact lens had a high sensitivity of $7.483 \text{ mV mm Hg}^{-1}$ with superior linearity on silicone eyeballs, and excellent biocompatibility when worn onto the rabbit eyes. The intraocular pressure monitoring/real-time display in the smartphone was also realized via the integration of the smart contact lens with Bluetooth modules. An SCL can also function as a theranostic device by delivering a drug to decrease the intraocular pressure when a threshold value is reached. Kim et al.²⁴² reported on a theranostic SCL (Figure 18, Physical signal) consisting of a sensitive gold hollow nanowire sensor, a flexible drug delivery system, wireless power and communication systems, and an integrated circuit chip for pressure monitoring and control. These functionalities were demonstrated in *in vivo* experiments on glaucoma-induced rabbits, thus highlighting SCLs for POC ophthalmic diagnosis and disease management.




Figure 9. Wearable device based on modified eyeglasses for the detection of glucose, ethanol, and vitamins in stimulated tears. (A) Pictures of the analytical device attached to the nose-bridge pad of the eyeglasses, and the portable potentiostat linked to the side of the wearable. Scheme showing the tears flow on the device, the enzymatic detection mechanism for ethanol, and the overall form of the obtained signal. (a) Dry device; (b) tears enter the device; (c) stable signal related to the analyte concentration; (d) tears leave the device, making it ready for the next measurement. (B) When performing ethanol detection, the signals were recorded after the volunteer had a drink. The device was calibrated through the simultaneous usage of a breathalyzer. (C) Detection mechanism of: (i) glucose using GOx and a PB-modified carbon electrode; (ii) Vitamins using SWV and a bare carbon electrode. (D) Assembly of the microfluidic device. 1—hydrophilic membrane; 2—double-sided tape; 3—filter paper; 4—screen-printed electrode; 5—hydrophilic membrane. The stimulation of tears using the menthol stick is shown in i–iii, while pictures of the assembly of the wearable device are shown from iv–vii. Reproduced with permission from Sempionatto et al.²³⁷ © 2019 Elsevier.

The main challenges for sensing biomarkers in tears are the small sample volumes combined with their rapid evaporation. The stimulation of tears can be difficult, and the obtained biofluid volume is commonly inconsistent as well as the concentration of analytes. However, these remain valuable for containing biomarkers for important diseases and exogenous compounds. Also, their characteristics facilitate analytical processing, such as transparency, lower cellular content, and fewer interfering components (compared to blood).

3.4. Wound Fluid

Wound biofluids, such as exudate and ISF, are important in the wound healing process and are rich in biomarkers.²⁴³ Chronic wounds, such as those caused by diabetes or prolonged pressure, pose significant challenges for traditional management due to their intricate healing processes and heightened vulnerability to infection.^{244,245} The composition of the wound fluid depends on the wound type (acute or chronic) and the phase of healing. For example, pH levels serve as key indicators, since chronic wounds typically exhibit an alkaline pH (7.15–8.93), indicating inflammation or infection, while healing wounds tend to be more acidic.^{246–249} Similarly,




Figure 10. Wearable sensor technologies for wound monitoring. (A) Microenvironments and biomarkers in wound biofluids for diagnosing and monitoring the wound healing status. Reproduced from Gao et al.²⁴³ Available under a CC BY 4.0 license. ©2021 American Association for the Advancement of Science. (B) Multiplexed biosensor for simultaneous monitoring of multiple wound biomarkers. Reproduced from Sani et al.²⁶⁸ Available under a CC BY 4.0 license. © 2023 American Association for the Advancement of Science. (C) Mechanisms of electrochemical sensors: amperometric enzymatic electrode sensors and potentiometric ion-selective sensors utilizing metal electrodes. Reproduced from Sani et al.²⁶⁸ Available under a CC BY 4.0 license. © 2023 American Association for the Advancement of Science. (D) Aptamer-functionalized electrochemical sensors with microfluidic wound collectors. Reproduced from Gao et al.²⁴³ Available under a CC BY 4.0 license. © 2021 American Association for the Advancement of Science. (E) Clinical validation of electrochemical sensor for wound monitoring. Reproduced from Gao et al.²⁴³ Available under a CC BY 4.0 license. © 2021 American Association for the Advancement of Science. (F) Colorimetric pH wound dressing integrating an optoelectronic sensor with green LED and PD for rapid detection and NFC wireless communication of wound status. Reproduced with permission from Cho et al.²⁴⁹ © 2024 John Wiley & Sons. (G) High-resolution pH colorimetric wound dressing leveraging fluorescent expression changes for precise diagnostics. Reproduced from Zhang et al.²⁸¹ Available under a CC BY 4.0 license. © 2024 John Wiley & Sons. (H) Wax paper-based dressing sensor enabling colorimetric diagnostics for five key biomarkers—uric acid, TMA, pH, temperature, and moisture—designed for advanced burn care management. Reproduced from Zheng et al.²⁵⁹ Available under a CC BY-NC license. © 2023 American Association for the Advancement of Science. (I) Colorimetric polymeric fiber-based wound fluid sensor integrating biomarkers for pH, glucose, and MMP, offering a multifunctional design for simultaneous wound condition diagnosis. Reproduced with permission from Giovannini et al.²⁸⁵ © 2024 John Wiley & Sons.

inflammatory cytokines such as TNF- α , IL-6, and IL-8 are elevated during the inflammatory phase of chronic wounds, signifying heightened immune activity and delayed healing.^{243,250–252} Elevated glucose levels in wound exudate may indicate poor metabolic control in diabetic wounds, while reduced levels may suggest bacterial consumption, pointing to a potential infection.^{253–257} Lactate, a byproduct of anaerobic metabolism, serves as a marker for tissue hypoxia and angiogenesis, providing insights into oxygenation and metabolic activity.²⁵⁸ Other biomarkers, such as reactive oxygen species (ROS) and uric acid, are associated with oxidative stress and inflammation; elevated levels may indicate excessive immune responses or impaired tissue repair.^{259–262} Additionally, monitoring the bacterial load, particularly the presence and concentration of specific pathogens, helps assess the infection severity and guide appropriate treatment strategies.^{263,264}

Figure 10 provides an overview of wearable sensor technologies for wound monitoring. These sensors commonly rely on noninvasive or minimally invasive sampling strategies, focusing on direct biofluid collection from the wound site.^{265–267} The multiplex sensing depicted in Figure 10B enables simultaneous monitoring of various biomarkers, permitting the assessment of wound healing and infection status.²⁶⁸ Wound monitoring can be performed with electrochemical and optical transduction techniques^{50,243,263,268–273} to quantify biomarkers with high sensitivity and specificity. Electrochemical sensors are used due to their sensitivity, compact size, and versatility,²⁶⁵ mostly being amperometric or potentiometric, as illustrated in Figure 10C. Cytokines and bacterial loads have been detected with aptamer-functionalized electrochemical sensors (Figure 10D).²⁴³ These sensors used graphene-gold nanoparticle composites to enhance electron transfer and signal fidelity and have been applied to monitor conditions associated with venous ulcers. Figure 10E shows the measured pH related to inflammation and infection, cytokine concentration related to immune activity, and bacterial load. Using such a multiplex sensor system permits identifying infections early, thus enabling timely intervention.

Wound monitoring can also be done with optical sensors, particularly colorimetric ones. Recent advancements include multifunctional wound dressings incorporating pH-sensitive dyes or natural extracts such as turmeric,^{135,274} cabbage,^{275,276} and blueberries.²⁷⁷ Traditional methods included pH sensors embedded in mammary glands or composite pH sensors fabricated with flexible polymers.^{278,279} However, these approaches often constrain patient mobility and require the involvement of medical professionals with specialized expertise, rendering them unsuitable for wearable diagnostic devices. Colorimetric sensors have been reported for lactic acid and glucose, which, as previously mentioned, are very relevant to monitor in wound exudates.^{280,281} Furthermore, fluorescent materials have been integrated into devices, aiding in the monitoring of the progression and pathological states of infected wounds.²⁵⁷ These materials show responses to glucose²⁸² and bacterial infection,²⁸³ for example, providing a visual indication of wound deterioration or healing. Diagnostic and therapeutic functions can be combined by incorporating drug layers into wound dressings.^{274,284}

Figure 10F illustrates a wound dressing with curcumin, a natural extract from turmeric, as a colorimetric indicator.²⁴⁹ The absorbance and reflectance of this multifunctional dressing are affected by structural changes in curcumin under different

pH conditions of the wound site. For nonmedically literate patients, the visible color changes provide an intuitive means to monitor wound healing or deterioration. These optoelectronic sensors utilize colored LEDs and photodiodes to measure changes in absorbance and reflection associated with the dressing's color. A POC strategy for pH determination of wound biofluids using holographic technology has been devised, leading to an adaptive holographic pH-sensing bandage.^{243,285} This dressing operates by detecting subtle shifts in optical peaks, enabling precise monitoring of pH variations. Figure 10G shows the wound dressing integrated with holograms to detect pH changes down to 0.25 intervals across a range from 6.50 to 8.75 in the wound environment.²⁸¹

The use of colorimetric sensors has been expanded to a wider range of biomarkers, as in the case of the wax paper in Figure 10H.²⁵⁹ This colorimetric wax paper can detect skin temperature, TMA (trimethylamine), uric acid, moisture, and pH. A breakthrough in polymeric fiber optics for chronic wound diagnosis and management, illustrated in Figure 10I, was achieved with simultaneous measurements of pH, concentration of glucose, and of metalloproteinases (MMPs) in the exudate.²⁸⁵ This multibiomarker sensor utilizes a fluorescent sensing approach by coating PDMS fibers with fluorescein isothiocyanate (FITC) and glucose oxidase-functionalized nanoparticles. This configuration enables precise detection of pH (range: 5–7), glucose (range: 0–10 mmol), and MMP concentrations (range: 0–10 $\mu\text{g/mL}$) in the wound exudate.

Research is expanding to include therapeutic functions, such as drug delivery,^{269,270} functional textiles,²⁴⁹ and electrical stimulation⁵⁰ into wearable sensors. By integrating such functionalities, these devices dynamically adapt to the wound environment to expedite healing.^{245,286–288} Wireless closed-loop smart wound care technologies may improve wound monitoring and treatment outcomes. For instance, a battery-free, wireless smart wound dressing was developed, which is capable of monitoring wound infection and delivering drugs on-demand using flexible electronics and NFC.²⁷⁰ This system includes sensors for temperature, pH, and uric acid, establishing a closed-loop mechanism that integrates real-time monitoring with therapeutic functions. Its effectiveness was validated through both *in vitro* and *in vivo* studies. Another study introduced a wireless, closed-loop smart bandage incorporating sensors and stimulators.⁵⁰ The bandage features a flexible bioelectronic system with skin-interfacing hydrogel electrodes for monitoring skin impedance and temperature, as well as delivering electrical stimulation in response to wound conditions. This approach led to accelerated wound healing and enhanced dermal recovery in preclinical models. Multifunctional platforms equipped with cloud connectivity and AI-driven analytics further enhance diagnostic precision, offering predictive insights and optimized care.²⁸⁹ Another interesting example is the bandages designed for diabetic wounds that adapt by monitoring glucose levels and releasing therapeutic agents as needed.^{286,290} In this case, hydrogels or porous membranes^{263,268} are employed for passive exudate absorption and integrated into wound dressings for continuous biofluid collection. Hydrogel-based dressings further contribute by mimicking the extracellular matrix, fostering a moist healing environment suitable for tissue regeneration.^{286,291} Water-powered bandages have also been fabricated to deliver electrical stimulation at low cost, which is essential for resource-limited settings.²⁹² Furthermore, integrated electrical

stimulation and photodynamic therapy enhance healing by promoting cell proliferation, angiogenesis, and antimicrobial activity. Electrical stimulation for wound healing has also been done with triboelectric nanogenerator (TENG)-based wearables.^{293,294}

Several challenges persist for the use of wearable sensors in wound monitoring. The heterogeneity of the wound biofluid composition across different wound types complicates sensor calibration and sensitivity optimization. Biofluid volume variability, such as excessive exudate in chronic wounds versus minimal fluid in healing wounds, impacts the sensor performance. Environmental factors such as temperature, humidity, and movement can also affect sensor stability, necessitating robust material designs. Furthermore, biocompatibility is a significant concern; prolonged sensor contact must not impede healing or provoke an adverse immune response.

Despite the current challenges, a few wearable technologies for wound healing have reached the market. Tools such as Woundchek Protease Status monitor protease activity in chronic wounds, delivering critical biochemical insights to guide treatment strategies. Devices like Natrox, which deliver oxygen to wounds, improve metabolism and tissue oxygenation, while SNaP (Smart Negative Pressure) provides a portable method for managing exudate and reducing bacterial loads. Grapheal's smart patch offers continuous, noninvasive monitoring of chronic wounds. This wearable patch uses a graphene core to detect and record bioparameters with high sensitivity, enabling caregivers to track healing progress remotely and receive early warnings of potential infections. Moreover, the patch can stimulate wound healing while conforming flexibly to various wound shapes. These systems are a demonstration of the seamless integration of monitoring and therapeutic functions, allowing real-time care adjustments.

3.5. Urine

Urine is an indispensable diagnostic medium due to its unique ability to reflect a spectrum of physiological and pathological conditions.^{295,296} It contains a variety of metabolites, electrolytes, proteins, and other analytes that diffuse from the bloodstream, providing a window into metabolic and systemic health. Biomarkers indicative of conditions, such as diabetes, kidney disorders, urinary tract infections, and metabolic imbalances, can be detected. However, the intrinsic dilution of analytes during urine production means that only those with sufficiently high plasma concentrations are readily detectable, unless supported by advanced, highly sensitive detection methods. The state of the art in wearable urine diagnostics integrates real-time, noninvasive analysis with user-friendly technologies.^{297,298} Biosensors embedded in diapers (see Section 5.8), for instance, enable detection of key biomarkers like glucose, bilirubin, and proteins. Using diapers is particularly beneficial for neonates, the elderly, and individuals requiring long-term care, where traditional sampling methods are impractical. Many of these devices now incorporate self-powered technologies, such as biofuel cells that generate electricity from urine itself.²⁹⁷ In addition to single-use wearable devices, smart urine analysis systems are evolving into sophisticated, multifunctional platforms, often integrating microfluidic channels for precise biomarker quantification alongside wireless communication modules for data transmission.

AI-based systems can analyze trends in urinary biomarkers, predict potential health risks, and provide tailored health

insights. Examples are smart toilets equipped with sensors and cameras that monitor urine flow, color, and volume while analyzing multiple biomarkers *in situ*.^{295,296,299–301} These devices can also collect stool samples and integrate their analysis with smartphone apps, offering a comprehensive approach to health tracking. Commercial diagnostic tools have expanded to include fluorescence-based sensors and immuno-chromatographic assays that address longstanding issues like interpretation variability and sensitivity. For example, optical readers integrated with immunoassay platforms have improved the reliability in detecting acute kidney injury biomarkers.³⁰² Systems such as nanoparticle-based CRISPR-Cas-amplified platforms³⁰³ or biosensors with volatile organic compound-based detection³⁰⁴ provide enhanced sensitivity, permitting the early detection of critical conditions like sepsis and cancer.²⁹⁵ Wearable devices with hybrid capabilities are also gaining traction; these systems not only monitor biomarkers but also deliver feedback to adjust therapeutic interventions in real time.²⁹⁵

The episodic nature of urination limits the temporal resolution of data collection, often restricting wearable devices to single-use designs. Variability in urine composition due to hydration, dietary intake, and circadian rhythms complicates the standardization of diagnostic thresholds.^{305–307} Additionally, embedding sensors into flexible and durable wearables requires overcoming significant engineering hurdles, including ensuring biocompatibility and user comfort.

3.6. Breath

Diseases of the respiratory system represent a major global health challenge, with increasing incidence rates of COPD, asthma, lung cancer, and other conditions placing significant burdens on healthcare systems.^{308,309} Exhaled breath analysis permits monitoring health systemically, including metabolic disorders, daily metabolic fluctuations, and overall physiological well-being. For instance, volatile organic compounds (VOCs) in exhaled breath, such as acetone and ammonia, provide insights into lipid metabolism, renal function, and liver health, making breath analysis useful for managing diabetes, kidney dysfunction, and cirrhosis.^{310–313} Traditional breath monitoring methods, such as pulmonary function tests and imaging studies, offer crucial diagnostic information, but they often require complex equipment and professional operation. In contrast, health monitoring can be done using exhaled breath aerosols (EBA) and exhaled breath condensate (EBC), which contain abundant biomarkers for both respiratory and systemic health conditions.^{314,315} Laboratory analysis of EBA involves particle collection and subsequent biochemical analysis. Particle collection typically utilizes filter membranes or electrostatic deposition techniques, followed by solvent extraction for mass spectrometry analysis or PCR-based determinations. For instance, an early COVID-19 diagnosis can be achieved by detecting SARS-CoV-2 nucleic acids in EBA. These laboratory analytical techniques provide high-sensitivity and high-resolution detection methods, permitting data validation and calibration of wearable devices.^{4,316}

EBA analysis has evolved from laboratory settings to real-time portable monitoring. Smart masks (see Section 5.7) serve as ideal platforms for wearable EBA analysis, incorporating features such as particle collection through integrated filter membranes or electrostatic deposition devices for real-time sample analysis via microfluidic technology. These devices can integrate immunosensors³¹⁷ or CRISPR sensing technology³³




Figure 11. Wearable smart mask system development for in situ breath aerosol and condensate analysis. (A) A wireless, battery-free, multifunctional pathogenic infection diagnosis system. Reproduced from Li et al.³¹⁸ Available under a CC BY 4.0 license. © 2023 Springer Nature. (B) A POC mask with microfluidics for COVID-19 test-based CRISPR technology. Reproduced with permission from Nguyen et al.³³ © 2021 Springer Nature. (C) A paper-based electrochemical sensor in a mask for H₂O₂ detection in exhaled breath aerosol. Reproduced from Maier et al.³¹⁹ Available under a CC-BY 4.0 license. © 2019 American Chemical Society.

for real-time pathogen detection, such as SARS-CoV-2 protein detection through immunoelectrochemical sensors, enabling rapid COVID-19 screening within minutes (Figure 11A).³¹⁸ Alternatively, a mask can incorporate a series of microfluidic chambers designed to sequentially extract, amplify, and detect viral DNA using CRISPR-based technology, providing an efficient approach for viral diagnostics (Figure 11B).³³ Moreover, wearable devices can monitor respiratory infection levels and treatment efficacy^{319,320} by measuring inflammation biomarkers in EBA (Figure 11C).

EBC is normally analyzed using LC-MS, nuclear magnetic resonance (NMR), and immunological detection methods (e.g., ELISA). Wearable devices for EBC analysis typically utilize passive cooling technologies, such as hydrogel evaporation and radiative cooling, to achieve efficient condensation under low power consumption (Figure 11C).³² These devices can integrate electrochemical, optical, and immunosensors for real-time detection of biomarkers, such as hydrogen peroxide, nitrite/nitrate, and ammonium. The collected data can be transmitted to smartphones or cloud platforms for real-time analysis and remote monitoring, with machine-learning algorithms enhancing analytical accuracy and efficiency. Besides particle collection, there are other innovations based on materials able to sample and analyze targets in breath, vapor, and/or gas.

3.6.1. Gas-Sampling and Sensing Mechanisms. We have already mentioned that exhaled gases contain biomarkers with clinical significance, including nitrogen oxides (NO₂, NO), hydrogen sulfide (H₂S), ammonia (NH₃), and VOCs like ethanol, acetone, and alkanes.^{32,321} These biomarkers are often related to specific diseases, including lung cancer, asthma, COPD, and tuberculosis.³² Current clinical breath analyzers are typically large and unable to provide continuous, real-time monitoring, which has led to growing interest in miniaturized, wearable “e-noses”. Electrochemical gas sensors that operate at room temperature are more suitable, with recent advances focusing on replacing the liquid electrolyte with flexible hydrogels. In the next few subsections, we discuss different hydrogel-based strategies to detect NO₂, H₂S, and O₂.

Figure 12A illustrates that hydrogels, as ion-conductive electrolytes, can be utilized to form a symmetrical electrode-hydrogel-electrode structure in electrochemical gas sensors when combined with metal electrodes. Under a direct current voltage, an anode-hydrogel interface and a cathode-hydrogel interface are established. Oxidizing gases such as NO₂ and O₂ are reduced at the cathode under the influence of the electric field, and the electrons transferred between the gas and the electrode generate a current response.^{322–327} Another transducing principle depicted in Figure 12B does not rely on constructing a special electrode-hydrogel interface, but requires




Figure 12. Gas-sensing mechanisms of hydrogel-based gas sensors. (A) Schematic illustrating that the NO_2 gas is reduced at the cathode-hydrogel interface of the electrochemical-type gas sensor under a DC external electric field. Reproduced with permission from Wu et al.³²² © 2021 John Wiley & Sons. (B) Schematic showing that soluble gas (NH_3) is absorbed by hydrogel and then hydrolyzed to generate movable ions, thereby reducing the resistance of hydrogel electrolyte. Reproduced with permission from Liu et al.³²⁸ © 2021 Elsevier. (C) Schematic of the self-powered electrochemical cell-type NO_2 sensor with two different electrodes. Reproduced with permission from Wu et al.³³⁰ © 2023 John Wiley & Sons. (D) Schematic diagram illustrating the device structure and H_2S -sensing mechanism of self-powered potentiometric gas sensor. Reproduced from Huang et al.³³¹ Available under a CC-BY 4.0 license. © 2023 Springer Nature.




Figure 13. Room-temperature gas sensing performance. (A) NO_2 sensor. (i, ii) Dynamic responses of the CaCl_2 -infiltrated hydrogel sensor with one and five turns of Ag coil electrode to 0.8–4 ppm of NO_2 gas at 33% RH. Reproduced with permission from Wu et al.³²² © 2021 John Wiley & Sons. (iii) Structure of the self-powered NO_2 sensor. (iv) Sensor response versus NO_2 concentration (C_{NO_2}) in a wide concentration range (200–10,000 ppb), showing its excellent linearity and high sensitivity. (v) Photographs presenting that the self-powered sensor in 10 ppm of NO_2 can generate sufficient current to turn on the LED and electronic timer. Reproduced with permission from Wu et al.³³⁰ © 2023 John Wiley & Sons. (B) H_2S sensor. (i) Dynamic open-circuit voltage change of Zn/Ag/hydrogel sensor to 40 and 20 ppb H_2S . (ii) Comparison of the responses of the Zn/Ag/hydrogel H_2S sensors to various gaseous chemicals, reflecting the high selectivity. (iii, iv) Response versus H_2S concentration curve of the Zn/Ag/hydrogel sensor to the exhaled breath of a healthy volunteer/halitosis sufferer and simulated halitosis gas. Reproduced from Huang et al.³³¹ Available under a CC-BY 4.0 license. © 2023 Springer Nature.

a small hydrogel volume and low initial conductivity, as well as strong adsorption and solubility of specific gases, such as NH_3 .^{328,329} The hydrolyzed gas produces ions, thereby reducing the sensor's impedance. Depending on the electrodes' potential difference and on the interaction between the gas and the interface, self-powered electrochemical cell-type sensors

can also be built. The reactive metals on one side of the hydrogel act as anodes, while the target gas (e.g., NO_2) spontaneously reduces at the cathode to generate current, as shown in Figure 12C.³³⁰ Another type of self-powered potentiometric gas sensor relies on the adsorption of gas onto electrodes. For example, H_2S can be adsorbed onto the

surface of an Ag electrode, altering its potential and enabling self-powered gas detection with minimized power consumption (Figure 12D).³³¹

3.6.2. Detection of NO₂. NO_x is an important biomarker for lung infections and bowel diseases.³³² The increasing recreational use of NO₂ also makes the development of rapid and simple techniques for its determination in breath relevant.³³³ Wu and co-workers found that ion-conductive polymer hydrogels could respond to NO₂.³²⁵ A CaCl₂-polyacrylamide (PAM) hydrogel enwound by two silver coils³²² showed high sensitivity (119.9%/ppm) at room temperature, in addition to high selectivity, short response and recovery times (29.8 and 41.0 s, respectively), good linearity, and a low LOD of 80 ppb. Because the electrical response mainly comes from the gas reaction at the hydrogel-electrode interface, mechanical deformation of the hydrogel, such as stretching and bending, has little impact on its sensing performance. In contrast, the electrolyte composition of the hydrogel affected sensor performance, and the addition of specific ionic salts to increase conductivity can promote electrochemical reactions at the electrode-hydrogel interface, thereby enhancing sensitivity. The high conductivity also suppresses electrical signal interference caused by ionization of dissolved gases due to hydrolysis, enhancing the selectivity. Increasing the number of silver wire coils was found to improve the sensor's response due to the larger reaction area (Figure 13Ai-ii).

To address long-term energy consumption, electrode corrosion, and humidity interference issues, an electrochemical self-powered sensor was designed, which featured heterogeneous anode/cathode materials, a recoverable electrode, and a hydrophobic hydrogel-electrode interface.¹¹ Compared with the silver coil structure mentioned above, the self-powered sensor configured as a flat-stacked Zn-hydrogel-carbon structure significantly increased the gas reaction surface area (Figure 13Aiii). The oxidized Zn electrode can be recovered through the charging process (reduced by applying a reversed DC voltage), ensuring long-term stability and a prolonged lifespan. With the addition of amphiphilic Zn(OTf)₂ to the hydrogel electrolyte, the hydrophobic electrode-gel interface created reduced humidity interference by repelling the adsorption of environmental moisture. With these optimizations, the self-powered sensor exhibited outstanding performance, including ultrahigh sensitivity (1.92%/ppb), linearity ($R^2 = 0.999$), an ultralow theoretical LOD (0.1 ppb), and humidity immunity (Figure 13Aiv). Notably, the energy generated from a 10 ppm of NO₂ reaction could even turn on an LED and electronic timer (Figure 13Av).

3.6.3. H₂S Sensors. H₂S is a typical biomarker produced by the metabolism of organic matter by accumulated bacteria, which can indicate oral diseases such as periodontal and mucosal diseases. A clinical study revealed that the average H₂S concentration exhaled from patients with oral halitosis is 20.64 ppb, much higher than the value of healthy volunteers (3.36 ppb). Wu and co-workers reported a self-driven and stretchable potentiometric H₂S sensor consisting of two different metal electrodes and a PAM/calcium alginate (CA) double-network hydrogel electrolyte.³³¹ Because adsorption and chemical reaction between H₂S and different metal electrodes differ, the open-circuit voltage (OCV) would change during exposure to different concentrations of H₂S. The chemical adsorption of H₂S on Ag electrodes was the strongest, with the open-circuit voltage being reversible when

H₂S was removed. This gas sensor exhibited high sensitivity, repeatability, and selectivity at room temperature (25 °C), being suitable for working under severe mechanical deformations or in aerobic/anaerobic environments. Furthermore, an organo-hydrogel obtained by replacing pure water with a water-glycerol binary solvent had an increased performance in terms of stability and environmental tolerance. Figure 13Bi shows that the self-driven sensor could detect ultralow concentrations of H₂S (20 ppb in the experiment and 0.79 ppb in theory), which meets the demand of detecting H₂S biomarkers released by bacteria. The sensor also had good selectivity, as indicated in Figure 13Bii, and could be used to diagnose halitosis (Figure 13Biii-iv).

3.6.4. O₂ Sensors. Detection of O₂ is important for health monitoring because the respiratory oxygen content or transcutaneous oxygen partial pressure (tcPO₂) is an indicator of cellular metabolism. Wearable O₂ sensors can be used in noninvasive diagnosis and rehabilitation of some chronic diseases, including respiratory diseases, diabetic foot, wound healing, and infection. Wearable electrochemical sensors have been developed by incorporating a hydrogel and a soft elastomer. For example, Wu and co-workers designed a waterproof and breathable Ecoflex-encapsulated hydrogel film, which can detect O₂ within a wide range of concentrations (from 5 ppm to 90% O₂).³²⁴ Notably, the waterproof and breathable elastomer endows good sweat and moisture interference. Therefore, not only could the sensor be assembled in a mask for breath detection, but it could also be adhered to the skin for monitoring tcPO₂.

Remaining challenges in wearable EBA and EBC analysis involve the standardization of sampling and analytical methods to ensure comparability and reliability. Sensors must achieve higher sensitivity and selectivity for accurate detection of low-concentration biomarkers, probably employing nanomaterials and novel sensing technologies. ML and big data methodologies may solve the problem of managing and analyzing the vast, complex data sets generated by these devices.^{4,334}

4. IMPLANTABLE AND INGESTIBLE BIOSENSORS

When monitoring is required for extended periods or when the desired biofluids are not accessible from the outside of the body, implantable and ingestible biosensors can be used. These biosensors can be applied for the detection of metabolites from different body fluids, such as the cerebrospinal fluid (CSF), gastric fluid, interstitial fluid, or blood, in real time.³³⁵ In the following, each biofluid will be discussed in further detail regarding its nature and bioanalytical application strategies.

4.1. Cerebrospinal Fluid

CSF is a clear liquid primarily produced in the choroid plexus of the ventricular system, playing a key role in cerebral blood regulation and in transporting molecules such as neurotransmitters and toxins.³³⁶ It flows constantly through the brain and spinal cord, providing mechanical support and acting as a shock absorber.³³⁷ In addition to monitoring physical parameters such as dielectric properties³³⁸ or flow dynamics,³³⁹ a primary area of advancement lies in the *in situ* quantification of specific chemical and biochemical species within CSF. Detecting biomarkers in CSF is important for diagnosing neurodegenerative diseases, cancers, and immune disorders.^{340–343} The collection of CSF has historically been done by performing either a needle biopsy or lumbar puncture (LP), which is extremely invasive and painful. In addition, LP leads




Figure 14. Overview of the implantable electrochemical device for the wireless and battery-free monitoring of catecholamines in CSF developed by Stuart et al. The image shows the device and its layers, as well as its operating principles and typical signals. Reproduced from Stuart et al.³⁵⁸ © 2022 American Chemical Society.

to multiple health-related complications for patients, including long-lasting headaches after the procedure.³⁴⁴ Furthermore, successive sampling can influence the concentration of some biomarkers, hampering the reliability of the sample.³⁴⁵ These reasons, combined with the possibility of real-time monitoring, are some of the most relevant for promoting the development of implantable devices for the analysis of CSF.

The practical application of implantable sensors is intricately linked to the anatomy of the central nervous system and the inherent challenges of accessing CSF-containing spaces. CSF resides primarily within the brain's ventricles, the subarachnoid space around the brain and spinal cord, and the central channel of the spinal cord.^{346,347} Accessing these locations typically requires neurosurgical procedures. Ventricular access often involves drilling a burr hole in the skull and inserting a catheter, while spinal access may involve procedures similar to lumbar puncture but utilizing a device intended for implantation.^{348,349} Each access route presents significant risks, including infection (meningitis, ventriculitis), hemorrhage, and potential damage to neural tissue. Therefore, sensor design must prioritize minimizing invasiveness. This translates into demands for small device size, flexibility to conform to surrounding anatomical structures, and the use of functional materials with established long-term biocompatibility to minimize tissue reaction and inflammation. Furthermore, the mechanical properties of the implant must be robust to

withstand physiological movements and CSF pressure fluctuations without fracturing or becoming dislodged.

Electrochemical methods are well-suited for such sensing devices due to their inherent potential for miniaturization, high sensitivity, selectivity, low power consumption, and suitability for digital integration. However, translating these methods into reliable *in vivo* sensors requires significant innovations in materials science and device engineering. Reliable molecular modification of electrode surfaces is essential for developing stable electrochemical sensors. Traditional gold–thiol (Au–S) linkages,^{350,351} such as those formed with cysteine or mercaptopropionic acid, are prone to displacement by endogenous thiols like glutathione (GSH), leading to signal drift or loss.³⁵² Stability can be enhanced by employing bidentate thiol linkers, which form dual coordination bonds with gold surfaces.³⁵³ This configuration demonstrated significantly lowered signal loss (5.3% after 2 h in GSH solution, compared to a 23.4% decrease observed with monodentate Au–S linkages). Further advancements include the use of gold–alkyne (Au–C≡C) bonds. Sensors utilizing Au–C≡C linkages exhibited superior stability, maintaining signal integrity even after 4 h in GSH solution.³⁵⁴ This improved performance is attributed to the higher bond energy and lower Gibbs free energy associated with Au–C≡C bonds, resulting in more robust surface assemblies. Additionally, conductive polymer coatings, such as poly(3,4-ethylenedioxythiophene) (PEDOT) integrated with polydopamine mela-

nin, have been explored to improve electrode stability and biocompatibility.³⁵⁵ These coatings not only enhanced electrochemical performance and mechanical stability but also supported cell proliferation and differentiation, making them suitable for long-term implantable applications.

A sonogel-carbon (SNGC) electrode modified with L-leucine was used for the electrochemical detection of homovanillic acid (HVA) in CSF.³⁵⁶ HVA, a key metabolite of dopamine and norepinephrine, serves as a biomarker for conditions like neuroblastoma.³⁵⁷ The fabrication of this SNGC-leucine sensor utilized a sol-gel process with methyltrimethoxysilane (MTMOS) and hydrochloric acid, followed by the incorporation of graphite powder and L-leucine. The resulting composite was packed into glass capillary tubes and polished to create the working electrode. The presence of L-leucine on the electrode surface enhanced the electrochemical detection of HVA, with the optimal voltammetric response being pH-dependent. DPV analysis demonstrated high analytical sensitivity, achieving a low LOD of 0.4 μM for HVA within a linear concentration range of 0.5 to 50 μM . The sensor's analytical performance was evaluated in artificial CSF samples.

Another example targeting neurotransmitter monitoring employed a miniaturized, wireless, and battery-free implantable platform that integrates electrochemical sensing (Figure 14).³⁵⁸ This device had carbon nanotube (CNT)-based sensors, taking advantage of its exceptional electrical conductivity and high surface area to detect neurotransmitters such as dopamine. A $\mu\text{-LED}$ was integrated on the same probe, enabling precise optogenetic control of nearby neuronal activity and allowing real-time investigation of stimulus-response dynamics. Wireless power transfer technology, combined with the absence of a battery, resulted in a compact (12 mm $\times$ 8.5 mm $\times$ 3.2 mm) and lightweight (<49 mg) device that minimized invasiveness and eliminated the need for battery replacement. A flexible serpentine interconnect ensured mechanical stability and electrical connectivity during implantation in moving subjects, including small animal models such as mice. Both *in vitro* and *in vivo* experiments demonstrated the platform's sensitivity and selectivity, highlighting its ability to monitor dopamine concentration changes in response to optogenetic stimulation and to track real-time dopamine levels following drug exposure (e.g., opioids and naloxone).

In addition to organic molecules, the monitoring of inorganic ions like calcium (Ca^{2+}) in CSF permits understanding neuronal signaling and pathological states.^{359,360} An all-solid-state potentiometric microelectrode, fabricated on the tip of an acupuncture needle (diameter <80 μm), was developed for *in vivo* Ca^{2+} monitoring in rat CSF.³⁶¹ This design, incorporating functional materials, circumvents the drawbacks associated with conventional liquid-filled reference electrodes. The calcium ion-selective electrode (ISE) was constructed by coating the acupuncture needle tip premodified with a PEDOT/poly(styrenesulfonate) (PSS) contact with a calcium ion-selective membrane. The resulting Ca^{2+} -ISE exhibited a Nernstian response over a wide concentration range (μM to mM levels) with a near-ideal slope (around 31 mV/decade) and a low LOD (0.12 μM Ca^{2+}). *In vivo* application of this microelectrode in rats following spinal cord transection revealed a sharp and significant increase in CSF calcium concentration from a baseline of approximately 21 to 133 μM , highlighting its capability for real-time monitoring of

dynamic ionic changes in the CNS with high temporal resolution. The versatility of this platform allows for the potential measurement of other ions by simply changing the ion-selective membrane, permitting investigation of various physiological and pathological conditions.

The detection of low-abundance protein biomarkers in CSF requires highly sensitive and specific recognition strategies. A sensor utilizing electrografted laser-induced graphene (E-LIG) electrodes was developed for the label-free detection of the poly(GP) dipeptide repeat, a biomarker for C9orf72-associated amyotrophic lateral sclerosis (ALS), directly in unprocessed CSF.³⁶² ALS is often linked to a hexanucleotide repeat expansion in the C9orf72 gene, leading to the production of toxic dipeptide repeat proteins, including poly(GP). Early detection of these biomarkers may allow for a timely intervention. The fabrication of the E-LIG electrodes involved a cost-effective laser-based conversion of a PI substrate to porous, conductive graphene. The graphene surface was then functionalized via electrografting and coated with antipoly-(GP) antibodies. Label-free detection of poly(GP) was achieved using electrochemical impedance spectroscopy (EIS), where the change in the charge transfer resistance (R_{ct}) upon antigen binding served as the detection signal. The E-LIG immunosensor demonstrated high sensitivity (LOD of 0.19 ng/mL in buffer and 0.27 ng/mL in CSF) and selectivity for poly(GP). The sensor quantified poly(GP) in CSF samples from ALS patients and distinguished between C9orf72-associated and non-C9orf72-associated ALS, thus confirming its potential for clinical diagnostics and POC applications.

In another approach for protein biomarker detection, a label-free electrochemical immunosensor for glial fibrillary acidic protein (GFAP),³⁶³ a biomarker for neurological conditions, was developed by using SPCEs modified with a nanocomposite.³⁶⁴ This sensor was applied to the GFAP determination in human CSF. The nanocomposite leverages the high conductivity of reduced graphene oxide (rGO), the peroxidase-like activity of molybdenum disulfide (MoS_2), and the electrocatalytic effect of silver nanoparticles (AgNPs) to enhance the amperometric detection of hydrogen peroxide, indirectly quantifying the antigen-antibody binding. The immunosensor was fabricated by covalently immobilizing a specific capture antibody for GFAP onto a nanocomposite-modified SPCE. The biorecognition event was detected by measuring the electrocatalytic reduction current of hydrogen peroxide. The immunosensor exhibited a linear response to GFAP concentrations from 0.6 to 100 ng/mL with a low LOD of 0.16 ng/mL. Its practical utility was confirmed by successfully determining the endogenous GFAP levels in CSF samples from patients diagnosed with encephalitis.

For Parkinson's disease diagnostics, a screen-printed electrochemical sensor modified with palladium nanoparticles (PdNPs) was developed for the detection of epinephrine and α -synuclein, two key biomarkers.³⁶⁵ A conductive ink based on carbon black in a poly(vinyl alcohol) (PVA) matrix was used to fabricate the three-electrode system. The deposition of PdNPs onto the electrode surface enhanced both electrical conductivity and reaction kinetics, with deposition parameters optimized using chemometric methods. The resulting sensor demonstrated a response to epinephrine in the range of 0.75 to 100 μM , with a LOD of 0.051 μM , showing promising results in synthetic CSF. Additionally, these disposable SPCEs were modified with specific antibodies for the EIS-based detection of α -synuclein. In phosphate buffer, this immunosensor

exhibited a linear response to α -synuclein concentrations from 1.5 to 15 $\mu\text{g mL}^{-1}$, with an LOD of 0.13 $\mu\text{g mL}^{-1}$.

An implantable sensor for measuring CSF dielectric properties as a potential biomarker for Alzheimer's disease contained a voltage-controlled oscillator and an implantable antenna designed for wideband operation with a low-quality factor (Q-factor) to maintain stability despite variations in the CSF.³⁶⁶ The implantable antenna transmitted signals that interact with CSF, and the received power level was measured externally. With this sensor, an increase in the loss tangent was correlated with the progression of Alzheimer's disease. Zarrin et al. developed an implantable sensor for monitoring CSF flow in patients with ventriculoperitoneal shunts.³⁶⁷ These shunts can divert CSF in conditions like hydrocephalus, but they often face high failure rates, and current clinical practices lack effective solutions for continuous, long-term monitoring of CSF flow rates. The sensor consisted of electrodes arranged in a cylindrical configuration surrounding the flow path. These electrodes generated an electric field across the CSF, inducing spatial variations in charge density within the CSF as it flowed through the shunts. The variations correlated with the flow rate, allowing the sensor to generate an electrical signal that corresponded to fluid movement. The recorded signals were then converted into readable flow data, enabling real-time monitoring of CSF dynamics. The ability to monitor CSF flow with high sensitivity was demonstrated, with an average error of 4.2% and only consuming 37.5 microjoules per flow measurement.³⁶⁷

A challenge hindering the long-term performance of implantable CSF sensors is biofouling, the nonspecific adsorption of biomolecules (primarily proteins like albumin, even at CSF's relatively low concentrations) followed by cellular adhesion (including microglia and astrocytes) onto the sensor.^{368,369} This process initiates a foreign body response, potentially leading to glial scar formation, which physically obstructs analyte diffusion to the sensing interface, consequently degrading signal sensitivity, accuracy, and overall device lifespan. To mitigate this, functional materials and surface engineering strategies are crucial. Passive approaches involve modifying the sensor interface with antifouling coatings, such as hydrophilic polymers like polyethylene glycol (PEG)³⁷⁰ or highly effective zwitterionic materials³⁷¹ (e.g., poly(sulfobetaine methacrylate)),³⁷² which creates a hydration layer to repel protein adsorption. Another avenue utilizes nanomaterials and fabrication techniques to create specific micro/nanotopographies on the sensor surface, bioinspired by structures like shark skin,³⁷³ to physically deter cellular attachment. Furthermore, active antifouling methods are being explored with electrical potentials applied to repel charged biomolecules (electrophoresis/dielectrophoresis).^{374,375} Achieving robust, long-term bioinertness in the complex CSF environment likely requires combinatorial approaches integrating these materials and surface modification techniques.

In addition to the challenges related to biointeractions, another hurdle lies in the disparity in mechanical properties between conventional rigid sensor materials (often with Young's moduli in the GPa range) and the extremely soft neural tissue surrounding CSF pathways (typically <1 kPa).^{376,377} This mechanical mismatch can cause chronic inflammation, tissue damage due to micromotion between the device and tissue, and impede seamless integration. Consequently, developing flexible and stretchable sensors using

functional materials is essential for long-term *in vivo* stability and biocompatibility within the CNS. Materials such as elastomeric polymers (e.g., PDMS),^{378,379} flexible polyimides,³⁸⁰ or parylene coatings serve as substrates or encapsulation layers.^{381,382} Integrating nanomaterials like carbon nanotubes,³⁸³ graphene,³⁸⁴ or metallic nanowires³⁸⁵ into these polymers can create conductive composites for electrodes and interconnects that retain functionality under strain. Device architectures incorporating thin-film deposition techniques and stretchable designs, such as serpentine interconnects, allow the sensor to conform intimately to the brain or spinal cord contours, minimizing stress concentration.^{386,387} However, fabricating complex electrochemical or optical sensing elements onto these nontraditional platforms while ensuring robust electrical connections and long-term hermetic sealing against the corrosive CSF environment remains a significant engineering challenge requiring interdisciplinary innovations in materials science, microfabrication, and analytical device design.

Another challenge is providing a sustainable power source for these sensors. Batteries have limited lifespans, which require replacement surgeries. While wireless power transfer is promising, its efficiency can decrease with the implant depth, often necessitating continuous external power sources. Biofuel cells (BFCs) present an alternative by harvesting energy directly from the biological environment (see Section 10.3).³⁸⁸ These devices utilize biological catalysts (enzymes or microbes) and endogenous fuels present in body fluids, such as lactate found in CSF.³⁸⁹ A typical enzymatic BFC consists of an anode and a cathode within a miniaturized housing, where the potential difference drives electrical current.³⁹⁰ Implantable BFCs may enable self-powered, autonomous sensors operating continuously for extended periods. However, achieving sufficient power density to operate the sensor and associated wireless communication circuitry remains a bottleneck. Current research focuses on developing more stable enzymes, improving electron transfer efficiency, designing BFC structures that optimize fuel/oxidant access while minimizing fouling, and enhancing overall biocompatibility. Hybrid systems combining BFCs with small rechargeable batteries or capacitors could also help to buffer power fluctuations and ensure reliable operation. Overcoming these challenges is crucial for realizing truly autonomous, long-term implantable diagnostic, and monitoring systems for neurological disorders.

4.2. Gastric Fluid

Gastric fluid and gastrointestinal (GI) microbiota are essential for maintaining human health, with deviations from the regular levels of homeostasis being associated with multiple diseases. The gut microbiota is formed by trillions of microbial cells that produce harmful and beneficial metabolites contributing to the development of diseases or providing protection from them.^{391,392} Beneficial metabolites, such as short-chain fatty acids (SCFAs) and γ -aminobutyric acid (GABA), aid digestion and gut-brain communication.^{393,394} On the other hand, metabolites such as p-cresol-sulfate, indoxyl sulfate, deoxycholic acid, lithocholic acid, tyramine, trimethylamine-N-oxide, trimethylaminuria, and N-phenylacetylglutamine can be associated with several diseases (e.g., neurodegenerative and cardiovascular diseases).^{395,396} Therefore, close and constant monitoring of the gut microbiota (GM) and their metabolites can provide important health information and indicate possible predispositions to different disorders.³⁹⁷




Figure 15. Edible electrochemical biosensor for the determination of glucose in gastric and intestinal fluids. (A) Schematics showing the preparation and application of the proposed device. (B) Examples of chronoamperograms in gastric fluid. (C) The same curve in the intestinal fluid. Reproduced with permission from Montiel et al.⁴⁰³ © 2019 Springer Nature.

The gastric fluid represents a uniquely hostile milieu for sensor development. Its extreme acidity (typically pH 1–3), driven by hydrochloric acid secreted by parietal cells, coupled with the presence of potent proteolytic enzymes like pepsin, creates a highly corrosive chemical environment.^{398,399} Furthermore, the constant mechanical agitation from peristalsis imposes significant physical stress.⁴⁰⁰ Consequently, designing sensors—whether implantable for applications like post-operative monitoring or ingestible for diagnostic screening—capable of withstanding these conditions while delivering accurate, reliable measurements constitutes a major materials science and engineering challenge. To be effective, these sensors must ideally function reliably while either free-floating or anchored or adhered near the gastric mucosa—the tissue lining the stomach that is the source of key chemical constituents. This necessitates robust materials, stable sensing mechanisms, sophisticated encapsulation strategies, and intelligent device designs.

Current methods to detect and measure metabolites in the gastric fluid include, for example, high-performance liquid chromatography and gas chromatography, which are not suitable for real-time analysis.^{401,402} Electrochemical biosensors used for gastrointestinal fluids include those made with food-grade materials, for example, olive oil and activated charcoal, which can serve as a binder and a conductor, respectively.⁴⁰³ In the latter work, glucose was monitored within the GI tract using a biosensor containing the enzyme GOx within the olive oil matrix and activated charcoal (Figure 15). The biosensors had a stable performance in strongly acidic conditions (pH 1.5) over 90 min, with a linear response to glucose concentration from 2 to 10 mM with precise control over sensor activation.⁴⁰³ Another ingestible biosensor made with a probiotic *Escherichia coli* (*E. coli*) was used to characterize gut metabolites related to major depressive disorder.²⁶ Upon exposure of gut metabolites associated with depression, such as indole, butyrate, tetrahydrofolate, hydrogen peroxide, and tetrathionate, the biosensor activates a CRISPR-DNA writing system, which records the detection event at the molecular level. Genetic recording allows for the analysis of metabolite exposure by sequencing bacterial DNA after it has been

excreted. The biosensor was administered orally in capsule form, ensuring safe delivery of the engineered bacteria to the gastrointestinal tract. To achieve spatial specificity, the system utilized riboflavin (vitamin B2) as a marker as its concentration increases in the large intestine. The *E. coli* biosensor was equipped with a riboflavin-responsive element that activates the detection mechanism specifically in the large intestine, thereby providing both temporal and spatial control over metabolite monitoring. With this biosensor, metabolite concentrations could be determined in a time- and dose-dependent manner.²⁶

Engineered bacteria are promising for sensing markers of stomach and bowel diseases, as they can be encapsulated within ingestible devices. This was done, for example, to detect colonic inflammation through luminescence and fluorescence outputs.³³⁵ Ingestible biosensors have also been developed to monitor gastrointestinal health. For instance, an ingestible microbioelectronic device was used for biomolecular detection *in situ* through luminescence readout electronics that communicate wirelessly with an external device. The device implanted in a porcine model could detect gastric bleeding by sensing levels of heme in real time.⁴⁰⁴

Gastric fluid pH, on the other hand, is also an important indicator of physiological status.⁴⁰⁵ The stomach's characteristic acidity is essential for digestion and pathogen defense. Significant deviations from this acidic range can signal important clinical events. For instance, a rise in pH following gastric surgery might indicate anastomotic leakage, where alkaline fluids from the intestine infiltrate the surgical site.^{354,406} Conversely, abnormally high pH could suggest impaired acid secretion (achlorhydria), linked to conditions like pernicious anemia or chronic gastritis.^{407,408} For temporary monitoring, such as during postoperative recovery periods (typically weeks), bioresorbable, wireless, passive sensors offer a particularly compelling approach. These devices eliminate the need for onboard batteries and avoid the requirement for subsequent surgical retrieval, degrading harmlessly over a predetermined time frame.

A prominent strategy for passive pH sensing utilizes inductor-capacitor (LC) resonant circuits.⁴⁰⁹ In this design,

the sensor itself contains no power source; instead, it relies on inductive coupling with an external reader device. The external reader generates a time-varying magnetic field that energizes the sensor's LC circuit, causing it to resonate at a specific frequency. By sweeping the reader's frequency and detecting the corresponding change in impedance reflected by the sensor, this resonant frequency can be determined from outside the body. Implementations often engineer the inductive component (L) to be pH-sensitive, as in a patterned zinc (Zn) foil structured into a serpentine geometry to accommodate deformation. This inductor was then encapsulated within a pH-responsive hydrogel. For example, a copolymer hydrogel like poly[2-(diisopropylamino)ethyl methacrylate]-*co*-poly(ethylene glycol)diacrylate (PDPAEMA-*co*-PEGDA) served this purpose well since it contained tertiary amine groups with a pK_a around 6.3. In the highly acidic environment ($pH < pK_a$), these amines become protonated, acquiring positive charges. The resulting electrostatic repulsion between these charges within the polymer network, combined with increased osmotic pressure, causes the hydrogel to swell. This swelling mechanically deforms the embedded Zn inductor, increasing its physical dimensions and, consequently, its inductance (L). Since the resonant frequency of an LC circuit can be affected by changes in inductance and capacitance, in general, the increase in L leads to a decrease in the resonant frequency. Therefore, a measurable change in resonant frequency provides a wireless, passive signal indicating a shift toward less acidic (or more alkaline) conditions. The selection of materials is paramount: zinc provides conductivity for the inductor and exhibits well-characterized bioresorption kinetics, while bioresorbable polymers like poly(lactic-*co*-glycolic acid) (PLGA) are used as dielectric materials for the capacitor (C). Meticulous encapsulation, often using biocompatible waxes or other polymers, is crucial to protect sensitive components, especially the capacitor, from premature degradation or signal drift due to biofluid permeation. This ensures that frequency shifts during the intended operational lifespan reliably reflect pH-induced inductance changes. Similar challenges related to material stability and response specificity in harsh biological fluids are encountered in sensor development for other environments, such as the urinary tract, where mineral encrustation poses significant problems.

Another important application is the simultaneous detection of iron species (Fe(II)/Fe(III)) in gastric fluid. While the small intestine is the primary site for iron absorption, understanding iron's chemical state (speciation) within the stomach is vital for assessing the bioavailability of dietary iron and the efficacy of oral iron supplements, as Fe(II) is generally more readily absorbed than Fe(III). Electrochemical sensors using nanocomposite-modified electrodes can achieve this simultaneous quantification. For instance, platforms incorporating nitrogen-doped carbon quantum dots (N-CQDs), AgNPs, and β -cyclodextrin (β -CD) demonstrated this potential.⁴¹⁰ In such systems, N-CQDs provide high electrical conductivity and electrocatalytic surfaces, enhancing signal transduction; AgNPs further improve electron transfer kinetics and increase the active surface area available for reaction; and β -cyclodextrin acts as a molecular recognition element, utilizing its hydrophobic inner cavity to selectively bind iron ions, thereby enhancing sensitivity and selectivity against potential interferents in the gastric matrix. Voltammetric techniques like DPV or SWV are typically employed because

they can resolve the distinct electrochemical signals (peak potentials) corresponding to the oxidation or reduction of Fe(II) and Fe(III). The ability to differentiate these species offers a more nuanced understanding of gastric iron chemistry, informing diagnostics for iron-related disorders and guiding the design of more effective supplements.

A critical area is real-time gas monitoring, particularly for hydrogen sulfide (H_2S), which is relevant in gastrointestinal homeostasis, including modulating mucosal defense mechanisms, regulating inflammation, and influencing gut motility. Aberrant H_2S levels have been implicated in gastritis and inflammatory bowel disease. Miniaturized, ingestible electrochemical sensors integrated into swallowable capsules allow for real-time H_2S monitoring directly within the GI tract.⁴¹¹ These sensors often employ working electrodes modified with materials selected for high selectivity and catalytic activity toward H_2S oxidation or reduction. For example, a gold (Au) electrode modified with a Nafion solid-polymer electrolyte layer was selective for H_2S . Integrated within a wirelessly communicating capsule, such sensors demonstrated linear responses to H_2S concentrations relevant to both physiological and pathological states (e.g., 0.21–4.5 ppm). Significant engineering challenges include ensuring efficient diffusion of the target gas to the sensing surface within the confined space of an encapsulated device, maintaining long-term sensor stability and calibration *in vivo*, and adequately addressing the power budget and data transmission requirements for ingestible electronic systems.

The adaptability of electrochemical strategies extends to other relevant analytes. Examples include detecting metabisulfite,⁴¹² a potentially sensitizing food additive whose presence and concentration in the stomach could be clinically relevant, or monitoring key ions like sodium (Na^+), potassium (K^+), and chloride (Cl^-) that influence osmotic balance and gastric secretions.⁴¹³ Furthermore, research is exploring alternative sensor form factors, such as flexible, thread-based electrochemical sensors.⁴¹⁴ These formats offer potential advantages in conforming intimately to the gastric mucosa, enabling more localized measurements at the tissue interface.

It has already been mentioned that any implantable or ingestible gastric sensor must possess exceptional robustness to survive in the highly aggressive gastric environment. Devices must reliably function despite exposure to extreme pH levels, which can plummet to as low as 1.5 during fasting and rise toward neutral after meals, creating large, rapid fluctuations. Furthermore, they must withstand potent enzymatic activity, particularly from pepsin, which actively digests proteins, potentially degrading sensor components or protective layers. Continuous mechanical stresses from gastric motility, including grinding and mixing motions (peristalsis), also impose significant physical demands. Consequently, robust protective coatings or encapsulation layers (e.g., using materials such as medical-grade silicones, parylene, or specialized polymers) are indispensable. These protective barriers must be impermeable enough to prevent damage while still allowing for controlled and predictable diffusion of the target analyte (like glucose, pH ions, or specific biomarkers) to the active sensing element.

Gastric sensors must also be compatible with biological tissues. All materials intended for contact with the gastric mucosa or internal environment must meet compatibility standards (e.g., ISO 10993)⁴¹⁵ to be nontoxic, nonimmunogenic, and elicit only a minimal inflammatory response. Failure to do so can lead to tissue damage, chronic inflammation, or

rejection. Also, a closely related and persistent challenge is biofouling. This process involves the rapid adsorption of biomolecules, primarily proteins and mucus components, onto the sensor surface immediately upon exposure to gastric fluid. This initial layer is often followed by the adhesion of cellular debris and bacteria, forming a complex biofilm. This biofouling layer presents a major hurdle because it can physically obstruct the path of the target analyte to the sensor, passivate reactive electrode surfaces, thereby reducing sensitivity, and alter the sensor's calibration and overall performance, ultimately limiting its functional lifespan. Therefore, effective antifouling strategies are required. This often involves surface modifications, such as grafting hydrophilic polymers like PEG⁴¹⁶ or zwitterionic materials,⁴¹⁷ applying specialized hydrogel coatings,⁴¹⁸ or designing micro/nanostructured surfaces that discourage adhesion,^{419,420} all aimed at maintaining a clean and accessible sensor interface for reliable long-term operation.

Further insights can be drawn from advancements like an ingestible robotic interface device designed for chronic electrostimulation.⁴²¹ This ingestible robotic interface device is released when its gelatin capsule dissolves in the stomach. Once deployed, a built-in magnet enables precise navigation to the target location using an external magnetic field. After reaching the desired site, its thin, flexible structure—made of circuits embedded within a soft elastomer—conforms closely to the stomach lining. A hydrogel adhesive activates upon contact with moisture, securing a strong and long-lasting attachment for several days. Finally, the device operates without a battery, receiving power and programmable stimulation commands wirelessly through near-field inductive coupling at a frequency of approximately 13.56 MHz. Adapting such principles—specifically, flexible conformal designs, robust bioadhesion for stable positioning and consistent fluid access, guided navigation, and wireless battery-free powering—offers a pathway to gastric fluid sensors capable of reliable, long-term *in situ* monitoring of crucial biomarkers within the challenging gastric lumen.

Sustained operation and data retrieval indeed pose engineering challenges for gastric fluid sensors (see Section 10 for more details on energy harvesting). Passive sensors, which require no internal battery and are powered wirelessly by an external reader (often via inductive coupling or NFC), offer simplicity but are typically limited by a very short communication range (millimeters to centimeters) and high sensitivity to the precise alignment between the sensor and reader. Active sensors, incorporating miniaturized batteries, possess greater operational flexibility and potentially longer communication range. However, they necessitate highly reliable, energy-dense, and safe power sources encased to prevent the leakage of potentially toxic contents into the body. Reliable wireless data transmission through the intervening biological tissues is also crucial. This requires power-efficient antenna designs and robust communication protocols optimized for low-power consumption and tissue penetration, such as the Medical Implant Communication Service (MICS) band,⁴²² Bluetooth Low Energy (BLE),⁴²³ or NFC,⁴²⁴ each with different trade-offs regarding range, data rate, and power usage (see Section 9 for more details on signal acquisition). An emerging, though still largely experimental, requirement being explored is the potential for energy harvesting directly from the gastric environment, for instance, using tiny generators harnessing mechanical energy from stomach movements, which could potentially eliminate battery limitations in the future.⁴²⁵

4.3. Interstitial Fluid

Interstitial fluid (ISF) is a major biological fluid that contains a rich and diverse variety of biomarkers and analytes. ISF is also one of the most epidermally accessible biofluids with a similar composition to blood, making it an ideal candidate for the detection of biomarkers.^{426–428} While blood remains the primary biofluid for the measurement of numerous biomarkers, its collection requires the use of invasive hypodermic needles, necessitating professional expertise for safe retrieval.^{429,430} ISF shows distinct advantages including greater volume, lower biointerference, and high clinical relevance. To fully utilize ISF sensors, researchers have developed biosensors capable of extracting biomarker data in a minimally invasive manner.⁴³¹

One important application is targeted to diabetes patients,^{432,433} to avoid needle injections or finger-pricking.^{434,435} Kim et al. developed a subcutaneously implantable electrochemical biosensor that detects minute changes in the dielectric permittivity, which are then correlated to changes in glucose levels.⁴³⁶ The device consists of a circuit that responds to glucose-induced dielectric changes by altering its resonant frequency, which is then wirelessly transmitted to an external reader. *In vitro* results demonstrated the ability to measure glucose concentrations from 50 to 500 mg/dL with a sensitivity of 2.3 kHz per mg/dL.⁴³⁶ Another subcutaneous implantable biosensor utilized electrical impedance spectroscopy to analyze the electrical properties of ISF, which depend on the glucose level.⁴³¹ The data from this implantable bioimpedance spectroscope are transmitted to an external reader for analysis. *In vivo* experiments were conducted on pigs to evaluate the sensor's performance. After a 14-day postsurgical recovery period, intravenous glucose tolerance tests were performed, during which 236 subcutaneous bioimpedance measurements were collected. Blood glucose levels ranged from 77.4 to 523.8 mg/dL during these tests. The study found a strong correlation between impedance measurements and blood glucose levels.⁴³¹

Continuous glucose monitoring with wearable devices via measurements collected in the subcutaneous ISF is revolutionizing diabetes management.⁴³⁷ Expanding such technologies beyond glucose to a broader range of clinically important molecules, such as biomarkers (e.g., creatinine for renal function or interleukin-6 for the detection of sepsis) and narrow-therapeutic-index drugs (e.g., toxic, “last-line-of-defense” antibiotics), could transform the personalization of clinical care. For example, the ability to measure drug concentrations *in situ* in the blood or ISF could enhance the accuracy with which drug doses are personalized to each patient. A problem with extending the technology underlying traditional glucose monitoring to other molecular analytes is that it relies on the transformation of glucose into an easily detectable, redox-active product via the action of a redox-active enzyme. Thus, while this approach has been expanded to the *in vivo* measurement of a limited number of other molecules, including lactate (using lactate dehydrogenase),⁴³⁸ acetyl choline (choline oxidase),⁴³⁹ and glutamate (glutamate oxidase),⁴⁴⁰ the absence of suitable redox transforming enzymes renders the vast majority of clinically relevant molecular species impossible to capture using such sensors. For example, while one can broaden enzymatic electrochemical sensors to a much wider range of molecular analytes by using potentiometric approaches (i.e., “ion-selective electrodes”) to detect the protons or other ionic species produced when hydrolytic enzymes degrade their targets, this approach has

seen almost no successful application to *in vivo* measurements. Specifically, the first such potentiometric, enzymatic sensor, which used urease to target urea, was reported more than 55 years ago,⁴⁴¹ but only a single potentiometric enzymatic sensor has ever been reported to work *in vivo*—a penicillinase-based penicillin sensor⁴⁴²—in work that has not been expanded upon in the 5 years since it was originally published.

Given the difficulty of expanding *in vivo* enzymatic sensors to any more than a handful of analytes, in recent decades, the biosensor field has expended significant effort on affinity-based sensors. That is, to say, sensors that rely on the recognition by a receptor rather than employing the chemical transformation of their analytes. Affinity-based sensors, however, face signal transduction challenges, as most biomolecular receptors (e.g., antibodies) do not produce a measurable signal upon binding their targets. Hence, the development of affinity-based sensors that support measurements in blood or ISF *in vitro*, much less *in vivo* (where, for example, exogenous reagents cannot be easily added), has proven a significant hurdle. For example, despite an enormous volume of literature describing sensors that detect binding to surface-attached receptors via changes in the optical properties of the surface (e.g., surface plasmon resonance), the mass or steric bulk loaded on the surface (e.g., microcantilevers and quartz crystal microbalances), and the electric field at the surface (e.g., field-effect transistors), only a vanishingly small number of such sensors have been reported to work in unprocessed bodily fluids *in vitro*, much less *in vivo*^{443–445} (for an exceedingly rare counter example, see ref 446). The main problem is, as previously mentioned in different sections of this review article, biofouling: the nonspecific adsorption that inevitably occurs when an artificial surface is placed in unprocessed bodily fluids also changes the optical properties, mass, and electric field on the sensor surface, producing a signal that is indistinguishable from that caused by analyte binding.

How, then, do we solve the “signal transduction problem” that has prevented affinity-based sensor approaches from deployment in blood and ISF? A glance at nature suggests an alternative to this quandary. Specifically, naturally occurring bioreceptors designed by evolution such that they undergo a significant physical change, such as a conformational change or a change in oligomerization state, upon target binding, which is used to transduce the binding event into an easily measurable output that is not “spoofed” by nonspecific binding.⁴⁴⁷ The question, then, is how to “re-engineer” normally nonresponsive biomolecules into receptors that undergo such a binding-induced physical change. To date, two approaches to this end have been shown to work *in vivo*.

The most well-explored means of re-engineering biomolecules to undergo a binding-induced conformational change is to destabilize them, so that they are unfolded in the absence of their target analyte but then fold upon binding.⁴⁴⁸ This couples recognition to an enormous change in polymer physics which, in turn, can be used to produce optical or electrochemical outputs. Because of the relative paucity of electroactive species (at any reasonable potential) in biological fluids, the latter approach is particularly amenable to measurements performed in the living body. The most successful application of binding-induced conformational changes in *in vivo* biosensors so far has been the electrochemical aptamer-based (EAB) sensor platform. First introduced some 20 years ago,⁴⁴⁹ EAB sensors uniquely blend the broad applicability of affinity-based sensors with the selectivity required to perform useful *in vivo*

measurements. At their core, these sensors utilize aptamers that are re-engineered to undergo a significant binding-induced conformational change.⁴⁵⁰ They are modified with an electrochemically active “redox reporter,” and anchored onto the surface of an interrogating electrode. The structural shift they undergo upon target binding alters the rate of electron transfer from the redox reporter, producing an easily detectable electrochemical signal. Compatible with various electrochemical analysis techniques—including SWV, chronoamperometry, and impedance spectroscopy,^{451–454} EAB sensors allow for a continuous, real-time molecular monitoring *in situ* within the living body.

Since first adapted to *in vivo* deployments in 2017,⁴⁵² a half-dozen research groups have described EAB sensors supporting the real-time, *in vivo* measurement of about a dozen different analytes.^{455–465} These have been shown to support real-time, multihour measurements *in situ* in the body, with the large majority focusing on intravenous and intracranial sensors for measuring plasma drug and metabolite concentrations. This includes seconds and even subseconds resolution measurements of a number of drugs.^{455–458,464,466} and metabolites⁴⁶⁵ *in situ* in the jugular veins and antibiotics and psychoactive drugs in the brains^{463,467} of live rats. The real-time concentration information provided by these sensors also provides unique opportunities for the high-precision control over *in vivo* drug and metabolite concentrations. Specifically, EAB sensors have been shown to support closed-loop feedback control over *in vivo* molecular concentrations in the blood,^{452,455} brain,⁴⁶⁷ and solid peripheral tissues.⁴⁶⁸ This provides an unprecedented ability to maintain constant *in vivo* drug concentrations, or to follow predefined, time-varying concentration profiles.

Of course, intravenous (much less intracranial) sensors do not easily translate into convenient, easily wearable technologies. Thus, recent years have seen increasing focus on the adaptation of EAB sensors to measurements performed in the dermal or subcutaneous ISF. These measurements have been performed using submillimeter microneedles,⁴⁶⁰ hollow, fluid-⁴⁶⁹ or hydrogel-filled microneedles,⁴⁷⁰ and millimeter-scale wire sensors,^{468,471} with the hollow-needle approach even having been deployed on human subjects. A problem with the measurement of molecular analytes in the ISF (or, indeed, most bodily fluids) is that it is plasma drug and biomarker concentrations, and not their concentrations in the ISF, that are the basis of clinical decision-making. Hence, it is essential that we learn how best to estimate plasma concentrations using minimally invasive measurements performed in the dermal ISF.⁴⁷² Fortunately, preliminary efforts to this end suggest that, at the very least, the estimation of systemic (i.e., plasma) drug exposure (as “area under the curve” on time–concentration plots) from measurements collected in the ISF will prove straightforward. Specifically, for the two drugs investigated to date, tobramycin and doxorubicin, the correlation between exposure in the ISF and plasma exposure is predictively strong ($R^2 > 0.95$).^{460,471}

As a scientific tool, EAB sensors also provide opportunities to measure drug transport between bodily compartments with unprecedented ease and time resolution, with recent examples including the transport of drugs between the plasma and both the brain⁴⁶³ and the subcutaneous ISF.⁴⁶⁸ Likewise, by providing researchers with a tool to manipulate molecular concentrations in the body with high temporal and concentration resolution, the use of EAB sensor-informed




Figure 16. Fabrication methods for wearable devices. Microfluidics, Reproduced from Kim et al.¹³⁹ Available under a CC-BY 4.0 license. © 2022 John Wiley & Sons; Microneedles, Reproduced from Kim et al.⁴⁷⁹ © 2024 American Chemical Society; Smart contact lenses, Reprinted with permission from Kim et al.²³⁸ © 2021 Springer Nature; Mouthguards, Reprinted from Ichikawa et al.²³⁰ Available under a CC-BY 4.0 license. © 2024 MDPI; Wristband/Watches, Reprinted with permission from Zhao et al.⁴⁸⁰ © 2020 American Association for the Advancement of Science; Textile, Reprinted with permission from Fan et al.⁴⁸¹ © 2020 American Association for the Advancement of Science.

feedback control over *in vivo* molecular concentrations could likewise significantly impact biomedical research. For example, using EAB sensors to perform feedback control, researchers have eliminated subject-to-subject pharmacokinetic variability as a confounding variable and have even reproduced human pharmacokinetic profiles in rat animal studies.⁴⁶⁶

A second, recently developed alternative to EAB sensors, “molecular pendulum” sensors, may provide a simple yet powerful method to translate the binding of larger molecular analytes (i.e., proteins) into an electrochemical signal via changes in hydrodynamic drag.⁴⁷³ This approach builds on somewhat similar optical⁴⁷⁴ and electrochemical⁴⁷⁵ approaches. It employs a short, electrode-bound, double-stranded DNA element to which a receptor, such as an antibody, nanobody, or aptamer, and a redox reporter have been covalently attached at the distal end. When a positive potential is applied to the electrode, the negatively charged DNA is attracted to its surface, with the rate of the resulting motion depending monotonically on the hydrodynamics of the pendulum. That is, signal transduction in this sensor architecture relies on the increase in drag that occurs when target binding increases the hydrodynamic radius of the pendulum. This converts binding into a change in electron transfer rate that is easily measurable using chronoamperometry, at least for higher molecular weight analytes. The approach has not been applied to lower molecular weight analytes, as these do not significantly alter the hydrodynamic radius of the pendulum. Still, by swapping the target-recognizing receptor, the approach has been used for the detection of 10 different proteins spanning a wide range of sizes and charges. Preliminary results suggest that this sensor architecture supports measurements *in situ* in the body. Specifically, the approach has been shown to work for the detection of the cytokines interleukin-6 and tumor necrosis factor α and in the subcutaneous interstitial fluid of a live rat with 40 min time resolution.⁴⁷⁶

5. WEARABLE DEVICE FORMATS

The production of wearable sensors and devices depends on a variety of engineering aspects, such as materials, design, and fabrication processes.^{477,478} Herein, we discuss the fabrication of various wearable devices to adjust to body parts, where they can acquire biosignals and monitor health. We focus on the fabrication methods to avoid being overly repetitive as several examples of the use of the different formats are given in other sections. Figure 16 illustrates some of the main fabrication methods and devices that will be discussed in the following subsections.

5.1. Microfluidics

Microfluidics is the science of precisely manipulating very small fluid volumes (nL to pL) using microscale systems. In wearable bioanalysis, microfluidic platforms serve for collecting fluid-type samples (e.g., sweat, blood, tears, saliva) using capillary forces to analyze biochemical markers in real time. In this section, sweat-based devices will be used as an example of different microfluidic applications due to the numerous reports and their variations. They are mostly produced with soft lithography utilizing polymeric substrates, such as PDMS, due to their relatively low cost, rapid prototyping, and compatibility with soft and flexible substrates.^{482–484} This method involves the patterning of a master mold, followed by polymer casting and bonding to create microchannels.^{139,485,486} Song et al. developed a microfluidic sweat sensor patch on a flexible printed circuit board platform that can monitor sweat biomarkers (e.g., pH and Na^+).⁴⁸⁷ To improve scalability, laser engraving of patterns onto flexible substrates has been employed in ambient conditions.^{126,488,489} These manufacturing techniques enable the formation of microfluidic channels directly, thereby reducing fabrication complexity. Additionally, 3D printing has been exploited to fabricate complex and customized microfluidic designs, offering precise control over geometry and feature size.^{488,489}

Microfluidics can be used, for example, to address the challenge of low sweat rates coupled with the low density of

sweat glands in developing wearable sweat sensors. The human skin has the highest density of sweat glands in palms and soles (~ 600 glands/cm²), but the average sweat gland density is only ~ 150 sweat glands/cm².²⁴ Moreover, each sweat gland produces anywhere between 1 and 20 nL/min sweat with a secretory pressure between 3 and 70 kPa, depending on an individual's age, diet, environment, and physiological state.^{490,491} Microfluidics also enables fluid handling via active/passive valves and pumps, stabilizes the naturally occurring pulsating sweat flow to a more laminar flow for reliable sensor signals, and supports inclusion of sample pretreatment capabilities. It can also be used to filter out undesired constituents that can negatively affect sensor response. For example, naturally produced sweat contains a small amount of sebum, an oily substance produced by the sebaceous glands present in the hair follicles. The oily constituents of the sebum can deposit onto sensor surfaces and cause signal drift and biofouling. Microfluidic devices have been developed to address this issue which incorporate porous polypropylene-based filters that absorb sebum before the sweat is introduced to the sensors.⁴⁹²

Key challenges for wearable microfluidics include ensuring mechanical durability under continuous wear and achieving reliable fluid flow control in ultrathin devices. A range of materials has been utilized in wearable microfluidics, including silicones,^{493–498} paper,^{499–502} fabrics,^{503,504} and flexible plastics⁵⁰⁵ fabricated using techniques such as soft lithography,⁴⁹³ laser cutting,^{505–507} and 3D printing.^{128,508–510} Wearable microfluidics have been employed to house a rich variety of electrochemical,^{508,509,511–514} colorimetric,^{27,132,135,146,515,516} fluorometric,^{123,158–160} and surface-enhanced Raman spectroscopic^{517–522} based transducers to detect biomarkers such as metabolites,^{494,520,523,524} electrolytes,^{118,492,500,525,526} proteins,^{102,527–529} and minerals.^{176,507,530,531} The simplest version of wearable microfluidics involves an inlet for the sweat to enter, an outlet to expel old sweat, a sensing chamber where the sensor is embedded, and a connective channel that links all the above three regions of the device. While simple in design, care must be taken to ensure that the dimensions and surface properties of the microfluidics allow rapid capture and transport of the sweat from the skin to the sensor and then ultimately out of the device. Here, advanced computer simulation models provide researchers with insights into designing the architecture and selecting the right materials for developing microfluidics that meet their needs.⁵³²

Simple microfluidic designs are mostly suitable for applications that involve high sweat rates, for example, in the case of active users. However, for the majority of applications, users will have low sweat rates. A strategy to solve this involves the use of locally administered sweat-inducing drugs^{197,533,534} or raising the local skin temperature via Joule heating to induce sweat synthesis.⁵²⁵ While such approaches facilitate a higher sweat rate, the reliance on external factors to induce sweating artificially and their effect on modulating the natural sweat composition remain issues to investigate. The efficient capture of naturally produced sweat may be done with multiple inlets^{494,524,525,527,535} and materials with capillary action (e.g., paper). One such embodiment involves a paper microfluidic array with 96 inlets distributed across 8 interconnected input channels around the collection zone for efficient collection of sweat.⁵⁰¹ The multiple circular and semicircular inlets, inlet density, number and length of inlet channels, and dimensions

of the harvesting zone were rationally optimized based on probability analysis to capture 100% of the sweat generated in the skin region where the device is applied.

Researchers are also taking inspiration from nature to reach an efficient sweat harvesting.⁵¹¹ For example, a device with open microfluidic structures inspired on tree frog toe pads was used for rapid collection of sweat at low flow rates (~ 0.1 $\mu\text{L}/\text{min}\cdot\text{cm}^2$).⁵³⁶ Taking cues from the hexagonal epithelial cells on the toe pads of the tree frog *Rhacophorus pardalis*, the team designed microfluidics with rigid hexagonal walls covered with soft elastomeric materials. The rigid walls helped maintain the open microfluidics structure, while the soft elastomeric covering enabled conformal attachment of the device to the human skin. The strong capillary force exerted by the microfluidic design allowed the device to capture sweat and perform sweat analysis with >75 nL of sweat collection in 45 s. Another nature-inspired microfluidic device was based on Ginkgo biloba veins to include microfluidic channels with a 5° wedge angle that facilitates 40% higher flow rate than traditional channels with a 0° wedge angle.⁵²³ The unique Ginkgo biloba mimicking microchannels achieve higher fluid transport due to the elevated capillary forces created by the geometry of the channels. Rapid and directional collection of sweat was also obtained with Janus-based microfluidic membranes.^{517,537–539} A typical approach for developing Janus-based membranes involves creating a mesh-like hydrophobic layer in direct contact with a highly mesh-like hydrophilic layer. The hydrophobic layer is mounted on the skin, while the sensor is interfaced with the hydrophilic side. The two layers are usually made via electrospinning, which yields large sheets with reproducible pore size and density. Computer simulations provide a valuable understanding of the working principle of these hybrid membranes.⁵¹⁷ Sweat droplet contacting the hydrophobic layer experiences opposing hydrostatic pressure (P_H) and Laplace pressure (P_L). When P_H exceeds P_L , the sweat droplet penetrates through the hydrophobic layer and comes in contact with the hydrophilic layer, where capillary pressure (P_C) dominates and serves as the driving force to rapidly pull the remainder of the fluid into the hydrophilic region and enable lateral spreading of the liquid within the hydrophilic layer. Once the sweat spreads across the hydrophilic layer, P_H decreases drastically, thus disincentivizing sweat from re-entering the hydrophobic layer. Janus-type architectures can also be fashioned in the form of yarns for rapid, pumpless capture and transport of sweat.⁵⁴⁰ Using osmotic pumps for capturing sweat represents yet another path for uses wherein the sweat rate is low.^{541,542}

In addition to enabling efficient sweat collection, wearable microfluidics also aid in advanced fluid handling, which is necessary for various sensing applications. In this context, innovations in microfluidic valves and pumps come into play. Capillary burst valves are among the most explored valve designs.^{494,543–546} The burst pressure (BP) for valves located within rectangular channels is based on the Young–Laplace equation

$$\text{BP} = -2\sigma \left[\frac{\cos \theta_I^*}{b} + \frac{\cos \theta_A}{h} \right]$$

where σ is the surface tension of the liquid, θ_A is the contact angle of the channel, θ_I^* is the $\min[\theta_A + \beta; 180^\circ]$, β is the diverging angle of the channel, and b and h are the width and height of the diverging section, respectively. The sweat can

open these valves and travel forward only when the sweat pressure is higher than BP. Such valves thus allow a simple geometry-guided passive valving methodology for controlling the direction of the sweat flow within the microfluidics. Selectively patterned hydrophobic and hydrophilic regions within the microfluidics can also be implemented as valves.^{506,547,548} For example, in one embodiment, hydrophobically patterned channels route incoming sweat toward specific sensing regions within the device.⁵⁴⁷ When these regions are filled, the sweat is directed toward a superabsorbent polymer-based gate located at the entrance of the sensing region, which rapidly expands and closes the sensing region from accepting more sweat. The sweat is then forced to travel to subsequent sensing regions where the entire process of sensing region filling and thereafter superabsorbent polymer-based gate closing occurs, thus allowing a time-sequenced analysis of sweat. Along similar lines, time-sequenced sweat routing can also be achieved by strategically depositing a water-soluble polymer within the microfluidic channels that initially stop the sweat from flowing past it, but over time, the polymer dissolves, thus allowing sweat to flow across and travel to sensing regions that were previously inaccessible to it.⁵⁴⁹

While the above-mentioned approaches of passive valves serve in developing simple, low-cost devices with fluid routing capabilities, a major challenge involves their inability to completely isolate the liquids across the valves, which can result in sample intermixing and contamination. A recent work attempts to address this issue by demonstrating liquid-bridge-cutting valves.⁵⁵⁰ It builds on the concept of capillary burst valves and leverages the liquid bridge breakup principle that allows the device to separate liquids and fill the resulting gap with air. Each sensing region is loaded with filter paper placed at a small offset from a capillary burst valve, creating an air gap between the paper and the valve. When sweat arrives at this valve, the front-end bulges as pressure builds up until the liquid front touches the tip of the filter paper, forming a liquid bridge. At this point, the capillary forces within the paper draw in the sample until the entire sensing chamber is filled. Upon complete filling of the chamber, the liquid bridge breaks and recreates the small air pocket between the paper and the valve, which stops the flow of any new fluid into the filled chamber. Such air pocket-based designs can also be used to control the flow rate, which can be helpful in obtaining more reliable data.⁵⁵⁰ Tesla valves represent another type of passive valve design for fluid flow control.^{551,552} Variations of Tesla valves that include single-stage, multistage, and derivatives of the original design can be employed to achieve unidirectional flow of liquid.⁵⁵³

Active valves embedded in wearable microfluidics that are activated by heat,^{554,555} electrical,⁵⁵⁶ and electromagnetic⁵⁵⁷ triggers have also been developed for more stringent control of sweat flow within the device. One example of a heat-triggered valve involves the use of proprietary thermoplastic microspheres encapsulating a gas (Expancel).⁵⁵⁵ The encapsulated gas reversibly expands when heated. The device had affixed microspheres to a thin layer of silicone and embedded at specific regions within the microfluidic channels. Thin-film heaters, powered by a wearable flexible electronic module and located beneath the microsphere-functionalized silicone layer, provide the heat necessary to expand the microspheres, which ultimately block the channel and stop the fluid flow. Electrical trigger-based electrowetting valves can also be employed for controlling sweat flow.⁵⁵⁶ A device exemplifying this approach

was made with inject-printed silver electrodes coated with a self-assembled monolayer of perfluorodecanethiol embedded at the floor of the microfluidic channels. This monolayer renders the electrodes hydrophobic, which hampers sweat from flowing over it, thus acting as a valve. However, applying 4 V for ~ 17 s reversibly transforms the hydrophobic coating into a hydrophilic layer, thus enabling the liquid to cross. Such valves can be used for retaining fluids for at least 9 h.⁵⁵⁶

Strategies for incorporating micropumps within wearable microfluidics have also been investigated. Passive capillary pumps may contain microposts within the microfluidic chambers.⁴⁹⁴ They can be integrated with inlet channels that possess continuously tapering cross-sectional areas to augment the sweat collection process.⁵⁵⁸ Other strategies include active mechanisms, as with finger-actuated pumps in which paper or silicone-based pull tabs attached to the soft microfluidics allow the user to exert suction forces to drive the collected sweat sample to specific regions.^{183,559,560} Piezoelectric materials have also been used in wearable pumps⁵⁶¹ by incorporating aluminum oxide/carbon nanotubes core-shell nanofillers within a ferroelectric polymer. The polymer composite was sandwiched between two electrodes that were biased at 160 V and were largely deformed (bending angle: 74°), thus enabling the system to behave like a piezoelectric pump. Another strategy involves a portable electrohydrodynamic pump ($10 \times 2 \times 1.05$ cm).⁵⁶² It includes three main components: an interdigitated electrode system, flow channels, and an insulating fluid. The electrical field applied across the electrodes drives the fluid forward. The device utilizes the insulating fluid due to its low conductivity (10^{-9} S/m), high dielectric withstand voltage (5–6 kV at a gap: 0.5 mm), and a dielectric constant of 6.1. Moreover, it has a boiling point of 76°C , high thermal conductivity, ultralow toxicity, zero ozone depletion potential, zero flash point, and nonflammability. To reduce the size of the electrohydrodynamic pumping system, triboelectric self-powered devices can be used.⁵⁶³ The device includes a rotatory triboelectric nanogenerator that collects ambient energy and converts it into a high-voltage power source for powering an electrohydrodynamic pump that can generate a maximum pressure of 4.49 kPa and a maximum flow rate of 502 mL/min.

Most chemical sensors, especially those relying on enzymes, aptamers, and antibodies, require specific incubation periods, washing steps, and multistep detection protocols. Wearable microfluidics are not designed to achieve these capabilities. To address this issue, a new class of wearable microfluidics that offers on-demand sensing has been reported.^{564,565} These devices employ the above-mentioned pumps and valves to introduce captured sweat to sensors at user-defined time points, thus allowing complete control over the incubation period. One of the earliest examples had a series of finger-actuated pumps and capillary burst valves strategically located within the microfluidic architecture. This allowed the user to initiate sensor reactions on-demand, thus opening the possibilities toward true lab-on-skin devices. Additional capabilities were introduced by embedding user-activated 3D-printed rotational routers that can direct collected sweat to different regions within the microfluidic patch for fluid handling and sensing applications.⁵⁶⁵

The many advances made in the field of wearable microfluidics as reported here are a testament to the collaborative efforts put together by a diverse set of researchers, including mechanical, chemical, and biomedical

engineers working closely with material scientists and clinicians. As articulated in this section, the field of wearable microfluidics has come a long way from simple designs reported less than a decade ago to advanced systems that can perform complex fluid handling in a miniaturized and low-power form factor, which is amenable to the unique needs of wearable sensing applications.

5.2. Microneedles

Microneedles (MNs) are needle-like vertical structures typically in the hundreds to thousands of micrometers range in height, with the aim to pierce tissues at defined depths. One of their biggest advantages when used in humans is that MNs application can be pain-free if their height is not enough to reach the nerves under the epidermis. Also, they provide a minimally invasive alternative for creating microchannels in the skin for sample collection or for the direct acquisition of electrical and chemical signals.⁵⁶⁶ Microneedles can be solid, coated, hollow, or dissolvable, depending on the desired application, being typically built using silicon, metals, ceramics, or polymers.⁵⁶⁷ Multiple construction techniques can be applied. For example, silicon (Si)-based microneedles are normally fabricated with photolithography and reactive ion etching (RIE), ensuring high mechanical strength and precise geometries.⁵⁶⁸ For flexible and biodegradable microneedles, polymers are processed using molding or 3D printing techniques.^{569–572} After defining 3D polymeric structures, thin metal layers can be deposited selectively on the microneedle tips through electroplating or laser micromachining, creating precise recording sites.⁵⁷³

A variety of applications exist for MNs. Solid ones can be used for drug administration if they are properly coated, whereas hollow MNs are applied for minimally invasive ISF low-volume sampling. Conductive MNs can be used for ISF sensing and biosensing inside skin. Biodegradable MNs can be designed to naturally dissolve and release chemical agents in the skin. Significant advancements have been made in ISF glucose sensing, particularly with commercial, needle-based continuous glucose monitoring devices that provide accurate, calibration-free, and real-time glucose measurements in subcutaneous tissue.⁵⁷⁴

Concerns regarding microneedles often include minimizing immune responses caused by their penetrating structures, which can lead to the inflammation or scarring of biological tissues. To mitigate these, strategies are used to reduce the elastic modulus of microneedles to levels comparable to those of biological tissues by using extremely soft materials.^{575,576} For example, Kim et al. utilized liquid metals (LMs) as microneedle materials sufficiently stiff to penetrate the skin when frozen and become softer upon penetration because of their low melting point, causing no significant immune responses on the skin after implantation.⁴⁷⁹

Despite their variety of possible uses, MNs are still not ubiquitous in medicine and other fields such as agriculture because of their cumbersome fabrication routes. When using photolithography, for example, clean rooms and multiple steps are required, resulting in high costs and low scalability. As previously mentioned, alternative methods, as additive manufacturing, have been implemented for MNs fabrication, but they are limited in terms of resolution, scalability, flexibility, and materials portfolio. Conductive MNs are challenging to fabricate with these techniques in a single step. Rosati et al.⁵⁶⁹ introduced a method to produce MNs in a

single step on a flexible plastic substrate. They used inkjet printing of AgNPs with a research-grade printer by controlling the kinetics of the curing of the AgNPs ink when in contact with the substrate. The MNs presented conductivity immediately after the fabrication process, which consisted of the layer-by-layer, precise deposition of picoliter drops of the AgNPs ink. Different shapes and sizes were tested, identifying the limits of the method (e.g., verticality of the structures, minimum base width, maximum height). The mechanical properties of the inkjet-printed real-time cured AgNP ink have been characterized, obtaining a Young's modulus and a hardness of 20 and 76% of those of bulk silver, respectively. The inkjet-printed MNs were tested as a proof-of-concept on plants, performing EIS measurements to characterize the plant hydration level. An increased signal-to-noise ratio and much lower impedance values were obtained in comparison with the same measurement performed by flat electrodes of the same material/size. This fabrication method is being further developed for use in patients, extending the portfolio of printing materials to gold nanoparticles and other nanomaterials, and implementing improvements to increase scalability and decrease the fabrication time per microneedle. A similar and even more scalable approach for the fabrication of fully-printed AgNP microneedles has been proposed by Baraniecki et al.,⁵⁷⁰ studying the use of aerosol jet printing, and validating them for the monitoring of total ion conductivity in plants.

The fundamental sensing mechanism integrated into microneedle tips often involves electrochemical detection, particularly enzymatic or affinity-based electrochemical biosensing methods. For enzymatic sensors, enzymes are immobilized on the microneedle surface, catalyzing biochemical reactions that produce electroactive intermediates measured through chronoamperometric or voltammetric techniques. In affinity-based biosensors, aptamers or antibodies anchored to microneedle facilitate selective binding of target molecules, generating measurable changes in electrical properties that correlate with analyte concentrations. This direct microneedle–ISF interface eliminates external fluid extraction or dilution steps, significantly enhancing sensing accuracy, reliability, and response time. Optical transduction is an alternative due to the possibilities of label-free and real-time detection along with its high sensitivity, specificity, affordability, and small size.⁵⁷⁷

Microneedle sensors have been applied in numerous health and clinical contexts, illustrating their versatility and potential. Enzyme-based microneedle glucose sensors integrate GOx onto microneedle tips, typically using drop-casting, electrodeposition or entrapping methods to immobilize the enzyme, catalyzing the oxidation of glucose from ISF and generating measurable current signals proportional to the biomarker concentration.^{578–580} Continuous glucose monitoring is one of the most advanced clinical applications, delivering precise and minimally invasive measurements, reducing patient discomfort compared to traditional finger-stick methods. Such devices are used in diabetes management, enabling real-time glycemic control.⁵⁸¹ Xiao et al. have expanded microneedle applications to monitor uric acid levels, using surface-enhanced Raman spectroscopy (SERS) integrated into MN arrays, with noninvasive assessment of health disorders. The 3D-printed microneedles are connected via a microfluidic channel to the SERS sensing chip. This sensor is based on nanogold arrays, which enhance the SERS signals, on a silicon wafer.⁵⁸²

The versatility of microneedle arrays further enables multiplexing by simultaneously monitoring different bio-




Figure 17. Ion-sensing microneedles. (A) Overview of the device, its wrist positioning, electrode modifications, and typical measurements. Optical and scanning electron microscopies from (B) Ca²⁺-, (C) K⁺-, and (D) Na⁺-sensing electrodes in different magnitudes. Reproduced from Huang et al.⁵⁸⁶ Available under a CC-BY 4.0 license. © 2023 Springer Nature.

markers. For instance, Tehrani et al. demonstrated integrated microneedle sensor systems capable of concurrently detecting glucose, lactate, or alcohol. These microneedles perform as electrodes and were fabricated using CNC milling and poly(methyl methacrylate) (PMMA) as a base material. They functionalized the electrodes by electrodepositing a conductive polymer (o-phenylenediamine) and drop-casting enzyme solutions (GOx, lactate oxidase (LOx), and AOX) in a chitosan matrix to enhance enzyme stability and sensor performance. The combination of a nonionic surfactant and PVC was used as a diffusion-controlled film in a microneedle system for real-time analysis in freely moving human subjects, highlighting practical clinical potential.⁵⁸³ Other multiplexed

microneedle arrays have been developed by Windmiller et al. for simultaneous monitoring of neurotransmitters like glutamate alongside glucose, offering valuable insights for neurological applications. They combine 3D-printed solid and hollow microneedles and employ electropolymeric entrapment of enzymes (glutamate oxidase and GOx) in poly(o-phenylenediamine) thin films within the space created between both microneedle structures to ensure enzyme stability and selective analyte detection.⁵⁸⁴ A noteworthy example is the work presented by Huang et al., who developed a wearable microneedle-based sensor system for continuous transdermal monitoring of critical physiological ions, such as calcium (Ca²⁺), potassium (K⁺), and sodium (Na⁺). This system




Figure 18. Overview of the smart contact lens for ophthalmic health monitoring. The digital image of the wireless SCL (in the center). Reproduced/adapted with permission from Song et al.⁵⁹⁵ © 2022 John Wiley & Sons. Visions, a neuromorphic bionic eye with filter-free color vision. Reproduced/adapted with permission from Long et al.⁶⁰³ © 2023 Springer Nature. Biomarker in tear. Reproduced/adapted with permission from Park et al.²³⁶ © 2024 Springer Nature. Physical signal. Reproduced/adapted with permission from Kim et al.²⁴² © 2022 Springer Nature. Eye tracking. Reproduced/adapted with permission from Zhu et al.⁶⁰⁴ © 2024 Springer Nature. Self-powered CL. Reproduced with permission from Liu et al.⁶⁰⁵ © 2024 Royal Society of Chemistry. Neuroprosthetic CL. Reproduced/adapted with permission from Liu et al.⁶⁰⁶ © 2024 Springer Nature.

integrates ion-selective microneedle electrodes into a three-dimensional array, employing a fabrication strategy involving the assembly of planar microneedle sheets produced through laser micromachining of stainless steel, followed by electrochemical deposition of a gold layer and a conductive polymer (PEDOT:PSS). Ion-selective membranes were then applied via dip-coating, creating ion-selective electrodes capable of potentiometric detection of the aforementioned ions (Figure 17).⁵⁸⁵

Therapeutic drug monitoring (TDM) is another area where microneedle sensors have been applied. Goud et al. illustrate the substantial clinical potential of wearable electrochemical microneedle sensors for continuous levodopa monitoring in Parkinson's disease management. The 3D-printed microneedle arrays were functionalized with enzymatic coatings via electrodeposition and incorporated two different electrochemical detection mechanisms (enzymatic-amperometric and nonenzymatic voltammetric detection), enabling continuous, real-time tracking of drug pharmacokinetics in ISF.⁵⁸⁷ Similar applications include microneedle-based sensors for real-time continuous monitoring of apomorphine in Parkinson's treatment, demonstrating high sensitivity, selectivity, and good operational stability. Hollow microneedles were applied as working electrodes, being packed with carbon paste modified by rhodium nanoparticles, and apomorphine was detected by measuring its electrochemical oxidation peak.⁵⁸⁸ In this same context, Lin et al. developed a wearable microneedle-based electrochemical aptamer biosensor patch. The microneedles were fabricated from clinical-grade acupuncture needles, modified through electrodeposition of gold nanoparticles to enhance aptamer immobilization and electrochemical sensitivity. This biosensor permitted continuous and

real-time monitoring of antibiotics, such as tobramycin and vancomycin.⁵⁸⁹

A wearable 3D-printed microneedle sensor was developed by Mishra et al.⁵⁹⁰ to detect organophosphate nerve agents. The hollow microneedles were packed with the corresponding electrode paste (carbon or Ag/AgCl ink) and functionalized with enzymes that generated an electrochemical signal. The same research group prepared similar microneedles but employed a computer numerical control (CNC) micromachining method. It facilitates continuous monitoring of opioids and nerve agents simultaneously, providing capabilities for emergency medical services and military personnel exposed to hazardous substances.⁵⁹¹ Dermatological and oncological monitoring further expands the microneedle applications. Using a similar protocol for microneedle preparation from the previous work, microneedle sensors developed by Ciui et al. were used for melanoma screening by detecting the tyrosinase enzyme directly from the skin and drop-casting its substrate (catechol) onto the surface. This method offers a minimally invasive, real-time diagnostic capability.⁵⁹²

Theranostic systems have also been developed with microneedle wearable sensors, combining real-time biomarker monitoring with precise drug delivery. For instance, Razzaghi et al. introduced a wearable theranostic microneedle system featuring 3D-printed hollow microneedles. The system integrates colorimetric sensors to monitor physiological parameters such as pH, glucose, and lactate, alongside an ultrasonic atomizer, enabling remote-controlled, on-demand drug delivery. The colorimetric sensors were integrated by drop-casting enzymes (GOx, lactate oxidase, and HRP) or chromogenic dye solution (for pH sensor) onto filter papers interfacing with microneedles. This combination demonstrated rapid, minimally invasive sampling and precise medication

administration, which is promising for chronic disease management and personalized remote healthcare.⁵⁹³

Several challenges must be addressed in microneedle-based wearable sensors, including ensuring consistent skin penetration, biofouling management, maintaining sensor accuracy under varying physiological conditions, and long-term biosensor stability. This requires integration of functionalized materials such as antifouling coatings and nanostructured surfaces and comprehensive clinical validation to establish accuracy and safety in diverse patient populations. Future developments will likely focus on enhancing sensor longevity and performance through improved material biocompatibility, incorporating self-powered systems, and integrating advanced wireless electronics. Such improvements will facilitate real-time remote health monitoring and fully autonomous operation, positioning microneedle-based wearable sensors at the forefront of personalized medicine, continuous health monitoring, and therapeutic management technologies.

5.3. Smart Contact Lenses (SCLs)

Smart contact lenses (SCLs) are wearable devices designed for varied functions, including the monitoring of physicochemical parameters (e.g., eye pressure) and biochemical species on tears. They have been fabricated with various methods that can integrate functional components with soft lens platforms.^{236,594,595} Photolithography, for example, is used for precise electrode patterning on the lens surface. However, its planar nature limits the lens functionality to 2D structures. Soft lithography with elastomeric stamps is an alternative to transfer functional components onto soft lens surfaces⁵⁹⁴ (e.g., for fabricating microfluidic components and sensor integration for real-time monitoring).⁵⁹⁶ For instance, Ku et al. integrated immunosensors into contact lenses with PDMS microfluidic channels to detect cortisol levels in tear fluid.⁵⁹⁷ Direct printing and inkjet printing are also used in SCL manufacturing.⁵⁹⁸ Kim et al. introduced a high-resolution 3D printing of LM to form stretchable free-standing 3D patterns of interconnects essential to electrically connect individual device components for their integration.²³⁸ Park et al. utilized inkjet printing for fabricating wirelessly chargeable 3D solid-state supercapacitors that serve as a power source for SCLs.⁵⁹⁹ The flexibility and precision offered by these methods enable the creation of custom molds and components on different substrates including flexible and stretchable materials. Electrospinning methods to produce ultrafine fibers from nanoparticles, nanorods, nanowires, nanotubes, and nanosheets in a polymer solution or melt can also be used to fabricate electrodes on the soft lens platform.⁶⁰⁰ Metal-based nanowires, nanofibers, and hybrid structures, such as silver nanowires (AgNWs) and silver nanofibers (AgNFs), have been researched for their enhanced electrical, mechanical, and optical properties in SCLs.

Various applications of SCLs are described in Section 3.3 when the detection and monitoring of biomarkers in tears were discussed. SCLs have also been utilized in other health-monitoring approaches, as depicted in Figure 18. They encompass six main aspects: vision repair, detection of biomarkers in tears, determination of physiological signals, eye movement tracking, power management of smart contact lenses, and neuroprosthetic contact lenses.

Human vision has basic features of a wide field of view, high resolution, and processing ability of visual information.⁶⁰¹ Image sensors constructed by a flexible photodetector array with the capability of photon detection have promoted imaging

technology, showing wide applications in the restoration of human visual function, medicine, optical communication, and the visual system of biorobotics. The photodetector can even achieve enhanced vision, like night vision, by adjusting the response spectral range from UV (ultraviolet, 10–300 nm), Vis (visible, 400–700 nm), IR (infrared, 700 nm–1 mm), to THz (Terahertz, 0.1 to 1 mm). The difficulty lies in the transformation of the planar photodetector to the hemispherical substrate with a diagonal visual field >100°, high resolution, in-device preprocessing signals, and optical adaptivity. Many attempts to find solutions have been made, as in Fan's research,⁶⁰² where a hemispherical electrochemical eye made with perovskite nanowires (NWs) served as an image sensor to mimic the photoreceptors on a human retina. A neuromorphic spherical bionic eye was endowed with color vision, in-device processing, and optical adaptivity.⁶⁰³ Figure 18 ("Vision") shows the magnified structures of the neuromorphic bionic eye, which includes a convex lens, electronic iris, artificial crystalline lens, artificial retina based on perovskite NWs self-driven image sensor array with color imaging ability, and artificial optical nerves.

The potential of eye-machine/computer interaction via SCLs requires efficient and wireless communication. Mao et al.⁶⁰⁴ proposed a frequency encoding strategy for *in situ* eye tracing and wireless eye-machine interaction. Four passive copper radio frequency (RF) tags with different working frequency coils were embedded into biocompatible contact lenses. The eye movement information was encoded in RF signals for detecting the movement and closure of the eye. Then, a portable sweeping-frequency reader was fixed on the glasses and opposite to the user's eyeball, collecting the tags' signal without embedding chips and batteries. Since the RF signals depend on the eye movements or positions, a high angular accuracy of eye tracking was obtained. By detecting and analyzing these frequency changes, the system (or machine) can determine the direction and degree of eye movement in real time, which provides multiple eye-machine interaction modes.

Energy storage devices can be integrated into SCLs for driving the sensing units employed in health monitoring.⁶⁰⁷ For instance, Pourshaban et al.⁶⁰⁸ reported an eye-tear activated Mg-Air battery driven by eye blinking motion. It contains 4 parts, which were the substrate, an Mg anode, a Pt cathode, and an eye-tear fluid as the electrolyte. When the eyes periodically blinked, the battery was activated to deliver a power density of 1.3 mW cm⁻² at a load of 740 Ω. However, the Mg-Air battery had low safety and poor stability. To meet the demands of wearable SCLs, supercapacitors are being considered owing to their high-power density, security, and stability. Park's group⁵⁹⁹ developed rechargeable solid-state supercapacitors directly printed on soft substrates. When connected to an antenna, rectifier, and light-emitting diode, they provided continuous power for SCLs. An eye-wearable Ti₃C₂T_x MXene-based supercapacitor was used for powering a sensor to measure the intraocular pressure, as shown in Figure 18 ("Self-powered CL").⁶⁰⁵ The supercapacitor delivers a high energy density of 10 mW h cm⁻², which ensures the stable work of the integrated sensor. For more information on self-powered devices, see Section 10.

Neuroprosthetic devices that rebuild the connection between the nervous system and organs are promising for disease monitoring and health management. For instance, a neuroprosthetic contact lens (NCL) was used for POC




Figure 19. Electrochemical biosensor embedded in a mouthguard. (A) Picture of the modified mouthguard, containing an integrated electrode, as well as schematics showing the layers of the working electrode and the enzymatic detection principle. (B) Chronoamperograms obtained in different lactate concentrations. Reproduced with permission from Kim et al.⁶¹⁶ © 2014 Royal Society of Chemistry.

monitoring and feedback of intraocular pressure, as shown schematically in Figure 18 (“Neuroprosthetic CL”).⁶⁰⁶ This NCL makes up for the lack of feedback on the abnormal intraocular pressure of the SCL and establishes the relationship between this abnormal pressure and the somatosensory cortex. The sensorimotor system consists of a pressure sensor, a temperature sensor, and a wireless transmission/display App. The NCL exhibited a high sensitivity of $12.52 \text{ mV mm Hg}^{-1}$ and excellent stability. *In vivo* experiments in rats proved that NCL can convert temperature-corrected changes in intraocular pressure into a pulse signal, which is transmitted to the somatosensory cortex through electric current, triggering graded motor cortex feedback. The corresponding signals generated in the motor cortex further activate the leg twitch responses, finally forming a closed loop of intraocular pressure signal generation-nerve-perception-motor activities. These NCL devices represent a solution for the early prevention and precision diagnosis/treatment of ophthalmic diseases.

Regarding the determination of biochemical biomarkers, Song et al. developed smart contact lenses for the real-time quantification of cholesterol by using electrochemical biosensors.⁶⁰⁹ The device is intended to monitor hyperlipidemia patients and requires only a smartphone to be operated, not obstructing vision and presenting good wearability. The electrodes were built through the combination of metal thin-film deposition and screen printing, and the working electrode was modified with Prussian blue and the cholesterol oxidase enzyme. The bioreceptor catalyzes the oxidation of cholesterol into H_2O_2 , which is in turn detected electrochemically. Stretchable antennas and integrated circuits for wireless communication were also integrated into the smart lenses, making it straightforward to use. *In vivo* experiments displayed a good correlation between cholesterol levels in tears and blood, showing their potential applicability in clinical settings.

5.4. Mouthguards

Since the 1960s, when Graf and Mühlemann developed a removable partial denture to monitor salivary fluoride ions, interest in intraoral wearable sensors has grown, driven by the potential of saliva as a noninvasive diagnostic medium.^{610,611} The emphasis on saliva arises from its capacity to transport molecules directly from the bloodstream via transcellular or paracellular pathways, effectively serving as a “mirror” of the body’s physiological state.^{610,612,613} These early devices

integrated oral sensors, miniature transmitters, and onboard power sources into removable dentures, enabling *in vivo* measurements while the denture was worn.⁶¹¹ Despite the novelty of this approach, initial prototypes did not undergo comprehensive safety assessments, which constrained their development. Recent advances in wearable technology, including miniaturization, mechanical flexibility, and wireless communication, have enabled more sophisticated, noninvasive access to salivary biomarkers and strengthened the technological foundation for physiological monitoring, notably via mouthguard-based sensors.

A mouthguard wearable sensor is typically fabricated from medical-grade polymers such as polyethylene terephthalate glycol (PETG), ethylene-vinyl acetate (EVA), and natural rubber.⁶¹⁴ To obtain a custom fit, the user’s oral anatomy is captured using a dental cast,⁶¹⁵ and the support is formed by pressure forming, vacuum forming, or 3D printing.⁶¹⁴ Commonly used in clinical settings and as protective gear against sports-related injuries, these devices seat securely over the teeth, are straightforward to wear and replace, and maintain continuous contact with relevant oral biofluids—features that are advantageous for real-time monitoring.⁶¹⁶ The first mouthguard biosensor was reported in 2014 by Joseph Wang’s group, who screen-printed a three-electrode system onto a flexible, commercial polyethylene terephthalate (PET) mouthguard using Prussian blue/graphite inks for the working and auxiliary electrodes and Ag/AgCl ink for the reference (Figure 19). Lactate oxidase was immobilized on the working electrode by electropolymerization within poly(o-phenylenediamine), which both entrapped the enzyme and mitigated interference from the complex salivary matrix.⁶¹³ Although the sensor provided reliable amperometric detection of lactate in saliva samples, it was not tested *in vivo*. The authors emphasized key limitations, calling for further miniaturization, integration of wireless circuitry, and rigorous safety and biocompatibility assessments prior to practical use.⁶¹⁶

Building on these advances, Arakawa and co-workers addressed several of these limitations.²²⁴ The mouthguard support was thermoformed from PETG by suction molding after being heat-softened. Platinum and silver electrodes, materials considered to be biocompatible and safe for human use, were deposited on PETG by parallel-plate sputtering and insulated with PDMS. A cellulose acetate membrane was

applied as an interference-rejection layer, combining electrostatic repulsion and size-exclusion effects. GOx was immobilized on the Pt electrode using poly(MPC-co-EHMA-co-MBP) (PMEHB), chosen for its biocompatibility and mechanical strength; a PMEHB overcoat further limited enzyme leakage and suppressed contaminants. A BLE wireless module and a battery were affixed near the sensing region by heat-welding with a heat gun. The mouthguard was then fitted to a fasting subject, and continuous signal changes were recorded wirelessly.

In addition to monitoring biomolecules in saliva for health purposes, mouthguard devices are also being used as physical sensors. Ichikawa and colleagues⁶¹⁷ developed a mouthguard wearable sensor to measure salivary turbidity, an indicator of oral hygiene. Their optical sensor, integrated with an LED, phototransistor, wireless module, and coin cells, was sealed inside a double-layered PETG mouthguard. The optical path between the LED and phototransistor measured the turbidity in the sensing area. In *in vivo* experiments with diluted milk, the mouthguard sensor successfully quantified turbidity in the range of 100–1000 FTU, demonstrating high linearity and repeatability, sufficient for standard salivary turbidity measurements. The same group⁶¹⁸ introduced a battery-free bite sensor for diagnosing intraoral and extraoral conditions. The device integrates an ultrathin sensor sheet, 82 μm thick, comprising a PET dielectric, an electret, and electrodes. It functions as an energy harvester, generating both power and signal from bite-induced deformation, thereby eliminating the need for a battery. The mouthguard is vacuum-formed on a dental model, and the sensor is embedded in the occlusal surface of a double-layered shield. *In vitro* tests demonstrated a dynamic range up to 2.5 MPa, corresponding to approximately 1 kN of bite force across the full dentition.

Mouthguard-based platforms have evolved from proof-of-concept biochemical sensors to multifunctional devices capable of both chemical and physical monitoring. Customizable, biocompatible materials and improved integration of sensing, sealing, and wireless modules now enable continuous data acquisition in the oral environment. Challenges in developing wearable mouthguard sensors include biofouling caused by saliva's high protein content. Furthermore, biomarker concentrations are often far lower than in blood, and correlations between salivary and serum levels are not yet well-defined. Additional factors, including background noise, matrix interference, viscosity, salivary flow rate, and recent food intake, can compromise accuracy and stability.^{610,613} Important requirements include sensor integration, wireless modules, and power sources within the limited form factor of a mouthguard to avoid altering natural occlusion. In this context, one needs to consider safety concerns in using batteries owing to the risk of accidental ingestion.⁶¹⁹

5.5. Wristband/Watches

Technological advances have transformed wrist-worn devices from simple timekeeping jewelry into networked computers. Modern devices now incorporate miniaturized batteries and processors, a variety of applications, and wireless communication modules, enabling continuous data acquisition and connectivity. Also, there is an expanding set of sensors for health-monitoring features. Commonly termed smartwatches, wristbands, or fitness trackers, these devices are currently among the most used wearable systems for health monitoring.^{5,7,620} The first generation of wrist-worn technologies was

marked by the advent of Android Wear and Apple Watch in the early 2010s.⁶²¹ These initial smartwatches were equipped mainly with motion sensors such as accelerometers, gyroscopes, and magnetometers, which detect changes in linear acceleration and rotational velocity.^{5,622} In practice, they were used predominantly for fitness monitoring, including step counting, estimation of energy expenditure, assessment of exercise intensity, and inference of sleep patterns.⁶²³ To this day, motion sensors remain among the most widely implemented technologies in wrist-worn wearable devices.

Over time, health-monitoring smartwatches have evolved from simple activity trackers into complex multisensor devices. Current models can monitor blood pressure and heart rate, sleep–wake cycles, skin temperature, and other vital signals. A central technology in this context is photoplethysmography (PPG), a fiber-free optical sensing method that probes physiological parameters through light–tissue interactions.⁶ On the wrist, PPG is used to track blood volume changes in the microvasculature, which are converted into electrical signals; the interval between successive pulsations (interbeat interval) can then be processed by algorithms for atrial fibrillation detection.⁶²⁴ Devices such as the Apple Watch combine PPG with single-lead electrocardiography. Beyond cardiac monitoring, PPG signals are also used to derive respiratory waveforms and to distinguish sleep–wake states, supporting the assessment of sleep quality and the detection of conditions such as sleep apnea.⁶²⁵ Skin temperature is another relevant physiological parameter, as it is associated with stress levels⁶²⁶ and reflects systemic status, including the presence of fever or infection. In smartwatches, this vital sign is typically monitored using thermistors, thermoelectric-based sensors, or optical approaches.⁶²⁷

In the literature, the terms *smartwatch* and *fitness tracker* usually refer to commercial wrist-worn devices designed to monitor physical health parameters. In contrast, *wristband* is often used for custom or prototype wrist-worn sensors developed in research settings that can measure both physical and biochemical markers. For biochemical monitoring, sweat is the most common biological matrix,^{202,628–630} owing to its noninvasive accessibility and the possibility of continuous sampling. Wristbands themselves can be constructed from diverse structural (flexible or rigid) and sensing materials, using both standard microfabrication and low-cost solution processing routes. Emaminejad and co-workers,²⁰² for example, reported a flexible wristband for sweat analysis of Na^+ , Cl^- , and glucose. In their design, the electronics are integrated on a flexible printed circuit board within a plastic housing, while the sensing and iontophoresis electrodes are microfabricated on an ultrathin PET film by photolithography and thermal evaporation. A key feature of this platform is the fully integrated, on-wrist sweat stimulation and analysis: iontophoretic electrodes contact the skin through a drug-loaded agarose hydrogel that locally induces sweating, and a thin rayon pad wicks sweat directly to the sensing membranes, enabling autonomous, exercise-independent, real-time measurements.

Cai and co-workers⁶²⁹ pursued a complementary strategy, implementing an integrated sweat-sensing wristwatch built on a rigid FR-4 commercial printed circuit board, with all electronics embedded in a 3D-printed resin case and held against the skin by an elastic silicone wristband. In this architecture, the ion sensors are fabricated directly on the FR-4: during the standard PCB manufacturing process, the sensing electrodes are defined as metallic pads on the board by




Figure 20. Sweat-sensing wristband developed by Ma et al. (A) Steps involved in the patterning of the device using double-sided photolithography; (B) flexible PET cloth used in the wristband; (C) sharp interface between PET and Cu cloth; (D) a picture of the cloth containing the patterned circuits; (E) electrodes in detail. Reproduced from Ma et al.⁶²⁸ Available under a CC-BY 4.0 license. © 2023 American Association for the Advancement of Science.

photolithography. Sweat handling is achieved by a laser-cut PET microfluidic layer bonded on top of the sensor array with medical adhesive; this PET film defines inlet and outlet ports and a network of shallow microchannels. During use, sweat generated naturally during exercise is pressed into the inlet by contact between the underside of the watch and the skin. From there, it is wicked through the PET microchannels by capillary forces, forming a continuous thin film that flows across the membrane-coated electrodes and enables a stable, on-wrist, real-time potentiometric readout of electrolyte concentrations. In a different approach, Ma and collaborators⁶²⁸ built a sweat-sensing wristband directly in a single piece of textile instead of on a rigid PCB (Figure 20). A commercial PET cloth serves simultaneously as a mechanical support and electronic substrate. To construct the sensor array, the textile is first converted into a conductive Cu-patterned fabric via polymer-assisted metal deposition (PAMD) and then patterned by double-sided photolithography and wet etching to define all electrodes and interconnects monolithically on the cloth. In this case, sweat is collected by exercise-induced perspiration using an epidermal microfluidic module: a PDMS channel layer molded from a 3D-printed template is laminated onto double-sided medical tape patterned with multiple sweat inlets and a reservoir over the textile electrodes so that sweat from a larger skin area is continuously drawn into the channels and across the sensing cavity by hydraulic/osmotic pressure and capillary flow. The textile-first architecture of this platform preserves air and moisture permeability, offering long-term wearing comfort and breathability that conventional flexible electronics on plastics cannot match.

Taken together, these examples illustrate how wrist-worn devices for health monitoring have evolved from motion-centric smartwatches to sophisticated platforms capable of biochemical sensing, built on flexible films, rigid PCBs or textiles. At the same time, they highlight that the main construction-related limitations of wrist-worn systems still lie in the difficulty of monolithically integrating batteries and electronics with robust skin–sensor interfaces and reliable

sweat collection/transport, while maintaining a form factor that is comfortable, manufacturable at scale and stable over time.

5.6. Textiles

Electronic textiles (e-textiles) represent a natural evolution of wearable technology, merging conventional fabrics with sensors or electronics. These smart textiles have social acceptability across sectors such as healthcare, fashion, sports, and personal defense by enabling real-time monitoring, adaptive functionality, and interactive communication between users and their environments.^{631–636} The intimate contact with large body areas makes them ideal for hosting distributed biosensors at different body locations to track the physiological, biochemical, and metabolic markers of the wearer more effectively than many conventional sensing methods. The all-day duration of wearability further makes e-textiles suitable for continuous monitoring of biomarkers for chronic conditions, rehabilitation, personal health, and lifestyle management.^{637–639} The widespread adoption of e-textiles is hindered by several challenges, especially in fabrication.^{640–643} For e-textiles to be adopted on a large scale, they must balance electronic performance with wearability, ensuring durability, washability, and long-term reliability.⁶⁴⁴ Scalable manufacturing must maintain high performance without excessive production complexity or cost. Because powering these textiles remains a significant hurdle, researchers are exploring solutions such as harvesting energy from body movement or fluid or using radio frequency waves to power wirelessly.^{645,646} This section explores the key fabrication techniques used in e-textile manufacturing, their associated challenges, and emerging wireless solutions that could pave the way for widespread adoption.

5.6.1. Fabrication of e-Textiles. A typical e-textile requires a supporting matrix to maintain its everyday-wear feature, functional components to promote the capacity of sensing, computing, and communication, and a protection layer to endorse durability under daily use and washing (Figure




Figure 21. E-textile fabrication architecture. (A) Overview of e-textile structure and definition. Reproduced from Lin et al.⁶⁷⁰ Available under a CC BY 4.0 license. © 2020 Springer Nature. (B) Supporting matrix of e-textile formation. Reproduced from Latko-Duralek et al.⁶⁵² Available under a CC BY 4.0 license. © 2023 MDPI and Wang et al.⁶⁵³ © 2019 John Wiley & Sons. (C) Functional parts and their integration into textile. It typically follows three main approaches. First, surface attachment. Attaching premanufactured electronic components directly onto textile surface. Reproduced from Wicaksono et al.⁶⁵⁴ Available under a CC BY 4.0 license. © 2020 Springer Nature. Second, in situ manufacturing. Directly fabricating functional elements onto textile surface using coating or printing technique. Reproduced with permission from La et al.⁶⁶⁰ © 2018 John Wiley & Sons. Third, fiber integration. Incorporating functional fibers, such as conductive fibers or fiber-based devices, into textile through methods like digital embroidery, conventional textile weaving, knitting, or bonding. Reproduced with permission from Shi et al.⁶⁷¹ © 2021 Springer Nature, Yan et al.⁶⁷² © 2022 Springer Nature, and Lin et al. © 2022 Springer Nature⁶⁷³ Lin et al. is available under a CC BY 4.0 license.




Figure 22. Fabrication of functional fibers and yarns in E-textile. (A) Coating-based methods by introducing functional materials on the surface of fiber or yarn. Reproduced with permission from Bi et al.⁶⁷⁵ © 2023 John Wiley & Sons. (B) Spinning-based methods by extruding liquid or molten polymer or polymer composite via a small hole to form a thin filament that solidifies into fiber mostly by coagulation bath, and can be further functionalized by functional materials bath coating. Reproduced from Fu et al.⁶⁸³ Available under a CC BY-NC license. © 2024 Springer Nature. (C) Thermal-drawing-based method by heating a prior prepared macroscopic preform to a viscous state and drawn into kilometer-long fiber. Reproduced with permission from Yan et al.⁶⁷² © 2022 Springer Nature.

21A).^{647–651} The supporting matrix is usually the textile itself, which is made from yarns/threads by weaving, knitting, crocheting or bonding. The base yarns or threads are usually made from natural fibers like cotton, wool, and silk, which can offer comfort and breathability, or synthetic fibers such as polyester and nylon, which provide durability and moisture-wicking properties (Figure 21B).^{642,652,653} These base materials are not inherently conductive, but can be trans-

formed into conductive materials for sensing through the incorporation of functional components. These functional components can be external attachments of classic electronic devices, such as LEDs, integrated circuits, and traditional batteries, that are attached to the surface of a traditional textile (Figure 21C).⁶⁵⁴ However, this approach tends to be less desirable as it can compromise the textile-like properties such as breathability, flexibility, and lightness. Thus, the primary

thrust in e-textile research is to integrate or embed the electronic function into textiles, maintaining the desired characteristics while offering advanced functionalities. One strategy is to directly coat, electroplate, or print functional elements on the surface of textile using various inks, bypassing rigid electronic devices (Figure 21C).^{655–660} The typical function imparted to textiles is electrical conductivity for applications such as physical and electrochemical sensing, wireless power transfer, or communication. The conductive inks are typically based on materials including metallic nanoparticles or nanowires, MXenes, carbon-based materials (carbon nanotubes, graphene), and conjugated polymers (PEDOT:PSS, PPy, PANI).^{4,661} Metallic ink provides the best conductivity, and silver-based inks, including nanoparticles, nanowires, and flakes, are most used because of the good balance between cost and performance, offering better oxidation resistance than copper and significantly lower cost than gold.^{657,662} Carbon-based and conjugated polymer-based inks provide higher electrochemical activity for electrochemical or physiological sensing and larger resistance for physical sensing (strain, pressure, temperature) compared to metal-based ones.^{663,664} Other functionalities, such as energy harvesting (triboelectric generators, piezoelectric and thermoelectric generators), electroluminescence, and biosensing, are often implemented at the single-fiber level and then integrated into textiles,^{641,652,665–669} which will be introduced in the next section.^{655–661}

Most methods for coating, electroplating, and printing used for flexible polymer substrates can be adapted for textiles. However, an extra layer for protection and insulating additives for modifying the ink is required to improve adhesion, mechanical robustness, and durability. This inevitably alters the textile nature with reduced breathability and increased stiffness. Additionally, the insulator additives lower the electrical conductivity to around 10% of pure metal, and the conductive path degrades over time due to washing and daily wear. Despite these limitations, coating and printing methods are advantageous in fabrication speed and simplicity, especially when fine pattern resolution is not required (less than 100 μm). They eliminate the need for complex photolithography, and the setup can be deployed rapidly. Coating is mostly done with techniques such as dip-coating, spray coating, electroless plating, electroplating, and vapor deposition (Figure 22A).⁶⁷⁴ Dip-coating, in particular, is simple and scalable, involving immersion of fibers into functional solutions followed by solvent evaporation and thermal curing. For example, conductive cotton yarns coated with inks containing conductive carbon paste (CCP), PEDOT:PSS, and sodium alginate (cross-linked with calcium chloride) achieved a conductivity of $19.46 \pm 0.240 \text{ S}\cdot\text{cm}^{-1}$, with minimal loss after washing.⁶⁷⁴ Similarly, MXene-coated Kevlar yarns reached a low linear resistance of $5.3 \pm 0.7 \Omega\cdot\text{cm}^{-1}$ using only $0.12 \text{ mg}\cdot\text{cm}^{-1}$ of MXene.⁶⁷⁵ Electroluminescent fibers, promising for textile displays and wireless interaction, have also been fabricated by coating ZnS phosphors on silver-plated yarns and encapsulating them with silicone, maintaining a luminance of $122 \text{ cd}\cdot\text{m}^{-2}$ after 10,000 folding cycles.⁶⁷¹

Electroplating enables highly conductive and precisely controlled metal or polymer coatings (e.g., Cu, Ag, Ni, Au, PEDOT:PSS) where high conductivity and flexibility are required, which cannot be provided by dip-coating or spray coating.⁶⁷⁶ The thickness of such a coating is usually less than 1 μm to lower the cost and minimize added stiffness. This

method requires conductive surfaces or pretreatment, limiting direct use on insulating textiles. Electroless plating chemically produces uniform metal coatings on insulating fibers without external current, but has slower deposition rates and stringent control of plating bath chemistry.⁶⁴⁸ Spinning is an interesting alternative, including wet spinning, melt spinning, electrospinning, and microfluidic spinning (Figure 22B).^{677,678} It integrates functional materials into polymer fibers in a single step at scale, enabling precise control over uniformity, conductivity, and mechanical properties. Wet spinning is the most adopted because of its low processing temperatures and simple process, which is preferable for thermally sensitive or nonthermoplastic functional materials.^{679–681} Fibers produced with wet spinning include those containing ATO@TiO_2 nanoparticles, which achieved a stable electrical conductivity of $2.5 \text{ S}\cdot\text{m}^{-1}$ under strains of 50 and 100% for 1000 cycles.⁶⁸² A self-healing electroluminescent fiber has also been developed using wet spinning, in which ZnS and Ni particles are dispersed within a high- k PVDF-HFP matrix to form a core-shell structure.⁶⁸³ A hydrogel layer served as the transparent conductive outer sheath. This fiber recovered over 98.6% of its initial luminance after complete severing and continued to operate reliably for more than 10 months.

While coating and spinning methods possess various advantages, they often require adhesion promoters or suffer from an overall conductivity reduction by insulating additives and limit architectural complexity. Thermal drawing presents an alternative for fabricating multifunctional fibers with integrated internal structures (Figure 22C). This process involves designing a macroscopic preform composed of polymers, metals, and semiconductors arranged in a precise geometry, which is then heated and drawn into fine microscale fibers. Thermal drawing preserves the geometric arrangement of materials across scales, allowing for the integration of diverse functions within a single fiber. For example, acoustic sensing fibers have been created using this method by embedding piezoelectric poly(vinylidene fluoride-co-trifluoroethylene) (P(VDF-TrFE)) and BaTiO_3 nanoparticles within an elastomeric cladding with internal copper electrodes.⁶⁷² These fibers have demonstrated applications in assistive technologies, including voice detection, acoustic communication, and real-time heart sound monitoring. Thermal drawing has also been used to develop fibers for computing,⁶⁸⁴ neural interfaces,⁶⁸⁵ and energy storage.⁶⁸⁶ Its compatibility with continuous manufacturing and scalability makes it highly attractive for commercial e-textile applications.

An alternative strategy for e-textile fabrication is to integrate premanufactured functional fibers or fiber-based devices directly into textiles (Figure 21C).^{668,687–691} This fiber-based approach offers advantages over on-site fabrication, particularly in its compatibility with existing textile manufacturing processes and improved durability, as the functional fibers are preinsulated before integration. It also maintains the intrinsic porosity of textiles, which is critical for comfort and long-term wearability. By functionalizing individual fibers with distinct sensing materials prior to assembly, multiple sensing modalities can be achieved within a single textile platform. A solid metal wire covered by textile yarns using a braiding machine provides robust conductivity to make e-textile where high conductivity is required, but faces limitations in comfort. Thus, composites made from conductive materials with yarns or synthetic fibers have become the preferred choice. Metallic ink, MXene, carbon-based materials (CNTs, graphene), or




Figure 23. System-level wireless e-textile design and fabrication. (A) Metamaterial textile network. Reproduced with permission from Tian et al.⁶⁹³ © 2019 Springer Nature. (B) Digital embroidery near-field clothing network. Reproduced from Lin et al.⁶⁷⁰ Available under a CC BY 4.0 license. © 2020 Springer Nature. (C) Reconfigurable NFC textile network. Reproduced with permission from Hajiaghajani et al.⁶⁹⁴ © 2021 Springer Nature. (D) Contactless e-textile. Reproduced from Nguyen et al.⁶⁹⁶ Available under a CC BY-NC license. © 2024 American Association for the Advancement of Science.

conductive polymer (e.g., PEDOT:PSS, PPy, PANI, etc.) have been adopted to integrate into textile fibers.^{4,661} Unlike film-based coatings that often fail under repeated deformation, these fiber-integrated textiles are better suited to conform to complex body movements without compromising performance. This method combines wearability, robustness, and scalability, making it highly suitable for healthcare applications that demand reliable, unobtrusive operation during daily use.

One common approach is weaving or knitting functional fibers directly into the textile. It is compatible with industrial-scale textile manufacturing and allows for integration at the yarn level and across large textile areas, preserving the textile nature and user comfort. Electrochemical textiles capable of real-time health monitoring have been developed by weaving sensing fibers into textiles to detect analytes such as glucose, Na^+ , K^+ , Ca^{2+} , and pH.^{671,692} Fiber integration can also be made with digital embroidery, where conductive or functional fibers are stitched onto existing textiles using programmable embroidery machines. One distinctive advantage of embroidery is the ability to place stitches in any direction—forward, backward, and sideways—allowing for flexible and complex designs. Silver-plated polyamide yarn encapsulated in transparent thermoplastic polyurethane tubing has been stitched onto regular textiles through this method to form wireless coil networks.^{670,673} It can sense physiological signals, including spinal cord posture, temperature, gaits, etc., across multiple body locations. While embroidery offers excellent design flexibility and preserves the porosity of unstitched areas, it may introduce local stiffness or thickness where dense patterns are applied.

5.6.2. Wireless e-Textiles. E-textile platforms have been produced that are capable of wirelessly receiving power, detecting physiological signals, and transmitting data. Especially for radio frequency network systems, they can detect signals from different locations on the body within one

wearable format, eliminating bulky batteries or rigid circuits for an unobtrusive form. These networks distribute sensing and communication tasks across multiple nodes, typically sharing a common reader or power source, which reduces the need for each sensor to be fully self-contained. By attaching prepatterned conductive fabrics (such as copper/nickel-coated polyester) to textiles via adhesives, Tian et al. introduced a metamaterial network through radio surface plasmon propagation along the patterned fabrics (Figure 23A).⁶⁹³ It enables wireless signals in radio-wave propagation confined around the body rather than radiation to the outside. The power transmission efficiency was enhanced by 3 orders of magnitude compared to that of a traditional radiative network. Additionally, signal confinement within 10 cm of the body enhances the communication security by limiting the risk of external interception. Using computer-controlled embroidery of conductive threads (silver-plated polyamide yarn encapsulated in transparent thermoplastic polyurethane) onto regular textiles, Lin et al. developed near-field-enabled clothing with battery-free sensors to monitor physiological signals continuously (Figure 23B).⁶⁷⁰ Multiple electromagnetically responsive thread patterns were integrated to form a body network with one reader hub, extending the short working distance in a few centimeters of near-field technology to meter-scale networks. Thus, wireless signal collection can be done from different sensors located at different locations on the body.

Hajiaghajani et al. proposed a reconfigurable NFC textile network via magneto-inductive wave propagation along flexible arrays of discrete resonators constructed from laser-cut metal coils (Figure 23C).⁶⁹⁴ The networks allow battery-free sensors to be added or relocated (“drag-and-drop”) anywhere along the coil arrays. The authors demonstrated real-world use across multiple garments and during user motion, highlighting the practicality of this distributed sensing system approach. Another innovation comes from Takahashi et al., who




Figure 24. Paper-based diaper biofuel cell developed by Shitanda et al. (A) Picture of the device embedded in the diapers, containing the electrodes and the wireless data transmitter. (B) Preparation process of the device. Reproduced from Shitanda et al.⁷²¹ © 2021 American Chemical Society.

developed a twin meander coil textile system for passive inductive telemetry (PIT).⁶⁹⁵ The large reader coil was fabricated using industrial knitting techniques, and interrogated spatially distributed sensor coils, each having a different resonant frequency. The meandered winding pattern alternates between clockwise and counterclockwise turns, which mitigates electromagnetic interference from the body. This enables accurate and spatially flexible sensing without requiring direct coupling or complex routing. Contactless sensing eliminates the need for physical contact between the textile and the skin, thus enhancing user comfort and long-term durability. Nguyen et al. developed a contactless metamaterial textile that was integrated into the furniture surface for ambient physiological monitoring.⁶⁹⁶ The textile comprised multilayer structures with patterned conductive top layers, intermediate nonconductive fabric layers, and conductive bottom isolation layers. The conductive layers were specifically patterned via laser cutting to support spoof surface plasmon (SSP) modes. Using the close-range interaction between SSP surface waves and the human body, it can monitor multisite physiological signals, including heartbeat, respiration, and blood pressure.

5.7. Facemasks

Respiratory monitoring masks have been used for personalized medicine and early disease diagnosis,^{697,698} with the integration of sensors and data processing technologies. These devices can be fabricated using several methods, including the direct integration of sensors into the facial region of the mask to ensure direct contact with the breath airflow. By embedding sensors such as airflow vibration sensors,⁶⁹⁹ pressure sensors,⁷⁰⁰ humidity sensors,⁷⁰¹ and temperature sensors⁷⁰² within the mask, researchers can collect multiple physical parameters of respiratory airflow, including the respiratory rate, airflow velocity, humidity, and temperature. A key challenge is reaching large-scale fabrication of facemasks incorporating wearable sensors. A possible solution is to employ standard computerized embroidery where conductive threads are produced via a scalable roll-to-roll process and optimized to maintain performance under mechanical stress, washing, and long-term use.⁷⁰³ Using these threads, sensors could be directly integrated into fabrics as demonstrated with a facemask for respiration monitoring by measuring moisture-induced conductivity changes in the textile. Indeed, smart facemasks may evolve into platforms for personalized, real-time health monitoring if sensors and electronic devices can be integrated.

For instance, a facemask embedded with a near-field electrochemical sensor was used for continuous respiratory assessment, which contained a lightweight, flexible, and battery-free sensor chip that could detect breathing patterns through humidity-induced impedance changes.⁷⁰⁴

Once processed by algorithms, the data collected with facemasks as indicated above can extract valuable respiratory health information, providing support for clinical diagnosis and daily health management.^{705,706} In addition to monitoring physical parameters, the chemical composition of exhaled breath also serves as a health indicator (see Section 3.6). The incorporation of electrochemical sensors allows the mask to detect specific chemical components in breath gases, such as oxygen, carbon dioxide, and nitric oxide.^{706–709} These sensors operate by detecting redox reactions of gas molecules on electrode surfaces or changes in electrolyte pH, enabling real-time reflection of the chemical composition and concentration of breath gases. Monitoring masks have also benefited from microfluidics to detect a broader range of biomarkers.^{32,710} For instance, through microfluidics, researchers can extract non-volatile substances such as inflammatory factors, proteins, alcohol, and lactate, providing new sample sources for studying metabolic conditions and other respiratory system diseases.³²

By integrating ML algorithms, smart respiratory monitoring masks can extract valuable health insights from vast amounts of respiratory data, helping users identify potential health issues in a timely manner. For example, algorithms such as support vector machines (SVMs) and convolutional neural networks (CNNs) have been widely used for the classification and pattern recognition of breath data, laying the foundation for personalized and precision medicine.³¹⁸ In a recent example, ML was used to determine glucose concentration through breath analysis with a facemask incorporating an organic electrochemical transistor (OECT).⁷¹¹

5.8. Diapers

Sensors embedded within diapers enable real-time analysis of urine biomarkers for diabetes management,²⁹⁸ neonatal jaundice monitoring,⁷¹² and long-term care.^{713–716} Additional discussion on urine biomarkers can be found in Section 3. The fabrication process consists of integrating flexible and functional materials into the diaper structure.^{297,717} Biosensors are typically embedded between the diversion and absorption layers, allowing direct contact with urine. Substrates such as PET or cellulose-based materials are chosen for their flexibility and compatibility with screen-printing technologies.^{298,306,718}




Figure 25. (A) Possible structures of PS (PAM/SF) and PSGP (PAM/SF/GO/PEDOT:PSS) hydrogels, highlighting the presence of SF nanofibrils and the intertwining of GO sheets through hydrogen bonding. (B) Diagram illustrating how the conductive components (GO and PEDOT:PSS) in PSGP sensors respond to compression and stretching, resulting in changes in pressure or strain. Reproduced from He et al.⁷²³ © 2020 American Chemical Society.

Conductive inks, including silver and carbon composites, are printed to form circuitry and electrodes, while functionalized layers sensitive to specific biomolecules, such as GOx for glucose detection or bilirubin oxidase for bilirubin analysis, provide precise biochemical detection. Self-powered diaper sensors have been made with biofuel cells that harness enzymatic reactions to generate electricity from urine glucose (see Section 10.3).^{297,719,720} Paper-based glucose biofuel cells were fabricated by printing MgO-templated porous carbon electrodes and graft-polymerizing active enzyme layers, ensuring both stability and high sensitivity (Figure 24).²⁹⁷ Data transmission modules, equipped with wireless communication technologies such as Bluetooth or Wi-Fi, transmit real-time data to smartphones or cloud platforms, facilitating remote monitoring (see Section 9). Structural integration is achieved through modified diaper layers that optimize fluid diversion and interaction with the biosensor, using hydrophilic polymers for efficient urine flow and absorption layers to maintain user comfort and hygiene.

Design principles for these sensors prioritize biocompatibility, ensuring that materials are safe for prolonged skin contact and hypoallergenic.⁷²² Flexibility is essential to maintain comfort and functionality during movement, while miniaturization ensures that the sensors are lightweight and do not add bulk. Methods such as screen printing and roll-to-roll processing can be used to obtain durable diapers capable of withstanding a wet environment. These design considerations ensure that diaper sensors meet the requirements of diverse user populations from neonates to elderly patients.

6. PHYSICAL SENSORS

Wearable physical sensors are used to monitor parameters such as heart rate, blood pressure, and body temperature and can be combined with chemical sensors to provide a comprehensive picture of the health of an individual. In the following sections, the recent advancements of the most relevant wearable physical sensors will be further explored.

6.1. Pressure and Strain Sensors

Pressure and strain sensors can be applied for monitoring multiple essential clinical parameters, including pulse and respiration, besides acting in fall detection, gait analysis, assessment of joint movements, and so on. The landscape of the literature on wearable sensors in Section 2 highlights the dominance of pressure and strain sensors due to the large number of publications (tens of thousands). An inspection of the clusters associated with these sensors indicated the relevance of advanced materials in their use, particularly nanomaterials and hydrogels. Because the topic contains such an extensive literature, we will not attempt a historical overview of material developments. Instead, we will present a few recent examples that we hope will illustrate these advancements.

The importance of hydrogen bonding and the incorporation of nanomaterials into hydrogels with tailored properties were highlighted by He et al.⁷²³ They developed a soft, conductive hydrogel composed of PEDOT:PSS, PAM, silk fibroin (SF), and graphene oxide (GO). Figure 25A depicts the possible structures of two hydrogels: PS (PAM/SF) and PSGP (PAM/SF/GO/PEDOT:PSS). The SF nanofibrils impart mechanical strength to the PS hydrogels, enabling them to function as elastic matrices. When GO sheets are incorporated into the PS matrix via hydrogen bonding and combined with PEDOT:PSS layers, the resulting PSGP hydrogels also exhibit conductivity. PSGP hydrogels were employed in strain and pressure sensors capable of monitoring physical signals from the human body, such as pulses, breathing, facial gestures, and joint movements. Figure 25B illustrates how the conductive components of PSGP sensors respond to stretching and compression.

Besides hydrogen bonding (H-bonding), other noncovalent interactions, including dipole–dipole, ionic, and van der Waals interactions, can impart additional properties to hydrogels. Liu et al.⁷²⁴ reported hydrogels consisting of a soft, homogeneous polymer network combined with a hard, dynamic network of Fe³⁺-cross-linked cellulose nanocrystals (CNCrys–Fe³⁺).




Figure 26. (A) (i) LED brightness decreases as the F-hydrogel is elongated, indicating a decrease in conductivity with strain. Step changes in resistance are shown in (Aii) and (Bii) for elongation and finger bending, respectively. (B) (i) Images showing the index finger bending at different angles. (C) Relative resistance changes during regular and rapid breathing, obtained using the wearable sensors shown schematically. (D) Image of a sensor placed on the wrist of a volunteer. (E) Schematic representation of the F-hydrogel sensors. (F) Response of the strain sensor before and

Figure 26. continued

after exercise, corresponding to a volunteer running. The measured heart rates were 78 beats/min at rest and 102 beats/min postexercise. Reproduced from Liu et al.⁷²⁴ © 2017 American Chemical Society.




Figure 27. Fabrication process of the DTPAM hydrogel, showing how dopamine (DA) molecules intercalated within talc layers (a) are incorporated into a PAM hydrogel (b). The distinct interactions in DTPAM are highlighted: (1) Amide groups in PDA chains, formed in the presence of talc, enable linkage with PAM chains; (2) π - π stacking between PDA chains; (3) H-bonding within the hydrogel. Reproduced from Jing et al.⁷²⁵ © 2018 American Chemical Society.




Figure 28. Illustration of the design strategy for producing the multiscale hydrogel. The image depicts microspheres that can be derived from one of three polymers: poly(acrylic acid) hydrogel (PAAM), poly(cyclodextrin) hydrogel (PCDM), and poly(cyclodextrin)-chitosan hydrogel (PCCM). The inset highlights the groups involved in hydrogen bonding and hydrophobic interactions. Reproduced from Zhao et al.⁷²⁶ © 2024 American Chemical Society.

These hydrogels exhibited excellent mechanical properties due to the synergistic interaction between the two networks. The dynamic CNCrys- Fe^{3+} coordination bonds dissipated energy under stress, while the homogeneous polymer network facilitated smooth stress transfer. Furthermore, the reorganization of CNCrys and Fe^{3+} through ionic coordination allowed the hydrogels to self-heal within 5 min without requiring external healing agents or stimuli. The functional network hydrogels were explored as strain sensors, with their conductivity varying in response to strain. This is illustrated in Figure 26Ai, where the brightness of an LED decreases as the strain increases from 0 to 150%. Step-like resistance

changes during loading–unloading cycles, which could be repeated more than 200 times, are shown in Figure 26Aii. The hydrogel sensor was also capable of monitoring finger joint motion, as depicted in Figure 26B. The monitoring of physiological signals, such as breathing and pulse, is demonstrated in Figure 26D–F. For pulse monitoring, in particular, reproducible signals corresponding to 78 beats/min at rest and 102 beats/min after exercise were obtained.

Another example involving the combination of H-bonding and other noncovalent interactions was reported in multifunctional hydrogels suitable for various applications, including strain sensors.⁷²⁵ The hydrogel, referred to as DTPAM, was




Figure 29. (A) Photograph of robotic fingers integrating the piezocapacitive sensor while handling a lightbulb. (B) Displacements of one finger and the corresponding capacitive responses, showing full recovery of the capacitance signal when no pressure is exerted. (C) Image of a sensor attached to the wrist of a volunteer. (D) Pulse signal recorded from the volunteer. (E) Comparison of the signals measured at the pulse and the palm, demonstrating that the measurement at the pulse corresponds to the pulse signal. (F) Capacitance response of a single pulse, featuring the characteristic two peaks. Reproduced from Kwon et al.⁷²⁸ © 2016 American Chemical Society.

synthesized by incorporating polydopamine-coated talc (PDA–talc) nanoflakes into a PAM hydrogel. The key intermolecular interactions are illustrated in the scheme in Figure 27, which outlines the fabrication process of DTPAM. In particular, intercalating dopamine molecules ensured higher dispersion of talc and preservation of catechol groups within the hydrogel. The final properties were governed by π – π stacking of PDA chains and H-bonding in the hydrogel. DTPAM exhibited characteristics such as stretchability, biocompatibility, self-healing, and adhesiveness, which allowed its use in strain sensors to monitor human motion.

The incorporation of polymer hydrogel microspheres into a matrix hydrogel enabled the exploitation of dynamic hydrogen bonds, hydrophobic interactions, and covalent bonds to produce a multiscale hydrogel.⁷²⁶ The design strategy, illustrated in Figure 28, involves immersing microspheres of one of three types of polymer into an aqueous solution of acrylamide/1-vinyl-3-butylimidazolium bromide (AM/[VBIIm]Br) and the cross-linking agent *N,N'*-methylene-bisacrylamide (MBA). Polymerization in the solution and within the microspheres resulted in a double network comprising a chemical network from PAM-[VBIIm]Br and physical entanglements with the microspheres as nodes. Additionally, the hydrophilic groups on PAM and the microspheres formed dynamic hydrogen bonds, enhancing physical cross-linking strength. These microspheres improved the hydrogel structure, resulting in superior sensing performance.

Dynamic cross-linking can also be essential for conductive hydrogels used in strain and pressure sensors, enabling them to withstand cold temperatures without freezing.⁷²⁷ A hydrogel

consisted of a double network of chitosan-poly(acrylamide-co-acrylic acid) [CS-P(AM-co-AA)], with dynamic cross-links within the chitosan physical network and through ionic coordination with Fe ions. While the mechanical properties of the hydrogel, such as tensile strength, compressibility, and self-recovery, were enhanced by immobilizing ions in the dynamic cross-links, the free ions in the hydrogel imparted a high conductivity and freezing tolerance.

The examples above illustrate how hydrogels can be designed to exploit intermolecular interactions and achieve variable properties. Other strategies also exist to enhance material properties for wearable sensors. For instance, microporous dielectric elastomers with great piezocapacitive effects have been used in tactile pressure sensors.⁷²⁸ The enhancement in piezocapacitive effect was obtained using a three-dimensional (3D) microporous Ecoflex elastomer, which is highly deformable upon compression. Wearable sensors with piezocapacitive effects were employed in various applications, as illustrated in Figure 29. A sensor was attached to the fingertips of a robot to measure the pressure exerted on a target object—specifically, a lightbulb, as shown in Figure 29A. The seizing motion at different displacements was reflected in the relative changes in capacitance, as depicted in Figure 29B. This figure also highlights the full recovery of the capacitance signal upon the release of the object. Due to the elastomeric nature of the sensor, it serves as a cushion, enabling the robotic finger to handle the lightbulb without breaking it. In a health-monitoring application, the sensor was adhered to a volunteer's wrist, as shown in Figure 29C, to measure the pulse signal presented in Figure 29D. This signal was confirmed to




Figure 30. Schematic illustration of the various uses for the PLA pressure sensors, which are fully degradable. The sensors can be used to detect various body movements and pulse signals. Reproduced from Qin et al.⁷³² © 2024 American Chemical Society.

originate from the wrist pulse by comparing it with the sensor signal recorded on the palm. As demonstrated in Figure 29E, the palm signal was significantly lower, comparable to the noise level of the pulse signal. Due to its fast response, the sensor was also capable of measuring a single pulse signal. The result, shown in Figure 29F, represents a typical single pulse signal of a healthy adult male, characterized by two peaks (P_1 and P_2) that provide information on arterial stiffness. For example, the radial artery augmentation index (AI_r) can be defined as $AI_r = P_2/P_1$. Furthermore, the pulse pressure P_1 was approximately 3 Pa, consistent with measurements from other wearable devices.

The porous nature of materials for pressure and strain sensors can be exploited in various ways. A porous network with graphene and PDMS was used for pressure and strain sensing to recognize walking states and measure wrist blood pressure.⁷²⁹ Capacitive pressure sensors were developed with a porous pyramid dielectric layer, which was insensitive to strain and temperature.⁷³⁰ A similar concept was employed to develop high-performance capacitive pressure sensors using tilted micropillar dielectric arrays.⁷³¹ In yet another example, electromechanical sensors were developed using poly(lactic acid) (PLA) electret films with a cellular structure.⁷³² This cellular structure was essential for increasing the charge stored in the electret film, thereby enhancing pressure sensitivity. The example of pressure sensors made from PLA films highlights the overarching principles that define wearable sensors. Specifically, the sensing design involves careful selection of materials, molecular and film architectures and fabrication conditions. In the case of PLA sensors, the polymer serves as a biodegradable electret that can be engineered with a cellular microstructure. Under optimized corona charging conditions, PLA cellular electrets exhibit enhanced polarization, resulting in high pressure sensitivity. The pressure sensitivity was measured at 1000 pC/kPa within a detection range of 0.03–62.4 kPa. These fully degradable sensors were utilized to monitor body movements and detect subtle physiological signals, as illustrated in Figure 30.

The combination of pyramidal and porous structures was also investigated to develop piezoresistive pressure sensors capable of achieving both high sensitivity and linear response.⁷³³ The unique design of the dual sensor is illustrated in Figure 31. The sensor is described as dual because it integrates a micropatterned (micropyramidal) PDMS component with a porous component made of carbon black and filter paper. These two parts, shown in Figure 31B, face an interdigitated electrode (IDE). The distinct ways the two components respond to pressure are depicted in Figure 31C,D. The complete sensor architecture, shown in Figure 31E, consists of these components sandwiched between two PET films, one of which contains the interdigitated electrodes. At low pressures (<1 kPa), the micropyramidal component provides high sensitivity, while at higher pressures, the conductive filter paper component ensures both high sensitivity and a linear response within the 1–20 kPa range.

This section has primarily focused on hydrogel-containing sensors as these are among the most widely used materials for such applications. They not only offer versatility in materials engineering to enhance sensing properties but also provide excellent mechanical and functional adaptability. Another class of widely used materials is nanomaterials, particularly graphene. Because of its exceptional mechanical and electrical properties, graphene has been incorporated into various strain sensors, often in combination with other materials. For instance, wearable sensors created via the layer-by-layer assembly of graphene and yarns from different sources have been employed for human motion monitoring.⁷³⁴ Other nanomaterials are also used for similar purposes, primarily to impart electrical conductivity. These include carbon nanotubes,⁷³⁵ MXenes,⁷³⁶ and silver nanowires.^{737,738} Another approach to creating conductive materials for strain and pressure sensors involves the incorporation of LMs. For example, a microfluidic sensor employing a wave-patterned LM demonstrated superelasticity, sensitivity, and low hysteresis, making it suitable for human activity monitoring.⁷³⁹




Figure 31. (A) Illustration of the pressure sensor placed on the pulse for monitoring the pulse signal. (B) Two-component sensor deposited on interdigitated electrodes, electrically connected to measure the piezoresistive signal. (C) Diagram showing how the micropyramidal sensor responds to pressure, exhibiting sensitivity to low pressures. (D) Diagram showing the response of the porous sensor component, which provides a linear response and is sensitive to higher pressures. (E) Architecture of the dual sensor, highlighting the distinct ways the two components respond under pressure. Reproduced from Jafarizadeh, B. et al.⁷³³ © 2024 American Chemical Society.

There are numerous advantages to integrating wearable sensors with textiles, as already discussed in Section 5.6. Examples of textile-containing sensors include a washable piezoresistive sensor made from a composite of AuNW-impregnated knitted cotton fabric.⁷⁴⁰ Another noteworthy example involves strain sensors created from knitted textiles combined with polymers and elastomer yarns.⁷⁴¹ Additionally, Lou et al.⁷⁴² demonstrated a triboelectric textile sensor produced from core-shell yarns (see Section 10.1 for triboelectric nanogenerators). Figure 32 shows textile pressure sensors incorporating different polymers, with stainless steel yarns used as electrode layers. The two polymers selected, nylon and polytetrafluoroethylene (PTFE), were chosen for their mechanical properties and high abrasion resistance. The textile sensors could be integrated into standard clothing or

attached to the human skin. Therefore, human movements could be monitored, including hand, knee, and elbow motions, with the bending speeds and angles. This type of monitoring helps to analyze patients with potential mobility disorders. Another health-monitoring application is for measuring the pulse wave at the carotid artery.

6.2. Blood Pressure Sensors

Blood pressure sensors are useful for cardiovascular health management as they can be integrated into daily routines and provide continuous, real-time monitoring.⁷⁴³ This section offers a concise overview of the field, highlighting the role of blood pressure monitoring, optimal on-body sensor locations, innovative device strategies, representative applications, and ongoing challenges that require further research and development.




Figure 32. (A) Yarns that were used to produce the sensors. (B) Photographs of different combinations of knitted textiles with different colors. (C) Schematic diagram of knitted textiles including stainless steel yarns. Reproduced from Lou et al.⁷⁴² © 2020 American Chemical Society.




Figure 33. Soft wearable blood pressure sensors. (A) A pair of resistive strain sensors with surface crack design captures the pulse wave velocity. Reproduced from Wu et al.⁷⁴⁹ © 2023 American Chemical Society. (B) A wireless capacitive pressure sensor (left) and the measured pulse pressure waveform of a human subject with a heart rate of 82 bpm (right). Reproduced with permission from Chen et al.⁷⁵⁸ © 2014 Springer Nature. (C) Pulse wave measured by a 3D-printed piezoelectric pressure sensor. The inset figures show the peaks representing systolic pressure and diastolic pressure, respectively. Reproduced with permission from Mandal et al.⁷⁶² © 2024 John Wiley & Sons. (D) A real-time pulse and blood pressure monitoring system developed using triboelectric nanogenerators. Reproduced with permission from Zhang et al.⁷⁶³ © 2025 John Wiley & Sons. (E) A microfluidic pressure sensor enabled by fluid-driven impedance changes. Reproduced with permission from Li et al.⁷⁶⁴ © 2014 John Wiley & Sons. (F) A multimodal wrist biosensor for wearable cuff-less blood pressure monitoring. Reproduced from Rachim et al.⁷⁶⁶ Available under a CC BY 4.0 license. © 2019 Springer Nature.

Blood pressure measures the internal pressure exerted by the heart muscles to pump blood throughout the body. Maintaining blood pressure within a specific range is essential for optimal health. Abnormal levels—whether elevated (hyper-

tension) or reduced (hypotension)—can signal dysfunctions in the cardiovascular system. Blood pressure is divided into two types: systolic blood pressure (SBP) and diastolic blood pressure (DBP).⁷⁴⁴ While SBP measures the contracted status

of the heart, DBP characterizes the relaxed status.⁷⁴⁵ Wearable blood pressure sensors for long-term monitoring have attracted much attention for the early detection and prevention of heart diseases, such as stroke, hypertension, coronary artery diseases, and myocardial ischemia.⁷⁴⁶ Early detection enables timely interventions, including lifestyle modifications and pharmacological treatments, which can significantly reduce the risk of heart attack, stroke, kidney disease, and other life-threatening complications.

The choice of on-body locations for soft wearable blood pressure sensors is crucial to the accuracy and reliability of the sensing systems. The sensor's placement affects the quality of the signal captured and impacts the user's comfort during use. The upper arm, traditionally the standard location for blood pressure measurement, offers proximity to the brachial artery, a major blood vessel that provides a strong and reliable signal.⁷⁴⁷ However, upper arm placement may be less convenient for continuous monitoring due to its potential interference with daily activities. The wrist, a more accessible location, has gained popularity for wearable blood pressure sensors. Wrist-worn devices allow for discrete monitoring and integration with existing smartwatch and fitness tracker designs. However, the radial artery at the wrist is smaller and more superficial than the brachial artery, potentially leading to less accurate readings.⁷⁴⁸ Alternative locations, such as the finger, earlobe, and ankle, are also under investigation. Each location presents advantages and challenges in terms of signal quality, user comfort, and sensor design.

The demand for noninvasive, continuous, and precise blood pressure monitoring has driven the development of soft wearable sensors with various sensing mechanisms. Examples include strain sensors, ultrasonic sensors, and optical sensors. Among these, strain sensors have received the most attention for their simplicity, low cost, and good conformability. Currently, mechanical sensors used in blood pressure monitoring are mainly categorized into four types:

1. **Resistive sensors** use the change in electrical resistance of a material under applied pressure. These offer a relatively simple and cost-effective approach with good sensitivity.⁷ When external pressure is applied, the geometric shape of the active material changes, causing restructuring or even breakage of the conducting network and hence significantly increasing the resistance.^{749,750} These sensors can be fabricated using a variety of flexible materials, including metal nanoparticles,⁷⁵¹ metal nanowires,⁷⁵² carbon-based nanomaterials,⁷⁵⁰ conductive polymers,⁷⁵³ and MOF compounds.^{754,755} A major challenge of resistive sensors, however, is the repeatability. Recently, cracks have been introduced in the active material to control the conducting pathway, offering an excellent combination of sensitivity, strain range, and repeatability.⁷⁴⁹ Figure 33A shows a set of two silver nanowire-based strain sensors with surface crack design to capture the pulse wave velocity, which can be further translated to blood pressure.⁷⁴⁹
2. **Capacitive sensors**, which measure the change in capacitance due to pressure-induced deformation, provide excellent linearity and repeatability. These sensors often employ flexible electrodes and dielectric materials, enabling conformal contact with the skin and accurate detection of pressure variations.⁷⁵⁶ Recent

advancements, such as the design of a microstructured dielectric layer,⁷⁵⁷ have improved sensitivity. Figure 33B shows a wireless capacitive pressure sensor and the measured pulse pressure waveform.⁷⁵⁸

3. **Piezoelectric sensors**, which generate electrical charges in response to mechanical stress, offer the advantage of self-powered operation.⁷⁵⁹ This type of sensor excels in capturing rapid pulse waveforms due to its millisecond-level response time.⁷⁶⁰ Materials such as polyvinylidene fluoride (PVDF) and its copolymers are commonly used in flexible piezoelectric sensors, providing a balance of piezoelectric properties, flexibility, and biocompatibility.⁷⁶¹ Figure 33C shows a 3D-printed PVDF-BaTiO₃ sensor using direct-ink writing. The sensor was calibrated against a standard cuff-based oscillometric device and measured the blood pressure of a subject by placing the sensor near the radial artery.⁷⁶²
4. **Triboelectric sensors** operate based on triboelectric effects and electrostatic induction principles. When materials come into contact and separate, a potential difference is generated, inducing transient currents that enable highly sensitive pressure detection. Triboelectric sensors are known for ultrahigh sensitivity and extremely low power consumption. They can even achieve fully self-powered monitoring by generating electricity through slight relative motion between the skin and sensor, offering a promising solution for long-term portable heartbeat monitoring. Figure 33D shows a flexible triboelectric nanogenerator-based wearable blood pressure sensor with high sensitivity (resolution of 1–6 kPa) and fast response time (8.5 ms).⁷⁶³

Other sensing mechanisms have been used for detecting blood pressure, including microfluidic channels,⁷⁶⁴ ultrasound technology,⁷⁶⁵ and optical sensors,⁷⁶⁶ together with advanced signal processing algorithms to enhance accuracy and reliability.⁷⁶⁷ Figure 33E shows a microfluidic pressure sensor that measures the impedance change of internal fluid due to external pressure.⁷⁶⁴ Flexible piezocomposite ultrasonic sensors have also been used to monitor continuous blood pressure through ultrasonic motion tracking of the blood vessel wall.⁷⁶⁸ Optical sensors, utilizing PPG, measure changes in light absorption or reflection caused by blood flow variations. This method involves the combination of LEDs with photodetectors. When light passes through the epidermis, it is absorbed and scattered by blood and surrounding tissues, resulting in changes in the reflected light intensity that synchronize with vascular pulsation (Figure 33F). PPG-based sensors are noninvasive and can be integrated into compact, wrist-worn devices. However, their accuracy can be affected by motion artifacts and variations in skin pigmentation.⁷⁶⁹

Soft wearable blood pressure sensors can be used in remote patient monitoring to facilitate timely interventions and reduce the need for frequent clinic visits.⁷⁷⁰ This is beneficial for elderly patients or those residing in rural areas with limited access to healthcare facilities. In sports medicine, wearable blood pressure sensors can monitor athletes' cardiovascular responses during training and competition.^{771,772} Furthermore, soft wearable blood pressure sensors hold promise for early diagnosis and prevention of preeclampsia in pregnant women, a life-threatening condition characterized by high blood pressure and organ damage.⁷⁷³ Wearable blood pressure sensors are also highly beneficial for infants, as they are



Figure 34. Soft wearable heart rate monitoring devices. (A) Chest strap ECG sensor. Reproduced from Romano et al.⁷⁹⁷ Available under a CC BY 4.0 license. © 2022 MDPI. (B) Smart clothing with PPG sensors. Reproduced from Wicaksono et al.⁶⁵⁴ Available under a CC BY 4.0 license. © 2020 Springer Nature. (C) Ear-Worn Piezoelectric Sensor. Reproduced from Park et al.⁸⁰⁴ Available under a CC BY 4.0 license. © 2015 MDPI. (D) Compact patch integrating ECG sensor, signal processing, and transmission module. Reproduced from Lee et al.⁷⁸⁹ Available under a CC BY 4.0 license. © 2018 Springer Nature. (E) Epidermal electronic heart rate device based on triboelectric Sensors. Reproduced with permission from Pang et al.⁷⁵⁰ © 2012 Springer Nature. (F) Wristband pulse sensor-based piezoresistive technology. Reproduced from Wang et al.⁸⁰¹ Available under a CC BY 4.0 license. © 2022 MDPI.

more vulnerable to the discomfort and risks associated with traditional monitoring devices.⁷⁷⁴

6.3. Sensors to Monitor Heart Rate

Heart rate (HR), defined as the frequency of cardiac pulsations, serves as an important biomarker for cardiovascular function. Its value is dynamically regulated by multiple factors, including age, gender, genetics, health status, circadian rhythm, and psychological stress.⁷⁷⁵ Derivative parameters such as normal resting HR and heart rate variability (HRV) can reveal cardiovascular status, with the HR rate reflecting long-term cardiovascular risk.^{748,776} HRV may be related to autonomic system integrity by analyzing fluctuations between adjacent heartbeat intervals. HRV abnormalities are associated with arrhythmias, heart failure, and the risk of sudden cardiac death.⁷⁷⁷ In sports science, maintaining training intensity within 70–85% of maximum heart rate can optimize cardiopulmonary adaptation and reduce the risk of myocardial overload.⁷⁷⁸ Heart rate monitoring is crucial for evaluating pharmaceutical efficacy. For instance, β -receptor blockers require dynamic dose adjustments based on patients' real-time heart rates.⁷⁷⁹ Thus, long-term, precise heart rate monitoring is key in preventive medicine and chronic disease management.

6.3.1. Electrocardiogram Monitoring. ECG is the gold standard for noninvasive heart rate determination.⁷⁸⁰ Blood pumping is regulated by coordinated contraction and relaxation of myocardial cells, which generate complex electrical patterns through bioelectric potential propagation.⁷⁸¹ ECG captures these electrical signals via electrodes. After the signal processing step, the heart rate is calculated by identifying cardiac cycles. Although ECG-based heart rate measurement is highly accurate, it is significantly affected by electrode adhesion and skin impedance.⁷⁸² The challenges of ECG signal acquisition are mainly due to motion interference and long-term wearability.⁷⁸³ During intense motion, the relative displacement between electrodes and skin and sweat accumulation may lead to motion artifact noise.⁷⁸⁴ Researchers

have developed epidermal dry electrodes with a porous structure and ultrathin thickness to address these issues.^{785–788}

These types of electrodes have the advantage of improved adhesion, breathability, wearing comfort, and long-term stability.⁷⁸⁹ Centers for ECG analysis obtained using wearable devices are arising, promoting more accessible and frequent healthcare.⁷⁹⁰

6.3.2. Photoplethysmography Monitoring (PPG). PPG achieves heart rate monitoring by detecting light intensity fluctuations caused by blood volume changes in the blood vessels.⁷⁹¹ The resulting waveform can be further analyzed for parameters such as systolic amplitude, pulse width, and peak-to-peak intervals, which are used to evaluate vascular disease and autonomic function.⁷⁹² With advancements in DL models, PPG signals can now be used to predict cardiovascular disease risks, expanding their diagnostic potential.⁷⁹³ However, the PPG signal quality is easily affected by motion artifacts, ambient light interference, and individual differences in skin reflectance.^{794,795}

6.3.3. Mechanical Sensing. When blood is pumped through the blood vessels, it induces mechanical waves and strain, which propagate throughout the skin. Detecting these mechanical waves and strain on the skin enables the capture of arterial pulsation signals or seismographic signals, indirectly inferring heart rate. Strain detection is primarily achieved through various mechanical (strain or pressure) sensors.⁷⁹⁶ Sensitivity and stability are the major concerns for this type of sensor.

6.3.4. Sensor Form Factors and Anatomical Considerations. The effectiveness of heart rate measurement largely depends on balancing the signal quality and wearing comfort. Traditional 12-LED ECG uses electrodes to capture cardiac electrical activity with extreme precision, but the large number of patches limits its practicality for daily applications. Detection methods based on blood flow can monitor arteries and capillary networks.⁷⁹⁶ To collect high-quality heart-rate-related signals, various sensors need to be stably worn on the




Figure 35. (A) Body locations for temperature sensing. Reproduced/adapted from Yu et al.⁸²² Available under a CC BY 4.0 license. © 2024 John Wiley & Sons. (B) Body temporal map before, after, and during the exercise. Reproduced/adapted from Alam et al.⁵¹⁵ © 2023 American Chemical Society. (C) Demonstration of integrated wearable T-sensor attached to skin and wound for continuous temperature monitoring.

skin surface. The typical wearable form factors include chest straps, wristbands, earpieces, clothing integration, and patch and electronic skin types:

6.3.5. Chest-Worn. Chest strap heart rate monitors, being close to the heart source, provide optimal signal quality with secure fixation and strong resistance to motion interference (Figure 34a).⁷⁹⁷ They are suitable for high-intensity exercise and precise monitoring.^{797,798} As one of the earliest commercialized heart rate monitoring devices, chest strap monitors use two or more electrodes applied to the chest skin to collect cardiac electrical signals, calculating heart rate after signal processing. These sensors possess high accuracy, high temporal resolution, and strong motion interference resistance, but as they require stable adhesion to the skin surface, they need adhesives or relatively high pressure, potentially causing skin discomfort with long-term wear.⁷⁹⁸ Modern chest-worn heart rate sensors are integrated with signal processing and wireless transmission modules into a compact patch (Figure 34d).^{789,799} The patch is attached to the skin using hypoallergenic adhesive, eliminating the need for preload pressure. This design enhances the long-term wearing comfort and stability.

6.3.6. Wristbands. Wristband heart rate sensors are known for their excellent portability and daily wearing comfort, with ease of use and high user acceptance.⁸⁰⁰ They are particularly suitable for daily activities and general fitness monitoring (Figure 34f).⁸⁰¹ As the most common form of heart rate monitoring today, wristband sensors are mainly used in smartwatches and smart bracelets. These devices can employ PPG or pressure sensors. PPG sensors have the advantages of convenience and comfort, while typically being integrated with accelerometers and gyroscopes to compensate and eliminate motion artifacts through algorithms.⁸⁰² Pressure-sensor-based wristbands capture strain produced by arterial pulsation; thus, they need to closely adhere to the wrist pulse area and achieve stable pressure sensing by applying preload pressure. By extracting key features of the pulse pressure waveform and

analyzing with preload pressure and skin temperature, additional physiological parameters such as blood pressure can be obtained, also ensuring good monitoring stability.⁸⁰³ However, because of its constraining effect on the skin, long-term wearing comfort presents challenges. The latest wristband heart rate sensors have begun to adopt multisensor fusion technology, combining PPG and pressure sensing and processing data using machine learning algorithms, which has improved monitoring accuracy and anti-interference capability.⁸⁰⁰

6.3.7. Ear-Worn. Ear-worn heart rate monitors feature minimal motion interference and excellent signal stability, making them ideal for during-exercise and daily continuous monitoring (Figure 34c).⁸⁰⁴ Ear-worn heart rate monitors are mainly integrated into wireless earphones or specialized ear-hook devices.⁸⁰⁵ Because the ear has rich capillaries and relatively small movement amplitude during daily activities, these devices provide signal stability significantly better than wrist monitoring, performing more reliably, especially during high-intensity exercise. The key advantage lies in seamless integration with daily audio equipment, offering a 'monitoring without awareness' experience. Additionally, they can simultaneously track other health indicators such as blood oxygen saturation.^{806–808} A major concern of this type of device is long-term comfort, especially the risk of otitis externa after wearing the earphones.⁸⁰⁹

6.3.8. Smart Clothing. Smart clothing with integrated heart rate sensors is designed to seamlessly blend with the skin, offering excellent comfort and wearability for long-term monitoring (Section 5.6).⁸¹⁰ These garments conform to large skin areas, providing stable monitoring and distributed signal redundancy, making them ideal for all-weather health management and professional sports training (Figure 34b).⁶⁵⁴ By integration of ECG, PPG, and pressure sensors, smart clothing enables multimodal monitoring, delivering comprehensive cardiovascular health data.⁸¹¹ The larger sensing area also supports multipoint sensing,⁸¹² capturing heart rate data




Figure 36. Advances in T-sensing device strategies based on sensing mechanism and materials engineering. (A) (i, ii) Effect of cooling/heating on PEO domains doped in PVDF capacitive layer with two-dimensional square hole structure and corresponding micrograph. (iii) Change in capacitance during cooling/heating cycles. (iv, v) Effect of different mesh sizes on repeatability and sensitivity. (vi) Flexibility of the capacitive layer of T-sensor. Printed with permission from Hu et al.⁸⁴³ © 2024 John Wiley & Sons. (B) (i) Schematic illustration of resistive T-sensor attached to the human skin and effect of PDA@rGO in thermal properties of ESM substrate simulation by finite elemental analysis. (ii, iii) Resistance and I–V response of bare ESM and doped ESM and corresponding structural changes in conductive pathways in the presence of graphene and PDA polymer during heat sensing. (iv, v) Response time of T-sensor from 30 to 40 °C and continuous response stability with five different temperature settings for 300 s. Modified from Zhang et al.⁸⁴⁴ © 2023 American Chemical Society. (C) (i) Wearable microfiber T-sensor described with PDMS bottom and PDMS/MoS₂ layer encapsulation. The detailed structure including optical microscope images. (ii, iii) Recorded spectrum and according shift of spectrum at different temperatures and linearity analysis. (iv) Response of T-sensor at forehead. Printed with permission from Wang et al.⁸⁴⁵ © 2024 Elsevier.

from different body parts and additional metrics like blood pressure. However, challenges remain in waterproofing, washability, and limited flexibility. Current solutions include washable electronic textiles such as conductive yarns, printed flexible circuits, and detachable modules.⁶⁵⁴

6.3.9. Electronic Tattoos. Electronic tattoos or epidermal electronic heart rate sensors achieve perfect skin conformity with ultrathin flexible structures, and the ability to be worn continuously for days without detachment (Figure 34e).^{750,813–815} These devices can be used in medical monitoring and continuous health data collection, employing

ultrathin flexible ECG electrodes to capture weak ECG signals or miniature pressure sensors to detect pulse waves.⁷⁵⁰ With thicknesses typically less than 10 μm and mechanical properties like or softer than skin (Young's modulus below 100 kPa), these sensors provide an imperceptible adhesion experience, reducing dynamic noise and providing high-quality signals. These parameters offer practical advantages, including long-term wear (>24 h) without skin discomfort. The ultrathin transparent nature leads to improved user compliance. Additionally, they can be customized in shape and function for different body parts and monitoring needs. Recent research has introduced biodegradable materials, allowing the sensors to decompose after completing their monitoring period without manual removal.^{816,817} Due to direct contact with human skin, epidermal electronics require high levels of biocompatibility and stricter regulation before being applied for commercial use.

6.4. Temperature Sensors

Temperature sensing is essential for monitoring physiological states, managing diseases, and ensuring the reliability of integrated sensors. An important parameter is the core body temperature (CBT), as it reflects health conditions and indicates physical, biological, or environmental stress.⁸¹⁸ An ideal CBT ranges between 36.5 and 37.5 $^{\circ}\text{C}$, with fever being detected if it rises to more than 37.5 to 38.3 $^{\circ}\text{C}$.⁸¹⁹ When infected with a virus or bacteria, the human body heats itself up to increase the immune cell activity.⁸²⁰ Inflamed parts or a wound will have its temperature increased, while blood circulation passes this heat to the skin, which is dissipated by sweating. Hence, the skin temperature is also related to CBT. Various body locations can be employed for CBT measurements based on requirements, comfort, or feasibility. Figure 35A shows locations for invasive and noninvasive measurements on the human body. According to medical science guidelines, the temperature of the pulmonary artery is the most accurate location to measure CBT, but it is invasive.⁸²¹ Oral, nasopharynx, esophagus, digestive track, urinary, and rectal are considered alternative invasive CBT locations.⁸²² Oral use is the most practiced in daily routine due to its feasibility. In contrast, the noninvasive methods of skin temperature are not accurate for determining CBT, though they are still suitable for detecting major body temperature changes. The skin over the forehead is considered the most accurate, compared to the axilla and other skin locations.⁸²³ Figure 35B shows body temperature maps for a person, before and after heating with physical exercises. Distinct temperature distributions across the body from 32.5 to 36.5 $^{\circ}\text{C}$ are detected.⁵¹⁵

Skin temperature provides a dynamic interface between the internal metabolic activity and external environmental conditions. Unlike CBT, which is relatively stable, the skin temperature fluctuates rapidly, reflecting blood flow, sweat evaporation, and thermal regulation processes. This variability makes skin temperature both a critical health marker and a parameter that must be carefully accounted for in wearable biosensors. Monitoring skin temperature enables the early detection of health issues. For example, fever, heat stress, or inflammation can manifest as changes in the peripheral temperature before other symptoms appear. Skin temperature monitoring is also valuable in chronic disease management, where conditions such as diabetes or vascular disorders influence local thermal profiles. Additionally, skin temperature is important for calibration, particularly for biosensors that are dependent on enzymatic reactions.

Continuous CBT monitoring has been performed for skin and wounds with wearable temperature sensors (T-sensors), as shown in Figure 35C.⁸²⁴ Such sensors are integrated with the Internet of Things (IoT) to collect long-term temperature profiles of patients and notifications to physicians for on-time and continuous treatments.^{825,826} To improve the safety standards, attempts were made to integrate sensors with energy storage device patches.⁸²⁶ Wearable T-sensors are now commercially available, with the heat response transformed into electronic or optical signals through capacitive, resistive/thermistor, thermoelectric, optical, pyroelectric, and colorimetric processes.^{827–832} Materials used in these sensors include metals and metal-oxide materials, semiconducting polymers, carbon-based materials/graphene, 2D materials, hydrogels, ionogels, and hybrid composites.^{833–838} The sensing units must be flexible, stretchable, and mechanically stable, thus requiring suitable substrates. The latter may be polyimide, PDMS, polyethylene naphthalate, thermoplastic polyurethane, textiles, and cellulose-based substrates.^{839–842}

Figure 36 summarizes recent reports of materials engineering to enable flexible wearable T-sensors with different sensing mechanisms. Hu et al. utilized thermotropic composites made with poly(ethylene oxide) (PEO)/PVDF/ H_3PO_4 as a dielectric layer in capacitive T-sensors.⁸⁴³ PEO acted as a thermotropic dopant to induce ion dissociation and affect the electric double-layer capacitance when the temperature changed, as depicted in Figure 36Ai. This doping strategy leads to highly disordered PEO, thus hampering device performance. Using two-dimensional square holes shown in Figure 36Aii helped to obtain ordered PEO, with which reproducible sensors could be achieved with reduced error performance (<2.2%). Figure 36Aiii–v illustrates the high performance for the flexible sensors shown in Figure 36Avi, which had a resolution of 0.05 $^{\circ}\text{C}$ and response speed (<12 s) within 35–43 $^{\circ}\text{C}$. Resistive sensors are also widely employed, especially with graphene. Zhang et al. produced a wearable T-sensor with polydopamine/graphene composite (PDA@rGO) with an eggshell membrane (ESM) to ensure flexibility and breathability.⁸⁴⁴ Figure 36Bi shows the heat simulation of ESM with and without PDA@rGO on skin (37 $^{\circ}\text{C}$), which confirms the high thermal conductivity sensing contribution from the composite. The resistive sensing mechanism and resistance change over 30 to 42 $^{\circ}\text{C}$ are described in Figure 36Bii,iii. Upon introducing PDA and rGO, the temperature coefficient of the resistance increases. The sensing response and stability at various temperatures are depicted in Figure 36Biv,v, with the best T-sensor exhibiting high sensitivity and resolution, good linearity, and rapid response (4–8 s).

Wearable optical T-sensors allow for continuous monitoring.⁸³¹ These can be achieved, for example, by using optical fibers whose optical properties depend on the temperature. These sensors are immune to electromagnetic interference, which makes them robust and accurate.⁸⁴⁶ Figure 36Ci shows a T-sensor encapsulated in PDMS and MoS_2 with a smaller diameter in the middle to provide a mismatch to the optical path.⁸⁴⁵ When tested between 30 and 45 $^{\circ}\text{C}$, the T-sensor spectrum red-shifted with increasing temperature (Figure 36Cii). The sensitivity was 1.10 nm/ $^{\circ}\text{C}$ with linearity approaching 0.99. The blue-shift analysis also achieved a linearity of 0.96 in Figure 36Ciii. The demonstration of a wearable sensor on the forehead was made to monitor temperature from 37 to 42 $^{\circ}\text{C}$, which matches the performance of commercial mercury T-sensors (Figure 36Civ). With its




Figure 37. Advanced strategies for flexible designs and integrated self-powered T-sensors. (A) (i) Schematic diagram of the T-sensor unit by the 3D direct writing method, polygonal design, components, and its flexibility, (ii) T-sensing performance at different strain conditions. (iii) T-sensing performance while rest and exercise. Reproduced with permission from Wang et al.⁸⁴⁷ © 2024 Elsevier. (B) (i) Schematic of the compact band for continuous monitoring of temperature on the forearm with a schematic of T-sensor with TEG self-powered unit. The integration components, demonstration of signal measurement and power management. (ii) Sensing performance of T-sensor mounted on forearm. (iii) Calculated and average CBT determination. Modified from Khan et al.⁸⁵¹ © 2024 American Chemical Society. (C) (i) Concept of integrated health monitoring incorporating T-sensor with self-powered stretchable power unit. (ii) Schematic of sensor unit on laser-induced graphene substrate, TENG with T-sensor and system working mechanism. (iii) Step-by-step nanofabrication of sensor components. (iv) Dynamic sensing response and calibration plot of the T-sensor. Modified from Ding et al.⁸⁵² © 2024 American Chemical Society.

fiber shape, low density, and high strength, this T-sensor can be integrated into wearable patches and facemasks to serve for oral temperature monitoring.

The fabrication process for a strain-insensitive wearable T-sensor is shown in Figure 37Ai.⁸⁴⁷ The sensor was obtained by direct-ink-writing and three-dimensional printing of PE-




Figure 38. (a) Application of T-sensor integrated in smart textile for handicap patient's health monitoring and thermal management. (b) Mounted smart textile band and PCB on forearm. (c) Infrared image of thermal regulation by heating textile. (d, e) Measured body temperatures and pulse waveforms simultaneously and their minimized cross effects. (f, g) Validation of closed-loop system on the leg of two volunteers, with measurements of temperature and pulse wave wirelessly on mobile phones. (h, i) The recorded typical body temperature and pulse waveform of person 1 and person 2. (j, k) Thermoregulated temperature by the system. (l) Possible applications of the smart-textile-based wireless closed-loop system. Reproduced/adapted with permission from Zhang et al.⁸⁵³ © 2024 John Wiley & Sons.

DOT:PSS and graphene ink on poly(vinyl alcohol). A polygonal design allowed for a strain-insensitive measurement of the temperature, as indicated in Figure 37Aii and Figure 37Aiii. The sensitivity was $0.01\text{ }^{\circ}\text{C}^{-1}$ in the range of $20\text{--}50\text{ }^{\circ}\text{C}$, and $0.001\text{ }^{\circ}\text{C}^{-1}$ in the range of $55\text{--}75\text{ }^{\circ}\text{C}$. Other designs, such as those obtained with micropatterning and kirigami approaches, are efficient in achieving flexible T-sensors.⁸⁴⁸ Furthermore, these sensors can also be based on thermoelectric, piezoelectric, and colorimetric principles of detection.^{849,850}

Self-powered T-sensors were produced with the incorporation of a thermoelectric generator (TEG) and a battery, as shown in Figure 37Bi (see Section 10 for more details on self-powered devices).⁸⁵¹ The powering unit was connected to a commercial T-sensor mounted on a flexible printed circuit board, targeted to monitor CBT on the forearm based on arterial blood heat reaching the skin. In this setup, the human volunteer was stabilized in a temperature-controlled room ($25\text{ }^{\circ}\text{C}$) to prevent thermoregulation. Figure 37Bii,iii shows the heat flux and an average skin temperature of $31.3\text{ }^{\circ}\text{C}$ for the volunteer measured for 600 s. The calculated and expected

CBT were 36.55 and $36.50\text{ }^{\circ}\text{C}$, respectively. Another integrated self-powered T-sensor is shown in Figure 37Ci, made with graphene/ Co_3O_4 nanocomposite. Figure 37Cii describes the strategy of TENG for power generation, micro supercapacitors for storage and power supply, and the T-sensor for CBT monitoring.⁸⁵² The ink printing process is depicted in Figure 37Ciii, while the sensing response with a sensitivity of $0.074\%\text{ }^{\circ}\text{C}^{-1}$ and a LOD of $0.042\text{ }^{\circ}\text{C}$ is shown in Figure 37Civ.

Most integrated systems utilize thermosensitive sensors that may be affected by changes in the strain or humidity. This limitation was overcome with the development of a fiber-based flexible T-sensor in a smart textile.⁸⁵³ A wireless closed-loop system was employed for decoupled multimodal health monitoring (T-sensing and human pulses) and personalized thermoregulation for precise pulse measurements. The T-sensor made with $\text{rGO}/\text{MnO}_2/\text{Kevlar}/\text{TPU}$ (thermoplastic polyurethane) is illustrated in Figure 38a. The measurements of pulse waves and temperature with the T-sensor of Figure 38b are shown in Figure 38c–e. The application of the system on the leg of two volunteers monitored wirelessly with mobile phones is depicted in Figure 38f–k. This type of T-sensing




Figure 39. (A) Schematic of the sensor network illustrates the configuration of the rGO temperature sensor, which is connected between pads A and B. As the temperature increases, the resistance value of the sensor decreases. In contrast, the silver sensor between pads B and D exhibits an increase in resistance value with rising temperature. Additionally, a humidity sensor, comprising a silver interdigitated electrodes, is positioned between pads C and D. This sensor assesses the alteration in conductivity of the substrate. Furthermore, the electrodes utilized for motion and seizure detection are situated between pads E and F. (B) Thermal images of the inkjet-printed device applied to the skin, captured at room temperature and during hot air exposure for calibration purposes. (C) The resulting series data for both resistors connected. The calibration curves comparing the response of the silver nanoparticle meander resistor and the rGO serpentine resistor. Reprinted with permission from Maroli et al.⁸⁵⁹ © 2024 Elsevier.

platform can be extended to multiple healthcare applications such as geriatrics, postpartum care, rehabilitation, and sports, as shown in Figure 38I. Other advances allow for monitoring of wound healing and ongoing treatments such as the diagnosis of kidney necrosis.^{854,855}

Wei Gao and collaborators demonstrated how integrating temperature sensing into sweat-based platforms improved the reliability of biomarker detection by accounting for temper-

ature-dependent aptamer sensors.⁸⁵⁶ These wearable devices employ various types of temperature sensors including thermistors, resistive temperature devices (RTDs), and IDEs. Thermistors, with their negative temperature coefficient (NTC) behavior, are widely used because of their simplicity and high sensitivity. However, their nonlinear resistance–temperature relationship can complicate calibration. IDEs leverage capacitance or resistance changes in response to the

temperature and are often used in multifunctional sensor platforms. RTDs, such as silver- or graphene-based devices, provide precise temperature measurements with adaptable designs, making them suitable for flexible and wearable applications.

Maroli et al.⁸⁵⁷ proposed the approach depicted in Figure 39 to wearable temperature sensing through an RTD using rGO. These resistors are characterized by their linear negative temperature coefficient, distinguishing them from conventional thermistors that exhibit a nonlinear response. The linearity is a significant advantage for wearable systems, simplifying calibration and ensuring more consistent temperature measurements across a wide operational range. The design optimizes the sensing area while maintaining a compact form factor, enhancing its sensitivity to thermal changes. The linearity of the rGO resistors enables their combination with other materials, such as silver-based resistors, to create composite systems that do not experience changes in their resistance value with temperature. The combination of these materials yielded a resistor whose value exhibited a variation of less than 0.22%, falling within the instrument's specified error range, across a temperature range of 40 to 100 °C. By leveraging the differing thermal coefficients of these materials, a methodology for compensating for temperature fluctuations was demonstrated, thereby ensuring stable and reliable sensor performance even in dynamic environments. In the latter work, inkjet printing was employed to fabricate the silver-based meander resistors, thereby enabling the precise deposition of conductive materials onto flexible substrates. This method is conducive to scalability and cost-effectiveness, rendering it an optimal choice for wearable technologies. As reported by Scroccarello et al.,⁸⁵⁸ the rGO sensor was fabricated by laser-induced reduction of graphene oxide and then transferred to the substrate using a press. The temperature values obtained with both sensors were found to be in good agreement with those observed with the thermal camera.

In summary, T-sensors are essential tools for wearable healthcare monitoring, with the area having advanced significantly over the past few years. One of the main potential challenges in developing such devices is the moderate accuracy of measuring CBT on the skin, which is often not as accurate as in invasive body locations. The skin temperature may be impacted by sweating and the outer environment temperature or humidity, which reduce the effectiveness of measured signals. Considering health monitoring, T-sensors also need high-end integration with electronics, IoT, and energy storage devices. This integration results in high costs due to the use of nanofabrication techniques and compact prototyping. Further, battery-utilizing devices bring health or safety hazards, whereas safe and biocompatible aqueous energy storage devices are power-deficient. As demonstrated, self-powered T-sensing modules are quite space-consuming as a unit and very specific in their body location for CBT sensing. The use of electronic components brings electromagnetic radiation interference and radio frequency emission, which are considerable challenges. Hence, further research should focus on developing hazard-free materials, safe energy storage methodologies, low-cost engineering, and devices with low-radiation interference.

7. DESIGN OF WEARABLE DEVICES FOR HEALTHCARE APPLICATIONS

7.1. Adhesion of Wearable Sensors to Human Skin

The functionality of many wearable sensors depends largely on the stable adhesion between the device and the target surface (e.g., the human skin).^{860–863} Without tailored skin-sensor interfaces, wearable sensors and devices are prone to motion artifacts, compromising reliable signal acquisition and user satisfaction (Figure 40).^{864–866} Secure device fixation is the




Figure 40. Schematic illustration of the relationship between adhesive types, wearable device functions, and requirements for stable adhesion for reliable sensing.

primary motivation because even minor sensor displacements can introduce significant noise into physiological measurements and continuous monitoring.^{867,868} Achieving stable adhesion requires careful consideration of several key aspects. First, conformability, the ability of the adhesive layer to follow the complex topography of the skin, is essential for maintaining intimate contact and minimizing interfacial impedance. Second, biocompatibility is required to prevent adverse skin reactions, including irritation, allergic responses, or damage during removal. Third, interference must be minimized; the adhesive should not introduce mechanical or electrical disturbances that compromise signal quality. Finally, exposure to sweat, water, temperature fluctuations, desquamation, and dynamic strains can substantially degrade adhesive performance.

Various material strategies have been explored to address these requirements. Hydrogel adhesives provide excellent adhesion and biocompatibility, but challenges remain regarding their mechanical durability and degradation under harsh environmental conditions. Dry adhesives, including medical pressure-sensitive adhesives (PSAs), offer reversible adhesion and are particularly suited for applications that require repeated attachment and detachment. Self-adhesive and ultrathin sensors enable intimate skin contact, allowing for seamless skin integration.

7.2. Conformal Contact of Wearable Devices on Substrates

Conformal contact refers to the intimate, gap-free interface between a device and a complex, dynamic skin surface.⁸⁶⁹ This




Figure 41. (A) Mechanism of progressive delamination of the electrode from the skin. Reprinted with permission from Shin et al.⁸⁷⁰ © 2024 John Wiley & Sons. (B) Hydrogel adhesive interface capable of providing anisotropic out-of-plane swelling, rapid, and robust adhesion on wet tissues. Reprinted with permission from Deng et al.⁸⁷¹ © 2020 Springer Nature. (C) Comparison of RMS noise values between commercial (black), PPd (red), Au (blue), and pristine PEDOT:PSS (green) electrodes and EMG signals during standing, walking, and running. Reprinted with permission from Shin et al.⁸⁷⁰ © 2024 John Wiley & Sons. (D) Comparison of the cell viability between the control group and the experimental group. Reproduced from Kim et al.⁸⁷² Available under a CC BY-NC license. © 2024 American Association for the Advancement of Science. (E) (Top) Schematic of surface-coating with/without causing VOCs and (bottom) lung histopathology analysis of mice after 21 days of exposure to adhesives. Hematoxylin/eosin staining and Masson's trichrome staining. Reproduced from Choi et al.⁸⁷³ © 2022 American Chemical Society.

attribute is essential for reducing skin-electrode impedance, minimizing motion artifacts, enhancing signal fidelity, and

improving the user comfort. A recurring challenge in wearable bioelectronics arises from the mechanical mismatch between

Robustness under various environmental conditions




Figure 42. (A) Long-term stability of ECG signal of ionogel electrode in the aquatic environment. Reproduced with permission from Yu et al.⁸⁸⁵ © 2021 John Wiley & Sons. (B) Comparison of hydrogel and organo-hydrogel after 24 h exposure to various environments and relative resistance responses of organo-hydrogel under different temperatures. Reproduced from Han et al.⁸⁸² Available under a CC BY 4.0 license. © 2025 John Wiley & Sons. (C) Stable adhesion of PSAs over million cycles of loading. Reproduced from Jeon et al.⁸⁸⁶ Available under a CC BY 4.0 license. © 2024 American Chemical Society.

soft, curvilinear biological surfaces and the typically planar, rigid, or semiflexible materials used in sensors.^{865,869} Dynamic skin deformation during daily activities, combined with the presence of sweat, hair, and sebaceous secretions, can disrupt electrode-skin interfaces, resulting in increased impedance, baseline drift, and signal degradation. Therefore, strategies to enhance conformal contact are needed for advancing the functionality and reliability of the wearable sensors.

Conventional metal electrodes are an example of devices that exhibit limited stretchability and often fail to conform to dynamic deformations of the skin. This mechanical sensor-skin incompatibility can lead to localized stress concentration,

promoting gradual microscale separation that ultimately progresses into permanent detachment under repeated deformations (Figure 41A).⁸⁷⁰ On the other hand, hydrogel materials and interfaces can achieve conformal contact with skin by absorbing moisture and swelling, allowing them to fill microscale gaps (Figure 41B).⁸⁷¹ This swelling-induced adaptation improves adhesion and reduces interfacial impedance, enhancing the mechanical stability and electrical performance of wearable sensors. Hydrogels with elastic properties matching those of skin offer a soft, compliant interface that maintains intimate contact even on irregular or dynamic skin surfaces. Among recent examples, Shin et al.

developed an ultrathin polymeric conductive adhesive (PPd) electrode by incorporating PVA and d-sorbitol into PE-DOT:PSS to modulate elasticity and adhesion.⁸⁷⁰ The PPd electrode minimized motion artifacts and connector-induced noise, ensuring stable signal acquisition under dynamic conditions during walking and running (Figure 41C).⁸⁷⁰

A similar focus on adhesion and mechanical compatibility is reflected in the work by Zhang et al., who developed a ~10- μ m-thick mesh-reinforced hydrogel sensor for continuous wearable bioelectronics.⁸⁷⁴ The ultrathin geometry of the hydrogel minimized interfacial air gaps, thus reducing skin-electrode impedance and improving the signal-to-noise ratio. The role of hydrogel interfaces was further exemplified by Deng et al., who introduced an electrical hydrogel adhesive match by utilizing a graphene nanocomposite hydrogel layer.⁸⁷¹ The adhesive formed robust, electrically conductive bonding with wet dynamic tissues, overcoming the limitations of van der Waals-based attachment or suturing. Kim et al. addressed conformal contact through multiscale physical adhesion mechanisms by creating an electrostatic-mechanical synergistic elastomer-hydrogel composite.⁸⁷⁵ Inspired by natural suction-based adhesion, these structures combined capillary and hydrogen-bonding interactions to achieve strong adhesion on wet tissues.

Overall, diverse strategies in achieving conformal contact—including ultrathin geometries, adhesive and hydrogel adhesive, and multiscale microstructures—demonstrate converging solutions toward this goal across multiple applications for long-term, high-fidelity health monitoring in a dynamic environment.

7.3. Biocompatibility Issues

Interface materials, such as adhesives, influence the physiological interaction with tissues, with biocompatibility being determinant for device safety.⁸⁶¹ Ensuring biocompatibility in wearable sensors requires careful selection of materials that minimize adverse biological responses.⁸⁷⁶ Adhesive and conductive components must be engineered to maintain skin safety while providing stable mechanical and electrical performance. In addition, the device's structural design, including penetration depth and interface mechanics, should align with the skin properties. Choi et al. demonstrated that the gel adhesives exhibits excellent biocompatibility by eliminating the release of VOCs, which are common causes of skin, eye, and respiratory irritation (Figure 41E).⁸⁷³ These VOC-free materials minimize potential toxicity and irritation during prolonged skin contact. For epidermal electrophysiological monitoring, Kim et al. have developed a stretchable micro-needle adhesive patch utilizing an electrically conductive adhesive composed of silver flakes and silicone (Figure 41D).⁸⁷² The materials exhibited no cytotoxic effects, confirming their intrinsic biocompatibility for skin contact applications. Microneedles penetrate only the stratum corneum without reaching pain receptors, minimizing irritation.

In implantable devices, Wu et al. provided evidence that an adhesive interface can prevent fibrous capsule formation by reducing the level of inflammatory cell infiltration.⁸⁶¹ These adhesive layers achieved conformal integration with diverse organ surfaces, eliminating observable fibrosis over 12 weeks and serving as active mediators of immune tolerance. Overall, designing adhesives for wearable bioelectronics requires a delicate balance of adhesion strength, mechanical compliance,

electrical conductivity, and biological tolerance for clinical, daily, and implantable applications.

7.4. Robustness under Various Environmental Conditions

Wearable biosensors must withstand a variety of conditions to ensure constant robust and reliable measurements. Biofluid accumulation, temperature changes, the presence of water and applied stress are some of the challenges. Sweat accumulation at the skin–sensor interface is critical, as it increases interfacial impedance, degrades adhesion, and introduces noise and signal drift as well as sweat-induced discomfort and skin irritation.^{874,877} To address these problems, Zhang et al. considered sweat removal by developing a three-dimensional liquid diode structure that actively transports sweat away from the skin through directional liquid transport channels under heavy sweating.⁸⁷⁷ Using a different approach, Wang et al. introduced a Janus membrane-based pH sensor combining a hydrophobic porous substrate with a hydrophilic nanofiber layer to achieve simultaneous sweat absorption, gas permeability, and self-adhesion.⁸⁷⁸ This configuration allowed reliable pH monitoring for over 7 h without detachment or discomfort. Lin et al. incorporated gel-free soft encapsulation and anisotropic conductive films into a wearable ultrasound system to maintain adhesion and coupling in moist environments without relying on traditional coupling gels.⁸⁷⁹ This design supported continuous deep tissue monitoring while minimizing moisture interference.

Regarding water resistance, Yu et al. developed a water-resistant ionogel electrode for underwater ambulatory ECG monitoring, enabling stable physiological signal detection in aquatic environments (Figure 42A). The ionogel achieves underwater adhesion by disrupting the hydration layer through its hydrophobic polymer network, allowing strong interfacial interactions even in wet conditions. Xiang et al. also created water-resistant devices, introducing a multifunctional conductive hydrogel combining dipole–dipole and hydrogen bonding interactions for underwater sensing.⁸⁸⁰ Their hydrogel material maintained stable adhesion, high conductivity, and EEG signal quality even under aquatic immersion, demonstrating robust wet-environment compatibility. As for dry-adhesive materials, Zhou et al. designed a bioinspired adhesive fabric by embedding LM beads within a PLA fiber network to enable reversible adhesion and photothermal responsiveness while maintaining strong bonding and excellent stability in humid or wet environments.⁸⁸¹

Adhesives must maintain conformal contact, adhesion strength, and functional integrity despite environmental extremes such as freezing or heat, to ensure continuous signal acquisition and user comfort. For these applications, Han et al. developed an ionic conductive hydrogel that maintains strong adhesion even at $-80\text{ }^{\circ}\text{C}$, enabled by robust dipole–dipole interactions, hydrogen bonding, and ionic bonding with various substrates (Figure 42B).⁸⁸² This stable adhesion under extreme cold ensures continuous skin contact, preventing signal degradation and device detachment. Shang et al. developed a mussel-inspired organo-hydrogel with robust wet adhesion and environmental tolerance ranging from -40 to $60\text{ }^{\circ}\text{C}$.⁸⁸³ The adhesive layer combined catechol chemistry and hydrophobic interactions to prevent adhesion failure. In another example, Zhou et al. reported glutinous-rice-inspired adhesives for antifreezing and moisture retention for 90 days.⁸⁸⁴

Finally, stable adhesion under dynamically applied stress also must be considered. Jeon et al. evaluated the vibration stability of double-sided PSAs under cyclic circular loading at 5 Hz and 2 N for up to one million cycles to mimic complex human body motions (Figure 42C).⁸⁸⁶ The PSAs exhibited minimal displacement without adhesive failure, maintaining mechanical integrity throughout the test. These findings demonstrate the long-term mechanical durability and reliable adhesion performance of PSAs for wearable devices in dynamic vibrational environments. Furthermore, Wang et al. achieved vibration-resistant adhesion by integrating mushroom-shaped microtips, mechanically isolated adhesive units, and a porous backing layer.⁸⁸⁷ This multiscale design enabled stress decoupling and energy absorption, ensuring stable adhesion under dynamic conditions.

In summary, the adhesive interfaces for wearable sensors are critical for maintaining a stable device performance under practical conditions. For sweat management, permeable structures or ultrathin adhesive layers are employed to enhance water vapor transmission and minimize moisture accumulation at the skin interface. In underwater applications, hydrogel-based adhesives provide reliable adhesion by retaining interfacial bonding even in wet conditions while also offering mechanical compliance. Similarly, hydrogels can maintain adhesion across a wide temperature range, ensuring conformal contact under limited moisture thermal variations, but can suffer from decreased electrical or mechanical performance over time. Dry adhesives may exhibit reduced strength under high humidity or heat, which is critical for maintaining a high sensitivity. Therefore, future strategies must address this trade-off by integrating materials and designs that simultaneously ensure robust adhesion and functional stability across a diverse range of environments.

8. SOFT ELECTRONICS

This topic is essential for wearable sensors since many applications involve embedded electronics, which should preferably be produced with soft electronics for seamless integration across on-body and in-body platforms. The field of soft electronics emerged from the transition from early electronics with inflexible integrated circuits (ICs) fabricated on glass or wafer substrates⁸⁸⁸ to organic semiconductors, plastic substrates, and printing techniques.^{889–891} Softness and flexibility are insufficient, however, for applications that require interfaces with living organisms. In such cases, electronics must also stretch, as defined by elastic responses to large strain deformations associated with non-Gaussian curvatures. These opportunities motivated research on materials and approaches to platforms that can accommodate uniaxial and multiaxial strains often exceeding 50% without compromising electrical performance.⁸⁹² Successful outcomes of these early efforts are in bioinspired, including hemispherical “electronic eye” cameras⁸⁹³ and skin-like, or “epidermal”, sensor systems.⁸⁹⁴ These devices retain function under complex deformations, enabling reliable operation on dynamically moving and deforming surfaces.

As shown in Section 7, to achieve accurate signal acquisition in wearable sensors, conformal contact between the electrode and the tissue is required, as signal quality diminishes with increasing distance between them (Figure 43). According to Coulomb's law, any geometric gap reduces the electric field's intensity significantly. Thus, minimizing this gap through the selection of appropriate materials and fabrication methods is




Figure 43. Impedance comparison between rigid and soft electrodes interfacing with tissues, along with their corresponding equivalent circuit models. R_{ei} : Electrode resistance, R_{Ti} : Tissue resistance, R_e : Interfacial resistance, C_e : Interfacial capacitance.

vital for achieving stable and effective bioelectronic performance. Furthermore, the intimate contact between the electrode and tissue can enhance the capacitance significantly, favoring the resistive equivalent circuit and shifting the phenomena in the interface from capacitive- to ohmic-dominated (simplified equivalent circuit models can be seen in Figure 43). Such an aspect also contributes to lowering the impedance of the system, potentially improving device performance. Electrodes should also conform to soft biological tissues without causing damage. Rigid components can induce inflammatory responses and scarring, leading to the migration and accumulation of astrocytes and microglia around the electrodes. This forms a 50–200 μm thick sheath, separating the recording site from the target tissue.^{895–897}

Soft and stretchable electrodes can be fabricated by using two primary strategies: (1) engineering rigid materials into geometrically deformable structures or (2) utilizing intrinsically soft and stretchable materials combined with conductive fillers. Soft and stretchable electrodes achieve mechanical compliance through structurally deformable designs, allowing conducting materials to stretch while maintaining an electrical performance. In contrast, intrinsically soft and stretchable electrodes leverage elastic polymers mixed with inherently soft conductive fillers, such as metal nanomaterials, LMs, and organic conductive materials. Their mechanical compliance can be further enhanced through additional structural and polymer engineering, as depicted in Figure 44A.

Conducting materials govern the electrical performance of stretchable devices and can include thin metal films, LMs, inorganic conductive fillers, and organic conductive fillers. Thin metal films offer high conductivity ($\sim 10^6$ S/cm) but are inherently brittle under lateral strain (while remaining flexible under vertical strain), requiring structural engineering to enhance mechanical compliance. Nanoscale inorganic conductive fillers (e.g., nanoparticles (NPs), NWs,⁸⁹⁸ CNTs,⁸⁹⁹ graphene, and nanosheets^{900,901}), provide tunable conductivity depending on their dispersion and alignment within soft elastic polymer matrices. LMs,⁹⁰² discussed in depth in Section 8.3, exhibit both high conductivity and deformability, and can be integrated with elastic polymers or used to form electrodes within elastic channels.⁹⁰³ Organic conductive fillers (e.g., PEDOT:PSS, PANI, and PPy) have lower conductivity ($\sim 10^{-2}$ to 10 S/cm) than metal films, LMs, and inorganic conductive fillers, but possess both ionic and electronic conductivity, making them particularly useful for bioelectrodes.

Soft polymeric materials include elastomers and gels. Elastomers, such as Ecoflex, PDMS, polyurethane (PU), and styrene-ethylene-butylene-styrene (SEBS), exhibit high elasticity ($>300\%$), robust covalent bonding, and excellent chemical resistance, allowing them to withstand repeated deformation and maintain long-term durability under chemically harsh environments without mechanical failure.⁹⁰⁴ In contrast, gels,




Figure 44. Materials and fabrication strategies for nanomaterials-based stretchable electrodes. (a) Schematic diagram illustrating various materials and fabrication strategies. (b, c) Stretchable electrodes using specially designed geometries. (d, e) Stretchable electrodes using cracked metal thin films. Intrinsically stretchable electrodes by blending conductive fillers and (f–h) elastomer, (i–k) gel, and (l, m) intrinsically stretchable electrodes by engineering polymer chains. (n) Conductivities and stretchabilities of stretchable electrodes fabricated via diverse materials and fabrication strategies.

including hydrogels⁹⁰⁵ and organogels,⁹⁰⁶ inherently exhibit high porosity and relatively weak mechanical properties, but offer high ionic conductivity, extremely soft mechanical properties, and high cell permeability. In particular, the exceptional softness of hydrogels (~kPa) allows for conformal contact with biological tissues, minimizing interfacial impedance and ensuring stable charge transfer at bioelectronic interfaces. Additionally, modifying polymer networks allows hydrogels to incorporate additional functionalities such as adhesiveness and self-healing capabilities, improving durability and prolonged operational stability.⁹⁰⁷

A summary of the characteristics of the most used materials for soft electronics is shown in Table 1.

Next, we review the two complementary strategies for achieving stretchability in soft electronic systems: those based

mainly on structures and those based mainly on materials. Together, these schemes provide routes to a full range of stretchable components, including electrodes, transistors, batteries, photodetectors, diodes, and LEDs, in some cases with integration densities and levels of performance approaching those of rigid systems.

8.1. Structural-Oriented Approaches

Various structural design strategies achieve soft electronics by combining hard, brittle electronic materials in strategic geometrical designs with soft elastomeric substrates and encapsulation layers. These approaches allow the resulting hard/soft constructs to maintain electrical integrity while responding elastically to large strain deformations. Traditional materials such as silicon (Si), germanium (Ge), and gallium

Table 1. Comparison of Electrode Materials for Wearable Electronics Applications

	Materials	Advantages	Disadvantages	Processing methods
Flexible Materials	Thin solid metal layer	High conductivity Compatible with existing processes	Rigid Limited flexibility Crack formation under strain	Photolithography
	Graphene	High conductivity Flexibility Biocompatibility	High cost Challenging large-scale production	Photolithography Soft lithography
	Silver nanowires/nanofibers (AgNW/NF)	High conductivity Excellent flexibility Transparency	Prone to oxidation Reduced durability under deformation	Electrospinning Photolithography
Soft Materials	Conductive polymers	Biocompatibility	Lower conductivity compared to metals	Photolithography
		Good mechanical properties	Limited durability under deformation	Solution casting
	Hydrogels	Stretchability	Sensitive to moisture	Solution processing
		Stretchability Biocompatibility	Limited fabrication methods; Low durability	3D printing
	Liquid Metals	Wet adhesion High conductivity Extreme softness Biocompatibility	Difficult handling	3D printing Injection molding

arsenide (GaAs) become highly flexible when rendered in thin geometries, simply due to the cubic dependence of the bending rigidity on thickness and an inverse dependence of the maximum strain on thickness for a given bending radius.⁹⁰⁸ Thin-film deposition and patterning methods can yield flexible nanomembranes, nanoribbons, nanowires, and nanomeshes in configurations that allow their transfer printing onto elastomers. In planar geometries, integration in this manner allows strains of only ~1%, limited by the fracture thresholds of the inorganic materials.⁹⁰⁹ To achieve stretchability, these thin materials can be configured into various geometries before or during the integration process, allowing effective end-to-end strains of the resulting structures of up to 100% or more without mechanical failures. Common designs include wavy, serpentine, helical, polygonal lattice, origami, and kirigami layouts (Figure 44B,C).

The formation of wavy structures follows from ultrathin films fully or selectively bonded to an elastomer substrate in a state of prestretch, followed by release. This buckling procedure imposes compressive forces on the films, thereby causing them to transform spontaneously into well-defined wavy shapes, sometimes assisted by thermal expansion or moisture-induced swelling.^{910–912} In advanced versions, this transformation of the initially planar structures yields well-defined 3D architectures. After full encapsulation, the results of this process can form the basis for soft electronics. In optimized layouts, these structures can sustain, for example, uniaxial stretching of up to 120% with an elastic response and an effective modulus comparable to that of the elastomer with little mechanical loading associated with the hard material component.⁹¹³ Other two-dimensional designs exploit island-bridge architectures, where rigid “islands” support the active components and flexible “bridges” form electrical interconnects to absorb mechanical strain.^{914,915} These structures can be formed by photolithography, transfer printing, multidirectional writing, laser ablation, and/or conformal coating, with the ability to establish robust adhesion to elastomeric substrates.^{916,917}

Among these options, serpentine interconnects, composed of periodically repeating half-circular arcs, enable high

stretchability through both in-plane rotation and out-of-plane buckling, especially when the traces are narrow, thin, and long.^{918,919} Using such structures, circuits can exhibit less than 0.2% maximum principal strain under an applied strain of 90%.⁹²⁰ Through deterministic compressive buckling methods similar to those mentioned previously, these 2D serpentine can transform into 3D helical structures, which more effectively distribute mechanical stress and reduce failure risk compared to their planar counterparts.^{921,922} Polygonal lattice and self-similar structures further support high-density integration of functional components while maintaining mechanical compliance.⁹²³ For example, a rotationally symmetric cellular lattice with 60% porosity and optimized arc angles can withstand up to 100% tensile strain under an external stress of 50 kPa while maintaining a low elastic modulus (~0.5 MPa).⁹²⁴ Designs inspired by the ancient arts of origami and kirigami can further extend mechanical compliance through folding and cutting techniques. Origami involves the folding of 2D sheets along predefined creases to achieve 3D structures that isolate integrated functional components from strains associated with deformations.^{925,926} Examples of this scheme are in stretchable lithium-ion batteries, soft robotics, and wearable sensors.⁹²⁷ Kirigami incorporates periodic cuts to distribute mechanical stress uniformly, enabling out-of-plane buckling motions and in-plane rotations to preserve electrical performance under large strain deformations.⁹²⁸ Such systems include GaAs solar trackers that dynamically respond to light direction at the substrate level without the need for costly, cumbersome structural supports, as an example of their potential.⁹²⁹

Ultrathin metal films (e.g., Au, Ag, and Pt) deposited on soft substrates such as SEBS and PDMS can form crack-based intrinsically stretchable electrodes (Figure 44D).^{930,931} The inherent mechanical mismatch between the metal films and elastomers induces nanoscale and/or microscale cracks, which help distribute mechanical stress and mitigate localized strain while maintaining electrical percolation networks, even under stretching. To further enhance stretchability, a prestretching strategy can be employed. When a metal thin film is deposited onto a prestretched elastomeric substrate, the release of

prestrain generates both wrinkled and cracked microstructures, which improve mechanical compliance (Figure 44E).⁹³² In addition, stacking ultrathin elastomeric layers with different thermal expansion coefficients facilitates controlled microcrack formation, enhancing stretchability.⁹³³ Interfacial engineering can also mitigate the delamination of cracked metal films and prevent fatigue-induced degradation of the electrodes.⁹³⁴

8.2. Material-Oriented Approaches

A second strategy to achieve soft electronic systems uses materials designed or engineered to exhibit elastic responses, enabling mechanical compliance without compromising the electrical performance. These materials generally fall into two primary categories: (1) intrinsically stretchable materials and (2) composites.

The first category commonly utilizes organic polymers with chemical structures that balance considerations in stretchability and electrical performance, primarily through molecular engineering of the polymer backbone or side chains. Conducting polymers typically exhibit high crystallinity and strong intermolecular interactions, which are essential for charge transport but limit mechanical stretchability. To achieve both electrical conductivity and mechanical compliance, it is necessary to modulate polymer chain interactions, control crystallinity, and introduce stress-dissipative mechanisms within the polymer network. Backbone engineering can reduce rigidity by introducing flexible segments, while tuning the strength of interchain interactions enables relaxation of the crystalline domains, thereby increasing chain mobility (Figure 44L).^{935,936} In addition, modulation of interchain interactions (e.g., π - π stacking)^{937–940} and incorporation of dynamic bonds⁹⁴¹—such as hydrogen bonds, metal–ligand coordination, or reversible covalent linkages—facilitates stress dissipation by allowing the network to temporarily reorganize under strain, preserving structural integrity and electrical continuity. Furthermore, blending small molecular additives such as plasticizers,^{942,943} ionic liquids,⁹⁴⁴ polar solvents,⁹⁴⁵ or surfactants⁹⁴⁶ into conducting polymers disrupts strong intermolecular forces, enhancing the mobility of polymer chains (Figure 44M). This physical softening approach provides a facile means of improving stretchability and conductivity without requiring significant changes to the polymer backbone structure.

One interesting example of such approaches involves incorporating conjugated comonomer units into high-molecular-weight diketopyrrolopyrrole (DPP)-based polymers to enhance their stretchability.⁹⁴⁷ By inducing controlled multiscale ordering, these materials decouple charge transport pathways from mechanical deformation. As a result, DPP terpolymers designed with modified polymer backbones can achieve reversible strains exceeding 100% while maintaining field-effect mobilities greater than $1 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$ even after 1,000 stretching cycles.⁹⁴⁸ Another strategy involves incorporating hydrogen-bonding sites into semiconducting polymers through a stepwise polymerization and thermal conversion process.⁹⁴⁹ These polymers can restore over 90% of their charge carrier mobility after mechanical damage without sacrificing stretchability. Precise submolecular control over the density and spatial distribution of hydrogen bonds enables the formation of supramolecular networks that dissipate mechanical stress and promote rapid recovery under cyclic strains of up to 50%.⁹⁵⁰ Molecular dynamics simulations and nanoconfinement studies provide further insight into how

dynamic cross-links regulate network evolution and facilitate efficient damage repair in these stretchable semiconductors.⁹⁵¹

Composites are equally important to the development of soft electronic systems, as the basis of the second category of materials. Blending various conductive fillers with elastic polymers can form intrinsically stretchable elastomeric nanocomposites with high conductivity and stretchability. Homogeneously dispersed conductive fillers form a 3D percolation network that maintains continuous electrical pathways even under stretching (Figure 44F).^{952,953} Conductivity and stretchability can be further enhanced with a heterogeneous spatial distribution of fillers. For instance, controlling the sedimentation rate of fillers induces a gradient distribution,⁹⁵⁴ and water-assisted self-assembly of nanowires forms locally bundled structures.⁹⁵⁵ Additionally, Marangoni flow-driven float-assembly enables selective surface embedding of fillers, resulting in nanomembranes with patterned conductivity (Figure 44G).⁹⁵⁶ Integrating conductive fillers with different dimensionalities improves network stability under strain. During stretching, 1D fillers tend to align along the strain direction and bridge conductive pathways, while 0D and 2D fillers occupy gaps, preventing disconnection and maintaining conductivity (Figure 44H).^{957,958} Combining functional fillers—such as low impedimetric materials,^{959,960} magnetic materials,^{961,962} or electrochemically sensitive materials⁹⁶³—allows for the fabrication of stretchable functional electrodes with tailored properties.

Gel-based nanocomposites, formed by incorporating conductive fillers into soft hydrogel matrices, are promising for bioelectronic interfaces due to their inherent softness, high permeability, and tissue-like mechanical properties.⁹⁶⁴ Optimized structural designs can simultaneously achieve high conductivity and stretchability in physiological environments for soft, tissue-integrated electronics (Figure 44I).⁹⁶⁵ The gel-based nanocomposites not only provide excellent mechanical compliance but also offer dual conduction pathways by integrating the ionic and electronic transport mechanisms. The hydrated polymer network facilitates efficient ion migration with continuous electronic channels provided by embedded conductive fillers. This synergy reduces interfacial impedance and enhances charge transfer efficiency between biological tissues and electronic devices (Figure 44J).⁹⁶⁶ Moreover, gels can be engineered to incorporate tissue adhesiveness and self-healing capabilities through reversible cross-links or supramolecular interactions. These dynamic features are essential for ensuring long-term mechanical integrity and reliable performance in wearable and implantable bioelectronics (Figure 44K).⁹⁶⁷

Regarding the materials used as fillers, nanomaterials should be highlighted. CNTs are important constituents in such systems as they form percolative conductive networks within elastomeric matrices, enabling reliable performance under large strains.^{968,969} In the liquid phase, CNTs can be dispersed uniformly into copolymer matrices via melt blending or solvent casting to create conductive, stretchable, and durable materials.^{970–972} Other nanostructured materials, such as AgNWs and AgNPs, represent additional options as conductive fillers in elastomeric composites. For instance, AgNWs dispersed in styrene–butadiene–styrene (SBS) matrices—combined with ligand engineering and nanowire welding techniques—offer exceptional electrical conductivity and mechanical resilience, achieving elongation at break exceeding 900%.^{973–976} In addition, careful control over deposition

techniques can lead to nanoscale fiber-like structures in the polymer semiconductor poly(3-hexylthiophene) (P3HT), with phase separation in SEBS composites to optimize charge transport and stretchability.⁹⁷⁷ Rubbery strain, pressure, and temperature sensors that withstand stretch up to 50% can be fabricated while maintaining quality measurement accuracy under deformation.⁹⁷⁸

The conductive components in composites can themselves offer high levels of mechanical compliance. An extreme example is LMs, such as eutectic gallium–indium (EGaIn), integrated by injection molding or by filling into microfluidic channels to achieve desired stretchable composites.⁹⁷⁹ EGaIn-filled microchannels accommodate large mechanical strains, while the surrounding elastomer matrix provides self-resealing capability after deformation.^{980,981} In another example, PEDOT:PSS is of interest in soft electronics as a conductive, biocompatible filler in composites.^{871,982,983} Blending PEDOT:PSS with polar solvents or plasticizers, such as ethylene glycol or d-sorbitol, disrupts hydrogen-bond interactions within the polymer network. For example, mixing PEDOT:PSS with ionic liquids resulted in improved stretchability of >100% and enhanced conductivity of >1000 S/cm.^{944,984}

Therefore, many materials and strategies exist for the production of soft electronics that are adequate for wearable sensing. The soft and stretchable devices obtained can exhibit very distinct conductivities and stretchabilities depending on the materials and fabrication strategies, which should be considered when designing a device. A summary of these parameters for different techniques and materials can be found in Figure 44N and Table 2.

Because of their relevance to soft electronics applied to healthcare, LMs will be highlighted in the following section.

8.3. Liquid Metals for Wearable Healthcare

Liquid metals (LMs) are metals or alloys kept in the liquid state at room temperature, typically possessing high electrical and thermal conductivity, excellent deformability, low toxicity, and intrinsic self-healing capabilities.¹⁰¹⁹ These attributes have positioned LMs, particularly gallium-based alloys, at the forefront of flexible sensor and wearable device innovation for healthcare.¹⁰²⁰ LMs can be integrated into soft, stretchable, and biocompatible systems, thus allowing for sensors that seamlessly conform to complex biological surfaces, such as human skin, joints, and organs.^{1021–1023} LMs have been utilized in mechanoelectrical sensors for motion analysis and strain detection and in biochemical sensors capable of detecting biomarkers.^{4,1024} They are ideal for electromechanical sensing, including body motion tracking, pressure and strain monitoring, and haptic interaction, where precise and reliable detection is essential.

8.3.1. LM-Based Healthcare Applications. Figure 45 illustrates the various uses of LMs in healthcare monitoring,^{1023,1025} which include sports and human–machine interfaces.^{1026,1027} LM-based sensitive elements can be employed in tactile sensors,^{1026,1028,1029} strain sensors,^{1030,1031} and pressure sensors.^{1032,1033} Highly conductive LMs can be used as electrodes and interconnections for wearable physicochemical sensors, encompassing ultrasound sensors,¹⁰²¹ electrophysiological sensors,¹⁰³⁴ and chemical sensors.^{1035,1036} Unlike rigid materials, LMs maintain mechanical integrity and electrical functionality during stretching and bending.¹⁰³⁷ Integrated into patches, gloves, or textiles, LM-based sensors enable precise monitoring of joint motion,¹⁰³⁸ muscle

Table 2. Conductivities and Stretchabilities of Stretchable Electrodes Fabricated via Diverse Materials and Fabrication Strategies

Categories	Conductivity (S/cm)	Stretchability (%)	ref
Geometrically designed metal films	~56,000	~130	985
	Pure gold conductivity	~50	986
	~450,000	~40	987
		~30	988
		~50	989
Crack-based metal films		~30	894
	~1250	~400	990
	~16,000	~300	933
	~20,000	~160	931
Elastomeric nanocomposites	~200	~500	991
	~6700	~150	899
	~6200	~300	958
	~100,000	~1000	956
	~11,000	~50	898
	~30,000	~840	953
	~45,000	~1200	992
	~2000	~1000	993
	~2100	~500	994
	~150	~600	995
	~90	~500	996
	~42,000	~270	958
	~31,000	~1000	957
	~120,000	~200	955
	~36,000	~50	997
Gel-based nanocomposites	~1500	~200	901
	~15,000	~100	954
	~11,000	~500	959
	~10	~400	965
	~4100	~800	944
	~0.01	~50	998
	~3300	~100	900
	~370	~250	999
	~700	~400	906
	~0.4	~40	1000
	~0.03	~600	1001
	~20	~350	964
	~0.2	~120	1002
	~80	~40	1003
	~0.02	~550	1004
	~10	~150	1005
	~0.1	~100	1006
	~0.5	~40	967
	~90	~800	1007
	~17,000	~250	1008
Conducting polymers	~170	~90	1009
	~250	~400	1010
	~400	~130	1011
	~6000	~100	1012
	~40	~35	1013
	~50	~20	983
	~30	~20	1014
	~300	~30	1015
	~550	~20	1016
	~670	~20	1017
	~40	~50	1018



Figure 45. Overview of LMs toward smart healthcare. Image source: Freepik.¹⁰³⁹

activity,¹⁰³⁶ and tactile interactions,^{1026,1028,1029} providing real-time biofeedback for early abnormality detection, rehabilitation, and enhanced human–machine interaction.

The fluidic properties of LMs are exploited in electro-mechanical sensors, such as pressure and strain sensors constructed by encapsulating LMs within microchannels. For instance, Yu et al. produced a stretchable tubular elastomeric piezoresistive (STEP) microfiber by embedding LMs into single elastomeric microtubes (Figure 46A).¹⁰³⁰ These STEP microfibers exhibit such characteristics as softness, thinness, flexibility, stretchability, and washability. By incorporating the STEP microfibers into targeted areas—e.g., ball and heel of socks—accurate measurement of plantar pressure can be performed during locomotion. The electrical signals generated by the STEP microfiber sensors provide real-time monitoring and visualization of the gait cycle, capturing phases such as heel strike, midstance, and toe-off. In addition to piezoresistive-based electromechanical sensors, LM sheath-core microfibers are suitable for electromechanical sensors utilizing triboelectricity. Zheng et al. reported stretchable, highly conductive LM sheath-core microfibers to work as self-powered electromechanical sensors. As shown in Figure 46B, when embedded into fabric gloves worn on a prosthetic hand, bending the fingers generates friction between the LM microfibers and polyurethane fibers in the glove, producing detectable voltage and current. Similarly, when the fiber is adhered to a human wrist, friction between the fiber and the skin produces triboelectric signals.¹⁰⁴⁰

Tactile sensors often utilize LM-based configurations, such as domed or patterned microchannels, which enable the detection of distributed mechanical events, surface roughness, and object stiffness. They can provide precise haptic interfacing, essential for applications in prosthetics, robotics, and virtual reality. For example, Chen et al. fabricated electrode arrays using a TPU and eGaIn-based composite, with a dielectric layer sandwiched between two layers of electrode arrays, forming a capacitive pressure sensor array. As

illustrated in Figure 46C, the integrated system demonstrated haptic interactive capabilities. Specifically, the tactile sensor was able to control LEDs through an external controller, enabling functionalities such as one-point or multipoint sliding, pressing, and precise location monitoring.¹⁰²⁶ A sensor array comprising bimodal sensing units that integrate resistive, strain, and pressure sensors was fabricated using eGaIn LM embedded within elastomer membranes, as shown in Figure 46D.¹⁰²⁸ To mimic human-like compliance perception, a 4×4 sensing matrix was designed to fit the dimensions of a human or robotic palm. This system not only directly captured the responses of strain and pressure sensors but also calculated compression and the corresponding compliance, providing a comprehensive representation of tactile interactions.

LMs are also employed in motion sensors, including tilt and inertial sensors. In these devices, LM droplets are confined within cavities or microchannels, where their movement under gravitational or inertial forces generates detectable changes in electrical properties. For example, Xu et al. reported a wearable body condition sensor system capable of wirelessly monitoring sleeping posture, respiration rate, and diaper moisture, with a built-in feedback alarm notification system. A key component of the system is a tilt sensor that utilizes an LM droplet confined within a cavity, enabling the detection of at least 18 slanting orientations. To continuously track sleep status, the system incorporates a multimodal flexible sensor sheet integrated with a wireless feedback signal processing unit, which is designed to be placed on a disposable diaper. This setup allows real-time multichannel signal detection and transmission to a smartphone interface, providing data on breathing patterns, diaper wetness, and sleeping posture as shown in Figure 46E.¹⁰²³ Similarly, Babatain et al. proposed a fully soft, laser-induced graphene (LIG) and LM-based inertial sensor integrated with additional capabilities for temperature, humidity, and breathing monitoring. The inertial sensor employs a design that confines a graphene-coated LM droplet within a fluidic channel, allowing the droplet to roll over LIG resistive electrodes. This configuration enables the precise detection of motion through resistance variations, which was validated by attaching the sensor to a human subject's chest during activities such as walking, jogging, lying down, and getting up. Real-time monitoring successfully captured resistance changes corresponding to each activity, demonstrating its capability for motion tracking and multifunctional health monitoring, as shown in Figure 46F.¹⁰²⁷

8.3.2. Liquid Metals as Conductive Electrodes in Wearable Healthcare Applications. LMs are used in electrophysiological sensors (ECG, EEG, and EMG)¹⁰³⁴ and flexible energy-harvesting systems.^{1041,1042} LM-based electrodes enhance the performance of wearable electrochemical sensors for real-time biomarker detection and offer improved durability in wireless power transmission and communication systems. By way of illustration, Hu et al. fabricated a wearable ultrasonic device for continuous, real-time, and direct cardiac function assessment (Figure 47A).¹⁰²¹ To improve the mechanical coupling between the device and human skin, allowing the left ventricle to be examined from different views during motion, eutectic gallium–indium LM was selected to fabricate electrodes and interconnections. A higher stretchability and fabrication resolution were obtained than with electrodes based on serpentine-shaped copper thin film. Similarly, Ding et al. developed skin-adhesive LM particles (ALMP) for fabricating epidermal electronics designed for




Figure 46. LMs as sensing elements in wearable healthcare applications. (A) LM-based piezoresistance sensor for monitoring plantar pressure during locomotion. Reproduced from Yu et al.¹⁰³⁰ © 2018 American Chemical Society. (B) LM-based triboelectric sensor for detecting hand and wrist motions. Conductance-stable LM sheath-core microfibers for stretchy smart fabrics and self-powered sensing. Reproduced with permission from Zheng et al.¹⁰⁴⁰ © 2021 American Association for the Advancement of Science. (C) LM-based sensor for tactile display interaction. Reproduced with permission from Chen et al.¹⁰²⁶ © 2023 John Wiley & Sons. (D) LM-based sensor to mimic human-like compliance perception. Reproduced from Chen et al.¹⁰²⁸ Available under a CC BY-NC license. © 2024 John Wiley & Sons. (E) LM-based tilt sensor. Reproduced with permission from Xu et al.¹⁰²³ © 2022 John Wiley & Sons. (F) LM-based inertial sensor for posture and body motion monitoring. Reproduced from Babatatin et al.¹⁰²⁷ © 2022 American Chemical Society.




Figure 47. Liquid metals as conductive electrodes in wearable healthcare applications. (A) Wearable cardiac ultrasound imager. Reproduced from Hu et al.¹⁰²¹ Available under a CC BY 4.0 license. © 2023 Springer Nature. (B) LM-based epidermal electronics designed for wear-resistant electrophysiological monitoring. Reproduced from Ding et al.¹⁰⁴³ © 2022 American Chemical Society. (C) LM-based stretchable electrochemical devices as epidermal biofuel cells. Reproduced with permission from Chen et al.¹⁰⁴⁵ © 2024 John Wiley & Sons. (D) LM-filled microstructured stretchable TENG fiber-based textile for self-powered breathing monitoring. Reproduced from Dong et al.¹⁰⁴¹ Available under a CC BY 4.0 license. © 2020 Springer Nature. (E) LM-based electronic textile systems with near-field wireless power and communication capabilities. Reproduced from Lin et al.⁶⁷³ Available under a CC BY 4.0 license. © 2022 Springer Nature. (F) LM-based electronic tattoo with an energy harvesting coil for wirelessly monitoring ECG signal. Reproduced from Alberto et al.¹⁰⁴⁹ Available under a CC BY 4.0 license. © 2020 Springer Nature.

wear-resistant electrophysiological monitoring. The ALMP demonstrated self-adhesive and ultraconformal properties, ensuring ultralow skin contact impedance and high-quality electrophysiological signal acquisition. Using this technology, they captured the flexion of the flexor carpi ulnaris with stable and reliable EMG signals (Figure 47B).¹⁰⁴³

In electrochemical sensing, LMs were used in printed stretchable devices.¹⁰⁴⁴ Stress-resistant composite silver (Ag) inks were combined with eutectic gallium–indium particles as dynamic electrical anchors, patterned into serpentine “bridges” connecting “island” electrodes. This architecture enabled high stretchability and stable electrochemical performance under deformation. A key application was epidermal biofuel cells,

which harvest bioenergy from sweat metabolites as shown in Figure 47C.¹⁰⁴⁴ However, few studies have explored LM-based electrochemical sensors, as challenges like LM's susceptibility to electrochemical oxidation and its reactivity in acidic or basic environments limit their applicability in broader electrochemical scenarios.¹⁰⁴⁵

In addition to conductive properties, LM-based energy-harvesting technologies have been developed to address the need for portable, wearable energy systems capable of storing and harvesting power for prolonged use of personal electronics. Examples include TENGs, piezoelectric nanogenerators, and TEGs, which will be discussed at length later on in this review paper (see Section 10). These systems harness energy from various body motions directly into electricity, enabling efficient and continuous operation of wearable devices. For example, an LM-filled microstructured stretchable TENG fiber-based textile was proposed to exhibit an OCV and a short-circuit transferred charge (Q_{sc}) as high as 490 V and 175 nC, respectively, as shown in Figure 47D.¹⁰⁴¹ LMs are also suitable for stretchable antennas in RF communication devices.^{673,1046–1048} By integrating LM-based antennas, healthcare devices can achieve wireless functionality for both data communication and power transfer. Indeed, electronic textile systems with near-field wireless power and communication capabilities were produced by digitally embroidering LM fibers onto clothing. With these textiles, robust and unobtrusive wireless power transfer can be exploited for near-field communication with nearby devices, such as health sensors and smartphones, even during strenuous exercise and daily activities. These wirelessly enabled garments can power bioelectronic implants in patients, detecting diseases in high-risk populations like the elderly. They can also monitor axillary temperature—a key indicator of exhaustion, physical performance, and various health conditions (Figure 47E).⁶⁷³ Alberto et al. reported a LM-based electronic tattoo capable of wirelessly harvesting energy through a passive antenna. This system employs a wireless power transfer (WPT) mechanism, integrating an inductive coil antenna with biopotential acquisition electrodes that interface directly with the skin. The design enables efficient wireless power transfer of approximately 300 mW across the skin and up to 100 mW beneath the skin, offering reliable energy support for wearable healthcare devices. This electronic tattoo can be incorporated into monitoring systems for real-time collection and transmission of biopotential signals, such as heartbeat rate, via Bluetooth. By combining energy harvesting, biopotential sensing, and wireless communication, this LM-based electronic tattoo may be used for long-term, remote patient monitoring and management (Figure 47F).¹⁰⁴⁹

8.4. Active Components in Soft Electronics

Active components in electronics, such as transistors and diodes, as well as LEDs and photodetectors (PDs), can be achieved using the strategies outlined in the previous section for broad classes of soft electronic systems. Stretchable organic field-effect transistors (OFETs) and organic electrochemical transistors (OECTs) play essential roles in low-power, stretchable logic circuits. OFETs are commonly fabricated through solution processing techniques such as inkjet printing, spin coating, and layer-by-layer deposition. The methods described previously support stretchability while preserving charge carrier mobilities in otherwise conventional device designs.^{1050–1052} OECTs operate using different principles,

commonly through redox-active PEDOT:PSS channels in a three-electrode configuration. Here, the application of gate voltage modulates the entire channel conductivity via electrochemical doping. This process results in high transconductance and signal amplification at low operating voltages.^{1053,1054} Both OFETs and OECTs have been built into soft electronic platforms and have been integrated into soft electronic devices, including displays, wearable sensors for chemical, physical, and biological monitoring, neural interfaces, electronic skin, and biomimetic systems for artificial perception.^{1055–1058} Due to their importance, thin film transistors (TFT) will be discussed in further detail in the following subsection.

Recent advances in stretchable solar cells, energy storage, and energy harvesting systems provide attractive options for power management (see Section 10). For instance, some of the best types of stretchable organic photovoltaic cells can maintain 80% of their initial power conversion efficiency under tensile strains of greater than 50%.¹⁰⁵⁹ Structural designs, such as those that involve weaving functional filaments into fabrics, facilitate conformal, self-powered wearable sensors. In energy storage, printed Ag_2O –Zn and Ag/AgCl batteries, as well as stretchable lithium-ion and emerging zinc-air chemistries, offer high energy densities while tolerating repeated cycles of mechanical deformation.^{1060,1061} The development of graphene-MXene composites serves as the basis for stretchable supercapacitors, as efficient secondary power sources for wearable systems.¹⁰⁶² Furthermore, stretchable energy-harvesting technologies, including TENGs, enzymatic biofuel cells, and wireless power transfer circuits based on inductive, capacitive, or radio-frequency (RF) mechanisms, enable continuous energy delivery.^{1063–1065} These systems support the sustainable, long-term operation of soft electronic platforms without the need for bulky or rigid power components.^{1066,1067}

Perovskites, quantum dots, and organic light-emitting conjugated polymers serve as key building blocks of nanocomposites for stretchable LEDs.^{1068,1069} Advanced fabrication techniques, including inkjet printing, transfer printing, and nanoconfinement, enable the integration of these materials into ultrathin mesh layouts on stretchable substrates, preserving both optical performance and mechanical resilience.^{1070,1071} Similarly, stretchable PDs can be constructed with a range of nanostructured materials, including AgNWs, semiconducting P3HT nanofibers, and hybrid systems that combine low-dimensional perovskites with wrinkled graphene. These PDs can sustain tensile strains exceeding 50% with minimal reductions in the photocurrent.¹⁰⁷² When embedded in soft, biocompatible substrates, these LEDs and PDs conform intimately to biological tissues, enabling a range of applications including epidermal pulse oximetry, UV dosimetry, and noninvasive biosensing and bioimaging.¹⁰⁷³

8.5. Thin-Film Transistors

Flexible and large-area electronics are important for various applications, from next-generation displays^{1074–1076} and large-area image sensors^{1077,1078} to wearable health monitors,^{627,1079–1081} digital microfluidics,^{1082,1083} and thin-film microprocessors.^{1084–1086} At the heart of these technologies lie TFTs, which enable lightweight, conformable, and scalable electronic systems. Flexible circuits typically use thin-film semiconductors deposited on plastic substrates (e.g., polyimide, PET). Key material platforms include silicon (single crystals, polycrystalline (LTPS), and amorphous), organic

Table 3. Properties of Semiconductor Materials for Flexible Electronic Devices

	Material	Max Fracture Strain (%)	Electron Mobility (cm ² /V·s)	Chemical Stability	Biocompatibility	Large-Area Fabrication	Miniaturization	CMOS-Compatibility
Silicon	Crystalline thin-film (Nanomembrane/wires)	~5	~1000	High	Medium-High	Low	Very High	Very High
	Amorphous Si (a-Si)	~1–2	~0.5–1	High	Medium-High	High	High	Very High
	LTPS	~0.5–1	~100	High	Medium-High	Medium	High	High
Organics		~20–30	~0.1–1	Low	Very High	Very High	Low–Medium	Low
Oxides		~2.9	~10–20	Medium-High	Medium	High	Medium	Medium
MoS ₂		~6–11	~10–100	Medium	Medium	Medium	Very High	High

semiconductors (e.g., conjugated polymers, small molecules), oxide semiconductors (e.g., amorphous Indium Gallium Zinc Oxide—*a*-IGZO), and 2D semiconductors (e.g., MoS₂, WSe₂).

In this section, we review and compare them, focusing on flexible-device demonstrations and benchmarks. Table 3 compares four semiconductor classes (silicon, crystalline and amorphous, organic polymers, *a*-IGZO, and 2D materials like MoS₂) across key metrics for flexible electronics. Amorphous and crystalline Si are brittle: bulk Si typically fractures at ~1% strain,¹⁰⁸⁷ *a*-Si is similar, whereas organic films can accommodate much larger deformation (organic light-emitting electrochemical cells, OLEC, devices stretch ~30% without cracking).¹⁰⁸⁸ Two-dimensional carbon shows the best stretchability—pristine graphene can sustain ~20–25% strain—but many 2D semiconductors are weaker, e.g., monolayer MoS₂ fails around 6–11%.^{1089,1090} Electron mobility likewise varies widely. Organic polymers generally have very low mobilities (<10 cm²/V·s and often <1 cm²/V·s),¹⁰⁹¹ whereas graphene's carriers are extraordinarily mobile when sandwiched between a proper dielectric like boron nitride, hBN (~2 × 10⁵ cm²/V·s).¹⁰⁹² Crystalline Si is intermediate (~1.35 × 10³ cm²/V·s for electrons).¹⁰⁹³ In practice, silicon's native oxide makes Si devices chemically stable, and graphene's inert sp² carbon network also resists corrosion, whereas organic films tend to degrade under environmental exposure (O₂, moisture).

Biocompatibility follows a similar trend: Si-based materials (including silicon oxide surfaces) and many bioengineered organics are known to be well tolerated *in vitro* and *in vivo*.¹⁰⁹⁴ In contrast, oxide materials (such as ZnO/IGZO) and some 2D semiconductors can leach ions or produce reactive byproducts; for example, ZnO nanowires show measurable cytotoxicity due to Zn²⁺ release.¹⁰⁹⁵ A key concern with *a*-IGZO materials is the potential biocompatibility and toxicity issues associated with indium, which can pose risks if released or degraded in biological environments. Careful material encapsulation and further biotoxicity studies are essential to ensure safe application in biointegrated devices.

Large-area manufacturability also differs. Organic semiconductors are solution-processable, enabling wafer-scale, low-cost films. In contrast, wafer-scale 2D film growth is still maturing: CVD and epitaxial methods exist, but defects and contamination are still problematic at the production scale.^{1096,1097} Finally, Si leads in CMOS scaling: conventional silicon processes allow aggressive lithography and device downscaling. Amorphous IGZO can be patterned below 1 μm and even integrated back-end-of-line on CMOS wafers.¹⁰⁹⁸ By comparison, 2D materials face integration challenges: their interfaces (dielectrics/contacts) and etching processes are not yet mature, making "CMOS+2D" hybrid integration challeng-

ing.¹⁰⁹⁶ Figure 48 illustrates these properties with a radar graph.




Figure 48. Benchmark of semiconductor materials for flexible electronics.

8.5.1. Silicon-Based TFTs (Si NMs, Si NWs, *a*-Si, Poly-Si). **8.5.1.1. Silicon Nanomembranes (Si NMs).** Ultrathin single-crystal Si layers (tens of nm) peeled from wafers and bonded onto polymer substrates combine the electronic quality of bulk Si with flexibility.^{894,1099} These Si NMs can be patterned into conventional CMOS devices, with built-in tensile strain (e.g., from Si/SiGe/Si stacks) to boost carrier mobility. In practice, Si NM FETs on plastic show electron mobilities of a few hundred cm²/V·s and hole mobilities ~100–150 cm²/V·s (comparable to SOI devices). Gate currents and interfaces behave like rigid Si MOSFETs (low flicker noise, negligible bias drift), with on/off ratios easily >10⁶ (due to Si's bandgap and doping) and subthreshold slopes around 100 mV/dec. Threshold voltages are similar to bulk Si (~0.5–1 V). In RF tests, even unstrained Si NM transistors (channel ~0.5 μm) achieved *f_T* ≈ 3.3 GHz. Strain-engineered Si NMs (e.g., Si/SiGe/Si trilayers) showed a ~47% mobility boost, pushing *f_T* toward 6 GHz.¹¹⁰⁰ Si NMs tolerate bending to millimeter-scale radii: theoretical fracture of a 20 μm device-on-polyimide happens at a radius of about 1 mm, and experiments have bent Si-NM circuits around a 2 mm radius without damage.¹¹⁰¹ Key trade-offs are cost and process complexity: SI NMs require Si wafer processing, oxide release to liberate the silicon from the SOI wafer, followed by a controlled and reliable transfer process, possibly at large scale. Often, NM circuits are cut into islands or meshes for stretchability (e.g., "nanomesh" Si has 25% stretch and μS₀ cm²/V·s) at the cost of extra steps.¹¹⁰²

Advantages of Si NMs include the performance of bulk Si on plastic, which is comparable with standard CMOS devices. Moreover, the designs and toolkit developed for silicon can be exploited to yield high-speed, low-power circuits on plastic. Very high speed (GHz) and transconductance far beyond a-Si or organics, with static off-currents <10 pA. Excellent analog stability and low power (due to low V_{th} and high μ). Also, large-area manufacturing is conceivable. However, the processing cost can be higher, and the flexibility is moderate since flat NMs fracture below $\sim$ mm-scale bend radii unless patterned (islands/meshes) for stretching.

8.5.1.2. Silicon Nanowires (Si NWs). Quasi-1D Si channels (diameter <100 nm) offer high performance with extreme deformability. Recent demonstrations grow aligned arrays of crystalline Si NWs via a solid–liquid–solid method, then transfer them onto elastomer (PDMS) or plastic (PET) substrates.^{1103,1104} The resulting Si NW FETs achieve high hole-carrier $\mu \approx 70 \text{ cm}^2/\text{V}\cdot\text{s}$ (similar for electrons in other work) and $I_{on}/I_{off} > 10^5$ – 10^6 . Subthreshold slopes are low ($SS \approx 134$ – $200 \text{ mV}/\text{dec}$), reflecting good interface quality. Because the nanowires are built on small “islands” of circuitry, these devices tolerate enormous strain: reported stretchability is up to 50% strain, withstanding >1000 cycles at 20% strain.¹¹⁰⁴ Even under such deformation, μ and I_{on}/I_{off} degrade only modestly. The Si NW FET arrays endure tiny bending radii—e.g., bending to $R \approx 0.5 \text{ mm}$ for 1000 cycles had negligible performance loss. Si NW devices can also be made into logic inverters on plastic, demonstrating scalability.

Si NWs are advantageous for their outstanding mechanical flexibility, as ultrathin Si wires on elastomer island networks yield stretchability (50%) and extreme bend radius (<1 mm). Performance is much better than a-Si or organics, and batch transfer growth allows wafer-scale integration. Devices have high I_{on}/I_{off} and work stably for >270 days without passivation. The disadvantages include a more complex device architecture and a lower mobility than bulk Si NMs (tens vs hundreds $\text{cm}^2/\text{V}\cdot\text{s}$).

8.5.1.3. Thin-Film Silicon (a-Si:H and LTPS) TFTs. Conventional display backplanes use hydrogenated amorphous Si (a-Si:H) or LTPS on glass or plastic. a-Si:H TFTs (PECVD-deposited at <300 °C) have very low carrier mobility (~ 0.5 – $1 \text{ cm}^2/\text{V}\cdot\text{s}$ for electrons, $\sim 0.01 \text{ cm}^2/\text{V}\cdot\text{s}$ holes), limiting speed to tens of kHz. Nevertheless, they reliably deliver $I_{on}/I_{off} > 10^6$, threshold <3 V, and $SS < 0.5 \text{ V}/\text{dec}$. These devices are intrinsically bendable if the substrate is thin enough: early work on 75 μm glass foils achieved a bending radius $\sim 5 \text{ mm}$ without loss of function.¹¹⁰⁵ Amorphous Si is stable and cheap (roll-to-roll PECVD), but has poor mobility and voltage stability under stress.^{1106,1107}

By contrast, LTPS (excimer laser annealed poly-Si) on plastic glass has much higher mobility (tens to hundreds $\text{cm}^2/\text{V}\cdot\text{s}$). For example, flexible LTPS devices have demonstrated μ up to $\sim 174 \text{ cm}^2/\text{V}\cdot\text{s}$ (on a 20 μm PI substrate). On/off ratios exceed 10^6 (low leakage $\sim 10 \text{ pA}$). Thresholds are typically a few volts; SS can be ~ 0.1 – $0.3 \text{ V}/\text{dec}$.^{1108,1109} Current LTPS displays (on glass) operate at tens of MHz; on plastic, the best bendable LTPS reached f_T similar to Si NM, but generally LTPS is modest (<100 MHz) due to larger dimensions. Mechanically, flexible LTPS has improved: early reports showed $R = 20 \text{ mm}$, but modern thin-film LTPS (e.g., on 8–20 μm PI) can bend to a few mm.

Large-area processing is a key feature of LTPS, while a-Si offers simpler fabrication, excellent uniformity, and stability for

low-cost flexible sensors and displays. Both technologies can integrate with existing fabs and are proven at scale. However, a-Si devices are slow (low μ , limited to $\sim$ kHz–MHz) and require thicker devices to maintain current, limiting ultimate flexibility. LTPS requires costly laser crystallization and often cannot use plastic substrates without substrate lifting or transfer. Both have poorer flexibility than nanostructured Si: even at $\sim 10 \mu\text{m}$ thickness, bending is limited to a few mm (much larger than 2D/0D Si devices). Device uniformity (for LTPS) and threshold stability (for a-Si under bias) remain concerns.

8.5.2. Organic Semiconductor TFTs. Organic TFTs (OTFTs) use conjugated polymers or small molecules as the active layer. Organic semiconductors are inherently flexible, can be solution-processed, and allow low-cost large-area manufacturing (printing, coating). Charge transport occurs via hopping between disordered molecular orbitals, typically yielding lower μ than inorganic semiconductors.^{1110–1112} Important advantages of OTFTs include device flexibility (bending radius down to a few mm or less), possible biocompatibility and biodegradability. Hence, OTFTs can be integrated with soft substrates and distinct form factors (e-skin, textile integration), also exhibiting low-cost fabrication.^{1113–1115} There are, however, two main disadvantages, namely, the low mobility (often <5 $\text{cm}^2/\text{V}\cdot\text{s}$), which limits speed, and the lack of environmental stability as many organics degrade in air or with electrical stress. Furthermore, multilayer encapsulation is often required.^{1116,1117} In summary, organic TFTs trade high flexibility and low-cost processing for inferior raw performance compared to oxide or silicon devices.^{1118,1119}

8.5.3. Oxide Semiconductor TFTs. Amorphous metal-oxide semiconductors (e.g., amorphous In–Ga–Zn–O, a-IGZO) emerged in the mid-1990s and have become mainstream in displays. They are typically n-type, transparent, and can be deposited at low temperature through radio frequency sputtering. Electron transport occurs via overlapping metal Ss orbitals, giving high mobility even in amorphous form.^{1120–1122} Oxide TFTs combine high speed and low leakage with simple fabrication. They are inherently transparent and uniform over large areas. Solution-processed strategies have been used, with performance not yet matching that achieved with vacuum systems.¹¹²³ The low off-current means oxide-based circuits can consume ultralow static power even if the absence of p-type oxide semiconductors makes complementary design currently challenging. They can operate at modest voltages (1–5 V). Device-to-device uniformity is good on rigid and on flexible substrates.¹¹²² An important limitation is that most oxide semiconductors are n-type only (lack a robust p-type counterpart), so complementary logic is hard. Another concern is stability, since threshold voltage drifts under bias/illumination/stress, especially on plastic, requiring compensation circuits.¹¹²⁴ Oxygen vacancies can cause trapping and noise. Mobility, though much higher than a-Si, remains far below crystalline Si, limiting ultimate f_T ($\sim 100 \text{ MHz}$). Also, oxide TFTs often require vacuum deposition (sputtering, ALD) of metal oxides, which is more expensive than printing or solution processes. Bending and long-term environmental stability (humidity/temperature) can degrade performance over time.

8.5.4. Two-Dimensional Semiconductors' TFTs. Atomically thin 2D semiconductors like MoS_2 , WS_2 , and WSe_2 provide an extreme form of thinness with a sizable bandgap (e.g., $\text{MoS}_2 \sim 1.8 \text{ eV}$ in monolayer) and inherently excellent


gate control. 2D layers are prepared by exfoliation or scalable CVD, then transferred onto flexible substrates with high- κ dielectrics.^{1125–1127} In RF testing, 2D MoS₂ transistors have shown remarkable speed. Recent work reports *intrinsic* current-gain cutoff frequencies up to ~40–42 GHz for optimized few-layer MoS₂ devices. For example, MoS₂ FET with self-aligned contacts achieved $f_T \approx 42$ GHz.¹¹²⁸ Even on flexible substrates, MoS₂ devices have reached ~13.5 GHz intrinsic f_T .¹¹²⁹ The intrinsic current gain (g_m/g_{ds}) can exceed 30, allowing voltage gain. Noise performance is favorable compared to graphene (because MoS₂ can pinch off current), and MoS₂ has excellent current saturation. 2D transistors tend to have very low gate leakage (thin dielectrics) and modest off-currents.¹¹³⁰ The main advantage of these transistors arises from the combination of atomic-scale thickness with relatively high speed. They yield large on/off ratios (unlike graphene) and have shown GHz-range amplification. The atomically thin channel gives excellent immunity to short-channel effects, enabling aggressive downscaling on a flexible substrate without losing gate control. They also allow remarkable flexibility: monolayers can withstand extreme bending. Unlike organics, 2D materials are inorganic crystals and are chemically stable (though some require encapsulation). Their electronic properties can be engineered by layer number, stacking, or doping. On the other hand, fabrication uniformity and scalability remain challenges for 2D materials. CVD-grown films still have grain boundaries and defect scattering. Carrier mobility in realistic devices (with contacts and substrates) is often 10–50 cm²/V·s, limiting performance below theoretical potential. ALD-grown 2Ds may provide more controllable large-scale uniformity and compatibility with 300 nm CMOS processing. However, this is an emerging technology with initial experiments showing modest mobility of the final layers. Contact resistance and thermal dissipation are concerns for short-channel devices. Integration of 2D devices often requires complex transfer or growth processes, which are currently less mature than thin-film deposition of oxides or organics.

These trade-offs are summarized in Table 4.

Therefore, thin-film flexible electronics span a broad range of materials, each with its own unique benefits. Oxide semiconductors (e.g., IGZO) offer high mobility and ultralow leakage for fast, low-power circuits but require improved p-type counterparts and stress stability. Organic semiconductors excel in flexibility and manufacturability and have steadily improved (some polymers now exceed amorphous silicon performance) but still lag in speed and stability. 2D semiconductors like MoS₂ combine the advantages of high mobility, large on/off ratio, and GHz-speed potential, at the cost of emerging fabrication challenges. Silicon nanomembranes bring true silicon performance to flexible form factors, enabling GHz silicon-like circuits, although with more complex processes. Overall, each technology occupies a different niche: organics for ultraconformable, low-cost large-area sensors; oxides for moderate-speed flexible displays and sensors; 2D materials for eventual high-speed flexible logic; and silicon nanomembranes for best-performance flexible ICs. Continued advances in materials (e.g., novel oxides, conjugated polymers, and 2D alloys) and processes (e.g., better encapsulation, printing techniques) are driving the field forward. Flexible electronics now combine years of materials innovation to approach the performance of rigid silicon in many respects while offering new mechanical possibilities.

Table 4. Comparison of Performance Metrics for Thin-Film Flexible Devices

	μ (cm ² /V s)	On/ Off Ratio	f_T	Stability	Power Consumption	Advantages	Disadvantages
Oxide (a-IGZO)	>10 (up to 70)	10^7	~10–100 MHz	V_{th} drift under bias/light	Low	High μ & very low leakage; transparent; low-T deposition; mature large-area processes; ultralow static power	n-type only (no p-device); threshold shifts; requires vacuum deposition; μ still below Si
Organic	0.1–3 (up to ~10 SC)	10^4 – 10^6	kHz–few MHz	Poor (moisture, bias stress)	Low	Intrinsically flexible/stretchable; solution-processable; low cost; biocompatible/biodegradable; novel form factors	Low μ & speed; high SS; significant bias-stress drift; environmental sensitivity; higher noise
2D Semiconductors	10–50 (up to 200)	10^6	~10–40 GHz	Substrate influence, interface sensitive	Low	Atomic-scale thickness; very high on/off; GHz-range speed; excellent gate control; flexible	Scalability challenges; grain boundaries and defect scattering; contact resistance; complex integration
Silicon Si NMJs	200–500	10^6	3–6 GHz	Excellent (CMOS-grade)	Very low	Bulk-Si performance on plastic; CMOS-compatible; ultralow noise; high speed and gm; mature Si processes	Complex wafer bonding/release; high cost; limited bend radius without patterning
Si NWs	~70	10^5 – 10^6	~GHz expected	Excellent	Very low	Exceptional stretchability (50%); ultraflexible; high I_{on}/I_{off} wafer-scale transfer; stable over cycles	Complex guided growth and transfer; lower μ vs Si NMJs; island-based architecture adds steps
a-Si	0.5–1	10^6	kHz–MHz	poor voltage bias stability	Low	Mature, low-cost large-area PECVD; stable; bendable on thin substrates; excellent uniformity	Very low μ & speed; large SS; voltage instability under stress; limited flexibility
LTPS	20–500	10^6	~100 MHz	some bias drift	Low	High μ for display backplanes; low leakage; good uniformity; moderate flexibility with thin PI	Costly laser crystallization; limited bend radius; threshold drift; process complexity



9. Signal Acquisition

We now focus on describing strategies for effective signal acquisition. The data acquisition process normally includes sensing, signal transmission, and display, according to the following steps: first, wearable sensors detect physical or chemical signals from the human body. These, then, undergo processing to minimize noise and enhance quality. As the signals output from sensors are analogical, they need to be converted to digital signals for effective transmission, which can be achieved by a wired interface or by wireless communication modules. Wired electronics hinder user comfort as the wires interconnecting sensors and read-out systems disrupt users' natural daily activities and pose a risk of disconnection. These limitations have motivated the incorporation of wireless systems to further improve users' comfort.^{1131,1132} The most used wireless communication technologies are NFC, Bluetooth, and Wireless Fidelity (Wi-Fi).^{236,1133,1134} They differ in many characteristics, such as the transmission distance, frequency, energy consumption, and application scalability. Figure 49 shows an overview of the methodologies for wireless signal acquisition in wearable sensors.

NFC is a short-range, point-to-point communication method with a transmission distance typically below 10 cm, operating in the high-frequency (HF) band at 13.56 MHz, which is lower than GHz-range wireless technologies such as Bluetooth and Wi-Fi (UHF at 2.4 GHz and SHF at 5–6 GHz). It can operate in passive mode, harvesting energy from the reader and requiring no onboard power source, unlike Bluetooth and Wi-Fi. The NFC system relies on transmitting

and receiving coils, where inductive coupling enables data transfer and facilitates battery-free operation. Indeed, Maroli et al. emphasized the battery-free operation and compact design enabled by the NFC,⁸⁵⁷ in contrast with Bluetooth-based systems such as that reported by Yang et al.¹¹³⁶ Koh et al. built a small NFC chip on a stretchable antenna for a microfluidic sensor monitoring lactate, chloride, pH, and glucose in sweat and displaying the concentration of each through color-responsive materials.¹¹⁷

The Silicon Craft Company developed a potentiostat on a single chip by combining NFC technology with an electrochemical interface. The system can perform electrochemical analyses, including cyclic voltammetry (CV), SWV, and chronoamperometry, directly on a wearable device. Data are transmitted wirelessly via NFC. The unification of potentiostat operations and wireless communication into a single chip allows for the achievement of exceptional miniaturization and simplicity, which, in turn, reduces both cost and complexity. This technology is advantageous for POC applications, where disposable sensors are placed on the skin or in contact with a sample (e.g., sweat or urine) and powered by the reader during the measurement.¹¹³⁵ As the transmission distance is quite limited, the safety of the data is at a high level. However, the limited operating range of the NFC reduces its practicality.

For long-distance wireless communication, Bluetooth or Wi-Fi systems have been adopted.^{102,1137} Although these systems require an external power source, they provide a wide communication range and reliable signal acquisition. Bluetooth is used in wearable sensors due to the much longer transmission distance (~10 m for classic, ~100 m for industrial use), higher frequency (2.4 GHz), and data transmission speed

(~3 Mbps). Compared with NFC electronics, a Bluetooth module generally needs a more complex design and larger size for the power management circuit and an efficient antenna.¹¹³⁸ Martin et al. demonstrated real-time sweat monitoring through a soft microfluidic sensor integrated with a Bluetooth communication module ($5.0 \times 2.4 \text{ cm}^2$).¹¹³⁹ Qazi et al. demonstrated remotely programmable wireless networks that can control multiple devices simultaneously.¹¹⁴⁰ This wireless system facilitates a large-scale collection of biosignals from various locations.

Wi-Fi possesses the largest transmission distance capabilities (~30 m indoors and ~100 m outdoors), frequency (2.4, 5, and 6 GHz), and data transmission speed (up to 1 Gbps). The unique benefit of Wi-Fi is the cloud backup, which is accessible to multiple devices and users. This feature enables collaborative monitoring and analysis of health data. In contrast, NFC is limited to point-to-point communication, and Bluetooth requires pairing or Bluetooth Mesh for mesh networking. Therefore, the energy consumption of Wi-Fi modules is higher than that of NFC and Bluetooth, leading to a higher requirement for efficient power resources and a challenge to the miniaturization of wearable sensors.¹¹⁴¹ Hybrid communication systems can be developed to meet diverse requirements such as transmission distance, energy consumption, and multiple signal channels. For example, Gaetani et al. stored the hand motion data in a cloud database through Wi-Fi for display and sent EMG data through BLE to control an armband.¹¹⁴² Such hybrid systems can leverage the strengths of different wireless technologies based on application needs, such as using NFC for low-energy, short-range interactions and Wi-Fi for long-range, high-data applications.

In a real scenario of wearable applications, the decision between NFC and BLE methodologies depends on the specific use case. NFC excels in close-proximity, energy-efficient systems, particularly for disposable or semidisposable devices like those developed by Maroli et al.⁸⁵⁹ The integration of the SIC4341 potentiostat ensures compact and robust electrochemical sensing for applications such as POC diagnostics, where portability and simplicity are paramount. In contrast, BLE is more suited for applications requiring long-range, continuous monitoring. The LMP91000 potentiostat paired with BLE, as shown in Yang et al.'s work,⁵³¹ provides greater flexibility for high-bandwidth data transmission and remote monitoring scenarios, such as fitness tracking or hospital-based multisensor networks. A comparison between these two methodologies is given in Table 5.

In summary, integrating wearable sensors with wireless communication modules such as NFC, Bluetooth, and Wi-Fi presents challenges in miniaturization, power consumption, security, and standardization. The foremost challenge lies in designing ultracompact wearable devices without compromising their functionality. Additionally, the high-power consumption of Bluetooth and Wi-Fi limits the battery life, making it difficult to maintain reliable data transmission under stringent power constraints. To address this, low-power communication protocols, such as BLE, and advanced energy-harvesting technologies should be investigated to prolong the battery life. Furthermore, signal interference in densely populated environments, coupled with environmental factors such as sweat and movement, can degrade the transmission reliability. Implementing interference management strategies, including frequency hopping and dual-band systems, may help mitigate these issues.^{1143,1144} Robust

Table 5. Comparison between NFC and BLE with Specific Types of Devices

Feature	SIC4341 (NFC)	BLE with LMP91000
Range	~10 cm	Up to 100 m
Power Source	Energy harvested from the reader	Battery required
Data Rate	~424 kbps	Up to 2 Mbps
Integration	Single-chip system for simplicity	Two-chip system (BLE + potentiostat)
Energy Consumption	Extremely low (no battery needed)	Ultralow (long battery life)
Cost and Complexity	Low	Moderate
Security	High (limited range)	Moderate to Low (extended range)
Use Cases	Disposable diagnostics, patch biosensors	Fitness trackers, remote health monitors

encryption protocols and secure authentication mechanisms are also essential to protect sensitive health data.¹¹⁴⁵ Developing standardized, modular designs and employing precertified modules can enhance compatibility with diverse healthcare systems and streamline adherence to regulatory requirements.

10. Energy Harvesting for Self-Powered Wearable Sensors

Harvesting energy from the human body or the environment to power wearable sensors is a promising solution for extending the operational time of the current devices. This approach eliminates the need for frequent battery replacement or charging, enhancing the convenience and willingness to use such devices. More importantly, it enables continuous and long-term collection of human health data, achieving around-the-clock health management. Several energy-harvesting technologies have been developed around the three primary energy sources available from the human body: mechanical energy from the movement of organs or limbs, thermal energy from body heat or breath, and chemical energy, such as from sweat glucose (Figure S0). Each of these technologies has its own unique application scenarios, which will be detailed below.




Figure S0. Energy sources from the human body.

10.1. Mechanical Energy Harvesting. 10.1.1. Triboelectric Nanogenerators. The TENG is a mechanical-electrical energy conversion device that combines contact electrification with electrostatic induction to achieve high material compatibility, high cost-effectiveness, and the ability to capture ultralow-frequency and atypical mechanical energy.^{1146–1152} Because TENGs are probably the most used devices for energy harvesting for self-powered sensors, a longer description will be provided compared to the other harvesting mechanisms. The working principle of TENGs mainly relies on the coupling of contact electrification and electrostatic induction, as shown in Figure S1. When two materials with




Figure S1. Schematics of the working principle of TENGs. Reproduced with permission from Xia et al.¹¹⁵³ © 2019 Elsevier.

different triboelectric polarities come into contact and then separate, electrons will transfer between them due to the difference in their ability to gain or lose electrons. This leads to the generation of an electric potential difference between the electrodes attached to these materials. When an external circuit is connected, electrons will flow through the circuit to balance the potential difference, thus generating an electrical current. For example, in a typical vertical contact-separation mode of a TENG, initially, there is no charge on the uncontacted surfaces of the dielectric materials (such as PMMA and commercial polyimide (Kapton)). Under external force, contact between these surfaces results in the generation of equal amounts of heterogeneous charge. When the external force is withdrawn, a potential difference forms between the two electrodes in the open-circuit state, and the voltage increases with the separation distance until saturation. When the external force is reapplied, and the materials gradually come back into contact, the heterogeneous charges on the surfaces are gradually neutralized, causing the potential difference between the two electrodes to disappear gradually.

There are four main operating modes for TENGs in common use today, as indicated in Figure S2A: vertical contact-separation mode, single-electrode mode, lateral sliding mode, and free-standing mode. TENGs for wearable sensors can have diverse structures depending on the specific application requirements. One common structure is the planar or sandwich structure, where the triboelectric layers are stacked one on top of the other with electrodes on either side to collect generated charges. In a simple pressure-sensing wearable TENG, a flexible polymer layer may act as one triboelectric

material and a thin metal film as the other, with the electrodes integrated into the substrate. Another structure is the fiber- or textile-based structure, which is suitable for seamless integration into clothing for wearable applications. This allows for comfortable and unobtrusive wear while being able to detect various body movements and pressures. There are also more complex structures such as multilayer or hybrid structures. Multilayer TENGs can enhance the electrical output by increasing the number of contact-separation interfaces. Hybrid structures may combine TENGs with other functional components such as piezoelectric elements or energy storage units to improve the overall performance and functionality of the wearable sensor. For example, integrating a piezoelectric layer with a TENG layer can enable the sensor to detect both the mechanical pressure and vibration simultaneously.

A wide range of materials can be used in TENGs for wearable sensors. Polymer materials are popular due to their flexibility, lightweight nature, and ease of processing. PTFE is a commonly used polymer with excellent triboelectric properties. Other polymers like PET, PI, and PDMS are also frequently employed. Inorganic materials are also used, including metals like aluminum and copper, and nanomaterials such as carbon nanotubes and graphene. They can help with improving the charge transfer efficiency and the overall performance of the sensors. Composite materials have also gained prominence. For instance, a composite of a polymer with embedded inorganic fillers can achieve a balance between flexibility and enhanced triboelectric performance. Furthermore, biobased and biodegradable materials are being explored for more environmentally friendly and biocompatible wearable TENGs, especially in applications with long direct contact with the skin, such as in healthcare monitoring.

10.1.2. Piezoelectric Nanogenerators. The operating principle of piezoelectric nanogenerators (PENGs) is the piezoelectric effect. When an external force acts on a piezoelectric material, the anions and cations in the lattice will be displaced from each other, resulting in the centers of the positive and negative charges no longer coinciding and thus generating an electric dipole moment. The superposition of the electric dipole moment is macroscopically manifested as a difference in the potential distribution of the entire material in the direction of the external force (Figure S2B). Typical piezoelectric materials include inorganic materials such as zinc oxide (ZnO), lead zirconate titanate (PZT), and barium titanate, as well as some organic materials, such as polyvinyl chloride, polyvinylidene fluoride, and its derivatives.¹¹⁵⁴

10.1.3. Electromagnetic Generators. The basic principle of electromagnetic generators (EMGe) is to convert kinetic energy into electrical energy by using the law of electromagnetic induction. Since the invention of the first alternator in 1866, EMGe has been essential for the progress of human society and is a core pillar of modern energy. EMGe is generally composed of a stator, rotor, end-caps, bearings, and other components. Driven by external forces, the rotor rotates in the stator, making the movement of cutting the magnetic inductance (Figure S2C) due to the law of electromagnetic induction to generate an induced electric potential, thus generating a current in the external circuit.¹¹⁵⁵

TENG, PENG, and EMGe are based on different operating principles, but they all share the common characteristic of converting mechanical energy to electrical energy. The output of TENG is characterized by high voltage and low current,




Figure 52. Emerging energy harvesting technologies. (A) Triboelectric nanogenerator. (B) Piezoelectric nanogenerator. (C) Electromagnetic generator. (D) Pyroelectric nanogenerator. (E) Thermoelectric generator. (F) Biofuel. (G) Hydrovoltaic effect generators. (H) Moisture effect generators.

while the output of EMGe is high current and low voltage, and the output voltages and currents of PENGs are both moderate. The three have similar application scenarios and complementary advantages, and either individually applied or composite devices have important roles in the field of wearable devices.

10.2. Thermal Energy Harvesting. 10.2.1. Pyroelectric Nanogenerators. The pyroelectric nanogenerator (PyNG) is an energy-harvesting device that utilizes materials with a pyroelectric effect to convert thermal energy into electrical energy.^{1156,1157} The pyroelectric effect refers to the change of the spontaneous polarization of certain crystals when they are heated to different temperatures. The working principle of the PyENG may be explained as follows. When the temperature is constant, the crystals have a stable strength of spontaneous polarization; when the temperature increases or decreases, it will lead to a decrease or increase in the strength of spontaneous polarization, which induces a change of charge on the electrodes and causes the generation of a pyroelectric current in the circuit (Figure 52D).

10.2.2. Thermoelectric Generators. Thermoelectric generators (TEGs) are based on the Seebeck effect, which converts the energy present in a temperature gradient into electrical energy. The Seebeck effect comes from the movement of carriers within a conductor under a temperature gradient, and when there is a temperature difference between the two ends of the loop formed by the two conductors, there is a potential difference and a current.^{1158,1159} A commonly used TEG, shown in Figure 52E, consists of an N-type semiconductor, a P-type semiconductor, and electrodes. The

temperature difference causes the directional movement of electrons and holes in the two semiconductors, which produces a DC output in the circuit.

Both PyENG and TEG can convert thermal energy into electrical energy, but the difference is that the former utilizes thermal energy that varies over time, and the latter utilizes thermal energy that varies over space. In addition, PyENG has a larger output voltage and smaller current, with its output depending on the rate of temperature change. This is suitable for environments with rapid temperature fluctuations. On the other hand, TEG has a larger current, a relatively high power density, and is easy to output in series, but it requires a stable heat source and heat dissipation conditions. Both technologies can be used to collect human body heat to power wearable devices.

10.3. Chemical Energy Harvesting. 10.3.1. Biofuel Cells. A BFC can generate electricity by obtaining biochemical energy from the environment of the living organisms. BFCs generate electricity through redox reaction mechanisms and can be categorized into enzyme fuel cells and microbial fuel cells, characterized by their mild actuation conditions and being environmentally friendly and sustainable. Figure 52F shows a typical glucose biofuel cell using GOx as a catalyst, where glucose is oxidized to gluconolactone at the anode, while oxygen reduction takes place at the cathode.¹¹⁶⁰ In addition to glucose, BFC can utilize substances such as lactic acid, urea, vitamin C, and ethanol, to generate power for wearable devices.¹¹⁶¹

10.3.2. Hydrovoltaic Effect Generators. The hydrovoltaic effect generator (HEG) was proposed in 2017 after the




Figure 53. Self-powered wearable devices for health monitoring. (A) Wearable sweat sensor system powered by TENG. Reproduced from Song et al.⁴⁸⁷ Available under a CC BY-NC license. © 2020 American Association for the Advancement of Science. (B) Hybrid nanogenerator-based sweat analysis system. Reproduced with permission from Gai et al.¹¹⁶⁵ © 2022 John Wiley & Sons. (C) Self-powered PENG sensor for measuring arterial pulses. Reproduced with permission from Park et al.¹¹⁶⁷ © 2017 John Wiley & Sons. (D) Self-powered TENG sensor for monitoring muscle activity. Reproduced from Wang et al.¹¹⁶⁸ © 2021 American Chemical Society. (E) TENG-based wearable knee rehabilitation sensor for tracking joint angle. Reproduced from Luo et al.¹¹⁶⁹ Available under a CC BY 4.0 license. © 2022 John Wiley & Sons.

discovery that the evaporation of water from a carbon black sheet could produce a sustained voltage of up to 1 V.¹¹⁶² The materials used for HEG expanded to include natural materials such as bacterial membranes, wood, and lotus leaves as well as various organic or inorganic materials such as MXene, ZnO, and Si NWs. It is now believed that the principle of the HEG is the kinetic effect. The pressure gradient caused by the evaporation of water in the nanomaterials drives the flow of water in the nanochannels, causing the diffusion layer in the overlapping bilayers to move from upstream to downstream, which in turn creates a potential difference (Figure 52G).

10.3.3. Moisture Effect Generators. The moisture effect generator (MEG) relies on the dissociation of free ions from functional groups of hygroscopic materials after absorbing moisture from the environment.¹¹⁶³ The ion concentration gradient prompts the directional diffusion of ions, generating a significant built-in electric field that results in the formation of a potential difference between the two electrodes (Figure 52H). The ionic concentration gradient is an essential principle of power generation. There are three ways to construct an ionic concentration gradient within a material: asymmetry in the chemical structure of the material, asymmetry in the distribution of moisture due to the asymmetry of the electrodes, and a nonhomogeneous structure of the assembled material.¹¹⁶⁴ Compared to HEG, MEG does not require bulk water sources and only utilizes moisture in the air, which may further increase the portability of wearable devices and broaden their application scenarios.

The above three energy-harvesting technologies for chemical energy all have superior DC output characteristics. Moreover, BFC can directly utilize chemical energy from human metabolites, and HEG and MEG can convert water energy in the environment to electrical energy for wearable devices, all

three of which have the potential for clean, safe, and sustainable development.

10.4. Self-Powered Sensing Strategies. The characteristics of the energy-harvesting technologies described should be considered in applying them to distinct wearable health-monitoring devices. Factors such as the monitored body part, environmental conditions, and performance must be accounted for. There are two basic self-powered-sensing strategies. The first approach is to construct self-powered sensing systems based on energy harvesting techniques and power management strategies. The second approach involves the direct construction of self-powered sensors using energy conversion devices. The strategy of the self-powered sensing system is to integrate an energy harvester, energy storage unit, and power management unit (PMU) to ensure a stable power supply for the sensors. This method is advantageous for active sensors that require higher or more stable power input. A PMU typically consists of components, such as rectifiers and voltage regulators. These components convert the alternating current generated by energy harvesters (such as TENGs and PENGs) into direct current and then store it in capacitors or batteries. The PMU also ensures that the stored energy is delivered to the sensor as needed, providing a steady power supply. This strategy is especially beneficial when the harvested energy is insufficient, as it helps smooth out energy fluctuations. By integrating energy storage and regulation, the PMU-based approach maintains the balance between energy production and consumption, ensuring efficient sensor operation over time.

A typical case of these strategies in wearable health monitoring is sweat analysis. For instance, Song et al. developed a wearable sweat sensor system powered by a TENG (Figure 53A).⁴⁸⁷ The system incorporates a PMU that stores harvested energy and supplies it to the sensor for the




Figure S4. Structure, composition, working principle, characteristics, and signal output of SUPS. (A) Schematic representation of the device, showing each material layer. Scanning electron microscopy (SEM) and atomic force microscopy (AFM) from (B) Cu and Kapton film and (C) Kapton film. (D) Schematics of the basic working mechanism of the device. (E) Picture showing the flexibility of the device; (F) Picture showing the size of the built device; and a sample signal when the sensor is placed on top of the (G) finger and (H) radial artery of a patient. Reproduced with permission from Ouyang et al.¹¹⁷¹ © 2017 John Wiley & Sons.

real-time monitoring of biomarkers like pH and sodium. The design of a low-power wireless sensor circuit and flexible PCB makes it suitable for continuous monitoring under comfortable wear. Gai et al. designed a wearable sweat analysis system that uses a hybrid nanogenerator module to harvest energy from human motion (Figure S3B).¹¹⁶⁵ The stored energy is used to power sensors for real-time analysis of biomarkers such as Na^+ and K^+ . Ding et al. demonstrated a wearable system powered by a biofuel cell that harvests energy from sweat.¹¹⁶⁶ The energy is stored in a microgrid system, which powers a range of sensors for continuous metabolic sensing, efficiently monitoring metabolites such as glucose and vitamin C.

Another strategy is to construct self-powered sensors that convert environmental and biological physical or chemical changes directly into electrical signals, without the need for external power. Most energy harvesters can also serve as self-powered sensors themselves and work in the form of passive sensors, directly converting the human body's respiration, pulse, body temperature, and sweat components into electrical signals that can be analyzed. The sensitivity of these sensors is positively correlated with the electrical output, which means the higher the output signal, the greater the sensitivity. Unlike other types of sensors, where higher sensitivity requires higher energy consumption, self-powered sensors achieve high sensitivity without additional power consumption. Therefore, the self-powered capability of these sensors is crucial for balancing ultrahigh sensitivity with low power usage, making them ideal for applications that require high performance with minimal energy demands. For example, Park et al. developed a

self-powered piezoelectric pulse sensor that converts mechanical energy from arterial pulse waves into electrical energy (Figure S3C).¹¹⁶⁷ This sensor, made from an inorganic piezoelectric material, measures pressure pulses in the radial and carotid arteries and wirelessly transmits the data to a mobile device. Wang et al. proposed a skin-mounted electrostatic sensor that can obtain muscle information from skin deformation by detecting signals from both the biceps and triceps (Figure S3D).¹¹⁶⁸ Luo et al. introduced a wearable self-powered knee rehabilitation sensor embedded in a knee brace to track muscle strength and joint angles in knee replacement patients (Figure S3E).¹¹⁶⁹ Zou et al. reported a stretchable self-powered respiratory sensor inspired by the shark gill with a PENG and a TENG combined in bionic structures.¹¹⁷⁰

Triboelectric nanogenerators offer a promising alternative as they can harvest energy from body movements and convert it into electrical signals for sensing purposes.^{1148–1152} This self-powering capability is crucial for wearable sensors as it eliminates the need for frequent battery replacements and enables seamless integration into clothing or accessories for continuous monitoring.

10.4.1. Pressure-Sensing Applications. TENGs have remarkable potential in wearable pressure-sensing applications because of their ability to convert mechanical pressure stimuli into electrical signals. For instance, Ouyang et al. (2017) developed a flexible self-powered ultrasensitive pulse sensor (SUPS) based on a triboelectric active sensor for cardiovascular disease diagnostics (Figure S4).¹¹⁷¹ The SUPS was designed with nanostructured Kapton and Cu films as




Figure 55. Scheme showing the preparation process of TAP-TENG. Reuse of (A) aluminum plastic bags and (B) tea leaves. (C) Processing of materials. Reproduced with permission from Xia et al.¹¹⁵³ © 2019 Elsevier.

triboelectric layers. When pressure was applied, such as when placed on the radial artery, contact electrification and electrostatic induction mechanisms came into play. As the external force was applied and released, electrons flowed back and forth through the external circuit, generating an electrical signal that was modulated by the force. The SUPS could detect the pulse pressure and provide reliable information on the cardiovascular system. It had excellent output performance with an open-circuit voltage of 1.52 V, a high peak signal-to-noise ratio of 45 dB, and a long-term stability over 107 cycles. Another example is the work of Xia and co-workers. (2019),¹¹⁵³ where a tea leaf triboelectric nanogenerator using aluminum plastic bags (TAP-TENG) was fabricated (Figure 55). The TAP-TENG was designed with a unique structure that could respond to different pressures *stimuli*. When used in an office environment, it could detect various human touching states like knocking, pressing, and stroking. The different pressure magnitudes and durations during these actions resulted in different electrical output signals. For example, when the honeycomb lantern TAP-TENG structure was integrated into a book, the turning of pages exerted different levels of pressure on the TENG, and it was able to harvest the mechanical energy generated by these actions and convert it to electrical energy. This indicates that TENGs can be applied in scenarios in which the detection of variable pressure inputs is required for interactive or monitoring purposes.

Materials and structures have been optimized to enhance the sensitivity and range of pressure detection using TENGs. For example, nanostructuring triboelectric layers can increase the effective contact area during pressure application, thereby enhancing the charge transfer and improving the electrical output signal strength. This enables the detection of even very subtle pressure changes typical of gentle touches in wearable haptic feedback systems or monitoring the light pressure exerted by body movements in smart clothing for health and fitness tracking.

10.4.2. Acoustic Sensing Applications. TENGs are promising for acoustic sensing in wearable devices. In audible sound sensing, TENGs can function as self-powered acoustic sensors by converting acoustic vibrations in the audible frequency range (20–20,000 Hz) into electrical signals.¹¹⁷² The contact-separation mode of TENGs plays a vital role here. When sound waves impinge on the triboelectric layers of the TENG, they cause periodic contact and separation of the layers. For instance, in a simple dielectric-dielectric contact-separation TENG setup, as sound waves exert pressure on the materials, the resulting mechanical deformations lead to the generation of charges due to triboelectrification and subsequent electrostatic induction-driven flow of electrons in the external circuit. This enables the detection of sound signals. TENG-based sensors have been investigated for hearing aids owing to their miniature structure and because they do not require an external power supply. They can capture ambient sounds and convert them into electrical signals to be further processed and enhance the hearing experience for individuals with hearing impairments. Compared with traditional piezoelectric-based acoustic transducers used in some hearing aids, TENGs have the advantages of higher output power density under certain conditions and are less likely to be affected by interference in some frequency ranges. This enables them to provide a more stable and clear sound amplification effect for hearing aid users. In cochlear implants, TENG-based systems can take advantage of the body's mechanical movements or acoustic vibrations to generate electricity.^{1173,1174} For example, when a person is talking, walking, or in an environment with ambient sound, the TENG integrated into the cochlear implant can convert these mechanical *stimuli* into electrical signals. The miniature structure of TENGs allows them to be implanted in a relatively small space within the ear, and their ability to work without an external power supply makes them more suitable for long-term use. The acoustic energy from the external environment can drive the contact separation or other working modes of the TENG, generating electrical signals that can be




Figure 56. (A) Schematic representation of the acoustic device implanted into the human body. (B) Structure of FUTUD and of (C) FEMC. Reproduced with permission from Sun et al.¹¹⁷⁶ © 2023 Elsevier.

further processed and used to stimulate the auditory nerve, thereby helping patients with hearing impairments to perceive sound.

10.4.3. Ultrasonic Sensing Application. TENGs can be used for ultrasonic sensing,¹¹⁷⁵ since frequencies higher than 20,000 Hz have unique characteristics to be harnessed by such technology. They may be used for noninvasive or minimally invasive monitoring, as in the diagnosis of certain internal organ conditions. Ultrasound waves are sent into the body, and the TENG-based sensors can detect the reflected signals. Yang et al. developed a flexible, ultrawideband triboelectric ultrasonic device (FUTUD),¹¹⁷⁶ as shown in Figure 56. The FUTUD was integrated with a flexible electromagnetic coil to transmit electrical signals containing physiological information from *in vivo* to *ex vivo* wirelessly as a passive wireless sensor. The mechanical vibrations caused by the ultrasound on the TENG's triboelectric layers result in charge generation and electrical signal output. By analyzing these signals, information about the structure and health of internal organs such as the liver, kidneys, or blood vessels can be obtained. This can aid in the early detection of diseases like tumors, cysts, or vascular disorders.

Ultrasonic sensing using TENGs can be integrated into patches or bands to monitor ultrasonic signals related to physiological processes. For instance, they can detect changes in blood flow in deep tissues by sensing ultrasonic echoes. This can provide valuable data on cardiovascular health and help in the long-term management of conditions such as hypertension or atherosclerosis. In addition to medical applications, in industrial and environmental monitoring scenarios, where wearable sensors are required, TENG-based ultrasonic sensors can be used. In industrial settings, they can detect flaws or cracks in materials by analyzing the ultrasonic reflections from the surfaces of the structures. Workers can wear these sensors as part of their personal protective equipment to quickly identify potential safety hazards in machinery and building components. In environmental monitoring, ultrasonic TENG sensors can be used to detect the presence of certain objects or changes in an underwater environment. For example, in marine research, they can help in tracking the movement of marine organisms or monitoring the health of coral reefs by sensing the ultrasonic signals that they emit or interact with. In conclusion, the ultrasonic sensing capabilities of TENGs in wearable sensors offer a wide range of applications that span

multiple domains, contributing to improved health monitoring, industrial safety, and environmental understanding.

11. The Use of Machine Learning and AI for Data Analysis

Health monitoring is not a new concept. Recording symptoms for diagnosis can be traced back to ancient times, primarily through visual and tactile observations. Over the centuries, various tools and instruments have been introduced, including noninvasive technologies such as ECG, pulse oximeters, and wearable devices. What is remarkable in recent years is the convergence of two technologies that have been revolutionizing health monitoring. As amply discussed in this paper, wearable devices enable continuous, real-time tracking of health conditions, while AI methodologies are applied to process and interpret the vast amounts of data generated by continuous monitoring. In this section, we present several examples of how ML and other AI techniques are employed for health monitoring. Before delving into these examples, we introduce the basic concepts of these AI tools along with related topics in data analytics.

11.1. Data Analytics and AI Tools. The analysis of the literature presented in Section 2 indicates that the use of ML and other AI methods for health monitoring has surged over the last 10 years, with most publications appearing within the past 5 years. However, ML has been employed to analyze sensing data for decades, often under the framework of artificial neural networks (e.g., electronic tongues¹¹⁷⁷). For wearable sensors, in particular, recent review papers have illustrated a variety of applications and challenges.⁵⁷ The integration of ML and AI in sensing and biosensing has been extensively discussed, with many authors emphasizing their potential for transformative advancements in both materials science and AI. For instance, the central role of sensing and biosensing in the technological revolution driven by ML and Big Data has been associated with ubiquitous sensing, which is expected to transform various sectors.¹¹⁷⁸ Additionally, the application of ML in sensor and biosensor data analysis could be extended to encompass knowledge discovery, aiming for a paradigm shift toward machine-generated knowledge.^{1179–1181} This concept is now being realized through tools powered by large-language models. Since the principles and applications of ML and AI in sensing have been reviewed comprehensively in several papers,⁵⁷ we provide only a concise overview of key definitions to help readers understand the descriptions and examples presented herein.

Data analytics refers to the systematic computational analysis and interpretation of data, focused on discovering, interpreting, and communicating meaningful patterns in data to guide decision-making. Data analytics directly intersects with several other key disciplines related to Mathematics, Statistics, and Computer Science, particularly ML in AI and Data Visualization.¹¹⁸² ML uses algorithms and statistical models to identify patterns in data and make decisions with minimal human intervention. As such, it gives computers the ability to learn descriptive and predictive models from data without being explicitly programmed. Tom Mitchell, in his seminal book,¹¹⁸³ states that “A computer program is said to learn from experience E with respect to some class of tasks T and performance measure P, if its performance at tasks in T, as measured by P, improves with experience E.”

Information visualization has been defined as the study of visual representations of abstract data to reinforce human cognition and derive insights.¹¹⁸⁴ It focuses on the creation and optimization of computer-based interactive graphical representations to explore and understand the data. *Multi-dimensional Data Visualization* specifically refers to techniques and methods for visually representing data described by more than three dimensions or variables (hence referred to as multidimensional, or high-dimensional data), allowing users to explore relationships between multiple attributes simultaneously.^{1185,1186} *Visual analytics* has been defined as the science of analytical reasoning facilitated by interactive visual interfaces. It focuses on combining techniques from data visualization, ML, and statistics to enable users to explore, understand, and interpret complex data through visual representations to support decision-making by allowing users to identify patterns, trends, and insights in large and complex data sets.¹¹⁸⁷

A *dimension reduction* technique is a mathematical method that transforms high-dimensional data into a lower-dimensional form while preserving meaningful structures and relationships in the data, making it more manageable for analysis, processing by algorithms, and visualization.¹¹⁸⁸ A *multidimensional projection* technique is a specific type of dimension reduction method that maps high-dimensional data onto a lower-dimensional (typically 2D or 3D) visual space while attempting to preserve relative distances or relationships between data points, making complex data structures visually interpretable.¹¹⁸⁹ The key relationship with dimension reduction is that multidimensional projection is essentially dimension reduction with a specific focus on creating meaningful visual representations, while dimension reduction can be used for various purposes, such as preprocessing or computation efficiency of ML algorithms.

In ML, *supervised learning* refers to an approach where an algorithm learns to make predictions from a data set containing input-output pairs where the desired outputs (called labels or targets) are known. When these outputs are categorical (e.g., spam/not spam or dog/cat/bird/fish), the task is called classification. When the outputs are continuous numerical values (e.g., temperature and house prices), the task is called regression. In both cases, the algorithm learns to discover patterns in the input features that allow it to map new, unseen inputs to their likely outputs, attempting to generalize from the training examples that it was shown. SVM is a particular algorithm for classification that works by finding an optimal hyperplane in an N-dimensional space to distinctly separate data points into different classes by maximizing the margin

between the nearest training points (called support vectors) of any class.¹¹⁹⁰ In *unsupervised learning*, an algorithm learns patterns, structures, or relationships without being given input-output pairs or predefined categories, discovering hidden patterns and groupings independently, directly from the data points.

Deep learning is a subset of machine learning that uses artificial neural networks with multiple layers (hence “deep”) to automatically learn representations and patterns from data. While loosely inspired by biological neural networks, modern deep learning architectures and algorithms are primarily mathematical/computational constructs that may work quite differently from the human brain. These models excel at discovering hierarchical features in data with each layer learning increasingly abstract representations. The approach has proved particularly effective for tasks like image and speech recognition, natural language processing, and computer vision components of autonomous driving systems.¹¹⁹¹ CNNs are a class of deep learning architectures specifically designed for processing structured grid data, such as images. They use convolutional layers to automatically detect features (e.g., edges and textures) at various levels of abstraction, enabling the model to learn spatial hierarchies. CNNs are widely used in image and video recognition, natural language processing, and other computer vision tasks.^{1192,1193}

Explainable Machine Learning (XAI) is the broad field focused on making AI systems understandable to humans. Explainable Machine Learning, in turn, is a subset of XAI that encompasses methods and techniques aimed at making ML systems’ decisions and behaviors understandable to humans. This includes both developing inherently interpretable models and creating tools to explain existing complex “black-box” models. The field addresses several key aspects: understanding how models arrive at specific decisions, identifying potential biases or failure modes, verifying compliance with regulations, and enabling meaningful human oversight. XAI techniques range from simple approaches like feature importance rankings to sophisticated methods that generate human-friendly explanations of model behavior through examples, rules, or visualizations.^{1194,1195} This is particularly relevant in practical application scenarios that require decisions to comply with fairness and safety principles, as well as with ethical and legal regulations.

Finally, *IoT* refers to the network of physical devices embedded with sensors, software, and other enabling technologies for collecting, exchanging, and processing data over the Internet. These interconnected devices allow for automation, remote monitoring, and improved decision-making across various industries, such as healthcare, transportation, and smart homes.^{1196,1197}

The connection between (bio)sensing and data analytics is crucial, as analytics transform raw sensor data into actionable insights. Sensor data are often unprocessed and require analytics to extract meaningful trends, patterns, and relationships. For example, a glucose sensor produces raw electrical signals that must be calibrated, processed, and interpreted to estimate the blood sugar levels. When large volumes of data need processing, dimensionality reduction techniques are often applied to high-dimensional data, e.g., measurements obtained from multimodal biosensors. Dimension reduction methods such as principal component analysis (PCA) or t-SNE are commonly employed for visualization and analysis, as the resulting visual representations depicted as point cloud

similarity views can give insight into the distribution, grouping patterns, and outliers in multidimensional data. In similar scenarios, it is also possible to apply classification and clustering algorithms to the data to obtain explicit categories or groups. Data analytics form the basis for automated or assisted decision-making in sensing systems. For instance, an air quality monitoring system may trigger alerts when pollutant levels exceed predefined thresholds. Analytics also help optimize sensor performance, such as compensating for drift or identifying faulty sensors. Key processes in sensor data analytics include data preparation, which involves aggregating, cleaning, and transforming data for analysis as well as handling missing values or noisy data points during preprocessing. Visualization and statistical methods are employed to explore the data, understand their characteristics, and identify anomalies or trends in sensor readings. Finally, statistical models and machine learning algorithms are applied to analyze the processed sensor data and obtain prediction models such as regression models to predict temperature trends or classification models for biosensors that detect specific biomarkers.

The ability to perform classification tasks, essential for diagnosis and health monitoring, is a significant feature of ML in analyzing sensing data. Additionally, ML methods are highly versatile, capable of handling diverse data types and excelling in scenarios where synergy is achieved through data fusion from multiple sources and modalities. Whether the data signals are electrical, optical, mechanical, or thermal—or whether ML algorithms process inputs like images, text, or other formats—their adaptability remains consistent. For health monitoring, ML can be combined effectively with time-series analysis and pattern recognition techniques, e.g., applied to images. The application of AI methods to process multidimensional data captured by multimodal sensing systems has been reviewed by Mahato et al.¹¹⁹⁸ The authors outline several potential applications, as illustrated in Figure 57, which also includes a flowchart depicting data processing and the development of intelligent systems.

11.2. Uses of AI in Health Monitoring. The use of AI in health monitoring, particularly ML, is growing at such a fast pace that it has become unfeasible to provide an exhaustive




Figure 57. Flowchart illustrating the integration of data fusion with Big Data and cloud computing methodologies in developing intelligent systems capable of making predictions for diagnosis and health monitoring. These systems ideally employ hand-held readers, such as smartphones. A nonexhaustive list of potential applications is displayed in the box on the right. Adapted with permission from Mahato et al.¹¹⁹⁸ © 2024 Springer Nature.

account of recent developments. We have thus selected a few representative examples, focusing primarily on ML and pattern recognition as the key subfields of AI. While we may refer to AI and ML interchangeably in this context, we remind readers that ML comprises a subarea of AI.

Various types of physical sensors, particularly wearable strain and pressure sensors, are used to analyze gait patterns, which are relevant for health monitoring as they provide insight into an individual's physical health, neurological condition, and overall functional ability. Gait patterns can serve as indicators of neurological diseases such as Parkinson's disease and multiple sclerosis, with changes in stride length, walking speed, or symmetry often signaling neurological impairments. Gait analysis is also instrumental in fall risk assessments and patient rehabilitation monitoring. In many of these applications, AI is essential in data analysis. For example, Xiang et al.¹¹⁹⁹ developed an AI-assisted insole sensing system for real-time gait monitoring. This system integrates silicone rubber with PMMA optical fibers (S-POF), enabling the conversion of pressure measured at key points into changes in light intensity. The insole system and its operating principle are depicted in Figure 58.

One particularly interesting aspect of this example is that the solution developed integrates combined pressure sensing data and image processing. As we will discuss in the Prospects and Outlook section, combining ML with image analysis of sensing units has the potential to transform health monitoring. Such approaches could enable almost instrument-free strategies, requiring only cameras or microscopes that may be embedded into routine devices such as smartphones. In the aforementioned study,¹¹⁹⁹ signal processing and data analysis were performed through several steps. The first module acquires data from the camera, transmits it to the computer, and converts the signal to grayscale. Three spots represent data from the channels corresponding to three positions on the insole. Adaptive thresholds and image segmentation are used within a multiplexing framework for the grayscale summation matrix to be processed. To extract gait information from each of the three channels individually, filtering and power-spectrum transformation are applied. These processed signals are then input into a neural network algorithm, which predicts and differentiates gait patterns.

Details of the identification of gait patterns are provided in Figure 59. Three regimes of walking speeds were defined, as follows: (i) normal walking speed, between 1.33 and 1.67 m/s; (ii) slow walking, below 1.33 m/s; and (iii) fast walking, above 1.67 m/s. For running, the authors also established three regimes: normal running speed, within the range of 2.67 to 3.33 m/s, and slow and fast running speeds, defined as below and above this range, respectively. Figure 59a illustrates the experimental setup, where participants were asked to walk specific distance ranges within a fixed time (30 s, as shown in the figure) for 20 repetitions in each case. This procedure ensured the acquisition of sufficient data for subsequent classification. The data set comprised 1,000 data instances, corresponding to 5 subjects $\times$ 10 routines $\times$ 20 repetitions. Each data instance included three arrays of frequency values and three arrays of pressure values, derived from time-domain and frequency-domain measurements. The histograms in Figure 59b compare the accuracy of predicting gait patterns using five classification models: decision trees, random forests, K-nearest neighbors (KNN), SVM, and Bayesian neural networks. Neural networks achieved the highest performance,




Figure 58. Insole system, positioned beneath the foot as shown in the top left of the figure, features 3 sensing channels (S-POF channels) embedded at strategic locations (heel and metatarsophalangeal joints). Plantar pressures are converted into light signals and transmitted via optical fibers to a smartphone camera, as illustrated in the central top portion of the figure. These image signals are processed in real time on a computer system, with adaptive thresholding applied as a preprocessing step. In the final stage, shown in the lower part of the figure, signals in both time and frequency domains are analyzed using neural networks to determine gait patterns (see further details below). Reproduced from Xiang et al.¹¹⁹⁹ © 2024 American Chemical Society.

with an average accuracy rate of 90.62%, as reflected in the confusion matrix shown in Figure 59c. These neural networks were subsequently employed as the classification model for predicting gait patterns, as illustrated in Figure 59d.

The potential for continuous monitoring enabled by wearable pressure and strain sensors has spurred significant research into processing data and correlating them with various types of motion and health conditions. For instance, screen-printed strain sensors composed of a carbon-based conductive network on flexible TPU films were used to detect EMG and ECG signals.¹²⁰⁰ Because these sensors are stretchable, they can conformally adhere to human skin, enabling long-term, continuous monitoring. The sensors are also capable of detecting various types of human motions, including joint movements, mouth opening, chewing, swallowing, breathing, and clenching. Figure 60 illustrates the electrical signals corresponding to several movements along with the confusion matrix generated from using an SVM algorithm to predict six types of knee movements.

Monitoring cardiovascular health has been a prominent topic, as highlighted in several of the previous sections. Due to the intrinsic nature of this type of monitoring—which involves multiple types of instruments and requires time-series analysis—ML is an obvious choice for developing next-generation diagnosis and monitoring systems. This concept has been explored by Wang et al.⁵⁸ and is illustrated in their chart (reproduced in Figure 61), where standard monitoring methods are compared to AI-based approaches. Figure 61A depicts the manual collection and processing of patient data in traditional settings, whereas the AI-based counterpart is shown

in Figure 61B. Significant differences are noted in data acquisition, as wearable devices afford continuous data collection, and in the data processing and diagnostic phases, where ML algorithms are applied. A variety of algorithms can be employed to classify or quantify samples, as outlined in the AI box in Figure 61, with the potential for expansion to include many other methods. Moreover, these algorithms can process diverse data types and are not restricted to those listed in the biosignal acquisition block. Most importantly, the outcomes from data processing can serve as inputs for intelligent systems, enabling functionalities beyond diagnostics, such as developing treatment plans, as exemplified in the AI-assisted diagnostics and treatment box in Figure 61.

In terms of data analysis, ML can be used in classifying data from varied sources, including pulse pressure, ECGs, and ultrasonography images. Figure 62 shows the design and functions of a soft wearable stethoscope (SWS) for cardiopulmonary auscultation.¹²⁰¹ The soft device whose photos are shown in Figure 62A comprises an elastomeric enclosure with a conductive hydrogel layer to auscultate cardiac and respiratory signals, as described in detail in Figure 62B. The latter shows the electronic components such as a microelectromechanical system (MEMS) microphone, a rechargeable battery, a Bluetooth for wireless transmission, and an electric circuit, in addition to the multiple layers of soft materials. Some of the stretching and bending characteristics of the SWS are shown in Figure 62C–F. The ML-based diagnosis system is depicted in Figure 62G, which indicates the flow for sound recording, preprocessing with denoising, and ML classification using CNNs.




Figure 59. (A) Schematic illustration of the scenario for collecting image data on walking and running. Human subjects were instructed to walk or run varying distances within a fixed time period of 30 s. (B) Histograms showing the prediction accuracy of gait patterns using five ML algorithms. (C) Confusion matrix for gait pattern predictions made by neural networks. (D) Flowchart illustrating the process of gait pattern prediction using neural networks. Reproduced from Xiang et al.¹¹⁹⁹ © 2024 American Chemical Society.

The combination of ML and image analysis is gaining popularity, with classifications not based on images of biological samples but rather on images of the sensing units themselves.¹²⁰² This approach offers the clear advantage of directly using cameras or optical microscopes for monitoring and diagnosis. A notable example is the integration of mechanoluminescent sensing and biosensing, which facilitated the detection of bacterial infections while simultaneously measuring the orthodontic force.¹²⁰³ The principles of detection are illustrated in Figure 63, featuring a mechanoluminescent elastomer composed of polyvinylidene fluoride-hexafluoropropylene (PVDF-HFP) embedded with $\text{SrAl}_2\text{O}_4:\text{Eu}^{2+}, \text{Dy}^{3+}$ (SAOED) phosphors to measure force. A biosensor consisting of a covalently cross-linked methyl acrylate-phenol red (MA-PR) and PAM film remains in direct contact with the teeth. Biosensing functions through the principle that lactic acid, a metabolic product of cariogenic bacteria, induces a color change upon reaction with MA-PR. The optical signal from this biosensor is captured by using a

smartphone camera, and the images are subsequently classified by a deep learning algorithm.

The color change induced in the invisible aligner attachment (IAA) of the orthodontic device is not easily observable with the naked eye. Therefore, the images were processed and classified using the ResNet-50 CNN model.¹²⁰⁴ Figure 64 illustrates the working principle and key results, with a high accuracy of 97.7% achieved in predicting infections caused by two bacteria, *Streptococcus mutans* and *Streptococcus sanguis*, for concentrations ranging from 10^1 to 10^5 CFU mL^{-1} (colony-forming units), as confirmed in the confusion matrix depicted in Figure 64C and further supported by the comparison of training and validation losses shown in Figure 64E. The 2D data visualizations of the image data (feature vectors) obtained with the multidimensional projection technique t-SNE (*t*-distributed Stochastic Neighbor Embedding) illustrate how the model learning capability: as shown in Figure 64F, the initial feature vectors of the images corresponding to distinct bacterial concentrations cannot be distinguished. In contrast,




Figure 60. (A) Typical responses from the pressure sensors for various types of human motion (i–ix). (B) Responses related to knee motions shown in (i) and (ii) were classified using an SVM algorithm with a high accuracy, as indicated in the confusion matrix in (iii). Reproduced from Zhu et al.¹²⁰⁰ © 2024 American Chemical Society.

Figure 64G demonstrates clear and accurate clustering of the images after 100 epochs of CNN training, clearly demonstrating the model learns to differentiate the images.

Speech analysis has been employed in diagnosis and health monitoring, traditionally performed by capturing sound with standard microphones. In recent years, it has been demonstrated that wearable pressure and strain sensors can also be utilized for such purposes. These applications require signal processing, followed by the establishment of speech patterns using ML algorithms. Although not specifically aimed at health monitoring, the work by Li et al.¹²⁰⁵ presented an intelligent system capable of achieving 95% accuracy in speech recognition. This was accomplished using a skin-like piezoresistive sensor that could also detect human pulses and robot clamping. The skin-like flexible pressure sensor (SFPS) featured a laser-induced graphene layer on a PDMS film, whose piezoresistive properties enabled a detection sensitivity of 1348 kPa^{-1} in the 0–2 kPa range, with a wide operational

range of up to 200 kPa. Figure 65a presents a schematic diagram of the speech recognition system, featuring a sensor (SFPS) attached to the throats of volunteers to measure the pressure of acoustic vibrations. The figure also depicts the structure of the CNN utilized in the system, consisting of three convolutional layers. The brief list of potential applications listed in the figure includes the development of voice-controlled robots, simultaneous translation, and speech reproduction. Figure 65b shows the current response curves of the SFPS corresponding to seven phrases. These phrases were classified with an accuracy of 95.3%, as demonstrated in the confusion matrix also presented in Figure 65b.

Most examples of ML applied to health monitoring and diagnosis using wearable sensors rely on black-box models (e.g., deep neural networks), i.e., it is not possible to understand which properties of the data contributed to the predictions, and how. The aforementioned examples fall into this category. Models that are highly accurate, but difficult to




Figure 61. Boxes in the top row (A) represent traditional methods for diagnosis and treatment planning, which rely entirely on human decisions and interventions. Patient visits are followed by examinations, the results of which guide medical doctors in diagnosing conditions and recommending treatments. These decisions are based on patient data and the doctors' experience. In contrast, the boxes in the bottom row (B) illustrate an AI-assisted approach to diagnosis and treatment planning. Data from biosignal acquisition are processed using AI tools, primarily ML, with the results informing AI systems to develop diagnoses and treatment plans. Reprinted with permission from Wang et al.⁵⁸ © 2024 Elsevier.

interpret have limited applicability in many practical scenarios related to health monitoring. However, there is a growing body of research focused on interpretable (or explainable) ML, which aims to design strategies that allow humans to understand the rationale behind a model's prediction. In the context of sensing data collected from various sources such as wearable devices, IoT systems, and medical instruments, interpretable ML is especially valuable. Sensor data is often complex, high-dimensional, and noisy, making the interpretability of ML models critical for fostering trust, ensuring transparency, and deriving actionable insights. Various strategies exist for achieving interpretable ML (or AI). Some involve the use of standard ML algorithms that incorporate rules, such as decision trees or random forests. This concept underpins the multidimensional calibration space (MCS) framework,¹²⁰⁶ which enhances the calibration and classification accuracy of sensors and biosensors by accounting for the roles of multiple interdependent variables. MCS enables the construction of interpretable models capable of mapping complex relationships between inputs and sensor outputs. Compared to traditional calibration methods—which typically consider single variables or a limited set and often overlook intricate dependencies in real-world systems—MCS offers a significant advantage.

Another strategy to achieve interpretability is to investigate the importance of individual features for the performance of standard ML algorithms. This approach was followed by Kimura et al.¹²⁰⁷ They combined lifestyle factors and wearable sensor data to predict high brain amyloid burden as determined by positron emission tomography. A high amyloid burden is associated with Alzheimer's disease. The researchers

used three ML algorithms to predict amyloid positivity: Elastic Net, kernel support vector machine, and logistic regression. Participants wore wristband sensors, and data were collected measuring variables such as conversation time, physical activity, sleep, and heart rate. After adding variables on lifestyle factors, demographic characteristics, and information on chronic diseases, 54 descriptive variables in total were input to the ML algorithms. The average prediction accuracy across the three algorithms was 0.79. In order to enhance model interpretability, the permutation importance estimate method was employed to identify the most relevant variables to distinguish individuals classified as positive or negative for amyloid across the three algorithms.

12. Business, Regulatory, and Ethical Considerations

The use of sensors for health monitoring is transitioning from the research stage into clinical trials and the use in humans for real-world applications. Several biosensors are already undergoing clinical testing and commercialization across different medical fields. For instance, GlySens developed implantable continuous glucose monitoring biosensors that offer long-term glucose tracking without frequent sensor replacement.¹²⁰⁸ For neurological disorders, various companies emerged over the past few decades, developing biosensors for early disease diagnosis, such as depression or seizures. Motif Neurotech, a neurotechnology company, is developing minimally invasive bioelectronics for mental health treatment, specifically to treat depression. Recently, the company has raised US\$18.75 million in Series A to support the development of a pea-sized brain implant designed to treat depression.¹²⁰⁹ Similarly, Amber Therapeutics has developed Picostim, an innovative brain biosensor that has shown promise in reducing seizures in




Figure 62. (A) Photos of the SWS and a schematic drawing of its position as attached to the human body. Also shown are the possible types of auscultation monitoring. (B) Representation of the SWS architecture, including the sensing parts and the signal acquisition component. (C) Image of the SWS interconnects when 20% stretched, whose properties can be simulated using finite element analysis (FEA) in (D). (E) Demonstration

Figure 62. continued

that the SWS can withstand 180° bending, with bending cycles being reproduced with FEA simulations. (G) Design of the ML-based diagnosis system, in which the sound detected using the SWS is preprocessed and then classified with ML algorithms. Reproduced from Lee et al.¹²⁰¹ Available under a CC BY-NC license. © 2022 American Association for the Advancement of Science.




Figure 63. Illustration of the dual sensing system to measure orthodontic force and detect bacteria. (A) Pictorial representation of the dual sensing device, which is used to measure the force as depicted in (B) and detect bacteria infection via pH changes (C). The reaction shown refers to lactic acid and MA-PR. (D) Schematic illustration of image processing using a smartphone. Reproduced with permission from Feng et al.¹²⁰³ © 2024 John Wiley & Sons.

epilepsy patients.¹²¹⁰ Another example in the space of brain-machine interface is Neuralink, founded by Elon Musk in 2019. Neuralink's biosensor platform aims to restore sensory and motor function and treat neurological disorders. The company began its first-in-human clinical trial in early 2024, successfully implanting its device in three patients. The first two recipients demonstrated the ability to control a computer mouse using thought alone. By the end of 2025, Neuralink plans to recruit 20–30 patients to advance commercialization. Additionally, the company has secured approval to commence clinical trials in Canada, targeting individuals with tetraparesis or tetraplegia, reflecting Neuralink's commitment to international expansion and regulatory compliance.^{1211,1212} Elon Musk is also developing TeslaBot, a humanoid robot designed to perform tasks that are unsafe, repetitive, or mundane for humans. Elon Musk envisions Optimus revolutionizing industries such as manufacturing and logistics by automating various functions.¹²¹³

From a regulatory perspective, biosensors fall under the broader category of digital health and care which is defined by the USA Food and Drug Administration (FDA) and EU to mean the “tools and services that use information and communication technologies (ICTs) to improve prevention, diagnosis, treatment, monitoring, and management of health

and lifestyle”, meaning biosensors are subject to regulations concerning medical devices, data protection, and cybersecurity. Within the medical devices section, biosensors may be considered medical devices if used for diagnosis, treatment, or monitoring, making them subject to strict regulations.¹²¹⁴ To comply with these regulations, the biosensor platforms need to be safe and reliable while maintaining privacy and data protection standards. In general, translating biosensors from the laboratory to clinical applications involves navigating a complex landscape of regulatory requirements to ensure their safety, efficacy, and reliability.

The adoption of biosensors for personalized medicine raises ethical, legal, and human rights concerns, particularly regarding access, fairness, and potential misuse. For example, animal use for testing newly developed biosensors is an important ethical aspect to consider. Traditional bioassays often require large amounts of animals or animal samples; on the other hand, biosensors have been presented as an alternative with the potential to reduce the number of animals that are needed to perfect and fully develop assays.¹²¹⁵ However, when running clinical trials with the newly developed biosensors, it is required to follow internationally recognized ethical guidelines, ensuring patient autonomy and well-being. These principles trace back to the Hippocratic Oath and have been reinforced




Figure 64. (A) Flowchart depicting the use of the ResNet-50 algorithm for image processing. (B) Color changes induced by different levels of proliferation of cariogenic bacteria. (C) Confusion matrix validating the classification accuracy for bacterial concentrations ranging from 10 to 10⁵ CFU mL⁻¹. (D) Accuracy progression as the number of training epochs increases. (E) Comparison of training and validation losses. (F) 2D visualization of image data using t-SNE for different bacterial concentrations before training. (G) Same as (F), but after 100 epochs of CNN training. (H) Illustrative diagram of the biosensing mechanism using the invisible aligner attachment, with images captured by a smartphone and processed using a CNN algorithm. Reproduced with permission from Feng et al.¹²⁰⁵ © 2024 John Wiley & Sons.

by the World Medical Association through the Geneva Declaration (1948) and the Physician's Pledge (2017). Additionally, key international frameworks, such as the Helsinki Declaration (1964) and the Universal Declaration

on Bioethics and Human Rights (2005), provide ethical guidance.¹²¹⁴

13. Main Challenges and Prospects

This Mega Review provided a comprehensive account of the latest developments in wearable sensors for health monitoring,




Figure 65. (A) Schematic of the speech analysis system, where signals captured by the skin-like flexible piezoresistive sensor (SFPS) are processed and classified using a three-layer CNN. Additional potential applications of the AI-assisted SFPS are also listed, including voice-controlled robots, speech reproduction, and simultaneous translation. (B) Current responses of the SFPS obtained from volunteers speaking seven phrases and (C) a confusion matrix demonstrate that the intelligent system can classify the phrases with an accuracy of 95.3%. Reproduced from Li et al.¹²⁰⁵ © 2024 American Chemical Society.

organized into numerous sections in which several important challenges and perspectives were already discussed. Given the breadth of the topic, it is clear that long-term progress will depend on interdisciplinary collaboration among materials scientists, engineers, data scientists, clinicians, and policy-makers. It is essential to incorporate various values¹²¹⁶ and preferences, which adds difficulty to the endeavor.¹²¹⁷ In particular, the most difficult challenge lies in developing products that can be effectively deployed in society. By integrating innovations in energy harvesting, biocompatible materials, and intelligent algorithms with equitable design and robust regulation, wearable sensors can evolve from prototypes into practical healthcare tools. Wearable sensors must satisfy competing requirements: high sensitivity and accuracy, mechanical flexibility, long-term stability, comfort, and affordable production. They must handle complex biological environments—sweat, temperature fluctuations, and motion—without signal degradation. High interindividual variability in sweat composition, for instance, presents additional challenges in interpreting and standardizing analysis,¹²¹⁸ which may require sweat rate normalization¹²¹⁹ and active sweat gland induction.²⁹ Furthermore, the explosion of data generated by wearables introduces new opportunities for AI and ML in healthcare, but also raises questions of privacy, transparency, and regulation. The challenges presented here focus, therefore,

on materials, energy systems, device integration, data management and processing, clinical validation, and societal adoption.

13.1. Materials. A central technical barrier in wearable sensor development lies in identifying materials that are simultaneously flexible, stretchable, durable, and biocompatible. Devices mounted on the skin or integrated into textiles are constantly exposed to mechanical deformation, moisture, temperature changes, and biological fluids. These factors can cause signal drift, degradation of active materials, or even user discomfort over prolonged use. Hydrogels, triboelectric polymers, 2D materials such as graphene, and LMs are promising candidates for soft electronics due to their mechanical adaptability and excellent electrical performance. However, hydrogels are particularly sensitive to humidity and temperature, which may lead to instability and poor reproducibility. Strategies to enhance drying and freezing resistance, or to introduce self-healing properties, are important for long-term stability. Similarly, LM-based sensors offer exceptional conductivity and flexibility but face issues of surface oxidation, electrochemical instability, and poor adhesion to rigid components. Developing robust encapsulation and interface engineering methods is essential for reliable operation in physiological environments. Current sensor designs encounter longevity and stability issues, especially due to the degradation of proteins, enzymes, and lipids in

wound environments.¹²²⁰ Also relevant is to employ sustainable materials^{1221–1224} and fabrication methods.^{788,1225–1229} Many current processes rely on nonrecyclable plastics and energy-intensive methods, increasing environmental impact and cost. Future research should prioritize biodegradable and recyclable materials, such as cellulose-based substrates or bioderived polymers, to align wearable technologies with sustainability goals.

13.2. Energy. Energy supply remains one of the most persistent challenges for wearable sensors. Traditional batteries impose limitations due to size, weight, rigidity, and finite lifespan. Frequent recharging or replacement reduces user compliance and practicality, especially for continuous health monitoring. Self-powered technologies—particularly TENGs, PENGs, and TEGs—are being explored to harvest mechanical and thermal energy from the human body or the environment. However, these systems face low and inconsistent energy outputs, as their performance depends on user motion or temperature gradients. Moreover, the generated signals are often noisy, requiring advanced filtering and signal processing algorithms to extract relevant information. Hybrid energy systems that combine multiple harvesting mechanisms, coupled with miniaturized energy storage devices such as micro supercapacitors and flexible batteries, offer a promising path toward continuous, autonomous operation. The integration of energy management circuits within flexible substrates is another engineering frontier. To achieve compact, lightweight, and comfortable designs, all components—including energy harvesters, power management units, sensors, and wireless communication modules—must be embedded seamlessly within soft materials. Achieving this level of miniaturization without compromising durability or comfort requires innovation in both materials and circuit design.

13.3. Device Integration. For wearable sensors to achieve widespread adoption, they must be comfortable, unobtrusive, and easy to use. The ideal wearable device should adapt to body movements, maintain stable contact with the skin, and operate reliably during daily activities such as exercise or sleep. Achieving this involves balancing mechanical compliance and structural integrity, as well as ensuring safe skin interfaces and breathable materials. Emerging solutions include textile-based sensors and soft robotics integration, where sensors are woven directly into fabrics or embedded within elastic substrates. However, the processing of conductive fibers, coatings, and adhesives remains complex, often incompatible with large-scale textile manufacturing. Electronic textiles (e-textiles) still lack a “killer application,” and their commercial success will likely begin with niche markets—such as athletic performance monitoring, medical garments, or military uniforms—before expanding to general use. Comfort and usability also extend to user interaction and data interpretation. Studies show that perceived complexity, especially among older adults, can discourage use. Therefore, wearable systems must provide intuitive interfaces and transparent feedback, potentially through smartphone integration or NFC systems that leverage existing user habits.

13.4. Data Processing and Data Management. Exceptional perspectives—though accompanied by equally tremendous challenges—exist in the use of AI, particularly ML. Wearable sensors can now generate vast quantities of time-series and multimodal data,^{803,1230–1232} whose processing requires sophisticated algorithms capable of distinguishing meaningful physiological changes from noise. For example,

ECG provides precise temporal information while PPG supplements hemodynamic parameters.¹²³³ Combined with ML algorithms, a more comprehensive cardiovascular function assessment can be achieved.¹²³⁴ AI and ML can detect subtle patterns, predict disease progression, and enable personalized interventions. In addition, multimodal fusion, combining electrical (ECG, EMG), optical (PPG), and mechanical signals, offers a richer understanding of cardiovascular and metabolic health. At the hardware level, advances in flexible electronics and micronano fabrication technologies have made sensors thinner, more flexible, and more integrated.¹²³⁵ At the software level, DL can eliminate various interferences and extract true health information from the multisource data,^{1236,1237} which can assist in precision medicine and personalized health management.¹²³⁸

However, several issues must be addressed for AI-driven insights to be clinically reliable. Data quality, standardization, and interoperability remain fundamental concerns as inconsistent sampling or sensor variability can compromise model accuracy. The lack of standardized biomarkers in nonblood fluids, such as sweat or saliva, further complicates clinical interpretation. Moreover, ensuring data privacy, cybersecurity, and algorithmic transparency is essential to building trust among patients and healthcare providers. A suitable e-health tool is expected to include clinical utility and safety, usability and human-centricity, functionality, and data management.¹²³⁹ A growing movement emphasizes the adoption of FAIR data principles: findable, accessible, interoperable, and reusable within healthcare ecosystems.¹²⁴⁰ While many industry-based wearable device providers store data streams on data clouds, healthcare institutions have started to prefer data storage on premises that includes full access to raw and processed data.¹²⁴¹ Implementing these principles requires collaboration between technologists, regulators, and clinicians to define consistent data formats, ethical frameworks, and secure infrastructures.

13.5. Clinical Validation. Despite impressive laboratory demonstrations, few wearable sensors have achieved clinical-grade validation or regulatory approval.^{1242,1243} The regulatory landscape for wearables remains fragmented: while the FDA and European CE systems have pathways for medical devices, many emerging sensors fall between categories or lack standardized testing protocols. Clinical translation is hindered by variability in biofluids, lack of multisite studies, and high validation costs. For successful clinical integration, wearable technologies must demonstrate not only technical performance but also clinical utility, usability, and safety. In fact, perceived complexity, particularly in interpreting device outputs, deters adoption among seniors.^{1244,1245} Even more difficult is to deploy technology such as POC devices in underserved communities.¹²⁴⁶ Collaborative frameworks involving healthcare institutions, academic researchers, and industry partners are necessary to conduct large-scale trials and define best practices. Cost is another major barrier—the estimated investment to bring a new medical device to market can exceed hundreds of millions of dollars when accounting for development failures and regulatory requirements. Nevertheless, some notable successes are emerging. Devices such as MC10s BioStamp¹²⁴⁷ and Sibel Health's ANNE One have achieved FDA clearance, validating the potential of soft, skin-conformal electronics for continuous monitoring. These examples illustrate how structural design, energy efficiency, and data integration can converge to produce clinically relevant

wearables. Also, even consumer health devices can provide valuable sensing capabilities, saving lives in scenarios like out-of-hospital cardiac arrest¹²⁴⁸ and detecting poor mental health states.¹²⁴⁹

13.6. Societal Adoption. Beyond technical and regulatory challenges, equity and accessibility are fundamental considerations for the evolution of wearable sensors. Adoption among elderly users and individuals in low-resource settings remains limited because of cost, connectivity barriers, and a lack of digital literacy. Future developments must therefore embrace user-centered and inclusive design, ensuring that devices are adaptable to diverse populations and economic conditions.

13.7. The Future. The evolution of wearable sensors for health monitoring signals more than technological progress; it points to a profound transformation in how knowledge about the human body is generated, interpreted, and applied. As discussed throughout this review, advances in materials science, soft electronics, energy systems, and artificial intelligence are converging to enable continuous, real-time interaction with physiological processes. However, the most significant implication of this convergence may lie beyond improved diagnostics or device performance. It suggests the emergence of what can be described as a fifth paradigm of knowledge generation,^{1180,1181} in which cognition is no longer exclusively human but is increasingly distributed across hybrid human–machine systems.

In this new paradigm, wearable sensors act as interfaces between the biological and digital, continuously translating physiological signals into data streams that can be processed, contextualized, and acted upon by intelligent algorithms. Unlike previous paradigms of knowledge generation, ranging from empirical observation to theory-driven science and data-intensive discovery, this new framework is characterized by autonomous, adaptive, and coevolving systems, where sensing, analysis, and decision-making occur in an integrated and dynamic loop. Wearable technologies thus become not only tools for measurement but also active agents in the production of biomedical knowledge. A key manifestation of this shift is the move toward continuous multimodal physiological mapping, enabling the construction of high-dimensional representations of human health. When coupled with AI methods, these data streams can reveal complex correlations and predictive patterns that are inaccessible through traditional clinical approaches. The role of AI in this context extends beyond data analysis: it becomes a cocreator of knowledge, capable of generating hypotheses, identifying anomalies, and guiding experimental or clinical interventions in real time and autonomously. This fundamentally challenges the classical separation between measurement and interpretation.

The integration of sensing and actuation further amplifies this transformation. Closed-loop wearable systems capable of responding to physiological changes blur the boundary between diagnosis and therapy. In such systems, knowledge is not only descriptive but operational, directly informing and executing interventions. This real-time coupling between perception and action reflects a cybernetic view of healthcare in which the human body is embedded within a responsive technological ecosystem.

AUTHOR INFORMATION

Corresponding Authors

Lais Canniatti Brazaca – São Carlos Institute of Chemistry, University of São Paulo, São Carlos, SP 13566-590, Brazil; National Institute of Science and Technology in Bioanalytics Lauro Kubota - INCTBio-LK, State University of Campinas (Unicamp), Campinas, SP 13083-970, Brazil; orcid.org/0000-0002-0456-7552; Email: lais.brazaca@usp.br

Juliane Sempionatto – Department of Electrical and Computer Engineering, Rice University, Houston, Texas 77005, United States; Email: jsemp@rice.edu

Oswaldo N. Oliveira, Jr. – São Carlos Institute of Physics, University of São Paulo, São Carlos, SP 13566-590, Brazil; orcid.org/0000-0002-5399-5860; Email: chu@ifsc.usp.br

Authors

Alvaro Moreno Lozano – Department of Bioengineering, Rice University, Houston, Texas 77030, United States

Beatriz Mayol – Department of Electrical and Computer Engineering, Rice University, Houston, Texas 77005, United States

Desiree Tamara Scheidt – São Carlos Institute of Chemistry, University of São Paulo, São Carlos, SP 13566-590, Brazil; National Institute of Science and Technology in Bioanalytics Lauro Kubota - INCTBio-LK, State University of Campinas (Unicamp), Campinas, SP 13083-970, Brazil; orcid.org/0000-0001-8504-6259

Domenico De Fazio – Department of Molecular Science and Nanosystems, Ca' Foscari University of Venice, 30172 Venezia, Italy; orcid.org/0000-0003-3327-078X

Enji Kim – Department of Materials Science and Engineering, Yonsei University, Seoul 03722, Republic of Korea; Center for Nanomedicine, Institute for Basic Science (IBS), Seoul 03722, Republic of Korea; orcid.org/0000-0002-4934-8951

Gabriel Maroli – Catalan Institute of Nanoscience and Nanotechnology (ICN2), CSIC and BIST, Bellaterra, Barcelona 08193, Spain

Giulio Rosati – Catalan Institute of Nanoscience and Nanotechnology (ICN2), CSIC and BIST, Bellaterra, Barcelona 08193, Spain; EURECAT - Technology Research Center, 08302 Mataró, Barcelona, Spain; orcid.org/0000-0002-0227-4561

Hatice Ceren Ates – Institute of Biomedical Engineering (MIBE), Department of Electrical Engineering, TUM School of Computation, Information and Technology, Technical University of Munich, 85748 Garching n. Munich, Germany

Hui Wang – School of Materials Science and Engineering, Nanyang Technological University, Singapore 639798, Singapore

Hye Jin Kim – Department of Biomedical Engineering, Yonsei University, Wonju 26493, Republic of Korea; Center for Nanoparticle Research, Institute for Basic Science (IBS), Seoul 08826, Republic of Korea; orcid.org/0000-0001-7070-4204

Ji-Hwan Ha – Department of Mechanical Engineering, Korea Advanced Institute of Science and Technology, Daejeon 34141, Republic of Korea; Department of Mechanical Engineering, Hanbat National University, Daejeon 34158, Republic of Korea

Jin Yang – Department of Optoelectronic Engineering, Key Laboratory of Optoelectronic Technology and Systems

- Ministry of Education, Chongqing University, Chongqing 400044, China; orcid.org/0000-0002-6606-8310
- Jisoo Jeon** – School of Materials Science and Engineering, Georgia Institute of Technology, Atlanta, Georgia 30332, United States; orcid.org/0000-0002-0969-0293
- Jörg Goldhahn** – Institute for Molecular and Translational Medicine, Department of Health Sciences and Technology, ETH Zurich, Zurich 8092, Switzerland
- Keval K. Sonigara** – Future Energy and Innovation Laboratory, Central European Institute of Technology, Brno University of Technology, Brno 61200, Czech Republic
- La Li** – School of Integrated Circuits and Electronics, Beijing Institute of Technology, Beijing 100081, China; orcid.org/0000-0001-7355-0407
- Mingyu Zhou** – Department of Biomedical Engineering, Northwestern University, Evanston, Illinois 60208, United States
- Seokjoo Cho** – Department of Mechanical Engineering, Korea Advanced Institute of Science and Technology, Daejeon 34141, Republic of Korea
- Shicheng Fan** – Mechanobiology Institute, National University of Singapore, Singapore 117411, Singapore; Institute for Health Innovation and Technology (iHealthtech), National University of Singapore, Singapore 117599, Singapore
- Shuang Wu** – Department of Mechanical and Aerospace Engineering, North Carolina State University, Raleigh, North Carolina 27695-7001, United States; Florida Institute of Technology, Melbourne, Florida 32901-6982, United States
- Vanessa Matos Gonçalves** – Institute for Molecular and Translational Medicine, Department of Health Sciences and Technology, ETH Zurich, Zurich 8092, Switzerland; orcid.org/0000-0001-7649-9678
- Vitor H. B. Dei Santi** – São Carlos Institute of Physics, University of São Paulo, São Carlos, SP 13566-590, Brazil; orcid.org/0009-0000-0315-517X
- Weixin Zhou** – Department of Mechanical and Aerospace Engineering, North Carolina State University, Raleigh, North Carolina 27695-7001, United States
- Wenzheng Heng** – Andrew and Peggy Cherg Department of Medical Engineering, Division of Engineering and Applied Science, California Institute of Technology, Pasadena, California 91125, United States
- Won Gi Chung** – Department of Materials Science and Engineering, Yonsei University, Seoul 03722, Republic of Korea; Center for Nanomedicine, Institute for Basic Science (IBS), Seoul 03722, Republic of Korea; orcid.org/0000-0001-9908-8492
- Xinming Li** – Querrey Simpson Institute for Bioelectronics, Northwestern University, Evanston, Illinois 60208, United States; Center for Bio-Integrated Electronics and Department of Biomedical Engineering, Northwestern University, Evanston, Illinois 60208, United States; orcid.org/0009-0007-8521-9317
- Yang Zou** – Beijing Institute of Nanoenergy and Nanosystems, Chinese Academy of Sciences, Beijing 101400, China; School of Medical Technology, Beijing Institute of Technology, Beijing 100081, China; orcid.org/0000-0003-2557-4230
- Yunxia Jin** – Institute for Health Innovation and Technology (iHealthtech) and Department of Biomedical Engineering, National University of Singapore, Singapore 117599, Singapore
- Zixuan Wu** – Research Center of Flexible Sensing Materials and Devices, School of Applied Physics and Materials, Wuyi University, Jiangmen 529020, China
- Amay J. Bandodkar** – Department of Electrical and Computer Engineering and Center for Advanced Self-Powered Systems of Integrated Sensors and Technologies (ASSIST), North Carolina State University, Raleigh, North Carolina 27606, United States; Lampe Joint Department of Biomedical Engineering, North Carolina State University and University of North Carolina at Chapel Hill, Chapel Hill, North Carolina 27695, United States; orcid.org/0000-0002-1792-1506
- Arben Merkoçi** – Catalan Institute of Nanoscience and Nanotechnology (ICN2), CSIC and BIST, Bellaterra, Barcelona 08193, Spain; Catalan Institution for Research and Advanced Studies (ICREA), Barcelona 08010, Spain; orcid.org/0000-0003-2486-8085
- Can Dincer** – Institute of Biomedical Engineering (MIBE), Department of Electrical Engineering, TUM School of Computation, Information and Technology, Technical University of Munich, 85748 Garching n. Munich, Germany; orcid.org/0000-0003-3301-1198
- Chwee Teck Lim** – Mechanobiology Institute, National University of Singapore, Singapore 117411, Singapore; Institute for Health Innovation and Technology (iHealthtech) and Department of Biomedical Engineering, National University of Singapore, Singapore 117599, Singapore; orcid.org/0000-0003-4019-9782
- Dae-Hyeong Kim** – Center for Nanoparticle Research, Institute for Basic Science (IBS), Seoul 08826, Republic of Korea; School of Chemical and Biological Engineering, Institute of Chemical Processes, Seoul National University (SNU), Seoul 08826, Republic of Korea; orcid.org/0000-0002-4722-1893
- Emanuel Carrilho** – São Carlos Institute of Chemistry, University of São Paulo, São Carlos, SP 13566-590, Brazil; National Institute of Science and Technology in Bioanalytics Lauro Kubota - INCTBio-LK, State University of Campinas (Unicamp), Campinas, SP 13083-970, Brazil; orcid.org/0000-0001-7351-8220
- Giovanni Antonio Salvatore** – Department of Molecular Science and Nanosystems, Ca' Foscari University of Venice, 30172 Venezia, Italy; orcid.org/0000-0002-8983-3257
- Guozhen Shen** – School of Integrated Circuits and Electronics, Beijing Institute of Technology, Beijing 100081, China; orcid.org/0000-0002-9755-1647
- Inkyu Park** – Department of Mechanical Engineering, Korea Advanced Institute of Science and Technology, Daejeon 34141, Republic of Korea; orcid.org/0000-0001-5761-7739
- Itthipon Jeerapan** – Division of Physical Science, Faculty of Science, Prince of Songkla University, Songkhla 90110, Thailand; Center of Excellence for Trace Analysis and Biosensor, Center of Excellence for Innovation in Chemistry, Prince of Songkla University, Songkhla 90110, Thailand; The iJE Electrochemistry for All Laboratory, Songkhla 90110, Thailand
- Jang-Ung Park** – Department of Materials Science and Engineering, Yonsei University, Seoul 03722, Republic of Korea; Center for Nanomedicine, Institute for Basic Science (IBS), Seoul 03722, Republic of Korea; Graduate Program of Nano Biomedical Engineering (NanoBME), Advanced Science Institute, Yonsei University, Seoul 03722, Republic of

Korea; Department of Neurosurgery, Yonsei University College of Medicine, Seoul 03722, Republic of Korea; orcid.org/0000-0003-1522-4958

Jin Wu — State Key Laboratory of Optoelectronic Materials and Technologies, Guangdong Province Key Laboratory of Display Material and Technology, School of Electronics and Information Technology, Sun Yat-sen University, Guangzhou 510275, China; orcid.org/0000-0002-3065-6858

John S. Ho — Institute for Health Innovation and Technology (iHealthtech) and Department of Electrical and Computer Engineering, National University of Singapore, Singapore 117599, Singapore; orcid.org/0000-0002-9458-9033

John Rogers — Department of Biomedical Engineering, Department of Mechanical Engineering, and Department of Materials Science and Engineering, Northwestern University, Evanston, Illinois 60208, United States; Rhaeos Inc., Evanston, Illinois 60201, United States; Department of Neurological Surgery, Feinberg School of Medicine, Northwestern University, Chicago, Illinois 60611, United States; orcid.org/0000-0002-2980-3961

Joseph Wang — Aiiso Yufeng Li Family Department of Chemical and Nano Engineering, University of California San Diego, La Jolla, San Diego, California 92093, United States; orcid.org/0000-0002-4921-9674

Kevin W. Plaxco — Department of Chemistry and Biochemistry, University of California, Santa Barbara, California 93106, United States; BioEngineering Department, University of California Santa Barbara, Santa Barbara, California 93106, United States

Maria Cristina Ferreira de Oliveira — Institute of Mathematical Sciences and Computing, University of Sao Paulo, Sao Carlos, SP 13566-590, Brazil

Martin Pumera — Future Energy and Innovation Laboratory, Central European Institute of Technology, Brno University of Technology, Brno 61200, Czech Republic; Advanced Nanorobots and Multiscale Robotics Laboratory, Faculty of Electrical Engineering and Computer Science, VSB—Technical University of Ostrava, Ostrava 70800, Czech Republic; Department of Chemical and Biomolecular Engineering, Yonsei University, Seoul 03722, Republic of Korea; orcid.org/0000-0001-5846-2951

Noé Brasier — Institute for Molecular and Translational Medicine, Department of Health Sciences and Technology, ETH Zurich, Zurich 8092, Switzerland; orcid.org/0000-0003-0186-0865

Omid Veisheh — Department of Bioengineering, Rice University, Houston, Texas 77030, United States; orcid.org/0000-0003-1153-8079

Pooi See Lee — School of Materials Science and Engineering, Nanyang Technological University, Singapore 639798, Singapore; orcid.org/0000-0003-1383-1623

Vladimir V. Tsukruk — School of Materials Science and Engineering, Georgia Institute of Technology, Atlanta, Georgia 30332, United States; orcid.org/0000-0001-5489-0967

Wei Gao — Andrew and Peggy Cherng Department of Medical Engineering, Division of Engineering and Applied Science, California Institute of Technology, Pasadena, California 91125, United States; orcid.org/0000-0002-8503-4562

Yong Zhu — Department of Mechanical and Aerospace Engineering, North Carolina State University, Raleigh, North Carolina 27695-7001, United States; orcid.org/0000-0002-3862-5757

Zhong Lin Wang — Beijing Institute of Nanoenergy and Nanosystems, Chinese Academy of Sciences, Beijing 101400, China

Zhou Li — Beijing Institute of Nanoenergy and Nanosystems, Chinese Academy of Sciences, Beijing 101400, China; School of Nanoscience and Technology, University of Chinese Academy of Sciences, Beijing 100049, China; orcid.org/0000-0002-9952-7296

Complete contact information is available at: <https://pubs.acs.org/10.1021/acsami.6c04520>

Funding

The Article Processing Charge for the publication of this research was funded by the Coordenacao de Aperfeicoamento de Pessoal de Nivel Superior (CAPES), Brazil (ROR identifier: 00x0ma614).

Notes

The authors declare no competing financial interest.

ACKNOWLEDGMENTS

The authors are thankful for the support of the following: CAPES, FAPESP (2023/10141-2, 2024/16261-2, 2025/27044-5, 2025/26623-1); CNPq (401977/2023-4, 385924/2025-9, 401615/2025-1, 408338/2024-5, 302652/2025-6); National Natural Science Foundation of China (22475243, 62401403); Independent Fund of the State Key Laboratory of Optoelectronic Materials and Technologies (Sun Yat-sen University) (OEMT-2022-ZRC-05); Foundation of the State Key Laboratory of Transducer Technology (No. SKT2301); ERDF/ESF project TECHSCALE (No. CZ.02.01.01/00/22_008/0004587); European Union (REFRESH—Research Excellence For REgion Sustainability and High-tech Industries, CZ.10.03.01/00/22_003/0000048, via the Operational Programme Just Transition); Horizon Europe EU Research and Innovation Program (MSCA grant agreement No. 101067295); Institute for Health Innovation and Technology (iHealthtech) (No. A-0001415-06-00); MechanoBioEngineering Laboratory at the Department of Biomedical Engineering of the National University of Singapore; National Natural Science Foundation of China (22475243, 62401403); Guangdong Basic and Applied Basic Research Foundation (2023A1515110386); Mechanobiology Institute (No. A-0003467-20-00) at the National University of Singapore (NUS); National Science Foundation (award no. 212284); National Research Foundation (NRF) (Grant SGSR CRE-ATE); National Research Foundation of Korea (NRF) (RS-2025-16063568); Grant Agency of the Czech Republic, GACR EXPRO (25-15484X); Ministry of Health of the Czech Republic (NW24-08-00473); ERDF/ESF Project TECHSCALE (CZ.02.01.01/00/22_008/0004587). The ICN2 is funded by the CERCA program/Generalitat de Catalunya. The ICN2 is supported by the Severo Ochoa Centres of Excellence program, Grant CEX2021-001214-S, funded by MCIU/AEI/10.13039.501100011033.

ABBREVIATIONS LIST

AFM - Atomic Force Microscopy
AI - Artificial intelligence
AIr - Radial artery augmentation index
a-IGZO - Amorphous indium gallium zinc oxide
AgNF - Silver nanofiber

- AgNP - Silver nanoparticle
AgNW - Silver nanowire
ALMP - Adhesive liquid metal particles
ALS - Amyotrophic lateral sclerosis
AMR - Antimicrobial resistance
AOx - Alcohol oxidase
BFC - Biofuel cell
BLE - Bluetooth low energy
BP - Burst pressure
CA - Calcium alginate
CBT - Core body temperature
CCP - Conductive carbon paste
CF - Cystic fibrosis
CNC - Computer numerical control
CNCry - Cellulose nanocrystal
CNN - Convolutional neural networks
CNT - Carbon nanotube
COPD - Chronic obstructive pulmonary disease
CQD - Carbon quantum dot
CRP - C-reactive protein
CSF - Cerebrospinal fluid
CV - Cyclic voltammetry
DA - Dopamine
DBP - Diastolic blood pressure
DL - Deep learning
DPV - Differential pulse voltammetry
DPP - Diketopyrrolopyrrole
EAB - Electrochemical aptamer-based
EBA - Exhaled breath aerosols
EBC - Exhaled breath condensate
ECG - Electrocardiogram
EGaIn - Eutectic gallium-indium
E-LIG - Electrografted laser-induced graphene
ELISA - Enzyme-linked immunosorbent assay
EMG - Electromyography
EMGe - Electromagnetic generator
ESM - Eggshell membrane
EVA - Ethylene-vinyl acetate
FDA - Food and Drug Administration
FITC - Fluorescein isothiocyanate
FUTUD - Ultra-wideband triboelectric ultrasonic device
GABA - γ -aminobutyric acid
GI - Gastrointestinal
GFAP - Glial fibrillary acidic protein
GM - Gut microbiota
GO - Graphene oxide
GOx - Glucose oxidase
GSH - Glutathione
HEG - Hydrovoltaic effect generator
HVA - Homovanillic acid
IAA - Invisible aligner attachment
ICT - Information and communication technologies
HR - Heart rate
HRV - Heart rate variability
IC - Integrated circuit
IDE - Interdigitated electrode
IoT - Internet of Things
IR - Infrared
ISE - Ion-selective electrode
ISF - Interstitial fluid
KNN - K-nearest neighbor
LC-MS - Liquid chromatography-mass spectrometry
LED - Light-Emitting Diode
LIG - Laser-induced graphene
LFA - Lateral flow assay
LM - Liquid metal
LOD - Limit of detection
LOx - Lactate oxidase
LP - Lumbar puncture
LTSP - Low temperature polycrystalline silicon
MBA - N,N'-methylene-bis-acrylamide
MCS - Multidimensional calibration space
MEG - Moisture effect generator
MEMs - Microelectromechanical system
MICS - Medical implant communication service
MIP - Molecularly imprinted polymer
ML - Machine Learning
MMP - Metalloproteinases
MN - Microneedle
MNC - Microbial nanocellulose
MTMOS - Methyltrimethoxysilane
NCL - Neuroprosthetic contact lenses
NFC - Near-field communication
NMR - Nuclear magnetic resonance
NP - Nanoparticles
NPL - Natural Language Processing
NTC - Negative temperature coefficient
NW - Nanowire
OCV - Open-circuit voltage
OECT - Organic electrochemical transistors
OFET - Organic field-effect transistors
OTFT - Organic thin film transistor
P3HT - Poly(3-hexylthiophene)
PAM - Polyacrylamide
PAMD - Polymer-assisted metal deposition
PANI - Polyaniline
PC - Capillary pressure
PCA - Principal component analysis
PCR - Polymerase chain reaction
PD - Photodetectors
PDA - Polydopamine
PdNP - Palladium nanoparticle
PDMS - Polydimethylsiloxane
PECVD - Plasma-enhanced chemical vapor deposition
PEDOT - Poly(3,4-ethylenedioxythiophene)
PEG - Polyethylene glycol
PENG - Piezoelectric nanogenerators
PEO - Polyethylene oxide
PET - Polyethylene terephthalate
PETG - Polyethylene terephthalate glycol
PIT - Passive inductive telemetry
PH - Hydrostatic pressure
PI - Polyimide
PL - Laplace pressure
PLA - Poly(lactic acid)
PLGA - Poly(lactic-co-glycolic acid)
PMEHB - Poly(MPC-co-EHMA-co-MBP)
PMMA - Polymethyl methacrylate
PMU - Power management unit
POC - Point of care
POU - Point of use
PPG - Photoplethysmography
PPy - Polymeric conductive adhesive
PSA - Pressure sensitive adhesive
PSS - Poly(styrenesulfonate)
PTFE - Polytetrafluoroethylene

PU - Polyurethane
 PVA - Poly(vinyl alcohol)
 PVC - Polyvinyl chloride
 PVDF - Polyvinylidene fluoride
 P(VDF-TrFE) - Poly(vinylidene fluoride-co-trifluoroethylene)
 PyNG - Pyroelectric nanogenerator
 PZT - Lead zirconate titanate
 Q_{sc} - Short-circuit transferred charge
 R_{ct} - Charge transfer resistance
 RIE - Reactive ion etching
 RF - Radio frequency
 RFID - Radio frequency identification
 rGO - Reduced graphene oxide
 ROS - Reactive oxygen species
 RTD - Resistive temperature device
 SBP - Systolic blood pressure
 SBS - Styrene-butadiene-styrene
 SCFA - Short-chain fatty acid
 SEBS - Styrene-ethylene-butylene-styrene
 SEM - Scanning Electron Microscopy
 SERS - Surface-enhanced Raman spectroscopy
 SF - Silk fibroin
 SFPS - Skin-like flexible pressure sensor
 SNGC - sonogel-carbon
 SPCE - Screen-printed carbon electrodes
 SSP - Spoof surface plasmon
 STEP - Stretchable tubular elastomeric piezoresistive
 SUPS - Self-powered ultrasensitive pulse sensor
 SVM - Support vector machine
 SWASV - Square wave anodic stripping voltammetry
 SWS - Soft wearable stethoscope
 SWV - Square wave voltammetry
 tcPO₂ - Transcutaneous oxygen partial pressure
 TDM - Therapeutic Drug Monitoring
 TEG - Thermoelectric generators
 TENG - Triboelectric nanogenerator
 TFT - Thin film transistors
 TPU - Thermoplastic polyurethane
 T-sensors - Temperature sensors
 t-SNE - t-distributed Stochastic Neighbor Embedding
 UV - Ultraviolet
 Vis - Visible
 VOC - Volatile organic compound
 WiFi - Wireless Fidelity
 WTF - Wireless power transfer
 XAI - Explainable Artificial Intelligence
 2-FMA - 2-fluoro-methamphetamine
 βCD - β-cyclodextrin
 μPAD - Microfluidic Paper-based analytical device

■ ADDITIONAL NOTE

^aThe 3D visualization at [CitationNetwork](#) reveals a dense cluster aggregation due to topic proximity, with some exceptions.

■ REFERENCES

- (1) Rayan, R. A.; Zafar, I. Monitoring Technologies for Precision Health. In *The Smart Cyber Ecosystem for Sustainable Development*; Wiley, 2021; pp 251–260 DOI: [10.1002/9781119761655.ch13](#).
- (2) Babu, M.; Lautman, Z.; Lin, X.; Sobota, M. H. B.; Snyder, M. P. Wearable Devices: Implications for Precision Medicine and the Future of Health Care. *Annu. Rev. Med.* **2024**, *75*, 401–415.
- (3) Peng, H. Wearable electronics. *Natl. Sci. Rev.* **2023**, *10*, No. nwac193, DOI: [10.1093/nsr/nwac193](#).
- (4) Ates, H. C.; Nguyen, P. Q.; Gonzalez-Macia, L.; et al. End-to-end design of wearable sensors. *Nat. Rev. Mater.* **2022**, *7*, 887–907.
- (5) Reeder, B.; David, A. Health at hand: A systematic review of smart watch uses for health and wellness. *J. Biomed. Inf.* **2016**, *63*, 269–276.
- (6) Heikenfeld, J.; Jajack, A.; Rogers, J.; et al. Wearable sensors: modalities, challenges, and prospects. *Lab Chip* **2018**, *18*, 217–248.
- (7) Köhler, C.; Bartschke, A.; Fürstenau, D.; Schaaf, T.; Salgado-Baez, E. The Value of Smartwatches in the Health Care Sector for Monitoring, Nudging, and Predicting: Viewpoint on 25 Years of Research. *J. Med. Internet Res.* **2024**, *26*, No. e58936.
- (8) Dörr, M.; Nohturfft, V.; Brasier, N.; et al. The WATCH AF Trial: SmartWATCHes for Detection of Atrial Fibrillation. *JACC Clin. Electrophysiol.* **2019**, *5*, 199–208.
- (9) Brasier, N.; Raichle, C. J.; Dörr, M.; et al. Detection of atrial fibrillation with a smartphone camera: first prospective, international, two-centre, clinical validation study (DETECT AF PRO). *EP Europace* **2019**, *21*, 41–47.
- (10) Rizas, K. D.; Freyer, L.; Sappier, N.; et al. Smartphone-based screening for atrial fibrillation: a pragmatic randomized clinical trial. *Nat. Med.* **2022**, *28*, 1823–1830.
- (11) Chu, M.; Zhang, S.; Gong, J.; et al. Telemedicine-based integrated management of atrial fibrillation in village clinics: a cluster randomized trial. *Nat. Med.* **2025**, *31* (4), 1276–1285, DOI: [10.1038/s41591-025-03511-2](#).
- (12) Shah, A. J.; Althobiani, M. A.; Saigal, A.; et al. Wearable technology interventions in patients with chronic obstructive pulmonary disease: a systematic review and meta-analysis. *npj Digit. Med.* **2023**, *6*, No. 222.
- (13) Althobiani, M. A.; Shuttleworth, R.; Conway, J.; et al. Supporting self-management for patients with Interstitial Lung Diseases: Utility and acceptability of digital devices. *PLoS Digit. Health* **2024**, *3*, No. e0000318.
- (14) Gashi, S.; Oldrati, P.; Moebus, M.; et al. Modeling multiple sclerosis using mobile and wearable sensor data. *npj Digit. Med.* **2024**, *7*, No. 64.
- (15) Teymourian, H.; Tehrani, F.; Longardner, K.; et al. Closing the loop for patients with Parkinson disease: where are we? *Nat. Rev. Neurol.* **2022**, *18*, 497–507.
- (16) Kourtis, L. C.; Regele, O. B.; Wright, J. M.; Jones, G. B. Digital biomarkers for Alzheimer's disease: the mobile/wearable devices opportunity. *npj Digit. Med.* **2019**, *2*, No. 9.
- (17) McGagh, D.; Song, K.; Yuan, H.; et al. Digital health technologies to strengthen patient-centred outcome assessment in clinical trials in inflammatory arthritis. *Lancet Rheumatol.* **2025**, *7*, e55–e63.
- (18) Partheniadis, K.; Stavrakis, M. Design and evaluation of a digital wearable ring and a smartphone application to help monitor and manage the effects of Raynaud's phenomenon. *Multimed. Tools Appl.* **2019**, *78*, 3365–3394.
- (19) Yang, A. F.; Patel, S.; Chun, K. S.; et al. Artificial Intelligence—Enabled Wearable Devices and Nocturnal Scratching in Mild Atopic Dermatitis. *JAMA Dermatol.* **2025**, *161*, 406–410, DOI: [10.1001/jamadermatol.2024.5697](#).
- (20) Fedor, S.; Lewis, R.; Pedrelli, P.; et al. Wearable Technology in Clinical Practice for Depressive Disorder. *N. Engl. J. Med.* **2023**, *389*, 2457–2466.
- (21) Fonseka, L. N.; Woo, B. K. P. Wearables in Schizophrenia: Update on Current and Future Clinical Applications. *JMIR mHealth uHealth* **2022**, *10*, No. e35600.
- (22) Sanches, C. A.; Silva, G. A.; Librantz, A. F. H.; Sampaio, L. M. M.; Belan, P. A. Wearable Devices to Diagnose and Monitor the Progression of COVID-19 Through Heart Rate Variability Measurement: Systematic Review and Meta-Analysis. *J. Med. Internet Res.* **2023**, *25*, No. e47112.
- (23) Brasier, N.; Wang, J.; Gao, W.; et al. Applied body-fluid analysis by wearable devices. *Nature* **2024**, *636*, 57–68.

- (24) Davis, N.; Heikenfeld, J.; Milla, C.; Javey, A. The challenges and promise of sweat sensing. *Nat. Biotechnol.* **2024**, *42*, 860–871.
- (25) Baker, L. B.; Wolfe, A. S. Physiological mechanisms determining eccrine sweat composition. *Eur. J. Appl. Physiol.* **2020**, *120*, 719–752.
- (26) Brasier, N.; Niederberger, C.; Salvatore, G. A. The sweat rate as a digital biomarker in clinical medicine beyond sports science. *Soft Science* **2024**, *4*, No. 6, DOI: 10.20517/ss.2023.39.
- (27) Spinelli, J. C.; Suleski, B. J.; Wright, D. E.; et al. Wearable microfluidic biosensors with haptic feedback for continuous monitoring of hydration biomarkers in workers. *npj Digit. Med.* **2025**, *8*, No. 76.
- (28) Brasier, N.; Frobert, O.; De Ieso, F.; et al. The potential of wearable sweat sensors in heart failure management. *Nat. Electron.* **2024**, *7*, 182–184.
- (29) Tu, J.; Min, J.; Song, Y.; et al. A wireless patch for the monitoring of C-reactive protein in sweat. *Nat. Biomed. Eng.* **2023**, *7*, 1293–1306, DOI: 10.1038/s41551-023-01059-5.
- (30) Brasier, N.; Sempionatto, J. R.; Bourke, S.; et al. Towards on-skin analysis of sweat for managing disorders of substance abuse. *Nat. Biomed. Eng.* **2024**, *8*, 925–929.
- (31) Niederberger, C.; Vermeersch, A.; Davidhi, F.; et al. Wearable sweat analysis to determine biological age. *Trends Biotechnol.* **2023**, *41*, 1113–1116.
- (32) Heng, W.; Yin, S.; Min, J.; et al. A smart mask for exhaled breath condensate harvesting and analysis. *Science* **2024**, *385*, 954–961.
- (33) Nguyen, P. Q.; Soenksen, L. R.; Donghia, N. M.; et al. Wearable materials with embedded synthetic biology sensors for biomolecule detection. *Nat. Biotechnol.* **2021**, *39*, 1366–1374.
- (34) Martens, T.; Beck, R. W.; Bailey, R.; et al. Effect of Continuous Glucose Monitoring on Glycemic Control in Patients With Type 2 Diabetes Treated With Basal Insulin. *JAMA* **2021**, *325*, 2262.
- (35) Tan, Y. Y.; Suan, E.; Koh, G. C. H.; Suhairi, S. B.; Tyagi, S. Effectiveness of continuous glucose monitoring in patient management of Type 2 Diabetes Mellitus: an umbrella review of systematic reviews from 2011 to 2024. *Arch. Public Health* **2024**, *82*, No. 231.
- (36) Brasier, N.; Ates, H. C.; Sempionatto, J. R.; et al. A three-level model for therapeutic drug monitoring of antimicrobials at the site of infection. *Lancet Infect. Dis.* **2023**, *23*, e445–e453.
- (37) Plackett, B. Why big pharma has abandoned antibiotics. *Nature* **2020**, *586*, S50–S52.
- (38) Antimicrobial resistance: a silent pandemic. *Nat. Commun.* **2024**, *15*, 6198, DOI: 10.1038/s41467-024-50457-z.
- (39) Cunha, C. B.; Opal, S. M. Antibiotic Stewardship. *Med. Clin. North Am.* **2018**, *102*, 831–843.
- (40) Robinson, A.; Lehmann, J.; Barriopedro, D.; Rahmstorf, S.; Coumou, D. Increasing heat and rainfall extremes now far outside the historical climate. *npj Clim. Atmos. Sci.* **2021**, *4*, No. 45.
- (41) Sun, Q.; Miao, C.; Hanel, M.; et al. Global heat stress on health, wildfires, and agricultural crops under different levels of climate warming. *Environ. Int.* **2019**, *128*, 125–136.
- (42) Watts, N.; Amann, M.; Ayeb-Karlsson, S.; et al. The Lancet Countdown on health and climate change: from 25 years of inaction to a global transformation for public health. *Lancet* **2018**, *391*, 581–630.
- (43) Koteswara Rao, K.; Lakshmi Kumar, T. V.; Kulkarni, A.; et al. Projections of heat stress and associated work performance over India in response to global warming. *Sci. Rep.* **2020**, *10*, No. 16675.
- (44) Flouris, A. D.; Dinas, P. C.; Ioannou, L. G.; et al. Workers' health and productivity under occupational heat strain: a systematic review and meta-analysis. *Lancet Planet. Health* **2018**, *2*, eS21–eS31.
- (45) Brasier, N.; Niederberger, C.; Zanella, M.; et al. The molecular signature of heat stress in sweat reveals non-invasive biomarker candidates for health monitoring. *Commun. Biol.* **2025**, *8*, No. 650, DOI: 10.1038/s42003-025-08080-1.
- (46) Chen, C.; Ding, S.; Wang, J. Digital health for aging populations. *Nat. Med.* **2023**, *29*, 1623–1630.
- (47) Xu, C.; Solomon, S. A.; Gao, W. Artificial intelligence-powered electronic skin. *Nat. Mach. Intell.* **2023**, *5*, 1344–1355.
- (48) Ghanim, R.; Kaushik, A.; Park, J.; Abramson, A. Communication protocols integrating wearables, ingestibles, and implantables for closed-loop therapies. *Device* **2023**, *1*, No. 100092.
- (49) Boughton, C. K.; Tripyla, A.; Hartnell, S.; et al. Fully automated closed-loop glucose control compared with standard insulin therapy in adults with type 2 diabetes requiring dialysis: an open-label, randomized crossover trial. *Nat. Med.* **2021**, *27*, 1471–1476.
- (50) Jiang, Y.; Trotsyuk, A. A.; Niu, S.; et al. Wireless, closed-loop, smart bandage with integrated sensors and stimulators for advanced wound care and accelerated healing. *Nat. Biotechnol.* **2023**, *41*, 652–662.
- (51) Hsu, T. C.; Whelan, P.; Gandrup, J.; et al. Personalized interventions for behaviour change: A scoping review of just-in-time adaptive interventions. *Br. J. Health Psychol.* **2025**, *30*, No. e12766, DOI: 10.1111/bjhp.12766.
- (52) Silva, F. N.; Amancio, D. R.; Bardosova, M.; Costa, L. da F.; Oliveira, O. N. Using network science and text analytics to produce surveys in a scientific topic. *J. Informetr.* **2016**, *10*, 487–502.
- (53) Brito, A. C. M.; Oliveira, M. C. F.; Oliveira, O. N.; Silva, F. N.; Amancio, D. R. Network Analysis and Natural Language Processing to Obtain a Landscape of the Scientific Literature on Materials Applications. *ACS Appl. Mater. Interfaces* **2023**, *15*, 27437–27446.
- (54) Brito, A. C. M.; Oliveira, M. C. F.; Oliveira, O. N.; Silva, F. N.; Amancio, D. R. History of *Chemistry of Materials* According to Topic Evolution Based on Network Analysis and Natural Language Processing. *Chem. Mater.* **2024**, *36*, 1–7.
- (55) OpenAlex. OpenAlex snapshot—OpenAlex API documentation. <https://docs.openalex.org/download-all-data/openalex-snapshot> (2024).
- (56) Aversano, L.; Iammarino, M.; Mancino, I.; Montano, D. A systematic review on artificial intelligence approaches for smart health devices. *PeerJ. Comput. Sci.* **2024**, *10*, No. e2232.
- (57) Xiao, X.; Yin, J.; Xu, J.; Tat, T.; Chen, J. Advances in Machine Learning for Wearable Sensors. *ACS Nano* **2024**, *18*, 22734–22751.
- (58) Wang, Y.; Zou, Y.; Li, Z. Emerging intelligent wearable devices for cardiovascular health monitoring. *Nano Today* **2024**, *59*, No. 102544.
- (59) Rotbei, S.; Tseng, W. H.; Merino-Barbancho, B.; et al. Evaluating impact of movement on diabetes via artificial intelligence and smart devices systematic literature review. *Expert Syst. Appl.* **2024**, *257*, No. 125058.
- (60) Zhou, S.; Song, D.; Pu, L.; Xu, W. Recent Advances in Machine Learning Assisted Hydrogel Flexible Sensing. *Z. Anorg. Allg. Chem.* **2024**, *650*, No. e202400051, DOI: 10.1002/zaac.202400051.
- (61) Cernat, A.; Groza, A.; Tertis, M.; et al. Where artificial intelligence stands in the development of electrochemical sensors for healthcare applications-A review. *TrAC, Trends Anal. Chem.* **2024**, *181*, No. 117999.
- (62) Sun, T.; Feng, B.; Huo, J.; et al. Artificial Intelligence Meets Flexible Sensors: Emerging Smart Flexible Sensing Systems Driven by Machine Learning and Artificial Synapses. *Nanomicro Lett.* **2024**, *16*, No. 14.
- (63) Friedel, M.; Thompson, I. A. P.; Kasting, G.; et al. Opportunities and challenges in the diagnostic utility of dermal interstitial fluid. *Nat. Biomed. Eng.* **2023**, *7*, 1541–1555.
- (64) Aps, J. K. M.; Martens, L. C. Review: The physiology of saliva and transfer of drugs into saliva. *Forensic Sci. Int.* **2005**, *150*, 119–131.
- (65) Gao, F.; Liu, C.; Zhang, L.; et al. Wearable and flexible electrochemical sensors for sweat analysis: a review. *Microsyst. Nanoeng.* **2023**, *9*, No. 1.
- (66) Bai, W.; Yu, X.; Li, Q.; et al. Recent advances in the study of tear biomarkers and contact lens-based biosensors. *Chem. Eng. J.* **2024**, *499*, No. 156540.
- (67) Wu, Z.; Qiao, Z.; Chen, S.; et al. Interstitial fluid-based wearable biosensors for minimally invasive healthcare and biomedical applications. *Commun. Mater.* **2024**, *5*, No. 33.

- (68) Svensby, A. U.; Nygren, E.; Gefen, A.; et al. The importance of the simulated wound fluid composition and properties in the determination of the fluid handling performance of wound dressings. *Int. Wound J.* **2024**, *21*, No. e14861.
- (69) Heikenfeld, J.; Jajack, A.; Feldman, B.; et al. Accessing analytes in biofluids for peripheral biochemical monitoring. *Nat. Biotechnol.* **2019**, *37*, 407–419.
- (70) Childs, A.; Mayol, B.; Lasalde-Ramírez, J. A.; et al. Diving into Sweat: Advances, Challenges, and Future Directions in Wearable Sweat Sensing. *ACS Nano* **2024**, *18*, 24605–24616.
- (71) Song, M.; Bai, H.; Zhang, P.; Zhou, X.; Ying, B. Promising applications of human-derived saliva biomarker testing in clinical diagnostics. *Int. J. Oral Sci.* **2023**, *15*, No. 2.
- (72) Akkurt Arslan, M.; Brignole-Baudouin, F.; Chardonnet, S.; et al. Profiling tear film enzymes reveals major metabolic pathways involved in the homeostasis of the ocular surface. *Sci. Rep.* **2023**, *13*, No. 15231.
- (73) Mahmoud, N. N.; Hamad, K.; Al Shibitini, A.; et al. Investigating Inflammatory Markers in Wound Healing: Understanding Implications and Identifying Artifacts. *ACS Pharmacol. Transl. Sci.* **2024**, *7*, 18–27.
- (74) Wang, X.; He, A.; Yu, B.; et al. Uncovering the Sweat Biofouling Components and Distributions in Electrochemical Sensors. *Anal. Chem.* **2022**, *94*, 14402–14409.
- (75) Baker, L. B. Physiology of sweat gland function: The roles of sweating and sweat composition in human health. *Temperature* **2019**, *6*, 211–259.
- (76) Gagnon, D.; Crandall, C. G. Sweating as a heat loss thermoeffector. *Handb. Clin. Neurol.* **2018**, *156*, 211–232, DOI: 10.1016/b978-0-444-63912-7.00013-8.
- (77) Chung, M.; Fortunato, G.; Radacs, N. Wearable flexible sweat sensors for healthcare monitoring: a review. *J. R. Soc. Interface* **2019**, *16* (159), No. 20190217, DOI: 10.1098/rsif.2019.0217.
- (78) Hu, Y.; Converse, C.; Lyons, M. C.; Hsu, W. H. Neural control of sweat secretion: a review. *Br. J. Dermatol.* **2018**, *178* (6), 1246–1256, DOI: 10.1111/bjd.15808.
- (79) Crandall, C. G. Mechanisms and controllers of eccrine sweating in humans. *Front. Biosci.* **2010**, *2*, 685–696, DOI: 10.2741/s94.
- (80) Taylor, N. A. S.; Machado-Moreira, C. A. Regional variations in transepidermal water loss, eccrine sweat gland density, sweat secretion rates and electrolyte composition in resting and exercising humans. *Extreme Physiol. Med.* **2013**, *2*, No. 4, DOI: 10.1186/2046-7648-2-4.
- (81) Niu, Z.; Goto, T. Effects of individual characteristics and local body functions on sweating response: A review. *Int. J. Biometeorol.* **2024**, *68* (11), 2185–2204, DOI: 10.1007/s00484-024-02758-7.
- (82) Périard, J. D.; Eijssvogels, T. M. H.; Daanen, H. A. M. Exercise under heat stress: thermoregulation, hydration, performance implications, and mitigation strategies. *Physiol. Rev.* **2021**, *101A*, DOI: 10.1152/physrev.00038.2020.
- (83) Sato, T.; Katayama, C.; Hayashida, Y.; Asanuma, Y.; Aoyama, Y. Role of basal sweating in maintaining skin hydration in the finger: A long-standing paradox in dry skin resolved. *Exp. Dermatol.* **2022**, *31* (12), 1891–1899, DOI: 10.1111/exd.14668.
- (84) Baker, P.; Huang, C.; Radi, R.; et al. Skin Barrier Function: The Interplay of Physical, Chemical, and Immunologic Properties. *Cells* **2023**, *12* (23), 2745, DOI: 10.3390/cells12232745.
- (85) Ali, S.; Yosipovitch, G. Skin pH: From Basic Science to Basic Skin Care. *Acta Dermato Venereol.* **2013**, *93* (3), 261–267, DOI: 10.2340/00015555-1531.
- (86) Baker, L. B. Sweating Rate and Sweat Sodium Concentration in Athletes: A Review of Methodology and Intra/Interindividual Variability. *Sports Medicine* **2017**, *47*, 111–128, DOI: 10.1007/s40279-017-0691-5.
- (87) Bates, G. P.; Miller, V. S. Sweat rate and sodium loss during work in the heat. *J. Occup. Med. Toxicol.* **2008**, *3*, No. 4, DOI: 10.1186/1745-6673-3-4.
- (88) Consolazio, C. F.; Nelson, R. A.; Matoush, L. O.; Harding, R. S.; Canham, J. E. Nitrogen Excretion in Sweat and Its Relation to Nitrogen Balance Requirements. *J. Nutr.* **1963**, *79* (4), 399–406, DOI: 10.1093/jn/79.4.399.
- (89) Kuan, W.-H.; Chen, Y.-L.; Liu, C.-L. Excretion of Ni, Pb, Cu, As, and Hg in Sweat under Two Sweating Conditions. *Int. J. Environ. Res. Public Health* **2022**, *19* (17), 4323, DOI: 10.3390/ijerph19074323.
- (90) Sears, M. E.; Kerr, K. J.; Bray, R. I. Arsenic, Cadmium, Lead, and Mercury in Sweat: A Systematic Review. *J. Environ. Public Health* **2012**, *2012*, No. 184745, DOI: 10.1155/2012/184745.
- (91) Monedeiro, F.; dos Reis, R. B.; Peria, F. M.; Sares, C. T. G.; De Martinis, B. S. Investigation of sweat VOC profiles in assessment of cancer biomarkers using HS-GC-MS. *J. Breath Res.* **2020**, *14*, No. 026009, DOI: 10.1088/1752-7163/ab5b3c.
- (92) Songsathitmetha, S.; Thaveesangsakulthai, I.; Chatdarong, K.; Kulsing, C. Sweat-Based Volatile Organic Compound Identification of SARS-CoV-2 Detection. In *4th International Electronic Conference on Biosensors*; MDPI, **2024**, DOI: 10.3390/proceedings2024104038.
- (93) Chawarski, M. C.; Fiellin, D. A.; O'Connor, P. G.; Bernard, M.; Schottenfeld, R. S. Utility of sweat patch testing for drug use monitoring in outpatient treatment for opiate dependence. *J. Subst. Abuse Treat.* **2007**, *33*, 411–415, DOI: 10.1016/j.jsat.2007.03.004.
- (94) Koster, R. A.; Alffenaar, J.-W. C.; Greijdanus, B.; VanDerNagel, J. E. L.; Uges, D. R. A. Application of Sweat Patch Screening for 16 Drugs and Metabolites Using a Fast and Highly Selective LC-MS/MS Method. *Ther. Drug Monit.* **2014**, *36* (1), 35–45, DOI: 10.1097/ftd.0b013e3182a04feb.
- (95) Baker, L. B.; et al. Exercise intensity effects on total sweat electrolyte losses and regional vs. whole-body sweat [Na⁺], [Cl[−]], and [K⁺]. *Eur. J. Appl. Physiol.* **2018**, *119* (2), 361–375, DOI: 10.1007/s00421-018-4048-z.
- (96) Delgado-Povedano, M. M.; Calderón-Santiago, M.; Priego-Capote, F.; de Castro, M. D. Study of sample preparation for quantitative analysis of amino acids in human sweat by liquid chromatography–tandem mass spectrometry. *Talanta* **2016**, *146*, 310–317, DOI: 10.1016/j.talanta.2015.07.066.
- (97) Harker, M.; Coulson, H.; Fairweather, I.; Taylor, D.; Daykin, C. A. Study of metabolite composition of eccrine sweat from healthy male and female human subjects by 1H NMR spectroscopy. *Metabolomics* **2006**, *2* (1), 105, DOI: 10.1007/s11306-006-0024-4.
- (98) Brunmair, J.; et al. Finger Sweat Analysis Enables Short Interval Metabolic Biomonitoring in Humans. *Nat. Commun.* **2021**, *12*, No. 5993, DOI: 10.1038/s41467-021-26245-4.
- (99) Delgado-Povedano, M. M.; Calderón-Santiago, M.; de Castro, M. D.; Priego-Capote, F. Metabolomics analysis of human sweat collected after moderate exercise. *Talanta* **2018**, *177*, 47–65, DOI: 10.1016/j.talanta.2017.09.028.
- (100) Alvi, S. N.; Abuhdeeb, K.; Hammami, M. M. Sweat Cortisol and Cortisone Determination in Healthy Adults: UHPLC-MS/MS Assay Validation and Clinical Application. *Adv. Pharmacol. Pharm. Sci.* **2022**, *2022*, No. 3133640, DOI: 10.1155/2022/3133640.
- (101) Hossain, N. I.; Noushin, T.; Tabassum, S. StressFit: a hybrid wearable physicochemical sensor suite for simultaneously measuring electromyogram and sweat cortisol. *Sci. Rep.* **2024**, *14*, No. 29667, DOI: 10.1038/s41598-024-81042-5.
- (102) Ye, C.; Wang, M.; Min, J.; et al. A wearable aptamer nanobiosensor for non-invasive female hormone monitoring. *Nat. Nanotechnol.* **2024**, *19*, 330–337.
- (103) Yu, Y.; Prassas, I.; Muytjens, C. M. J.; Diamandis, E. P. Proteomic and peptidomic analysis of human sweat with emphasis on proteolysis. *J. Proteomics* **2017**, *155*, 40–48, DOI: 10.1016/j.jpro.2017.01.005.
- (104) Hussain, J. N.; Mantri, N.; Cohen, M. M. Working Up a Good Sweat – The Challenges of Standardising Sweat Collection for Metabolomics Analysis. *Clin. Biochem. Rev.* **2017**, *38*, 13.
- (105) Saha, T.; Del Caño, R.; la Paz, E.; Sandhu, S. S.; Wang, J. Access and Management of Sweat for Non-Invasive Biomarker Monitoring: A Comprehensive Review. *Small* **2022**, *19* (51), No. 2206064, DOI: 10.1002/smll.202206064.

- (106) Nunes, M. J.; Moura, J. J. G.; Noronha, J. P.; et al. Evaluation of Sweat-Sampling Procedures for Human Stress-Biomarker Detection. *Analytica* **2022**, *3*, 178–194, DOI: 10.3390/analytica3020013.
- (107) Mishra, A.; Greaves, R.; Massie, J. The Relevance of Sweat Testing for the Diagnosis of Cystic Fibrosis in the Genomic Era. *Clin. Biochem. Rev./Austr. Assoc. Clin. Biochem.* **2005**, *26*, 135.
- (108) Hammond, K. B.; Turcios, N. L.; Gibson, L. E. Clinical evaluation of the macroduct sweat collection system and conductivity analyzer in the diagnosis of cystic fibrosis. *J. Pediatr.* **1994**, *124* (2), 255–260, DOI: 10.1016/s0022-3476(94)70314-0.
- (109) Collins, M. N.; Brawley, C. B.; McCracken, C. E.; et al. Risk Factors for Quantity Not Sufficient Sweat Collection in Infants 3 Months or Younger. *Am. J. Clin. Pathol.* **2014**, *142* (1), 72–75, DOI: 10.1309/ajcplhg2buvbtsly.
- (110) McColley, S. A.; Elbert, A.; Wu, R.; et al. Quantity not sufficient rates and delays in sweat testing in US infants with cystic fibrosis. *Pediatr. Pulmonol.* **2020**, *55* (11), 3053–3056, DOI: 10.1002/ppul.25027.
- (111) Morris, D.; Coyle, S.; Wu, Y.; et al. Bio-sensing textile based patch with integrated optical detection system for sweat monitoring. *Sens. Actuators, B* **2009**, *139* (1), 231–236, DOI: 10.1016/j.snb.2009.02.032.
- (112) Coyle, S.; Lau, K. T.; Moyna, N.; et al. BIOTEX—Biosensing Textiles for Personalised Healthcare Management. *IEEE Trans. Inf. Technol. Biomed.* **2010**, *14* (2), 364–370, DOI: 10.1109/titb.2009.2038484.
- (113) Bhandodkar, A. J.; Hung, V. W. S.; Jia, W.; et al. Tattoo-based potentiometric ion-selective sensors for epidermal pH monitoring. *Analyst* **2013**, *138* (1), 123–128, DOI: 10.1039/c2an36422k.
- (114) Guinovart, T.; Bhandodkar, A. J.; Windmiller, J. R.; Andrade, F. J.; Wang, J. A potentiometric tattoo sensor for monitoring ammonium in sweat. *Analyst* **2013**, *138* (22), 7031–7038, DOI: 10.1039/c3an01672b.
- (115) Huang, X.; Liu, Y.; Chen, K.; et al. Stretchable, Wireless Sensors and Functional Substrates for Epidermal Characterization of Sweat. *Small* **2014**, *10* (15), 3083–3090, DOI: 10.1002/sml.201400483.
- (116) Rose, D. P.; Ratterman, M. E.; Griffin, D. K.; et al. Adhesive RFID Sensor Patch for Monitoring of Sweat Electrolytes. *IEEE Trans. Biomed. Eng.* **2015**, *62* (6), 1457–1465, DOI: 10.1109/tbme.2014.2369991.
- (117) Koh, A.; Kang, D.; Xue, Y.; et al. A soft, wearable microfluidic device for the capture, storage, and colorimetric sensing of sweat. *Sci. Transl. Med.* **2016**, *8*, No. 366ra165.
- (118) Madhu, S.; Alam, M. S.; Ramasamy, S.; Choi, J. Wearable, fabric-based microfluidic systems with integrated electrochemical and colorimetric sensing array for multiplex sweat analysis. *Chem. Eng. J.* **2025**, *504*, No. 158979.
- (119) Li, M.; Wang, L.; Liu, R.; et al. A highly integrated sensing paper for wearable electrochemical sweat analysis. *Biosens. Bioelectron.* **2021**, *174*, No. 112828.
- (120) Bian, P.; Wang, Y.; McCarthy, T. J. Rediscovering Silicones: The Anomalous Water Permeability of “Hydrophobic” PDMS Suggests Nanostructure and Applications in Water Purification and Anti-Icing. *Macromol. Rapid Commun.* **2020**, *42* (5), No. 2000682, DOI: 10.1002/marc.202000682.
- (121) Hardy, B. S.; Uechi, K.; Zhen, J.; Pirouz Kavehpour, H. The deformation of flexible PDMS microchannels under a pressure driven flow. *Lab Chip* **2009**, *9* (7), 935–938, DOI: 10.1039/b813061b.
- (122) Qian, X.; Zhang, W.; Peng, C.; et al. Characterizing the Deformation of the Polydimethylsiloxane (PDMS) Membrane for Microfluidic System through Image Processing. *Micromachines* **2016**, *7* (5), 92 DOI: 10.3390/mi7050092.
- (123) Yang, D. S.; Wu, Y.; Kanatzidis, E. E.; et al. 3D-printed epidermal sweat microfluidic systems with integrated microcuvettes for precise spectroscopic and fluorometric biochemical assays. *Mater. Horiz.* **2023**, *10*, 4992–5003.
- (124) Wu, C.-H.; Ma, H. J. H.; Baessler, P.; Balanay, R. K.; Ray, T. R. Skin-interfaced microfluidic systems with spatially engineered 3D fluidics for sweat capture and analysis. *Sci. Adv.* **2023**, *9*, No. eadg4272, DOI: 10.1126/sciadv.adg4272.
- (125) Reeder, J. T.; Choi, J.; Xue, Y.; et al. Waterproof, electronics-enabled, epidermal microfluidic devices for sweat collection, biomarker analysis, and thermography in aquatic settings. *Sci. Adv.* **2019**, *5*, No. eaau6356, DOI: 10.1126/sciadv.aau6356.
- (126) Yang, Y.; Song, Y.; Bo, X.; et al. A laser-engraved wearable sensor for sensitive detection of uric acid and tyrosine in sweat. *Nat. Biotechnol.* **2020**, *38*, 217–224.
- (127) Wang, M.; Yang, Y.; Gao, W. Laser-engraved graphene for flexible and wearable electronics. *Trends Chem.* **2021**, *3* (11), 969–981, DOI: 10.1016/j.trechm.2021.09.001.
- (128) Chen, C.; Fu, Y.; Sparks, S. S.; et al. 3D-Printed Flexible Microfluidic Health Monitor for *In Situ* Sweat Analysis and Biomarker Detection. *ACS Sens.* **2024**, *9*, 3212–3223.
- (129) Sakurada, K.; Akutsu, T.; Fukushima, H.; Watanabe, K.; Yoshino, M. Detection of dermcidin for sweat identification by real-time RT-PCR and ELISA. *Forensic Sci. Int.* **2010**, *194* (1–3), 80–84, DOI: 10.1016/j.forsciint.2009.10.015.
- (130) Dai, X.; Okazaki, H.; Hanakawa, Y.; et al. Eccrine Sweat Contains IL-1 α , IL-1 β and IL-31 and Activates Epidermal Keratinocytes as a Danger Signal. *PLoS One* **2013**, *8* (7), No. e67666, DOI: 10.1371/journal.pone.0067666.
- (131) Hooton, K.; Han, W.; Li, L. Comprehensive and Quantitative Profiling of the Human Sweat Submetabolome Using High-Performance Chemical Isotope Labeling LC–MS. *Anal. Chem.* **2016**, *88* (14), 7378–7386, DOI: 10.1021/acs.analchem.6b01930.
- (132) Zhang, K.; Zhang, J.; Wang, F.; Kong, D. Stretchable and Superwetttable Colorimetric Sensing Patch for Epidermal Collection and Analysis of Sweat. *ACS Sens.* **2021**, *6*, 2261–2269.
- (133) Zhou, Y.; Han, H.; Naw, H. P. P.; et al. Real-time colorimetric hydration sensor for sport activities. *Mater. Des.* **2016**, *90*, 1181–1185, DOI: 10.1016/j.matdes.2015.06.078.
- (134) Promphet, N.; Rattanawaleedirojn, P.; Siralermukul, K.; et al. Non-invasive textile based colorimetric sensor for the simultaneous detection of sweat pH and lactate. *Talanta* **2019**, *192*, 424–430, DOI: 10.1016/j.talanta.2018.09.086.
- (135) Ha, J.-H.; Jeong, Y.; Ahn, J.; et al. A wearable colorimetric sweat pH sensor-based smart textile for health state diagnosis. *Mater. Horiz.* **2023**, *10*, 4163–4171.
- (136) Currano, L. J.; Sage, F. C.; Hagedon, M.; et al. Wearable Sensor System for Detection of Lactate in Sweat. *Sci. Rep.* **2018**, *8*, No. 15890, DOI: 10.1038/s41598-018-33565-x.
- (137) Wu, X.; Jin, D.; Li, M.; Luo, X.; Wang, L. Wearable analytical platform for colorimetric detection of sweat lactate using spherical colloidal photonic crystal hydrogel. *Microchem. J.* **2024**, *200*, No. 110327, DOI: 10.1016/j.microc.2024.110327.
- (138) Cho, S.; Shaban, S. M.; Song, R.; et al. A skin-interfaced microfluidic platform supports dynamic sweat biochemical analysis during human exercise. *Sci. Transl. Med.* **2024**, *16*, No. eado5366, DOI: 10.1126/scitranslmed.ado5366.
- (139) Kim, J.; Wu, Y.; Luan, H.; et al. A Skin-Interfaced, Miniaturized Microfluidic Analysis and Delivery System for Colorimetric Measurements of Nutrients in Sweat and Supply of Vitamins Through the Skin. *Adv. Sci.* **2022**, *9*, No. 2103331, DOI: 10.1002/advs.202103331.
- (140) Beygnezhad, F.; Dashtian, K.; Zare-Dorabei, R.; Zangouei, Z. Biological Ce/Fe-metal organic framework multifunctional nanozyme-based colorimetric sensor array for multi-analyte detection in human sweat. *Sens. Actuators, B* **2025**, *423*, No. 136769, DOI: 10.1016/j.snb.2024.136769.
- (141) Mukhopadhyay, M.; Subramanian, S. G.; Durga, K. V.; Sarkar, D.; DasGupta, S. Laser printing based colorimetric paper sensors for glucose and ketone detection: Design, fabrication, and theoretical analysis. *Sens. Actuators, B* **2022**, *371*, No. 132599, DOI: 10.1016/j.snb.2022.132599.
- (142) Wu, Y.; Li, X.; Madsen, K. E.; et al. Skin-interfaced microfluidic biosensors for colorimetric measurements of the

concentrations of ketones in sweat. *Lab Chip* **2024**, *24* (18), 4288–4295, DOI: 10.1039/d4lc00588k.

(143) Zhang, Y.; Guo, H.; Kim, S. B.; et al. Passive sweat collection and colorimetric analysis of biomarkers relevant to kidney disorders using a soft microfluidic system. *Lab Chip* **2019**, *19* (9), 1545–1555, DOI: 10.1039/c9lc00103d.

(144) Yang, H.; Ji, Y.; Shen, K.; Qian, Y.; Ye, C. Simultaneous detection of urea and lactate in sweat based on a wearable sweat biosensor. *Biomed. Opt. Express* **2023**, *15* (1), 14–27, DOI: 10.1364/boe.505004.

(145) Salatiello, S.; Spinelli, M.; Cassiano, C.; et al. Sweat urea bioassay based on degradation of Prussian Blue as the sensing architecture. *Anal. Chim. Acta* **2022**, *1210*, No. 339882, DOI: 10.1016/j.aca.2022.339882.

(146) Yue, X.; Xu, F.; Zhang, L.; et al. Simple, Skin-Attachable, and Multifunctional Colorimetric Sweat Sensor. *ACS Sens.* **2022**, *7*, 2198–2208.

(147) Li, F.; et al. Flexible microfluidic colorimetric detection chip integrated with ABTS+ and Co@MnO₂ nanozyme catalyzed TMB reaction systems for bio-enzyme free detection of sweat uric acid. *Anal. Chim. Acta* **2024**, *1299*, No. 342453, DOI: 10.1016/j.aca.2024.342453.

(148) Xiao, J.; Luo, Y.; Su, L.; et al. Hydrophilic metal-organic frameworks integrated uricase for wearable detection of sweat uric acid. *Anal. Chim. Acta* **2022**, *1208*, No. 339843, DOI: 10.1016/j.aca.2022.339843.

(149) Xiao, J.; Liu, Y.; Su, L.; et al. Microfluidic Chip-Based Wearable Colorimetric Sensor for Simple and Facile Detection of Sweat Glucose. *Anal. Chem.* **2019**, *91* (23), 14803–14807, DOI: 10.1021/acs.analchem.9b03110.

(150) Vaquer, A.; Barón, E.; de la Rica, R. Detection of low glucose levels in sweat with colorimetric wearable biosensors. *Analyst* **2021**, *146* (10), 3273–3279, DOI: 10.1039/d1an00283j.

(151) Garcia-Rey, S.; Gil-Hernandez, E.; Basabe-Desmonts, L.; Benito-Lopez, F. Colorimetric Determination of Glucose in Sweat Using an Alginate-Based Biosystem. *Polymers*. (2023) DOI: 10.3390/polym15051218.

(152) Dalirirad, S.; Steckl, A. J. Aptamer-based lateral flow assay for point of care cortisol detection in sweat. *Sens. Actuators, B* **2019**, *283*, 79–86, DOI: 10.1016/j.snb.2018.11.161.

(153) Tu, E.; Pearlmuter, P.; Tiangco, M.; et al. Comparison of Colorimetric Analyses to Determine Cortisol in Human Sweat. *ACS Omega* **2020**, *5* (14), 8211–8218, DOI: 10.1021/acsomega.0c00498.

(154) Kim, S. B.; Koo, J.; Yoon, J.; et al. Soft, skin-interfaced microfluidic systems with integrated enzymatic assays for measuring the concentration of ammonia and ethanol in sweat. *Lab Chip* **2020**, *20* (41), 84–92, DOI: 10.1039/c9lc01045a.

(155) Sánchez, M.; González, A.; Sabio, L.; et al. Photochromic polyoxometalate-based enzyme-free reusable sensors for real-time colorimetric detection of alcohol in sweat and saliva. *Mater. Today Chem.* **2021**, *21*, No. 100491, DOI: 10.1016/j.mtchem.2021.100491.

(156) Hladek, M. D.; Szanton, S. L.; Cho, Y. E.; et al. Using sweat to measure cytokines in older adults compared to younger adults: A pilot study. *J. Immunol. Methods* **2018**, *454*, 1–5, DOI: 10.1016/j.jim.2017.11.003.

(157) Marques-Deak, A.; Cizza, G.; Eskandari, F.; et al. Measurement of cytokines in sweat patches and plasma in healthy women: Validation in a controlled study. *J. Immunol. Methods* **2006**, *315* (1–2), 99–109, DOI: 10.1016/j.jim.2006.07.011.

(158) Sekine, Y.; Kim, S. B.; Zhang, Y.; et al. A fluorometric skin-interfaced microfluidic device and smartphone imaging module for *in situ* quantitative analysis of sweat chemistry. *Lab Chip* **2018**, *18*, 2178–2186.

(159) Cho, S. H.; Cho, S.; Lv, Z.; et al. Soft, wearable, microfluidic system for fluorometric analysis of loss of amino acids through eccrine sweat. *Lab Chip* **2025**, *25*, 1647–1655.

(160) Kim, S.; Lee, B.; Reeder, J. T.; et al. Soft, skin-interfaced microfluidic systems with integrated immunoassays, fluorometric

sensors, and impedance measurement capabilities. *Proc. Natl. Acad. Sci. U.S.A.* **2020**, *117*, 27906–27915.

(161) Mukherjee, S.; Pietrosemoli Salazar, S.; Saha, T.; Dickey, M. D.; Velev, O. D. Capillary-osmotic wearable patch based on lateral flow assay for sweat potassium analysis. *Sens. Actuators, B* **2024**, *419*, No. 136383, DOI: 10.1016/j.snb.2024.136383.

(162) Mou, L.; Xia, Y.; Jiang, X. Epidermal Sensor for Potentiometric Analysis of Metabolite and Electrolyte. *Anal. Chem.* **2021**, *93* (33), 11525–11531, DOI: 10.1021/acs.analchem.1c01940.

(163) Lai, M.; Zhong, L.; Liu, S.; et al. Carbon fiber-based multichannel solid-contact potentiometric ion sensors for real-time sweat electrolyte monitoring. *Anal. Chim. Acta* **2024**, *1287*, No. 342046, DOI: 10.1016/j.aca.2023.342046.

(164) Li, C.; Liu, Y.; Liu, G.; et al. Development of a low-cost flexible potentiometric detector and its integrated system for electrochemical sensing of electrolytes in human sweat. *Sens. Actuators Rep.* **2025**, *9*, No. 100286, DOI: 10.1016/j.snr.2025.100286.

(165) Kim, J.; Jeerapan, I.; Imani, S.; et al. Noninvasive Alcohol Monitoring Using a Wearable Tattoo-Based Iontophoretic-Biosensing System. *ACS Sens.* **2016**, *1* (8), 1011–1019, DOI: 10.1021/acssensors.6b00356.

(166) Lin, P.-H.; Sheu, S.-C.; Chen, C.-W.; Huang, S.-C.; Li, B.-R. Wearable hydrogel patch with noninvasive, electrochemical glucose sensor for natural sweat detection. *Talanta* **2022**, *241*, No. 123187, DOI: 10.1016/j.talanta.2021.123187.

(167) Mugo, S. M.; Robertson, S. V.; Lu, W. A molecularly imprinted electrochemical microneedle sensor for multiplexed metabolites detection in human sweat. *Talanta* **2023**, *259*, No. 124531, DOI: 10.1016/j.talanta.2023.124531.

(168) Pei, S.; Ji, W.; Yang, Y.; et al. Flexible Dual-Channel Molecularly Imprinted Electrochemical Sensor for Simultaneously Monitoring Sweat Cortisol and Lactate Levels. *Anal. Sens.* **2024**, *4* (5), No. e202400003, DOI: 10.1002/anse.202400003.

(169) Nah, J. S.; Barman, S. C.; Zahed, M. A.; et al. A wearable microfluidics-integrated impedimetric immunosensor based on Ti₃C₂T MXene incorporated laser-burned graphene for noninvasive sweat cortisol detection. *Sens. Actuators, B* **2021**, *329*, No. 129206, DOI: 10.1016/j.snb.2020.129206.

(170) Yuan, Z.; et al. A multi-modal sweat sensing patch for cross-verification of sweat rate, total ionic charge, and Na⁺ concentration. *Lab Chip* **2019**, *19* (19), 3179–3189, DOI: 10.1039/c9lc00598f.

(171) Munje, R. D.; Muthukumar, S.; Jagannath, B.; Prasad, S. A new paradigm in sweat based wearable diagnostics biosensors using Room Temperature Ionic Liquids (RTILs). *Sci. Rep.* **2017**, *7*, No. 1950, DOI: 10.1038/s41598-017-02133-0.

(172) Lee, H.-B.; Meeseepong, M.; Trung, T. Q.; Kim, B.-Y.; Lee, N.-E. A wearable lab-on-a-patch platform with stretchable nano-structured biosensor for non-invasive immunodetection of biomarker in sweat. *Biosens. Bioelectron.* **2020**, *156*, No. 112133, DOI: 10.1016/j.bios.2020.112133.

(173) Aydın, R.; Akkaya, A.; Kahveci, O.; Şahin, B. Nanostructured CuO Thin-Film-Based Conductometric Sensors for Real-Time Tracking of Sweat Loss. *ACS Omega* **2023**, *8* (22), 20009–20019, DOI: 10.1021/acsomega.3c02232.

(174) Gao, W.; Emaminejad, S.; Nyein, H. Y. Y.; et al. Fully integrated wearable sensor arrays for multiplexed *in situ* perspiration analysis. *Nature* **2016**, *529*, 509–514.

(175) Hui, X. et al. High-Performance Flexible Electrochemical Heavy Metal Sensor Based on Layer-by-Layer Assembly of Ti₃C₂T_x/MWNTs Nanocomposites for Noninvasive Detection of Copper and Zinc Ions in Human Biofluids. *ACS Appl. Mater. Interfaces* (2020) DOI: 10.1021/acsami.0c12239.

(176) Gao, W.; Nyein, H. Y. Y.; Shahpar, Z.; et al. Wearable Microsensor Array for Multiplexed Heavy Metal Monitoring of Body Fluids. *ACS Sens.* **2016**, *1*, 866–874.

(177) Silva, R. R.; Raymundo-Pereira, P. A.; Campos, A. M.; et al. Microbial nanocellulose adherent to human skin used in electrochemical sensors to detect metal ions and biomarkers in sweat. *Talanta* **2020**, *218*, No. 121153.

- (178) Tai, L.; Gao, W.; Chao, M.; et al. Methylxanthine Drug Monitoring with Wearable Sweat Sensors. *Adv. Mater.* **2018**, *30*, No. 1707442, DOI: 10.1002/adma.201707442.
- (179) Tai, L.-C.; Ahn, C. H.; Nyein, H. Y. Y.; et al. Nicotine Monitoring with a Wearable Sweat Band. *ACS Sens.* **2020**, *5* (6), 1831–1837, DOI: 10.1021/acssensors.0c00791.
- (180) Panneer Selvam, A.; Muthukumar, S.; Kamakoti, V.; Prasad, S. A wearable biochemical sensor for monitoring alcohol consumption lifestyle through Ethyl glucuronide (EtG) detection in human sweat. *Sci. Rep.* **2016**, *6*, No. 23111, DOI: 10.1038/srep23111.
- (181) Koh, E. H.; Lee, W. C.; Choi, Y. J.; et al. A Wearable Surface-Enhanced Raman Scattering Sensor for Label-Free Molecular Detection. *ACS Appl. Mater. Interfaces* **2021**, *13* (2), 3024–3032, DOI: 10.1021/acsami.0c18892.
- (182) Alizadeh, A.; Burns, A.; Lenigk, R.; et al. A wearable patch for continuous monitoring of sweat electrolytes during exertion. *Lab Chip* **2018**, *18* (17), 2632–2641, DOI: 10.1039/c8lc00510a.
- (183) Reeder, J. T.; Xue, Y.; Franklin, D.; et al. Resettable skin interfaced microfluidic sweat collection devices with chemesthetic hydration feedback. *Nat. Commun.* **2019**, *10*, No. 5513.
- (184) Zhong, B.; Qin, X.; Xu, H.; et al. Interindividual- and blood-correlated sweat phenylalanine multimodal analytical biochips for tracking exercise metabolism. *Nat. Commun.* **2024**, *15*, No. 624, DOI: 10.1038/s41467-024-44751-z.
- (185) Klous, L.; de Ruiter, C. J.; Scherrer, S.; Gerrett, N.; Daanen, H. A. M. The (in)dependency of blood and sweat sodium, chloride, potassium, ammonia, lactate and glucose concentrations during submaximal exercise. *Eur. J. Appl. Physiol.* **2021**, *121* (3), 803–816, DOI: 10.1007/s00421-020-04562-8.
- (186) Gokdemir, Y.; Karadag, B. T. Sweat Testing and Recent Advances. *Front. Pediatr.* **2021**, *9*, No. 649904, DOI: 10.3389/fped.2021.649904.
- (187) Cirilli, N.; Southern, K.; Barben, J.; et al. Standards of care guidance for sweat testing; phase two of the ECFS quality improvement programme. *J. Cystic Fibrosis* **2022**, *21* (3), 434–441, DOI: 10.1016/j.jcf.2022.01.004.
- (188) Vairo, D.; Bruzzese, L.; Marlinge, M.; et al. Towards Addressing the Body Electrolyte Environment via Sweat Analysis: Pilocarpine Iontophoresis Supports Assessment of Plasma Potassium Concentration. *Sci. Rep.* **2017**, *7*, No. 11801, DOI: 10.1038/s41598-017-12211-y.
- (189) Moyer, J.; Wilson, D.; Finkelstein, I.; Wong, B.; Potts, R. Correlation Between Sweat Glucose and Blood Glucose in Subjects with Diabetes. *Diabetes Technol. Ther.* **2012**, *14*, 398–402.
- (190) Nyein, H. Y. Y.; Bariya, M.; Kivimäki, L.; et al. Regional and correlative sweat analysis using high-throughput microfluidic sensing patches toward decoding sweat. *Sci. Adv.* **2019**, *5*, No. eaaw9906, DOI: 10.1126/sciadv.aaw9906.
- (191) Karpova, E. V.; Shcherbacheva, E. V.; Galushin, A. A.; et al. Noninvasive Diabetes Monitoring through Continuous Analysis of Sweat Using Flow-Through Glucose Biosensor. *Anal. Chem.* **2019**, *91*, 3778–3783.
- (192) Takei, N.; Inaba, T.; Morita, Y.; et al. Differential patterns of sweat and blood lactate concentration response during incremental exercise in varied ambient temperatures: A pilot study. *Temperature* **2024**, *11* (3), 247–253, DOI: 10.1080/23328940.2024.2375693.
- (193) Gamella, M.; Campuzano, S.; Manso, J.; et al. A novel non-invasive electrochemical biosensing device for in situ determination of the alcohol content in blood by monitoring ethanol in sweat. *Anal. Chim. Acta* **2014**, *806*, 1–7, DOI: 10.1016/j.aca.2013.09.020.
- (194) Buono, M. J. Sweat Ethanol Concentrations are Highly Correlated with Co-Existing Blood Values in Humans. *Exp. Physiol.* **1999**, *84* (2), 401–404, DOI: 10.1111/j.1469-445x.1999.01798.x.
- (195) Adelaars, S.; Konings, C. J.; Cox, L.; et al. The correlation of urea and creatinine concentrations in sweat and saliva with plasma during hemodialysis: an observational cohort study. *Clin. Chem. Lab. Med.* **2024**, *62* (6), 1118–1125, DOI: 10.1515/cclm-2023-1285.
- (196) Torrente-Rodríguez, R. M.; Tu, J.; Yang, Y.; et al. Investigation of Cortisol Dynamics in Human Sweat Using a Graphene-Based Wireless mHealth System. *Matter* **2020**, *2* (4), 921–937, DOI: 10.1016/j.matt.2020.01.021.
- (197) Kim, J.; Oh, S.; Yang, D. S.; et al. A skin-interfaced, miniaturized platform for triggered induction, capture and colorimetric multicomponent analysis of microliter volumes of sweat. *Biosens. Bioelectron.* **2024**, *253*, No. 116166.
- (198) Wang, M.; Yang, Y.; Min, J.; et al. A wearable electrochemical biosensor for the monitoring of metabolites and nutrients. *Nat. Biomed. Eng.* **2022**, *6* (11), 1225–1235, DOI: 10.1038/s41551-022-00916-z.
- (199) Saha, T.; Khan, M. I.; Sandhu, S. S.; et al. A Passive Perspiration Inspired Wearable Platform for Continuous Glucose Monitoring. *Adv. Sci.* **2024**, *11*, No. 2405518, DOI: 10.1002/advs.202405518.
- (200) Ok, J.; Park, S.; Jung, Y. H.; Kim, T. Wearable and Implantable Cortisol-Sensing Electronics for Stress Monitoring. *Adv. Mater.* **2024**, *36*, No. 2211595, DOI: 10.1002/adma.202211595.
- (201) Luan, Y.; Zhou, Y.; Li, C.; et al. Wearable Sensing Device Integrated with Prestored Reagents for Cortisol Detection in Sweat. *ACS Sens.* **2024**, *9* (4), 2075–2082, DOI: 10.1021/acssensors.4c00112.
- (202) Emaminejad, S.; Gao, W.; Wu, E.; et al. Autonomous sweat extraction and analysis applied to cystic fibrosis and glucose monitoring using a fully integrated wearable platform. *Proc. Natl. Acad. Sci. U.S.A.* **2017**, *114*, 4625–4630.
- (203) Premadasa, R.; Wan, Z.; Almasi, P.; et al. CFTrack: Advanced Diagnostic, Monitoring, and Tracking Device for Cystic Fibrosis Care. *ACS Sens.* **2024**, *9* (11), 5915–5925, DOI: 10.1021/acssensors.4c01669.
- (204) Ray, T. R.; Ivanovic, M.; Curtis, P. M.; et al. Soft, skin-interfaced sweat stickers for cystic fibrosis diagnosis and management. *Sci. Transl. Med.* **2021**, *13*, No. eabd8109, DOI: 10.1126/scitranslmed.abd8109.
- (205) Chen, Z.; Wang, W.; Tian, H.; et al. Wearable intelligent sweat platform for SERS-AI diagnosis of gout. *Lab Chip* **2024**, *24* (7), 1996–2004, DOI: 10.1039/d3lc01094e.
- (206) Rosati, G.; Gherardi, G.; Grigoletto, D.; et al. Lactate Dehydrogenase and Glutamate Pyruvate Transaminase biosensing strategies for lactate detection on screen-printed sensors. Catalysis efficiency and interference analysis in complex matrices: from cell cultures to sport medicine. *Sens. Biosensing Res.* **2018**, *21*, 54–64.
- (207) Ibáñez-Redín, G.; et al. Wearable potentiometric biosensor for analysis of urea in sweat. *Biosens. Bioelectron.* **2023**, *223*, No. 114994.
- (208) Raymundo-Pereira, P. A.; Gomes, N. O.; Machado, S. A. S.; Oliveira, O. N. Wearable glove-embedded sensors for therapeutic drug monitoring in sweat for personalized medicine. *Chem. Eng. J.* **2022**, *435*, No. 135047.
- (209) Ghaffari, R.; Aranyosi, A. J.; Lee, S. P.; Model, J. B.; Baker, L. B. The Gx Sweat Patch for personalized hydration management. *Nat. Rev. Bioeng.* **2023**, *1* (1), 5–7, DOI: 10.1038/s44222-022-00005-5.
- (210) Sempionatto, J. R.; Moon, J.-M.; Wang, J. Touch-Based Fingertip Blood-Free Reliable Glucose Monitoring: Personalized Data Processing for Predicting Blood Glucose Concentrations. *ACS Sens.* **2021**, *6* (5), 1875–1883, DOI: 10.1021/acssensors.1c00139.
- (211) Cardoso, A. G.; Viltres, H.; Ortega, G. A.; et al. Electrochemical sensing of analytes in saliva: Challenges, progress, and perspectives. *TrAC, Trends Anal. Chem.* **2023**, *160*, No. 116965.
- (212) Proctor, G. B. The physiology of salivary secretion. *Periodontology 2000* **2016**, *70*, 11–25.
- (213) Surdu, A.; Foia, L. G.; Luchian, I.; et al. Saliva as a Diagnostic Tool for Systemic Diseases—A Narrative Review. *Medicina* **2025**, *61*, 243.
- (214) Kumari, S.; Samara, M.; Ampadi Ramachandran, R.; et al. A Review on Saliva-Based Health Diagnostics: Biomarker Selection and Future Directions. *Biomed. Mater. Devices* **2024**, *2*, 121–138.
- (215) Desai, P.; Donovan, L.; Janowitz, E.; Kim, J. Y. The Clinical Utility of Salivary Biomarkers in the Identification of Type 2 Diabetes Risk and Metabolic Syndrome. *Diabetes Metab. Syndr. Obes.* **2020**, *13*, 3587–3599.

- (216) Bastias, D.; Maturana, A.; Marín, C.; Martínez, R.; Niklander, S. E. Salivary Biomarkers for Oral Cancer Detection: An Exploratory Systematic Review. *Int. J. Mol. Sci.* **2024**, *25*, 2634.
- (217) Yousif, G.; Murugesan, S.; Djekidel, M. N.; et al. Distinctive blood and salivary proteomics signatures in Qatari individuals at high risk for cardiovascular disease. *Sci. Rep.* **2025**, *15*, No. 4056.
- (218) Chojnowska, S.; Ptasińska-Sarosiak, I.; Kępa, A.; Knaś, M.; Waszkiewicz, N. Salivary Biomarkers of Stress, Anxiety and Depression. *J. Clin. Med.* **2021**, *10*, 517.
- (219) de Oliveira Neto, N. F.; Caixeta, R. A. V.; Zerbinati, R. M.; et al. The Emergence of Saliva as a Diagnostic and Prognostic Tool for Viral Infections. *Viruses* **2024**, *16*, 1759.
- (220) Cunha-Cruz, J.; Scott, J.; Rothen, M.; et al. Salivary characteristics and dental caries. *J. Am. Dental Assoc.* **2013**, *144*, e31–e40.
- (221) Timpel, J.; Klinghammer, S.; Riemenschneider, L.; et al. Sensors for in situ monitoring of oral and dental health parameters in saliva. *Clin. Oral Invest.* **2023**, *27*, 5719–5736.
- (222) Liao, C.; Chen, X.; Fu, Y. Salivary analysis: An emerging paradigm for non-invasive healthcare diagnosis and monitoring. *Interdiscip. Med.* **2023**, *1* (3), No. e20230009, DOI: 10.1002/INMD.20230009.
- (223) Kim, J.; Imani, S.; de Araujo, W. R.; et al. Wearable salivary uric acid mouthguard biosensor with integrated wireless electronics. *Biosens. Bioelectron.* **2015**, *74*, 1061–1068.
- (224) Arakawa, T.; Tomoto, K.; Nitta, H.; et al. A Wearable Cellulose Acetate-Coated Mouthguard Biosensor for In Vivo Salivary Glucose Measurement. *Anal. Chem.* **2020**, *92*, 12201–12207.
- (225) Mannoor, M. S.; Tao, H.; Clayton, J. D.; et al. Graphene-based wireless bacteria detection on tooth enamel. *Nat. Commun.* **2012**, *3*, No. 763.
- (226) Lim, H.-R.; Lee, S. M.; Park, S.; et al. Smart bioelectronic pacifier for real-time continuous monitoring of salivary electrolytes. *Biosens. Bioelectron.* **2022**, *210*, No. 114329.
- (227) García-Carmona, L.; Martín, A.; Sempionatto, J. R.; et al. Pacifier Biosensor: Toward Noninvasive Saliva Biomarker Monitoring. *Anal. Chem.* **2019**, *91*, 13883–13891.
- (228) Ma, H.; Khazaei Nejad, S.; Vargas Ramos, D.; et al. Lab-on-a-lollipop (LoL) platform for preventing food-induced toxicity: all-in-one system for saliva sampling and electrochemical detection of vanillin. *Lab Chip* **2024**, *24*, 4306–4320.
- (229) Constantin, V.; Luchian, I.; Goriuc, A.; et al. Salivary Biomarkers Identification: Advances in Standard and Emerging Technologies. *Oral* **2025**, *5*, 26.
- (230) Ichikawa, K.; Iitani, K.; Kawase, G.; et al. Mouthguard-Type Wearable Sensor for Monitoring Salivary Turbidity to Assess Oral Hygiene. *Sensors* **2024**, *24*, 1436.
- (231) de Castro, L. F.; de Freitas, S. V.; Duarte, L. C.; et al. Salivary diagnostics on paper microfluidic devices and their use as wearable sensors for glucose monitoring. *Anal. Bioanal. Chem.* **2019**, *411*, 4919–4928.
- (232) Sharma, A.; Hossain, N. I.; Thomas, A.; Sonkusale, S. Saliva-Sensing Dental Floss: An Innovative Tool for Assessing Stress via On-Demand Salivary Cortisol Measurement with Molecularly Imprinted Polymer and Thread Microfluidics Integration. *ACS Appl. Mater. Interfaces* **2025**, *17*, 25083–25096.
- (233) Kim, J.; Kim, M.; Lee, M. S.; et al. Wearable smart sensor systems integrated on soft contact lenses for wireless ocular diagnostics. *Nat. Commun.* **2017**, *8*, No. 14997.
- (234) Brazaca, L. C.; Imamura, A. H.; Blasques, R. V.; et al. The use of biological fluids in microfluidic paper-based analytical devices (μ PADs): Recent advances, challenges and future perspectives. *Biosens. Bioelectron.* **2024**, *246*, No. 115846.
- (235) Park, J.; Kim, J.; Kim, S. Y.; et al. Soft, smart contact lenses with integrations of wireless circuits, glucose sensors, and displays. *Sci. Adv.* **2018**, *4*, No. eaap9841, DOI: 10.1126/sciadv.aap9841.
- (236) Park, W.; Seo, H.; Kim, J.; et al. In-depth correlation analysis between tear glucose and blood glucose using a wireless smart contact lens. *Nat. Commun.* **2024**, *15*, No. 2828.
- (237) Sempionatto, J. R.; Brazaca, L. C.; García-Carmona, L.; et al. Eyeglasses-based tear biosensing system: non-invasive detection of alcohol, vitamins and glucose. *Biosens. Bioelectron.* **2019**, *137*, 161–170.
- (238) Kim, J.; Park, J.; Park, Y. G.; et al. A soft and transparent contact lens for the wireless quantitative monitoring of intraocular pressure. *Nat. Biomed. Eng.* **2021**, *5*, 772–782.
- (239) Yang, C.; Wu, Q.; Liu, J.; et al. Intelligent wireless theranostic contact lens for electrical sensing and regulation of intraocular pressure. *Nat. Commun.* **2022**, *13*, No. 2556.
- (240) Liu, Z.; Wang, G.; Ye, C.; et al. An Ultrasensitive Contact Lens Sensor Based On Self-Assembly Graphene For Continuous Intraocular Pressure Monitoring. *Adv. Funct. Mater.* **2021**, *31* (29), No. 2010991, DOI: 10.1002/adfm.202010991.
- (241) Duan, Z.; Yuan, M.; Liu, Z.; et al. An Ultrasensitive $\text{Ti}_3\text{C}_2\text{T}_x$ MXene-based Soft Contact Lens for Continuous and Nondestructive Intraocular Pressure Monitoring. *Small* **2024**, *20* (24), No. 2309785, DOI: 10.1002/smll.202309785.
- (242) Kim, T. Y.; Mok, J. W.; Hong, S. H.; et al. Wireless theranostic smart contact lens for monitoring and control of intraocular pressure in glaucoma. *Nat. Commun.* **2022**, *13*, No. 6801.
- (243) Gao, Y.; Nguyen, D. T.; Yeo, T.; et al. A flexible multiplexed immunosensor for point-of-care in situ wound monitoring. *Sci. Adv.* **2021**, *7*, No. eabg9614, DOI: 10.1126/sciadv.abg9614.
- (244) Wang, C.; Shirzaei Sani, E.; Shih, C. D.; et al. Wound management materials and technologies from bench to bedside and beyond. *Nat. Rev. Mater.* **2024**, *9*, 550–566.
- (245) Joorabloo, A.; Liu, T. Smart theranostics for wound monitoring and therapy. *Adv. Colloid Interface Sci.* **2024**, *330*, No. 103207.
- (246) Lee, D. U.; Kim, S.-C.; Choi, D. Y.; Jung, W.-K.; Moon, M. J. Basic amino acid-mediated cationic amphiphilic surfaces for antimicrobial pH monitoring sensor with wound healing effects. *Biomater. Res.* **2023**, *27* (1), No. s40824-023-00355-0, DOI: 10.1186/s40824-023-00355-0.
- (247) Tang, N.; Zheng, Y.; Jiang, X.; et al. Wearable Sensors and Systems for Wound Healing-Related pH and Temperature Detection. *Micromachines* **2021**, *12*, 430.
- (248) Rahimi, R.; Brener, U.; Ochoa, M.; Ziaie, B. Flexible and transparent pH monitoring system with NFC communication for wound monitoring applications. In *2017 IEEE 30th International Conference on Micro Electro Mechanical Systems (MEMS)*; IEEE, 2017; pp 125–128 DOI: 10.1109/MEMS.2017.7863356.
- (249) Cho, S.; Ha, J.; Ahn, J.; et al. Wireless, Battery-Free, Optoelectronic Diagnostic Sensor Integrated Colorimetric Dressing for Advanced Wound Care. *Adv. Funct. Mater.* **2024**, *34* (29), No. 2316196, DOI: 10.1002/adfm.202316196.
- (250) Burian, E. A.; Enevold, C.; Karlsmark, T.; Ågren, M. S. A simplified method for monitoring cytokines in wound fluid. *Wound Repair Regen.* **2023**, *31*, 47–55.
- (251) Gohel, M. S.; Windhaber, R. A. J.; Tarlton, J. F.; Whyman, M. R.; Poskitt, K. R. The relationship between cytokine concentrations and wound healing in chronic venous ulceration. *J. Vasc. Surg.* **2008**, *48*, 1272–1277.
- (252) Barbieri, B.; Silva, A.; Morari, J.; et al. Wound fluid sampling methods and analysis of cytokine mRNA expression in ulcers from patients with diabetes mellitus. *Regener. Ther.* **2024**, *26*, 425–431.
- (253) Zhu, Y.; Zhang, J.; Song, J.; et al. A Multifunctional Pro-Healing Zwitterionic Hydrogel for Simultaneous Optical Monitoring of pH and Glucose in Diabetic Wound Treatment. *Adv. Funct. Mater.* **2020**, *30* (6), No. 1905493, DOI: 10.1002/adfm.201905493.
- (254) Zerr MBA, K. J.; Furnary MD, A. P.; Grunkemeier PhD, G. L.; et al. Glucose Control Lowers the Risk of Wound Infection in Diabetics After Open Heart Operations. *Ann. Thorac. Surg.* **1997**, *63*, 356–361.
- (255) Mirani, B.; Hadisi, Z.; Pagan, E.; et al. Smart Dual-Sensor Wound Dressing for Monitoring Cutaneous Wounds. *Adv. Healthcare Mater.* **2023**, *12* (18), No. 2203233, DOI: 10.1002/adhm.202203233.

- (256) McMurry, J. F. Wound Healing with Diabetes Mellitus: Better Glucose Control for Better Wound Healing in Diabetes. *Surg. Clin. North Am.* **1984**, *64*, 769–778.
- (257) Fu, T.; Stupnitskaia, P.; Matoori, S. Next-Generation Diagnostic Wound Dressings for Diabetic Wounds. *ACS Meas. Sci. Au* **2022**, *2*, 377–384.
- (258) Garland, N. T.; Song, J. W.; Ma, T.; et al. A Miniaturized, Battery-Free, Wireless Wound Monitor That Predicts Wound Closure Rate Early. *Adv. Healthcare Mater.* **2023**, *12*, No. 2301280, DOI: 10.1002/adhm.202301280.
- (259) Zheng, X. T.; Yang, Z.; Sutarlie, L.; et al. Battery-free and AI-enabled multiplexed sensor patches for wound monitoring. *Sci. Adv.* **2023**, *9*, No. eadg6670, DOI: 10.1126/sciadv.adg6670.
- (260) Dong, Y.; Wang, Z. ROS-scavenging materials for skin wound healing: advancements and applications. *Front. Bioeng. Biotechnol.* **2023**, *11*, No. 1304835, DOI: 10.3389/fbioe.2023.1304835.
- (261) Arcangeli, D.; Gualandi, I.; Mariani, F.; et al. Smart Bandaid Integrated with Fully Textile OECT for Uric Acid Real-Time Monitoring in Wound Exudate. *ACS Sens.* **2023**, *8*, 1593–1608.
- (262) Dunnill, C.; Patton, T.; Brennan, J.; et al. Reactive oxygen species (ROS) and wound healing: the functional role of ROS and emerging ROS-modulating technologies for augmentation of the healing process. *Int. Wound J.* **2017**, *14*, 89–96.
- (263) Xiong, Z.; Achavananthadith, S.; Lian, S.; et al. A wireless and battery-free wound infection sensor based on DNA hydrogel. *Sci. Adv.* **2021**, *7*, No. eabj1617, DOI: 10.1126/sciadv.abj1617.
- (264) Zeng, Q.; Shi, G.; Zhang, M. Real-Time Monitoring of Wound States via Rationally Engineered Biosensors. *Adv. Sens. Res.* **2024**, *3* (3), No. 2200018, DOI: 10.1002/adsr.202200018.
- (265) Gao, F.; Liu, C.; Zhang, L.; et al. Wearable and flexible electrochemical sensors for sweat analysis: a review. *Microsyst. Nanoeng.* **2023**, *9*, No. 1.
- (266) Mota, F. A. R.; Passos, M. L. C.; Santos, J. L. M.; Saraiva, M. L. M. F. S. Comparative analysis of electrochemical and optical sensors for detection of chronic wounds biomarkers: A review. *Biosens. Bioelectron.* **2024**, *251*, No. 116095.
- (267) Farahani, M.; Shafiee, A. Wound Healing: From Passive to Smart Dressings. *Adv. Healthcare Mater.* **2021**, *10* (16), No. 2100477, DOI: 10.1002/adhm.202100477.
- (268) Shirzaei Sani, E.; Xu, C.; Wang, C.; et al. A stretchable wireless wearable bioelectronic system for multiplexed monitoring and combination treatment of infected chronic wounds. *Sci. Adv.* **2023**, *9*, No. eadf7388, DOI: 10.1126/sciadv.adf7388.
- (269) Meng, L.; Liu, S.; Borsa, B. A.; Eriksson, M.; Mak, W. C. A conducting polymer-based array with multiplex sensing and drug delivery capabilities for smart bandages. *Commun. Mater.* **2024**, *5*, No. 28.
- (270) Xu, G.; Lu, Y.; Cheng, C.; et al. Battery-Free and Wireless Smart Wound Dressing for Wound Infection Monitoring and Electrically Controlled On-Demand Drug Delivery. *Adv. Funct. Mater.* **2021**, *31*, No. 2100852, DOI: 10.1002/adfm.202100852.
- (271) Song, J. W.; Ryu, H.; Bai, W.; et al. Bioresorbable, wireless, and battery-free system for electrotherapy and impedance sensing at wound sites. *Sci. Adv.* **2023**, *9*, No. eade4687, DOI: 10.1126/sciadv.ade4687.
- (272) Ochoa, M.; Rahimi, R.; Zhou, J.; et al. Integrated sensing and delivery of oxygen for next-generation smart wound dressings. *Microsyst. Nanoeng.* **2020**, *6*, No. 46.
- (273) Hu, Y.; Wang, F.; Ye, H.; et al. MXene-based flexible electronic materials for wound infection detection and treatment. *npj Flexible Electron.* **2024**, *8*, No. 30.
- (274) Ha, J.; Kim, J. Y.; Kim, D.; et al. Multifunctional Micro/Nanofiber Based-Dressing Patch with Healing, Protection, and Monitoring Capabilities for Advanced Wound Care. *Adv. Mater. Technol.* **2023**, *8*, No. 2201765, DOI: 10.1002/admt.202201765.
- (275) Li, C.; Chang, F.; Gao, F.; et al. Chitosan-based composite film dressings with efficient self-diagnosis and synergistically inflammation resolution for accelerating diabetic wound healing. *Appl. Surf. Sci.* **2024**, *642*, No. 158578.
- (276) Khadem, E.; Kharaziha, M.; Salehi, S. Colorimetric pH-responsive and hemostatic hydrogel-based bioadhesives containing functionalized silver nanoparticles. *Mater. Today Bio* **2023**, *20*, No. 100650.
- (277) Li, Y.; Hu, Z.; Huo, R.; Cui, Z. Preparation of an indicator film based on pectin, sodium alginate, and xanthan gum containing blueberry anthocyanin extract and its application in blueberry freshness monitoring. *Heliyon* **2023**, *9*, No. e14421.
- (278) Mariani, F.; Serafini, M.; Gualandi, I.; et al. Advanced Wound Dressing for Real-Time pH Monitoring. *ACS Sens.* **2021**, *6*, 2366–2377.
- (279) Yoon, J. H.; Kim, S. M.; Park, H. J.; et al. Highly self-healable and flexible cable-type pH sensors for real-time monitoring of human fluids. *Biosens. Bioelectron.* **2020**, *150*, No. 111946.
- (280) Ha, J.; Ko, J.; Ahn, J.; et al. Nanotransfer Printing of Functional Nanomaterials on Electrospun Fibers for Wearable Healthcare Applications. *Adv. Funct. Mater.* **2024**, *34* (33), No. 2401404, DOI: 10.1002/adfm.202401404.
- (281) Zhang, Y.; Hu, Y.; Montelongo, Y.; et al. A Conformable Holographic Sensing Bandage for Wound Monitoring. *Adv. Funct. Mater.* **2024**, *34* (16), No. 2308490, DOI: 10.1002/adfm.202308490.
- (282) Giovannini, G.; Cinelli, P.; Boesel, L. F.; Rossi, R. M. Thioflavin-modified molecularly imprinted hydrogel for fluorescent-based non-enzymatic glucose detection in wound exudate. *Mater. Today Bio* **2022**, *14*, No. 100258.
- (283) Lopez, A. J.; Jones, L. M.; Reynolds, L.; et al. Detection of bacterial fluorescence from in vivo wound biofilms using a point-of-care fluorescence imaging device. *Int. Wound J.* **2021**, *18*, 626–638.
- (284) Wang, Z.; Ou, X.; Guan, L.; et al. Pomegranate-inspired multifunctional nanocomposite wound dressing for intelligent self-monitoring and promoting diabetic wound healing. *Biosens. Bioelectron.* **2023**, *235*, No. 115386.
- (285) Giovannini, G.; Sharma, K.; Boesel, L. F.; Rossi, R. M. Lab-on-a-Fiber Wearable Multi-Sensor for Monitoring Wound Healing. *Adv. Healthcare Mater.* **2024**, *13* (4), No. 2302603, DOI: 10.1002/adhm.202302603.
- (286) Mishra, A.; Kushare, A.; Gupta, M. N.; Ambre, P. Advanced Dressings for Chronic Wound Management. *ACS Appl. Bio Mater.* **2024**, *7*, 2660–2676.
- (287) Wang, X.; Zhong, B.; Lou, Z.; Han, W.; Wang, L. The advancement of intelligent dressings for monitoring chronic wound infections. *Chem. Eng. J.* **2024**, *484*, 149643.
- (288) Mishra, A.; Kushare, A.; Gupta, M. N.; Ambre, P. Advanced Dressings for Chronic Wound Management. *ACS Appl. Bio Mater.* **2024**, *7*, 2660–2676.
- (289) Huang, S. W.; Wu, Y. F.; Ahmed, T.; Pan, S. C.; Cheng, C. M. Point-of-care detection devices for wound care and monitoring. *Trends Biotechnol.* **2024**, *42*, 74–90.
- (290) Wang, X.; Yang, Y.; Zhao, W.; Zhu, Z.; Pei, X. Recent advances of hydrogels as smart dressings for diabetic wounds. *J. Mater. Chem. B* **2024**, *12*, 1126–1148.
- (291) Wang, X.; Zhong, B.; Lou, Z.; Han, W.; Wang, L. The advancement of intelligent dressings for monitoring chronic wound infections. *Chem. Eng. J.* **2024**, *484*, No. 149643.
- (292) Kaveti, R.; Jakus, M. A.; Chen, H.; et al. Water-powered, electronics-free dressings that electrically stimulate wounds for rapid wound closure. *Sci. Adv.* **2024**, *10*, No. eado7538, DOI: 10.1126/sciadv.ado7538.
- (293) Liu, H.; Li, Y.; Sun, Q.; et al. Triboelectric wearable devices for accelerated wound healing. *Chem. Eng. J.* **2024**, *497*, No. 154628.
- (294) Chen, X.; Peng, Y. Wearable bioelectronic system for wound healing and management. *Biomaterials Translational* **2023**, *4* (2), 65.
- (295) Haghayegh, F.; Norouziyazad, A.; Haghani, E.; et al. Revolutionary Point-of-Care Wearable Diagnostics for Early Disease Detection and Biomarker Discovery through Intelligent Technologies. *Adv. Sci.* **2024**, *11* (36), No. 2400595, DOI: 10.1002/adv.202400595.
- (296) Çelik, H.; Caf, B. B.; Geyik, C.; Çebi, G.; Tayfun, M. Enhancing urinalysis with smartphone and AI: a comprehensive

review of point-of-care urinalysis and nutritional advice. *Chem. Pap.* **2024**, *78*, 651–664.

(297) Shitanda, I.; Fujimura, Y.; Takarada, T.; et al. Self-Powered Diaper Sensor with Wireless Transmitter Powered by Paper-Based Biofuel Cell with Urine Glucose as Fuel. *ACS Sens.* **2021**, *6*, 3409–3415.

(298) Zhang, J.; Liu, J.; Su, H.; et al. A wearable self-powered biosensor system integrated with diaper for detecting the urine glucose of diabetic patients. *Sens. Actuators, B* **2021**, *341*, No. 130046.

(299) Zhang, Z.; Shi, Q.; He, T.; et al. Artificial intelligence of toilet (AI-Toilet) for an integrated health monitoring system (IHMS) using smart triboelectric pressure sensors and image sensor. *Nano Energy* **2021**, *90*, No. 106517.

(300) Park, S.-m.; min.; et al. A mountable toilet system for personalized health monitoring via the analysis of excreta. *Nat. Biomed. Eng.* **2020**, *4*, 624–635.

(301) Bae, J. H.; Lee, H. K. User Health Information Analysis with a Urine and Feces Separable Smart Toilet System. *IEEE Access* **2018**, *6*, 78751–78765.

(302) Han, S.; Yamamoto, S.; Jung, C. Y.; et al. Wearable sensors for monitoring chronic kidney disease. *Commun. Mater.* **2024**, *5*, No. 153.

(303) Hao, L.; Zhao, R. T.; Welch, N. L.; et al. CRISPR-Cas-amplified urinary biomarkers for multiplexed and portable cancer diagnostics. *Nat. Nanotechnol.* **2023**, *18*, 798–807.

(304) Suthat Na Ayutaya, V.; Tantisatirapoon, C.; Aekgawong, S.; et al. Urinary cancer detection by the target urine volatile organic compounds biosensor platform. *Sci. Rep.* **2024**, *14*, No. 3551.

(305) Luo, X.; Tan, H.; Wen, W. Recent Advances in Wearable Healthcare Devices: From Material to Application. *Bioengineering* **2024**, *11*, 358.

(306) Lin, J. H.; Shiu, B. C.; Lou, C. W.; Chang, Y. J. Design and Fabrication of Smart Diapers with Antibacterial Yarn. *J. Healthcare Eng.* **2017**, *2017*, No. 8046134, DOI: [10.1155/2017/8046134](https://doi.org/10.1155/2017/8046134).

(307) Ates, H. C.; Brunauer, A.; von Stetten, F.; et al. Integrated Devices for Non-Invasive Diagnostics. *Adv. Funct. Mater.* **2021**, *31*, No. 2010388.

(308) Momtazmanesh, S.; Moghaddam, S. S.; Ghamari, S. H.; et al. Global burden of chronic respiratory diseases and risk factors, 1990–2019: an update from the Global Burden of Disease Study 2019. *EClinicalMedicine* **2023**, *59*, No. 101936.

(309) Ebrahimi, H.; Aryan, Z.; Saeedi Moghaddam, S.; et al. Global, regional, and national burden of respiratory tract cancers and associated risk factors from 1990 to 2019: a systematic analysis for the Global Burden of Disease Study 2019. *Lancet Respir. Med.* **2021**, *9*, 1030–1049.

(310) Tassopoulos, C. N.; Barnett, D.; Russell Fraser, T. BREATH-ACETONE AND BLOOD-SUGAR MEASUREMENTS IN DIABETES. *Lancet* **1969**, *293*, 1282–1286.

(311) Wang, C.; Mbi, A.; Shepherd, M. A Study on Breath Acetone in Diabetic Patients Using a Cavity Ringdown Breath Analyzer: Exploring Correlations of Breath Acetone With Blood Glucose and Glycohemoglobin A1C. *IEEE Sens. J.* **2010**, *10*, 54–63.

(312) Narasimhan, L. R.; Goodman, W.; Patel, C. K. N. Correlation of breath ammonia with blood urea nitrogen and creatinine during hemodialysis. *Proc. Natl. Acad. Sci. U.S.A.* **2001**, *98*, 4617–4621.

(313) Dadamio, J.; Van den Velde, S.; Laleman, W.; et al. Breath biomarkers of liver cirrhosis. *J. Chromatogr. B* **2012**, *905*, 17–22.

(314) Johnson, G. R.; Morawska, L. The Mechanism of Breath Aerosol Formation. *J. Aerosol Med. Pulm. Drug Delivery* **2009**, *22*, 229–237.

(315) Vasilescu, A.; Hrincenko, B.; Swain, G. M.; Petcu, S. F. Exhaled breath biomarker sensing. *Biosens. Bioelectron.* **2021**, *182*, No. 113193.

(316) Heng, W.; Yin, S.; Chen, Y.; Gao, W. Exhaled Breath Analysis: from Laboratory Test to Wearable Sensing. *IEEE Rev. Biomed. Eng.* **2024**, 1–25.

(317) Li, X.; Sun, R.; Pan, J.; et al. Rapid and on-site wireless immunoassay of respiratory virus aerosols via hydrogel-modulated resonators. *Nat. Commun.* **2024**, *15*, No. 4035.

(318) Li, H.; Gong, H.; Wong, T. H.; et al. Wireless, battery-free, multifunctional integrated bioelectronics for respiratory pathogens monitoring and severity evaluation. *Nat. Commun.* **2023**, *14*, No. 7539.

(319) Maier, D.; Laubender, E.; Basavanna, A.; et al. Toward Continuous Monitoring of Breath Biochemistry: A Paper-Based Wearable Sensor for Real-Time Hydrogen Peroxide Measurement in Simulated Breath. *ACS Sens.* **2019**, *4*, 2945–2951.

(320) Adrover-Jaume, C.; Alba-Patiño, A.; Clemente, A.; et al. Paper biosensors for detecting elevated IL-6 levels in blood and respiratory samples from COVID-19 patients. *Sens. Actuators, B* **2021**, *330*, No. 129333.

(321) Luo, Y.; Li, J.; Ding, Q.; et al. Functionalized Hydrogel-Based Wearable Gas and Humidity Sensors. *Nano-Micro Lett.* **2023**, *15*, No. 136.

(322) Wu, Z.; Rong, L.; Yang, J.; et al. Ion-Conductive Hydrogel-Based Stretchable, Self-Healing, and Transparent NO₂ Sensor with High Sensitivity and Selectivity at Room Temperature. *Small* **2021**, *17* (52), No. 2104997, DOI: [10.1002/sml.202104997](https://doi.org/10.1002/sml.202104997).

(323) Liang, Y.; Wu, Z.; Wei, Y.; et al. Self-Healing, Self-Adhesive and Stable Organohydrogel-Based Stretchable Oxygen Sensor with High Performance at Room Temperature. *Nano-Micro Lett.* **2022**, *14*, No. 52.

(324) Wu, Z.; Ding, Q.; Wang, H.; et al. A Humidity-Resistant, Sensitive, and Stretchable Hydrogel-Based Oxygen Sensor for Wireless Health and Environmental Monitoring. *Adv. Funct. Mater.* **2024**, *34* (6), No. 2308280, DOI: [10.1002/adfm.202308280](https://doi.org/10.1002/adfm.202308280).

(325) Wu, J.; Wu, Z.; Han, S.; et al. Extremely Deformable, Transparent, and High-Performance Gas Sensor Based on Ionic Conductive Hydrogel. *ACS Appl. Mater. Interfaces* **2019**, *11*, 2364–2373.

(326) Ding, Q.; Zhou, Z.; Wang, H.; et al. Self-healable, recyclable, ultrastretchable, and high-performance NO₂ sensors based on an organohydrogel for room and sub-zero temperature and wireless operation. *SmartMat* **2023**, *4* (1), No. e1141, DOI: [10.1002/smm2.1141](https://doi.org/10.1002/smm2.1141).

(327) Wei, Y.; Wang, H.; Ding, Q.; et al. Hydrogel- and organohydrogel-based stretchable, ultrasensitive, transparent, room-temperature and real-time NO₂ sensors and the mechanism. *Mater. Horiz.* **2022**, *9*, 1921–1934.

(328) Liu, L.; Fei, T.; Guan, X.; Zhao, H.; Zhang, T. Humidity-activated ammonia sensor with excellent selectivity for exhaled breath analysis. *Sens. Actuators, B* **2021**, *334*, No. 129625.

(329) Liu, L.; Fei, T.; Guan, X.; et al. Room temperature ammonia gas sensor based on ionic conductive biomass hydrogels. *Sens. Actuators, B* **2020**, *320*, No. 128318.

(330) Wu, Z.; Wang, H.; Ding, Q.; et al. A Self-Powered, Rechargeable, and Wearable Hydrogel Patch for Wireless Gas Detection with Extraordinary Performance. *Adv. Funct. Mater.* **2023**, *33* (21), No. 2300046, DOI: [10.1002/adfm.202300046](https://doi.org/10.1002/adfm.202300046).

(331) Huang, W.; Ding, Q.; Wang, H.; et al. Design of stretchable and self-powered sensing device for portable and remote trace biomarkers detection. *Nat. Commun.* **2023**, *14*, No. 5221.

(332) Puckett, J. L.; George, S. C. Partitioned exhaled nitric oxide to non-invasively assess asthma. *Respir. Physiol. Neurobiol.* **2008**, *163*, 166–177.

(333) Jiménez, D. D.; Vinckenbosch, F.; Busink, J.; et al. Ex vivo detection of recreationally consumed nitrous oxide in exhaled breath. *Sci. Rep.* **2025**, *15*, No. 2901.

(334) Ates, H. C.; Dincer, C. Wearable breath analysis. *Nat. Rev. Bioeng.* **2023**, *1*, 80–82.

(335) Rivnay, J.; Raman, R.; Robinson, J. T.; et al. Integrating bioelectronics with cell-based synthetic biology. *Nat. Rev. Bioeng.* **2025**, *3* (4), 317–332, DOI: [10.1038/s44222-024-00262-6](https://doi.org/10.1038/s44222-024-00262-6).

(336) Simon, M. J.; Iliff, J. J. Regulation of cerebrospinal fluid (CSF) flow in neurodegenerative, neurovascular and neuroinflammatory disease. *Biochim. Biophys. Acta, Mol. Basis Dis.* **2016**, *1862*, 442–451.

- (337) Czarniak, N.; Kamińska, J.; Matowicka-Karna, J.; Koper-Lenkiewicz, O. Cerebrospinal Fluid—Basic Concepts Review. *Biomedicines* **2023**, *11*, 1461.
- (338) Lyu, H.; Jian, Z.; Liu, X.; Sun, Y.; Babakhani, A. Towards the Implementation of a Wirelessly Powered Dielectric Sensor With Digitized Output for Implantable Applications. *IEEE Sens. Lett.* **2019**, *3* (3), 1–4, DOI: 10.1109/lsens.2019.2896185.
- (339) Garrett, A.; Soler, G. J.; Diluna, M. L.; et al. A passive, biocompatible microfluidic flow sensor to assess flows in a cerebral spinal fluid shunt. *Sens. Actuators, A* **2020**, *312*, No. 112110, DOI: 10.1016/j.sna.2020.112110.
- (340) Krause, K.; Wulf, M.; Sommer, P.; et al. CSF Diagnostics: A Potentially Valuable Tool in Neurodegenerative and Inflammatory Disorders Involving Motor Neurons: A Review. *Diagnostics* **2021**, *11*, 1522.
- (341) Hatami-Fard, G.; Anastasova-Ivanova, S. Advancements in Cerebrospinal Fluid Biosensors: Bridging the Gap from Early Diagnosis to the Detection of Rare Diseases. *Sensors* **2024**, *24*, 3294.
- (342) Meldolesi, J. News about the Role of Fluid and Imaging Biomarkers in Neurodegenerative Diseases. *Biomedicines* **2021**, *9*, 252.
- (343) Vernau, W.; Vernau, K. A.; Sue Bailey, C. Cerebrospinal Fluid. In *Clinical Biochemistry of Domestic Animals*; Elsevier, 2008; pp 769–819 DOI: 10.1016/B978-0-12-370491-7.00026-X.
- (344) Ahmed, S. V.; Jayawarna, C.; Jude, E. Post lumbar puncture headache: diagnosis and management. *Postgrad. Med. J.* **2006**, *82*, 713–716.
- (345) Li, J.; Llano, D. A.; Ellis, T.; et al. Effect of human cerebrospinal fluid sampling frequency on amyloid- β levels. *Alzheimer's Dementia* **2012**, *8*, 295–303.
- (346) Nográdi, A.; Vrbová, G. Anatomy and Physiology of the Spinal Cord. In *Transplantation of Neural Tissue into the Spinal Cord*; Springer, 2006 DOI: 10.1007/0-387-32633-2_1.
- (347) Menéndez González, M. CNS Compartments: The Anatomy and Physiology of the Cerebrospinal Fluid. In *Liquorpheresis*; Springer Nature, 2023 DOI: 10.1007/978-3-031-43482-2_1.
- (348) KOENIG, M. Cerebrospinal Fluid Examination, Monitoring, and Diversion Techniques. In *Cerebrospinal Fluid in Clinical Practice*; Elsevier, 2009 DOI: 10.1016/b978-141602908-3.50010-8.
- (349) Wright, B. L. C.; Lai, J. T. F.; Sinclair, A. J. Cerebrospinal fluid and lumbar puncture: a practical review. *J. Neurol.* **2012**, *259* (8), 1530–1545, DOI: 10.1007/s00415-012-6413-x.
- (350) de Oliveira, F. C. M.; Serrano, S. H. P. Electrochemically Active L-Cysteine Gold Modified Electrodes. *Electrochim. Acta* **2014**, *125*, 566–572, DOI: 10.1016/j.electacta.2014.01.079.
- (351) Kim, K.; Kwak, J. Faradaic impedance titration of pure 3-mercaptopropionic acid and ethanethiol mixed monolayers on gold. *J. Electroanal. Chem.* **2001**, *512* (1–2), 83–91, DOI: 10.1016/S0022-0728(01)00588-5.
- (352) Lobo Maza, F.; Méndez De Leo, L.; Rubert, A. A.; et al. New Insight into the Interface Chemistry and Stability of Glutathione Self-Assembled Monolayers on Au(111). *J. Phys. Chem. C* **2016**, *120* (27), 14597–14607, DOI: 10.1021/acs.jpcc.5b12643.
- (353) Qian, Y.; Zhang, L.; Tian, Y. Highly Stable Electrochemical Probe with Bidentate Thiols for Ratiometric Monitoring of Endogenous Polysulfide in Living Mouse Brains. *Anal. Chem.* **2021**, *94* (2), 1447–1455, DOI: 10.1021/acs.analchem.1c04894.
- (354) Zhang, F.; Li, X.; Xu, X.; Cai, D.; Zhang, J. Relationship Between the pH of Enema Solutions and Intestinal Damage in Rabbits. *Biol. Res. Nurs.* **2014**, *17* (1), 78–86, DOI: 10.1177/1099800414527154.
- (355) Huang, W.-C.; Hung, C. H.; Lin, Y. W.; et al. Electrically Copolymerized Polydopamine Melanin/Poly(3,4-ethylenedioxythiophene) Applied for Bioactive Multimodal Neural Interfaces with Induced Pluripotent Stem Cell-Derived Neurons. *ACS Biomater. Sci. Eng.* **2022**, *8* (11), 4807–4818, DOI: 10.1021/acsbomaterials.2c00822.
- (356) Achache, M.; García-Guzmán, J. J.; Seddik, N. B.; et al. A novel sonogel-Leucine electrochemical sensor for the detection of neuroblastoma biomarker in human urine and synthetic cerebrospinal fluid. *Talanta* **2025**, *286*, No. 127452, DOI: 10.1016/j.talanta.2024.127452.
- (357) Williams, C. M.; Greer, M. Homovanillic Acid and Vanilmandelic Acid in Diagnosis of Neuroblastoma. *JAMA* **1963**, *183* (10), 836–840, DOI: 10.1001/jama.1963.03700100062010.
- (358) Stuart, T.; et al. Wireless, Battery-Free Implants for Electrochemical Catecholamine Sensing and Optogenetic Stimulation. *ACS Nano* **2022**, *17* (1), 561–574, DOI: 10.1021/acsnano.2c09475.
- (359) Bezprozvanny, I. Calcium signaling and neurodegenerative diseases. *Trends Mol. Med.* **2009**, *15* (3), 89–100, DOI: 10.1016/j.molmed.2009.01.001.
- (360) Pikor, D.; Hurla, M.; Slowikowski, B.; et al. Calcium Ions in the Physiology and Pathology of the Central Nervous System. *Int. J. Mol. Sci.* **2024**, *25*, No. 13133, DOI: 10.3390/ijms252313133.
- (361) Zhai, J.; Zhang, Y.; Zhao, D.; Kou, L.; Zhao, G. In vivo monitoring of calcium ions in rat cerebrospinal fluid using an all-solid-state acupuncture needle based potentiometric microelectrode. *Anal. Chim. Acta* **2022**, *1191*, No. 339209, DOI: 10.1016/j.aca.2021.339209.
- (362) Adil, O.; Adeyeye, C.; Shamsi, M. H. Electrografted Laser-Induced Graphene: Direct Detection of Neurodegenerative Disease Biomarker in Cerebrospinal Fluid. *ACS Sens.* **2024**, *9* (9), 4748–4757, DOI: 10.1021/acssensors.4c01150.
- (363) Simrén, J.; Weninger, H.; Brum, W. S.; et al. Differences between blood and cerebrospinal fluid glial fibrillary acidic protein levels: The effect of sample stability. *Alzheimer's Dementia* **2022**, *18* (12), 1988–1992, DOI: 10.1002/alz.12806.
- (364) García-Rodrigo, L.; Ramos-López, C.; Sánchez-Tirado, E.; et al. Label-free electrochemical immunosensing of glial fibrillary acidic protein (GFAP) at synthesized rGO/MoS₂/AgNPs nanocomposite. Application to the determination in human cerebrospinal fluid. *Talanta* **2024**, *270*, No. 125597, DOI: 10.1016/j.talanta.2023.125597.
- (365) Orzari, L. O.; Silva, L. R. G. E.; de Freitas, R. C.; Brazaca, L. C.; Janegitz, B. C. Lab-made disposable screen-printed electrochemical sensors and immunosensors modified with Pd nanoparticles for Parkinson's disease diagnostics. *Microchim. Acta* **2024**, *191* (1), 76 DOI: 10.1007/s00604-023-06158-3.
- (366) Manoufali, M.; Mobashsher, A. T.; Mohammed, B.; et al. Implantable Sensor for Detecting Changes in the Loss Tangent of Cerebrospinal Fluid. *IEEE Trans. Biomed. Circuits Syst.* **2020**, *14*, 452–462.
- (367) Zarrin, D. A.; Jafari, M.; Kim, W.; Colby, G. P. A novel ventriculoperitoneal shunt flow sensor based on electrically induced spatial variation in cerebrospinal fluid charge density. *Front. Bioeng. Biotechnol.* **2024**, *11*, No. 1339831, DOI: 10.3389/fbioe.2023.1339831.
- (368) Lichtenberg, J. Y.; Ling, Y.; Kim, S. Non-Specific Adsorption Reduction Methods in Biosensing. *Sensors* **2019**, *19* (11), 2488 DOI: 10.3390/s19112488.
- (369) Szunerits, S.; Pagneux, Q.; M'Barek, Y. Ben.; Vassal, S.; Boukherroub, R. Do not let electrode fouling be the enemy of bioanalysis. *Bioelectrochemistry* **2023**, *153*, No. 108479, DOI: 10.1016/j.bioelechem.2023.108479.
- (370) Lowe, S.; O'Brien-Simpson, N. M.; Connal, L. A. Antibiofouling polymer interfaces: poly(ethylene glycol) and other promising candidates. *Polym. Chem.* **2015**, *6* (2), 198–212, DOI: 10.1039/c4py01356e.
- (371) Wen, J.; Huang, S.; Hu, Q.; et al. Recent advances in zwitterionic polymers-based non-fouling coating strategies for biomedical applications. *Mater. Today Chem.* **2024**, *40*, No. 102232, DOI: 10.1016/j.mtchem.2024.102232.
- (372) Wellens, J.; Deschaume, O.; Putzeys, T.; et al. Sulfobetaine-based ultrathin coatings as effective antifouling layers for implantable neuroprosthetic devices. *Biosens. Bioelectron.* **2023**, *226*, No. 115121, DOI: 10.1016/j.bios.2023.115121.
- (373) Lin, Y.-T.; Ting, Y.-S.; Chen, B.-Y.; Cheng, Y.-W.; Liu, T.-Y. Bionic shark skin replica and zwitterionic polymer brushes functionalized PDMS membrane for anti-fouling and wound dressing

applications. *Surf. Coat. Technol.* **2020**, *391*, No. 125663, DOI: 10.1016/j.surfcoat.2020.125663.

(374) Latip, E. N. A.; Coudron, L.; McDonnell, M. B.; et al. Protein droplet actuation on superhydrophobic surfaces: a new approach toward anti-biofouling electrowetting systems. *RSC Adv.* **2017**, *7* (78), 49633–49648, DOI: 10.1039/c7ra10920b.

(375) Geng, H.; Cho, S. K. Anti-biofouling droplet manipulation by slippery liquid infused porous surface (SLIPS) integrated with electrowetting and liquid-dielectrophoresis. In *2018 IEEE Micro Electro Mechanical Systems (MEMS)*; IEEE, 2018 DOI: 10.1109/memsys.2018.8346534.

(376) Lecomte, A.; Descamps, E.; Bergaud, C. A review on mechanical considerations for chronically-implanted neural probes. *J. Neural Eng.* **2018**, *15* (3), No. 031001, DOI: 10.1088/1741-2552/aa8b4f.

(377) Jiang, Z.; Zhu, M.; Chen, X. Interfacing Neuron-Motor Pathways with Stretchable and Biocompatible Electrode Arrays. *Acc. Chem. Res.* **2024**, *57* (16), 2255–2266, DOI: 10.1021/acs.accounts.4c00215.

(378) Jeong, J.; Chou, N.; Kim, S. Long-term characterization of neural electrodes based on parylene-caulked polydimethylsiloxane substrate. *Biomed. Microdevices* **2016**, *18* (3), 42 DOI: 10.1007/s10544-016-0065-z.

(379) Barshutina, M. N.; Kirichenko, S.; Wodolajsky, V.; et al. PDMS-CNT composite for soft bioelectronic neuronal implants. *Composites, Part B* **2022**, *247*, No. 110286, DOI: 10.1016/j.compositesb.2022.110286.

(380) Cheng, M.-Y.; et al. A flexible polyimide cable for implantable neural probe arrays. *Microsyst. Technol.* **2012**, *19* (8), 1111–1118, DOI: 10.1007/s00542-012-1707-6.

(381) Castagnola, V.; Descamps, E.; Lecestre, A.; et al. Parylene-based flexible neural probes with PEDOT coated surface for brain stimulation and recording. *Biosens. Bioelectron.* **2015**, *67*, 450–457, DOI: 10.1016/j.bios.2014.09.004.

(382) Malekshoaraie, M. H.; Wu, B.; Krahe, D. D.; et al. Fully flexible implantable neural probes for electrophysiology recording and controlled neurochemical modulation. *Microsyst. Nanoeng.* **2024**, *10* (1), 91 DOI: 10.1038/s41378-024-00685-6.

(383) Liu, X.; Xu, Z.; Fu, X.; et al. Stable, long-term single-neuronal recording from the rat spinal cord with flexible carbon nanotube fiber electrodes. *J. Neural Eng.* **2022**, *19*, No. 056024, DOI: 10.1088/1741-2552/ac9258.

(384) Gao, L. et al. Free-Standing Carbon Nanotube Embroidered Graphene Film Electrode Array for Stable Neural Interfacing. *Nano Lett.* **2023** DOI: 10.1021/acs.nanolett.3c03421.

(385) Lienemann, S.; Zötterman, J.; Farnebo, S.; Tybrandt, K. Stretchable gold nanowire-based cuff electrodes for low-voltage peripheral nerve stimulation. *J. Neural Eng.* **2021**, *18*, No. 045007, DOI: 10.1088/1741-2552/abfebb.

(386) Tringides, C. M.; Mooney, D. J. Materials for Implantable Surface Electrode Arrays: Current Status and Future Directions. *Adv. Mater.* **2022**, *34*, No. 2107207, DOI: 10.1002/adma.202107207.

(387) Oldroyd, P.; Hadwe, S. El.; Barone, D. G.; Malliaras, G. G. Thin-film implants for bioelectronic medicine. *MRS Bull.* **2024**, *49* (10), 1045–1058, DOI: 10.1557/s43577-024-00786-7.

(388) Jeerapan, I.; Sempionatto, J. R.; Wang, J. On-Body Bioelectronics: Wearable Biofuel Cells for Bioenergy Harvesting and Self-Powered Biosensing. *Adv. Funct. Mater.* **2019**, *30* (29), No. 1906243, DOI: 10.1002/adfm.201906243.

(389) Stephani, C.; Choi, A. H. K.; Moerer, O. Point-of-care detection of lactate in cerebrospinal fluid. *Intensive Care Med. Exp.* **2021**, *9* (1), 18 DOI: 10.1186/s40635-021-00385-9.

(390) Wang, J.; Chakravarthy, K.; Jeerapan, I. Implantable, biofuel cells for self-charging medical devices. (2019).

(391) Tang, H.; Huang, W.; Yao, Y.-F. The metabolites of lactic acid bacteria: classification, biosynthesis and modulation of gut microbiota. *Microbial Cell* **2023**, *10*, 49–62.

(392) Lebovich, M.; Andrews, L. B. Surveying the Genetic Design Space for Transcription Factor-Based Metabolite Biosensors:

Synthetic Gamma-Aminobutyric Acid and Propionate Biosensors in *E. coli* Nissle 1917. *Front. Bioeng. Biotechnol.* **2022**, *10*, No. 938056, DOI: 10.3389/fbioe.2022.938056.

(393) Serebrinsky-Duek, K.; Barra, M.; Danino, T.; Garrido, D. Engineered Bacteria for Short-Chain-Fatty-Acid-Repressed Expression of Biotherapeutic Molecules. *Microbiol. Spectr.* **2023**, *11*, No. e00049-23, DOI: 10.1128/spectrum.00049-23.

(394) Forner, E.; Ezenarro, J.; Pérez-Montero, M.; et al. Electrochemical biosensor for aerobic acetate detection. *Talanta* **2023**, *265*, No. 124882.

(395) Panyod, S.; Wu, W. K.; Chen, C. C.; et al. Modulation of gut microbiota by foods and herbs to prevent cardiovascular diseases. *J. Tradit. Complement. Med.* **2023**, *13*, 107–118.

(396) Song, Z.; Cheng, L.; Liu, Y.; et al. Plant-derived bioactive components regulate gut microbiota to prevent depression and depressive-related neurodegenerative diseases: Focus on neurotransmitters. *Trends Food Sci. Technol.* **2022**, *129*, 581–590.

(397) Ngashangva, L.; Chattopadhyay, S. Biosensors for point-of-care testing and personalized monitoring of gastrointestinal microbiota. *Front. Microbiol.* **2023**, *14*, No. 1114707, DOI: 10.3389/fmicb.2023.1114707.

(398) Fry, C. Secretions of the salivary glands and stomach. *Surgery* **2009**, *27* (12), 503–506, DOI: 10.1016/j.mpsur.2009.10.014.

(399) Engevik, A. C.; Kaji, I.; Goldenring, J. R. The Physiology of the Gastric Parietal Cell. *Physiol. Rev.* **2020**, *100* (2), 573–602, DOI: 10.1152/physrev.00016.2019.

(400) Cannon, W. B. The nature of gastric peristalsis. *Am. J. Physiol.-Legacy Content* **1911**, *29* (2), 250–266, DOI: 10.1152/ajplegacy.1911.29.2.250.

(401) Chen, Z.; Xie, M.; Zhao, F.; Han, S. Application of Nanomaterial Modified Aptamer-Based Electrochemical Sensor in Detection of Heavy Metal Ions. *Foods* **2022**, *11*, 1404.

(402) Diaz-Amaya, S.; Lin, L.-K.; Deering, A. J.; Stanciu, L. A. Aptamer-based SERS biosensor for whole cell analytical detection of *E. coli* O157:H7. *Anal. Chim. Acta* **2019**, *1081*, 146–156.

(403) Ruiz-Valdepeñas Montiel, V.; Sempionatto, J. R.; Campuzano, S.; et al. Direct electrochemical biosensing in gastrointestinal fluids. *Anal. Bioanal. Chem.* **2019**, *411*, 4597–4604.

(404) Mimee, M.; Nadeau, P.; Hayward, A.; et al. An ingestible bacterial-electronic system to monitor gastrointestinal health. *Science* **2018**, *360*, 915–918.

(405) FIDDIAN-GREEN, R. G. Gastric intramucosal pH, tissue oxygenation and acid-base balance. *Br. J. Anaesth.* **1995**, *74* (5), 591–606, DOI: 10.1093/bja/74.5.591.

(406) Walshaw, J.; Hugh, K.; Helliwell, J.; Burke, J.; Jayne, D. Perianastomotic pH Monitoring for Early Detection of Anastomotic Leaks in Gastrointestinal Surgery: A Systematic Review of the Literature. *Surg. Innovation* **2025**, *32* (2), 180–195, DOI: 10.1177/15533506241313168.

(407) Cook, G. C. Infective gastroenteritis and its relationship to reduced gastric acidity. *Scand. J. Gastroenterol.* **1985**, *20*, 17–22, DOI: 10.3109/00365528509093751.

(408) Sipponen, P.; Kekki, M.; Seppala, K.; Siurala, M. The relationships between chronic gastritis and gastric acid secretion. *Aliment. Pharmacol. Ther.* **1996**, *10*, 103–118, DOI: 10.1046/j.1365-2036.1996.2164011.x.

(409) Li, S.; Lu, D.; Li, S.; et al. Bioresorbable, wireless, passive sensors for continuous pH measurements and early detection of gastric leakage. *Sci. Adv.* **2024**, *10* (16), No. ead0268, DOI: 10.1126/sciadv.ad0268.

(410) Ma, X.; Yu, J.; Wei, L.; et al. Electrochemical sensor based on N-CQDs/AgNPs/ β -CD nanomaterials: Application to simultaneous selective determination of Fe(II) and Fe(III) ions released from iron supplement in simulated gastric fluid. *Talanta* **2023**, *253*, No. 123959, DOI: 10.1016/j.talanta.2022.123959.

(411) Stine, J. M.; et al. Miniaturized Capsule System Toward Real-Time Electrochemical Detection of H₂S in the Gastrointestinal Tract. *Adv. Healthcare Mater.* **2023**, *13* (5), No. 2302897, DOI: 10.1002/adhm.202302897.

- (412) Durairaj, S.; Guo, Q.; Wang, Q.; Chen, A. Sensitive electrochemical detection of metabisulphite in gastrointestinal fluids. *Analyst* **2022**, *147* (23), 5508–5517, DOI: 10.1039/d2an01352e.
- (413) Hollander, F. The Significance of Sodium and Potassium in Gastric Secretion: A review of the problem. *Gastroenterology* **1961**, *40* (4), 477–490, DOI: 10.1016/s0016-5085(61)80130-3.
- (414) Asci, C.; Sharma, A.; Del-Rio-Ruiz, R.; Sonkusale, S. Ingestible pH sensing device for gastrointestinal health monitoring based on thread-based electrochemical sensors. *Microchim. Acta* **2023**, *190*, No. 385, DOI: 10.1007/s00604-023-05946-1.
- (415) Thangaraju, P.; Varthya, S. B. ISO 10993: Biological Evaluation of Medical Devices. In *Medical Device Guidelines and Regulations Handbook*; Springer International Publishing, 2022 DOI: 10.1007/978-3-030-91855-2_11.
- (416) Ozcelik, B.; et al. Poly(ethylene glycol)-Based Coatings Combining Low-Biofouling and Quorum-Sensing Inhibiting Properties to Reduce Bacterial Colonization. *ACS Biomater. Sci. Eng.* **2016**, *3* (1), 78–87, DOI: 10.1021/acsbiomaterials.6b00579.
- (417) He, M.; Gao, K.; Zhou, L.; et al. Zwitterionic materials for antifouling membrane surface construction. *Acta Biomater.* **2016**, *40*, 142–152, DOI: 10.1016/j.actbio.2016.03.038.
- (418) Chan, D.; Chien, J.; Axpe, E.; et al. Combinatorial Polyacrylamide Hydrogels for Preventing Biofouling on Implantable Biosensors. *Adv. Mater.* **2022**, *34*, No. 2109764, DOI: 10.1002/adma.202109764.
- (419) Jaggessar, A.; Shahali, H.; Mathew, A.; Yarlagadda, P. K. D. V. Bio-mimicking nano and micro-structured surface fabrication for antibacterial properties in medical implants. *J. Nanobiotechnol.* **2017**, *15* (1), No. 64, DOI: 10.1186/s12951-017-0306-1.
- (420) Liu, Y.; He, X.; Yuan, C.; Cao, P.; Bai, X. Antifouling applications and fabrications of biomimetic micro-structured surfaces: A review. *J. Adv. Res.* **2024**, *59*, 201–221, DOI: 10.1016/j.jare.2023.08.019.
- (421) Nan, K.; Wong, K.; Li, D.; et al. An ingestible, battery-free, tissue-adhering robotic interface for non-invasive and chronic electrostimulation of the gut. *Nat. Commun.* **2024**, *15* (1), No. 6749, DOI: 10.1038/s41467-024-51102-5.
- (422) Islam, M. N.; Yuce, M. R. Review of Medical Implant Communication System (MICS) band and network. *ICT Express* **2016**, *2* (4), 188–194, DOI: 10.1016/j.icte.2016.08.010.
- (423) Tosi, J.; Taffoni, F.; Santacatterina, M.; Sannino, R.; Formica, D. Performance Evaluation of Bluetooth Low Energy: A Systematic Review. *Sensors* **2017**, *17* (12), 2898 DOI: 10.3390/s17122898.
- (424) Kang, S.-G.; Song, M.-S.; Kim, J.-W.; Lee, J. W.; Kim, J. Near-Field Communication in Biomedical Applications. *Sensors* **2021**, *21* (3), 703 DOI: 10.3390/s21030703.
- (425) Xu, J.; Xu, B.; Yue, H.; et al. Origami-Inspired Bionic Soft Robot Stomach with Self-Powered Sensing. *Adv. Healthcare Mater.* **2024**, *13* (22), No. 2302761, DOI: 10.1002/adhm.202302761.
- (426) Kim, Y.; Prausnitz, M. R. Sensitive sensing of biomarkers in interstitial fluid. *Nat. Biomed. Eng.* **2021**, *5*, 3–5.
- (427) Friedel, M.; Thompson, I. A. P.; Kasting, G.; et al. Opportunities and challenges in the diagnostic utility of dermal interstitial fluid. *Nat. Biomed. Eng.* **2023**, *7*, 1541–1555.
- (428) Heikenfeld, J.; Jajack, A.; Feldman, B.; et al. Accessing analytes in biofluids for peripheral biochemical monitoring. *Nat. Biotechnol.* **2019**, *37*, 407–419.
- (429) Samant, P. P.; Niedzwiecki, M. M.; Raviele, N.; et al. Sampling interstitial fluid from human skin using a microneedle patch. *Sci. Transl. Med.* **2020**, *12* (571), No. eaaw0285, DOI: 10.1126/scitranslmed.aaw0285.
- (430) Frank, R.; Hargreaves, R. Clinical biomarkers in drug discovery and development. *Nat. Rev. Drug Discovery* **2003**, *2*, 566–580.
- (431) Wu, Z.; Qiao, Z.; Chen, S.; et al. Interstitial fluid-based wearable biosensors for minimally invasive healthcare and biomedical applications. *Commun. Mater.* **2024**, *5*, No. 33.
- (432) Fox, C. S.; Golden, S. H.; Anderson, C.; et al. Update on Prevention of Cardiovascular Disease in Adults With Type 2 Diabetes Mellitus in Light of Recent Evidence: A Scientific Statement From the American Heart Association and the American Diabetes Association. *Diabetes Care* **2015**, *38*, 1777–1803.
- (433) Cui, Y.; Zhang, L.; Zhang, M.; et al. Prevalence and causes of low vision and blindness in a Chinese population with type 2 diabetes: the Dongguan Eye Study. *Sci. Rep.* **2017**, *7*, No. 11195.
- (434) Klonoff, D. C. Continuous Glucose Monitoring. *Diabetes Care* **2005**, *28*, 1231–1239.
- (435) Yoo, E.-H.; Lee, S.-Y. Glucose Biosensors: An Overview of Use in Clinical Practice. *Sensors* **2010**, *10*, 4558–4576.
- (436) Kim, S.; Malik, J.; Seo, J. M.; Cho, Y. M.; Bien, F. Subcutaneously implantable electromagnetic biosensor system for continuous glucose monitoring. *Sci. Rep.* **2022**, *12*, No. 17395.
- (437) Olczuk, D.; Priefer, R. A history of continuous glucose monitors (CGMs) in self-monitoring of diabetes mellitus. *Diabetes & Metabolic Syndrome: Clinical Research & Reviews* **2018**, *12*, 181–187.
- (438) Saputra, H. A.; Karim, M. M. Enzymatic and Enzyme-Free Electrochemical Lactate Sensors: A Review of the Recent Developments. *Electrochem. Sci. Adv.* **2024**, *5*, No. e202400021.
- (439) Rajagopalan, V.; Venkataraman, S.; Rajendran, D. S.; et al. Acetylcholinesterase biosensors for electrochemical detection of neurotoxic pesticides and acetylcholine neurotransmitter: A literature review. *Environ. Res.* **2023**, *227*, No. 115724.
- (440) Wang, Q.; Yang, C.; Chen, S.; Li, J. Miniaturized Electrochemical Sensing Platforms for Quantitative Monitoring of Glutamate Dynamics in the Central Nervous System. *Angew. Chem., Int. Ed.* **2024**, *63*, No. e202406867.
- (441) Guilbault, G. G.; Montalvo, J. G. A Urea-Specific Enzyme Electrode. *J. Am. Chem. Soc.* **1969**, *91*, 2164–2165.
- (442) Rawson, T. M.; Gowers, S. A. N.; Freeman, D. M. E.; et al. Microneedle biosensors for real-time, minimally invasive drug monitoring of phenoxymethylpenicillin: a first-in-human evaluation in healthy volunteers. *Lancet Digit. Health* **2019**, *1*, e335–e343.
- (443) Vaisocherová, H.; Brynda, E.; Homola, J. Functionalizable low-fouling coatings for label-free biosensing in complex biological media: advances and applications. *Anal. Bioanal. Chem.* **2015**, *407*, 3927–3953.
- (444) Qu, J. H.; Dillen, A.; Saeys, W.; Lammertyn, J.; Spasic, D. Advancements in SPR biosensing technology: An overview of recent trends in smart layers design, multiplexing concepts, continuous monitoring and in vivo sensing. *Anal. Chim. Acta* **2020**, *1104*, 10–27.
- (445) Breault-Turcot, J.; Masson, J. F. Microdialysis SPR: diffusion-gated sensing in blood. *Chem. Sci.* **2015**, *6*, 4247–4254.
- (446) Zhao, C.; Cheung, K. M.; Huang, I. W.; et al. Implantable aptamer-field-effect transistor neuroprobes for in vivo neurotransmitter monitoring. *Sci. Adv.* **2021**, *7*, No. eabj7422.
- (447) Plaxco, K. W.; Soh, H. T. Switch based biosensors: a new approach towards real-time, in vivo molecular detection. *Trends Biotechnol.* **2011**, *29*, 1.
- (448) Vallée-Bélisle, A.; Plaxco, K. W. Structure-switching biosensors: inspired by Nature. *Curr. Opin. Struct. Biol.* **2010**, *20*, 518–526.
- (449) Xiao, Y.; Lubin, A. A.; Heeger, A. J.; et al. Label-Free Electronic Detection of Thrombin in Blood Serum by Using an Aptamer-Based Sensor. *Angew. Chem., Int. Ed.* **2005**, *44*, 5456–5459.
- (450) White, R. J.; Rowe, A. A.; Plaxco, K. W. Re-engineering aptamers to support reagentless, self-reporting electrochemical sensors. *Analyst* **2010**, *135*, 589–594.
- (451) White, R. J.; Plaxco, K. W. Exploiting binding-induced changes in probe flexibility for the optimization of electrochemical biosensors. *Anal. Chem.* **2010**, *82*, 73–76.
- (452) Arroyo-Currás, N.; Ortega, G.; Copp, D. A.; et al. High-Precision Control of Plasma Drug Levels Using Feedback-Controlled Dosing. *ACS Pharmacol. Transl. Sci.* **2018**, *1*, 110–118.
- (453) Verrinder, E.; Leung, K. K.; Erdal, M. K.; Sepunaru, L.; Plaxco, K. W. Comparison of voltammetric methods used in the interrogation of electrochemical aptamer-based sensors. *Sens. Diagnostics* **2024**, *3*, 95–103.

- (454) Roehrich, B.; Leung, K. K.; Gerson, J.; et al. Calibration-Free, Seconds-Resolved In Vivo Molecular Measurements using Fourier-Transform Impedance Spectroscopy Interrogation of Electrochemical Aptamer Sensors. *ACS Sens.* **2023**, *8*, 3051–3059.
- (455) Dauphin-Ducharme, P.; Ploense, K. L.; Arroyo-Currás, N.; Kippin, T. E.; Plaxco, K. W. Electrochemical Aptamer-Based Sensors: A Platform Approach to High-Frequency Molecular Monitoring In Situ in the Living Body. *Methods Mol. Biol.* **2022**, 2393, 479–492.
- (456) Li, H.; Li, S.; Dai, J.; et al. High frequency, calibration-free molecular measurements in situ in the living body. *Chem. Sci.* **2019**, *10*, 10843–10848.
- (457) Idili, A.; Arroyo-Currás, N.; Ploense, K. L.; et al. Seconds-resolved pharmacokinetic measurements of the chemotherapeutic irinotecan in situ in the living body. *Chem. Sci.* **2019**, *10*, 8164–8170.
- (458) Chamorro-García, A.; Gerson, J.; Flatebo, C.; et al. Real-Time, Seconds-Resolved Measurements of Plasma Methotrexate in Situ in the Living Body. *ACS Sens.* **2023**, *8*, 150–157.
- (459) Seo, J. W.; Fu, K.; Correa, S.; et al. Real-time monitoring of drug pharmacokinetics within tumor tissue in live animals. *Sci. Adv.* **2022**, *8*, No. eabk2901.
- (460) Lin, S.; Cheng, X.; Zhu, J.; et al. Wearable microneedle-based electrochemical aptamer biosensing for precision dosing of drugs with narrow therapeutic windows. *Sci. Adv.* **2022**, *8*, No. eabq4539.
- (461) Li, S.; Dai, J.; Zhu, M.; et al. Implantable Hydrogel-Protective DNA Aptamer-Based Sensor Supports Accurate, Continuous Electrochemical Analysis of Drugs at Multiple Sites in Living Rats. *ACS Nano* **2023**, *17*, 18525–18538.
- (462) Shaver, A.; Mahlum, J.; Scida, K.; et al. Optimization of Vancomycin Aptamer Sequence Length Increases the Sensitivity of Electrochemical, Aptamer-Based Sensors In Vivo. *ACS Sens.* **2022**, *7*, 3895–3905.
- (463) Gerson, J.; Erdal, M. K.; Dauphin-Ducharme, P.; et al. A high-precision view of intercompartmental drug transport via simultaneous, seconds-resolved, in situ measurements in the vein and brain. *Br. J. Pharmacol.* **2024**, *181*, 3869–3885.
- (464) Alkhamis, O.; Canoura, J.; Wu, Y.; et al. High-Affinity Aptamers for In Vitro and In Vivo Cocaine Sensing. *J. Am. Chem. Soc.* **2024**, *146*, 3230–3240.
- (465) Plaxco, K. W.; Idili, A.; Gerson, J.; Kippin, T. Seconds-resolved, in situ measurements of plasma phenylalanine disposition kinetics in living rats. *Anal. Chem.* **2021**, *93*, 4023–4032.
- (466) Arroyo-Currás, N.; Dauphin-Ducharme, P.; Ortega, G.; et al. Subsecond-Resolved Molecular Measurements in the Living Body Using Chronoamperometrically Interrogated Aptamer-Based Sensors. *ACS Sens.* **2018**, *3*, 360–366.
- (467) Gerson, J.; Erdal, M. K.; McDonough, M. H.; et al. High-precision monitoring of and feedback control over drug concentrations in the brains of freely moving rats. *Sci. Adv.* **2023**, *9*, No. eadg3254, DOI: 10.1126/sciadv.adg3254.
- (468) Emmons, N. A.; Duman, Z.; Erdal, M. K.; et al. Feedback Control over Plasma Drug Concentrations Achieves Rapid and Accurate Control over Solid-Tissue Drug Concentrations. *ACS Pharmacol. Transl. Sci.* **2025**, *8*, No. 26.
- (469) Friedel, M.; Werbovetz, B.; Drexelius, A.; et al. Continuous molecular monitoring of human dermal interstitial fluid with microneedle-enabled electrochemical aptamer sensors. *Lab Chip* **2023**, *23*, 3289–3299.
- (470) Bakhshandeh, F.; Zheng, H.; Barra, N. G.; et al. Wearable Aptalyzer Integrates Microneedle and Electrochemical Sensing for In Vivo Monitoring of Glucose and Lactate in Live Animals. *Adv. Mater.* **2024**, *36*, No. 2313743.
- (471) Emmons, N. A.; Gibson, J. M.; McDonough, M. H.; et al. Simultaneous, Seconds-Resolved Doxorubicin Measurements in the Blood and Subcutaneous Interstitial Fluid Identify Quantitative Pharmacokinetic Relationships between the Two. *ACS Pharmacol. Transl. Sci.* **2025**, *8* (5), 1347–1358, DOI: 10.1021/acsptsci.5c00062.
- (472) Erdal, M. K.; Kippin, T. E.; Hespanha, J. P.; Plaxco, K. W. Methods of Estimating Plasma Pharmacokinetics From Minimally Invasive, High Temporal Resolution Measurements of Interstitial Fluid Drug Concentrations. *Clin. Transl. Sci.* **2025**, *18* (7), No. e70248, DOI: 10.1111/cts.70248.
- (473) Das, J.; Gomis, S.; Chen, J. B.; et al. Reagentless biomolecular analysis using a molecular pendulum. *Nat. Chem.* **2021**, *13*, 428–434.
- (474) Langer, A.; Hampel, P. A.; Kaiser, W.; et al. Protein analysis by time-resolved measurements with an electro-switchable DNA chip. *Nat. Commun.* **2013**, *4*, No. 2099.
- (475) Plaxco, K. W.; Cash, K. J.; Ricci, F. An electrochemical sensor for the detection of protein-small molecule interactions directly in serum and other complex matrices. *J. Am. Chem. Soc.* **2009**, *131*, 6955–6957.
- (476) Zargartalebi, H.; Mirzaie, S.; GhavamiNejad, A.; et al. Active-reset protein sensors enable continuous in vivo monitoring of inflammation. *Science* **2024**, 386 (6726), 1146–1153, DOI: 10.1126/SCIENCE.ADN2600.
- (477) Yang, Y.; Gao, W. Wearable and flexible electronics for continuous molecular monitoring. *Chem. Soc. Rev.* **2019**, *48*, 1465–1491.
- (478) An, B. W.; Shin, J. H.; Kim, S. Y.; et al. Smart Sensor Systems for Wearable Electronic Devices. *Polymers.* **2017**, *9*, 303.
- (479) Kim, S.; Lee, J.; Chung, W. G.; et al. Three-Dimensional Electrodes of Liquid Metals for Long-Term, Wireless Cardiac Analysis and Modulation. *ACS Nano* **2024**, *18*, 24364–24378.
- (480) Zhao, Y.; Wang, B.; Hojajji, H.; et al. A wearable freestanding electrochemical sensing system. *Sci. Adv.* **2020**, *6*, No. eaaz0007, DOI: 10.1126/sciadv.aaz0007.
- (481) Fan, W.; He, Q.; Meng, K.; et al. Machine-knitted washable sensor array textile for precise epidermal physiological signal monitoring. *Sci. Adv.* **2020**, *6*, No. eaay2840, DOI: 10.1126/sciadv.aay2840.
- (482) Zhang, Y.; Duan, H.; Li, G.; et al. Construction of liquid metal-based soft microfluidic sensors via soft lithography. *J. Nanobiotechnol.* **2022**, *20*, No. 246.
- (483) Yeo, J. C.; Kenry, K.; Lim, C. T. Emergence of microfluidic wearable technologies. *Lab Chip* **2016**, *16*, 4082–4090.
- (484) Jeon, S.; Menard, E.; Park, J.; et al. Three-Dimensional Nanofabrication with Rubber Stamps and Conformable Photomasks. *Adv. Mater.* **2004**, *16*, 1369–1373.
- (485) Niculescu, A.-G.; Chircov, C.; Bircă, A. C.; Grumezescu, A. M. Fabrication and Applications of Microfluidic Devices: A Review. *Int. J. Mol. Sci.* **2021**, *22*, 2011.
- (486) SINGH, A.; MALEK, C. K.; KULKARNI, S. K. DEVELOPMENT IN MICROREACTOR TECHNOLOGY FOR NANO-PARTICLE SYNTHESIS. *Int. J. Nanosci.* **2010**, *09*, 93–112.
- (487) Song, Y.; Min, J.; Yu, Y.; et al. Wireless battery-free wearable sweat sensor powered by human motion. *Sci. Adv.* **2020**, *6*, No. aay9842, DOI: 10.1126/sciadv.aay9842.
- (488) Padash, M.; Enz, C.; Carrara, S. Microfluidics by Additive Manufacturing for Wearable Biosensors: A Review. *Sensors* **2020**, *20*, 4236.
- (489) Song, Y.; Tay, R. Y.; Li, J.; et al. 3D-printed epifluidic electronic skin for machine learning–powered multimodal health surveillance. *Sci. Adv.* **2023**, *9*, No. eadi6492, DOI: 10.1126/sciadv.adi6492.
- (490) Schulz, I. J. Micropuncture studies of the sweat formation in cystic fibrosis patients. *J. Clin. Invest.* **1969**, *48*, 1470–1477.
- (491) Legner, C.; Kalwa, U.; Patel, V.; Chesmore, A.; Pandey, S. Sweat sensing in the smart wearables era: Towards integrative, multifunctional and body-compliant perspiration analysis. *Sens. Actuators, A* **2019**, *296*, 200–221.
- (492) Yang, M.; Sun, N.; Lai, X.; et al. Paper-Based Sandwich-Structured Wearable Sensor with Sebum Filtering for Continuous Detection of Sweat pH. *ACS Sens.* **2023**, *8*, 176–186.
- (493) Yin, Y.; Tan, Z.; Zhu, W.; et al. A wearable microfluidic system for efficient sweat collection and real-time detection. *Talanta* **2024**, *274*, No. 125967.
- (494) Sun, T.; Hui, J.; Lin, B.; et al. Sequential biofluid sampling microfluidic multi-sensing patch for more accurate sweat analysis under sedentary condition. *Appl. Mater. Today* **2023**, *34*, No. 101910.

- (495) Yuan, Q.; Fang, H.; Wu, X.; et al. Self-Adhesive, Biocompatible, Wearable Microfluidics with Erasable Liquid Metal Plasmonic Hotspots for Glucose Detection in Sweat. *ACS Appl. Mater. Interfaces* **2024**, *16*, 66810–66818.
- (496) Zhang, Y.; Zeng, X.; Wang, C.; et al. An integrated wearable microfluidic biosensor for simultaneous detection of multiple biomarkers in sweat. *Talanta* **2025**, *285*, No. 127404.
- (497) Zhang, Y.; Jin, C.; Wang, C.; et al. Fe/Pt dual-atom catalyst-enabled wearable microfluidic patch for superior uric acid detection in sweat. *Biosens. Bioelectron.* **2025**, *271*, No. 117001.
- (498) Hu, M.; Zhu, K.; Wei, J.; et al. Wearable microfluidic SERS patch based on silk fibroin for the non-invasive monitoring of sweat cortisol and pH. *Sens. Actuators, B* **2025**, *427*, No. 137152.
- (499) Garg, M.; Guo, H.; Maclam, E.; et al. Molecularly Imprinted Wearable Sensor with Paper Microfluidics for Real-Time Sweat Biomarker Analysis. *ACS Appl. Mater. Interfaces* **2024**, *16*, 46113–46122.
- (500) Fiore, L.; Mazzaracchio, V.; Antinucci, A.; et al. Wearable electrochemical device based on butterfly-like paper-based microfluidics for pH and Na⁺ monitoring in sweat. *Microchim. Acta* **2024**, *191*, No. 580.
- (501) Yu, Q.; Wu, J.; Xiao, W.; et al. An integrated, wearable paper microfluidic array device for multicomponent electrochemical sweat sensing. *Cell Rep. Phys. Sci.* **2025**, *6*, No. 102389.
- (502) Lorestani, F.; Zhang, X.; Abdullah, A. M.; et al. A Highly Sensitive and Long-Term Stable Wearable Patch for Continuous Analysis of Biomarkers in Sweat. *Adv. Funct. Mater.* **2023**, *33* (52), No. 2306117, DOI: 10.1002/adfm.202306117.
- (503) Kuswandi, B.; Irsyad, L. H.; Puspiningtyas, A. R. Cloth-based microfluidic devices integrated onto the patch as wearable colorimetric sensors for simultaneous sweat analysis. *BioImpacts* **2023**, *13*, 347–353.
- (504) Hanze, M.; Piper, A.; Hamed, M. M. Stitched textile-based microfluidics for wearable devices. *Lab Chip* **2024**, *25*, 28–40.
- (505) Shen, Y.; Chen, G.; Chen, Y.; et al. Wearable microfluidic electrochemical sensor integrated with iontophoresis for non-invasive sweat ketone monitoring. *Sens. Actuators, B* **2024**, *421*, No. 136518.
- (506) Shajari, S.; Salahandish, R.; Zare, A.; et al. MicroSweat: A Wearable Microfluidic Patch for Noninvasive and Reliable Sweat Collection Enables Human Stress Monitoring. *Adv. Sci.* **2023**, *10*, No. 2204171, DOI: 10.1002/advs.202204171.
- (507) Zhang, Q.; Ma, S.; Zhan, X.; et al. Smartphone-based wearable microfluidic electrochemical sensor for on-site monitoring of copper ions in sweat without external driving. *Talanta* **2024**, *266*, No. 125015.
- (508) Ying, Z.; Qiao, L.; Liu, B.; Gao, L.; Zhang, P. Development of a microfluidic wearable electrochemical sensor for the non-invasive monitoring of oxidative stress biomarkers in human sweat. *Biosens. Bioelectron.* **2024**, *261*, No. 116502.
- (509) Mei, X.; Yang, J.; Yu, X.; et al. Wearable molecularly imprinted electrochemical sensor with integrated nanofiber-based microfluidic chip for in situ monitoring of cortisol in sweat. *Sens. Actuators, B* **2023**, *381*, No. 133451.
- (510) Mei, X.; Yang, J.; Liu, J.; Li, Y. Wearable, nanofiber-based microfluidic systems with integrated electrochemical and colorimetric sensing arrays for multiplex sweat analysis. *Chem. Eng. J.* **2023**, *454*, No. 140248.
- (511) Niu, J.; Lin, S.; Chen, D.; et al. A Fully Elastic Wearable Electrochemical Sweat Detection System of Tree-Bionic Microfluidic Structure for Real-Time Monitoring. *Small* **2024**, *20* (11), No. 2306769, DOI: 10.1002/sml.202306769.
- (512) Qiao, X.; Cai, Y.; Kong, Z.; Xu, Z.; Luo, X. A Wearable Electrochemical Sensor Based on Anti-Fouling and Self-Healing Polypeptide Complex Hydrogels for Sweat Monitoring. *ACS Sens.* **2023**, *8*, 2834–2842.
- (513) Hu, M.; Wang, Z.; Zhang, L.; Lin, S.; Liao, J. A microfluidic patch for wireless wearable electrochemical detection of sweat metabolites. *Sens. Actuators, B* **2025**, *422*, No. 136604.
- (514) Liu, G.; Gong, Z.; Dou, X.; et al. An Integrated Electrochemical Sensor with Flexible Microfluidic Structures for Human Sweat Analysis. *Biochip J.* **2025**, *19*, 50–61.
- (515) Alam, Md. S.; Kim, J. K.; Choi, J. Multifunctional Wearable System for Mapping Body Temperature and Analyzing Sweat. *ACS Sens.* **2023**, *8*, 1980–1988.
- (516) Liu, W.; Cheng, H.; Wang, X. Skin-interfaced colorimetric microfluidic devices for on-demand sweat analysis. *npj Flexible Electron.* **2023**, *7*, No. 43.
- (517) Niu, Q.; Yuan, R.; Zhan, H.; Jin, J.; Qiao, X. Janus Membrane-Based Wearable Dual-Channel SERS Sensor for Sweat Collection and Monitoring of Lactic Acid and pH Levels. *Anal. Chem.* **2025**, *97*, 6083–6091.
- (518) Li, Y.; Guo, Y.; Chen, H.; et al. Flexible Wearable Plasmonic Paper-Based Microfluidics with Expandable Channel and Adjustable Flow Rate for Portable Surface-Enhanced Raman Scattering Sweat Sensing. *ACS Photonics* **2024**, *11*, 613–625.
- (519) Yang, K.; Wang, Z.; Zhu, K.; et al. A wearable SERS-microfluidic patch for continuous monitoring of kidney health-related biomarkers in sweat. *Talanta* **2025**, *293*, No. 128039.
- (520) Hu, M.; Zhu, K.; Wei, J.; et al. Silk fibroin-based wearable SERS biosensor for simultaneous sweat monitoring of creatinine and uric acid. *Biosens. Bioelectron.* **2024**, *265*, No. 116662.
- (521) Lin, L. L.; Alvarez-Puebla, R.; Liz-Marzán, L. M.; et al. Surface-Enhanced Raman Spectroscopy for Biomedical Applications: Recent Advances and Future Challenges. *ACS Appl. Mater. Interfaces* **2025**, *17*, 16287–16379.
- (522) Xiong, S.; Zhu, C.; Wang, C.; Dong, P.; Wu, X. SERS-based pump-free microfluidic chip sensor for highly sensitive competitive immunoassay of cortisol in human sweat. *Lab Chip* **2024**, *24*, 5384–5397.
- (523) Liao, C.; Li, S.; Yang, C.; et al. Wearable epidermal sensor patch with biomimetic microfluidic channels for fast and time-sequence monitoring of sweat glucose and lactate. *Talanta* **2025**, *287*, No. 127683.
- (524) Garland, N. T.; Schmieder, J.; Johnson, Z. T.; et al. Wearable Flexible Perspiration Biosensors Using Laser-Induced Graphene and Polymeric Tape Microfluidics. *ACS Appl. Mater. Interfaces* **2023**, *15*, 38201–38213.
- (525) Li, X.; He, X.; Yang, X.; et al. A wearable sensor patch for joule-heating sweating and comfortable biofluid monitoring. *Sens. Actuators, B* **2024**, *419*, No. 136399.
- (526) Wei, L.; Lv, Z.; He, Y.; et al. In-situ admittance sensing of sweat rate and chloride level in sweat using wearable skin-interfaced microfluidic patch. *Sens. Actuators, B* **2023**, *379*, No. 133213.
- (527) Singh, N. K.; Chung, S.; Chang, A. Y.; Wang, J.; Hall, D. A. A non-invasive wearable stress patch for real-time cortisol monitoring using a pseudoknot-assisted aptamer. *Biosens. Bioelectron.* **2023**, *227*, No. 115097.
- (528) Weng, X.; Fu, Z.; Zhang, C.; Jiang, W.; Jiang, H. A Portable 3D Microfluidic Origami Biosensor for Cortisol Detection in Human Sweat. *Anal. Chem.* **2022**, *94*, 3526–3534.
- (529) Khumngern, S.; Jeerapan, I. Advances in wearable electrochemical antibody-based sensors for cortisol sensing. *Anal. Bioanal. Chem.* **2023**, *415*, 3863–3877.
- (530) Kim, J.; de Araujo, W. R.; Samek, I. A.; et al. Wearable temporary tattoo sensor for real-time trace metal monitoring in human sweat. *Electrochem. Commun.* **2015**, *51*, 41–45.
- (531) Yang, Q.; Rosati, G.; Abarintos, V.; et al. Wearable and fully printed microfluidic nanosensor for sweat rate, conductivity, and copper detection with healthcare applications. *Biosens. Bioelectron.* **2022**, *202*, No. 114005.
- (532) Nava-Granados, J.; Vasquez, K.; Medina, B. U.; et al. Flow by design: a guided review of microfluidics for wearable biosensors. *Lab Chip* **2026**, *26*, 1080–1107, DOI: 10.1039/D5LC000628G.
- (533) Chen, J.; Lou, Y.; Sun, L.; et al. Precise and efficient release of pilocarpine for sweat wearables. *Sens. Actuators, B* **2024**, *419*, No. 136403.

- (534) Zhang, Y.; Li, Z.; Fan, X.; et al. A fully integrated, non-invasive, and multimodal wearable device for sweat stimulation, collection and multiple physiological signals analysis. *Chem. Eng. J.* **2025**, *S05*, No. 159209.
- (535) Silverio, A. A.; Ho, E. T. W.; Ang, J.; Esguerra, K. Wearable microfluidic sweat collection platform with a calorimetric flow rate sensor for realtime and long-term sweat rate measurements. *J. Micromech. Microeng.* **2024**, *34*, No. 035008.
- (536) Lee, Y. S.; Shin, S.; Kang, G. R.; et al. Spatiotemporal molecular tracing of ultralow-volume biofluids via a soft skin-adaptive optical monolithic patch sensor. *Nat. Commun.* **2025**, *16*, No. 3272.
- (537) Hong, X.; Wu, H.; Wang, C.; et al. Hybrid Janus Membrane with Dual-Asymmetry Integration of Wettability and Conductivity for Ultra-Low-Volume Sweat Sensing. *ACS Appl. Mater. Interfaces* **2022**, *14*, 9644–9654.
- (538) Mei, X.; Zhu, L.; Peng, X.; Yang, J.; Li, Y. Wearable Janus nanofiber membrane-based colorimetric sensors for directional transfer and analysis of sweat. *Biomed. Anal.* **2024**, *1*, 64–72.
- (539) Li, D.; Liu, W.; Peng, T.; et al. Janus Textile: Advancing Wearable Technology for Autonomous Sweat Management and Beyond. *Small* **2025**, *21* (13), No. 2409730, DOI: 10.1002/sml.202409730.
- (540) Guo, X.; Zhang, Q.; Zhang, C.; et al. Pumpless microfluidic sweat sensing yarn. *Biosens. Bioelectron.* **2024**, *266*, No. 116713.
- (541) Saha, T.; Fang, J.; Mukherjee, S.; Dickey, M. D.; Velev, O. D. Wearable Osmotic-Capillary Patch for Prolonged Sweat Harvesting and Sensing. *ACS Appl. Mater. Interfaces* **2021**, *13*, 8071–8081.
- (542) Shay, T.; Dickey, M. D.; Velev, O. D. Hydrogel-enabled osmotic pumping for microfluidics: towards wearable human-device interfaces. *Lab Chip* **2017**, *17*, 710–716.
- (543) Choi, J.; Kang, D.; Han, S.; Kim, S. B.; Rogers, J. A. Thin, Soft, Skin-Mounted Microfluidic Networks with Capillary Bursting Valves for Chrono-Sampling of Sweat. *Adv. Healthcare Mater.* **2017**, *6* (5), No. 1601355, DOI: 10.1002/adhm.201601355.
- (544) Zhang, H.; Qiu, Y.; Yu, S.; et al. Wearable microfluidic patch with integrated capillary valves and pumps for sweat management and multiple biomarker analysis. *Biomicrofluidics* **2022**, *16*, No. 044104, DOI: 10.1063/5.0092084.
- (545) He, Y.; Wei, L.; Xu, W.; Wu, H.; Liu, A. Laser-Cutted Epidermal Microfluidic Patch with Capillary Bursting Valves for Chronological Capture, Storage, and Colorimetric Sensing of Sweat. *Biosensors* **2023**, *13*, 372.
- (546) Bandodkar, A. J.; Choi, J.; Lee, S. P.; et al. Soft, Skin-Interfaced Microfluidic Systems with Passive Galvanic Stopwatches for Precise Chronometric Sampling of Sweat. *Adv. Mater.* **2019**, *31*, No. 1902109, DOI: 10.1002/adma.201902109.
- (547) Kim, S. B.; Zhang, Y.; Won, S. M.; et al. Super-Absorbent Polymer Valves and Colorimetric Chemistries for Time-Sequenced Discrete Sampling and Chloride Analysis of Sweat via Skin-Mounted Soft Microfluidics. *Small* **2018**, *14*, No. 1703334, DOI: 10.1002/sml.201703334.
- (548) Zhang, Y.; Chen, Y.; Huang, J.; et al. Skin-interfaced microfluidic devices with one-opening chambers and hydrophobic valves for sweat collection and analysis. *Lab Chip* **2020**, *20*, 2635–2645.
- (549) Vaquer, A.; Barón, E.; de la Rica, R. Dissolvable Polymer Valves for Sweat Chrono-Sampling in Wearable Paper-Based Analytical Devices. *ACS Sens.* **2022**, *7*, 488–494.
- (550) Lin, R.; Guo, W.; Chen, Y.; et al. Liquid Bridge Cutting Valves for Microfluidic Passive Distribution and Sequential Reaction. *Small* **2025**, *21*, No. 2411708, DOI: 10.1002/sml.202411708.
- (551) Shi, H.; Cao, Y.; Zeng, Y.; et al. Wearable tesla valve-based sweat collection device for sweat colorimetric analysis. *Talanta* **2022**, *240*, No. 123208.
- (552) Zhao, P.; Wang, H.; Wang, Y.; et al. A time sequential microfluid sensor with Tesla valve channels. *Nano Res.* **2023**, *16*, 11667–11673.
- (553) Purwidyantri, A.; Prabowo, B. A. Tesla Valve Microfluidics: The Rise of Forgotten Technology. *Chemosensors* **2023**, *11*, 256.
- (554) Lin, H.; Tan, J.; Zhu, J.; et al. A programmable epidermal microfluidic valving system for wearable biofluid management and contextual biomarker analysis. *Nat. Commun.* **2020**, *11*, No. 4405.
- (555) Yang, J. H.; David, U.; Noh, Y. S.; Koh, A. Dual-valved skin-interfaced microfluidic device for programmed time-control sweat collection. *Sens. Actuators, B* **2023**, *395*, No. 134441.
- (556) Naik, A. R.; Warren, B.; Burns, A.; et al. Electrowetting valves for sweat-based microfluidics. *Microfluid. Nanofluidics* **2021**, *25*, No. 2.
- (557) Li, H.; Wang, W.; Xu, T.; Huang, Y.; Fang, W. High performance electromagnetic micropump with bio-inspired synchronous valves for integrated microfluidics. *Sens. Actuators, A* **2023**, *360*, No. 114568.
- (558) Xu, W.; Cao, Y.; Shi, H.; et al. Skin-interfaced sweat monitoring patch constructed by flexible microfluidic capillary pump and Cu-MOF sensitized electrochemical sensor. *Talanta* **2025**, *291*, No. 127895.
- (559) Bariya, M.; Davis, N.; Gillan, L.; et al. Resettable Microfluidics for Broad-Range and Prolonged Sweat Rate Sensing. *ACS Sens.* **2022**, *7*, 1156–1164.
- (560) Ning, Q.; Feng, S.; Sun, Q.; et al. Finger-actuated wireless-charging wearable multifunctional sweat-sensing system for levodopa and vitamin C. *Nano Res.* **2024**, *17*, 3096–3106.
- (561) Zhao, F.; Chen, X.; Zhang, J.; et al. A wearable, nozzle-diffuser microfluidic pump based on high-performance ferroelectric nanocomposites. *Sens. Actuators, B* **2021**, *347*, No. 130611.
- (562) Kuwajima, Y.; Yamaguchi, Y.; Yamada, Y.; et al. Pocketable and Smart Electrohydrodynamic Pump for Clothes. *ACS Appl. Mater. Interfaces* **2024**, *16*, 1883–1891.
- (563) Li, F.; Sun, S.; Wan, X.; et al. A self-powered soft triboelectric-electrohydrodynamic pump. *Nat. Commun.* **2025**, *16*, No. 1315.
- (564) Mishra, N.; Garland, N. T.; Hewett, K. A.; et al. A Soft Wearable Microfluidic Patch with Finger-Actuated Pumps and Valves for On-Demand, Longitudinal, and Multianalyte Sweat Sensing. *ACS Sens.* **2022**, *7*, 3169–3180.
- (565) Gama, M. I. H.; Abdullahi, S.; Omer, M. O. M.; et al. A novel manual rotating fluid control mechanism in a microfluidic device with a finger-actuated pump for dual-mode sweat sampling. *Lab Chip* **2025**, *25* (13), 3094–3108, DOI: 10.1039/D5LC00171D.
- (566) Cho, Y. H.; Park, Y.; Kim, S.; Park, J. 3D Electrodes for Bioelectronics. *Adv. Mater.* **2021**, *33* (47), No. 2005805, DOI: 10.1002/adma.202005805.
- (567) Aldawood, F. K.; Andar, A.; Desai, S. A Comprehensive Review of Microneedles: Types, Materials, Processes, Characterizations and Applications. *Polymers* **2021**, *13*, 2815.
- (568) Ji, H.; Wang, M.; Wang, Y.; et al. Skin-integrated, biocompatible, and stretchable silicon microneedle electrode for long-term EMG monitoring in motion scenario. *npj Flexible Electron.* **2023**, *7*, No. 46.
- (569) Rosati, G.; Deroco, P. B.; Bonando, M. G.; et al. Introducing all-inkjet-printed microneedles for in-vivo biosensing. *Sci. Rep.* **2024**, *14*, No. 29975.
- (570) Baraniecki, D.; Rosati, G.; Miarka, A.; et al. Fully-printed microneedles meet plants: a pathway towards easy-to-use NFC monitoring of total ionic conductivity in precision agriculture. *Bios. Bioelectron.* **2026**, *306*, No. 118667.
- (571) Sadeqi, A.; Kiaee, G.; Zeng, W.; Rezaei Nejad, H.; Sonkusale, S. Hard polymeric porous microneedles on stretchable substrate for transdermal drug delivery. *Sci. Rep.* **2022**, *12*, No. 1853.
- (572) Battisti, M.; Vecchione, R.; Casale, C.; et al. Non-invasive Production of Multi-Compartmental Biodegradable Polymer Microneedles for Controlled Intradermal Drug Release of Labile Molecules. *Front. Bioeng. Biotechnol.* **2019**, *7*, 296 DOI: 10.3389/fbioe.2019.00296.
- (573) Wang, K.; Achour, B.; et al. Direct-Print 3D Electrodes for Large-Scale, High-Density, and Customizable Neural Interfaces. *Adv. Sci.* **2025**, *12* (3), No. 2408602, DOI: 10.1002/adv.202408602.
- (574) Teymourian, H.; Tehrani, F.; Mahato, K.; Wang, J. Lab under the Skin: Microneedle Based Wearable Devices. *Adv. Healthcare Mater.* **2021**, *10* (17), No. 2002255, DOI: 10.1002/adhm.202002255.

- (575) Chung, W. G.; Jang, J.; Cui, G.; et al. Liquid-metal-based three-dimensional microelectrode arrays integrated with implantable ultrathin retinal prosthesis for vision restoration. *Nat. Nanotechnol.* **2024**, *19*, 688–697.
- (576) Lee, S.; Chung, W. G.; Jeong, H.; et al. Electrophysiological Analysis of Retinal Organoid Development Using 3D Microelectrodes of Liquid Metals. *Adv. Mater.* **2024**, *36* (35), No. 2404428, DOI: 10.1002/adma.202404428.
- (577) Hu, J. W.; Chatzilakou, E.; Pan, Z.; Traverso, G.; Yetisen, A. K. Microneedle Sensors for Point-of-Care Diagnostics. *Adv. Sci.* **2024**, *11* (12), No. 2306560.
- (578) Valdés-Ramírez, G.; Li, Y. C.; Kim, J.; et al. Microneedle-based self-powered glucose sensor. *Electrochem. Commun.* **2014**, *47*, 58–62.
- (579) Zhang, B. L.; Yang, Y.; Zhao, Z. Q.; Guo, X. D. A gold nanoparticles deposited polymer microneedle enzymatic biosensor for glucose sensing. *Electrochim. Acta* **2020**, *358*, No. 136917.
- (580) Kim, J.; Kim, H.; Park, S.; et al. Individually-addressable composite microneedle electrode array by mold-and-place method for glucose detection. *Sens. Actuators, B* **2024**, *401*, No. 134884.
- (581) Wang, Y.; Wu, Y.; Lei, Y. Microneedle-based glucose monitoring: a review from sampling methods to wearable biosensors. *Biomater. Sci.* **2023**, *11*, 5727–5757.
- (582) Xiao, J.; Zhang, S.; Liu, Q.; Xu, T.; Zhang, X. Microfluidic-based plasmonic microneedle biosensor for uric acid ultrasensitive monitoring. *Sens. Actuators, B* **2024**, *398*, No. 134685.
- (583) Tehrani, F.; Teymourian, H.; Wuerstle, B.; et al. An integrated wearable microneedle array for the continuous monitoring of multiple biomarkers in interstitial fluid. *Nat. Biomed. Eng.* **2022**, *6*, 1214–1224.
- (584) Windmiller, J. R.; Valdés-Ramírez, G.; Zhou, N.; et al. Bicomponent Microneedle Array Biosensor for Minimally-Invasive Glutamate Monitoring. *Electroanalysis* **2011**, *23*, 2302–2309.
- (585) Huang, X.; Zheng, S.; Liang, B.; et al. 3D-assembled microneedle ion sensor-based wearable system for the transdermal monitoring of physiological ion fluctuations. *Microsyst. Nanoeng.* **2023**, *9*, 1–16.
- (586) Huang, X.; Zheng, S.; Liang, B.; et al. 3D-assembled microneedle ion sensor-based wearable system for the transdermal monitoring of physiological ion fluctuations. *Microsyst. Nanoeng.* **2023**, *9*, No. 25.
- (587) Goud, K. Y.; Moonla, C.; Mishra, R. K.; et al. Wearable Electrochemical Microneedle Sensor for Continuous Monitoring of Levodopa: Toward Parkinson Management. *ACS Sens.* **2019**, *4*, 2196–2204.
- (588) Goud, K. Y.; Mahato, K.; Teymourian, H.; et al. Wearable electrochemical microneedle sensing platform for real-time continuous interstitial fluid monitoring of apomorphine: Toward Parkinson management. *Sens. Actuators, B* **2022**, *354*, No. 131234.
- (589) Lin, S.; Cheng, X.; Zhu, J.; et al. Wearable microneedle-based electrochemical aptamer biosensing for precision dosing of drugs with narrow therapeutic windows. *Sci. Adv.* **2022**, *8* (38), No. eabq4539.
- (590) Mishra, R. K.; Vinu Mohan, A. M.; Soto, F.; Chrostowski, R.; Wang, J. A microneedle biosensor for minimally-invasive transdermal detection of nerve agents. *Analyst* **2017**, *142*, 918–924.
- (591) Mishra, R. K.; Goud, K. Y.; Li, Z.; et al. Continuous Opioid Monitoring along with Nerve Agents on a Wearable Microneedle Sensor Array. *J. Am. Chem. Soc.* **2020**, *142*, 5991–5995.
- (592) Ciui, B.; Martin, A.; Mishra, R. K.; et al. Wearable Wireless Tyrosinase Bandage and Microneedle Sensors: Toward Melanoma Screening. *Adv. Healthcare Mater.* **2018**, *7*, No. 1701264.
- (593) Razzaghi, M.; Ninan, J. A.; Azimzadeh, M.; et al. Remote-Controlled Sensing and Drug Delivery via 3D-Printed Hollow Microneedles. *Adv. Healthcare Mater.* **2024**, *13* (23), No. 2400881, DOI: 10.1002/ADHM.202400881.
- (594) Seo, H.; Chung, W. G.; Kwon, Y. W.; et al. Smart Contact Lenses as Wearable Ophthalmic Devices for Disease Monitoring and Health Management. *Chem. Rev.* **2023**, *123*, 11488–11558.
- (595) Song, H.; Shin, H.; Seo, H.; et al. Wireless Non-Invasive Monitoring of Cholesterol Using a Smart Contact Lens. *Adv. Sci.* **2022**, *9*, No. 2203597, DOI: 10.1002/adv.202203597.
- (596) Ma, X.; Ahadian, S.; Liu, S.; et al. Smart Contact Lenses for Biosensing Applications. *Adv. Intell. Syst.* **2021**, *3* (5), No. 2000263, DOI: 10.1002/aisy.202170047.
- (597) Ku, M.; Kim, J.; Won, J. E.; et al. Smart, soft contact lens for wireless immunosensing of cortisol. *Sci. Adv.* **2020**, *6* (28), No. eabb2891, DOI: 10.1126/sciadv.abb2891.
- (598) An, B. W.; Kim, K.; Lee, H.; et al. High-Resolution Printing of 3D Structures Using an Electrohydrodynamic Inkjet with Multiple Functional Inks. *Adv. Mater.* **2015**, *27*, 4322–4328.
- (599) Park, J.; Ahn, D. B.; Kim, J.; et al. Printing of wirelessly rechargeable solid-state supercapacitors for soft, smart contact lenses with continuous operations. *Sci. Adv.* **2019**, *5* (12), No. eaay0764, DOI: 10.1126/sciadv.aay0764.
- (600) Jang, J.; Kim, J.; Shin, H.; et al. Smart contact lens and transparent heat patch for remote monitoring and therapy of chronic ocular surface inflammation using mobiles. *Sci. Adv.* **2021**, *7* (14), No. eabf7194, DOI: 10.1126/sciadv.abf7194.
- (601) Li, L.; Shen, G. MXene based flexible photodetectors: progress, challenges, and opportunities. *Mater. Horiz.* **2023**, *10*, 5457–5473.
- (602) Gu, L.; Poddar, S.; Lin, Y.; et al. A biomimetic eye with a hemispherical perovskite nanowire array retina. *Nature* **2020**, *581*, 278–282.
- (603) Long, Z.; Qiu, X.; Chan, C. L. J.; et al. A neuromorphic bionic eye with filter-free color vision using hemispherical perovskite nanowire array retina. *Nat. Commun.* **2023**, *14*, No. 1972.
- (604) Zhu, H.; Yang, H.; Xu, S.; et al. Frequency-encoded eye tracking smart contact lens for human–machine interaction. *Nat. Commun.* **2024**, *15*, No. 3588.
- (605) Liu, W.; Du, Z.; Duan, Z.; et al. Eye-wearable $\text{Ti}_3\text{C}_2\text{T}_x$ MXene-based micro-supercapacitor as a power unit for intraocular pressure applications. *J. Mater. Chem. A* **2024**, *12*, 16457–16465.
- (606) Liu, W.; Du, Z.; Duan, Z.; Li, L.; Shen, G. Neuroprosthetic contact lens enabled sensorimotor system for point-of-care monitoring and feedback of intraocular pressure. *Nat. Commun.* **2024**, *15*, No. 5635.
- (607) Mirzajani, H.; Mirlou, F.; Istif, E.; Singh, R.; Beker, L. Powering smart contact lenses for continuous health monitoring: Recent advancements and future challenges. *Biosens. Bioelectron.* **2022**, *197*, No. 113761.
- (608) Pourshaban, E.; Banerjee, A.; Karkhanis, M. U.; et al. Eye Tear Activated Mg-Air Battery Driven by Natural Eye Blinking for Smart Contact Lenses. *Adv. Mater. Technol.* **2023**, *8* (1), No. 2200518, DOI: 10.1002/admt.202200518.
- (609) Song, H.; Shin, H.; Seo, H.; et al. Wireless Non-Invasive Monitoring of Cholesterol Using a Smart Contact Lens. *Adv. Sci.* **2022**, *9* (28), No. 2203597, DOI: 10.1002/adv.202203597.
- (610) Kim, J.; Campbell, A. S.; de Ávila, B. E.-F.; Wang, J. Wearable biosensors for healthcare monitoring. *Nat. Biotechnol.* **2019**, *37*, 389–406.
- (611) Graf, H.; Mühlemann, H. R. Oral telemetry of fluoride ion activity. *Arch. Oral Biol.* **1969**, *14*, 259–263.
- (612) Pfaffe, T.; Cooper-White, J.; Beyerlein, P.; Kostner, K.; Punyadeera, C. Diagnostic Potential of Saliva: Current State and Future Applications. *Clin. Chem.* **2011**, *57*, 675–687.
- (613) Moonla, C.; Lee, D. H.; Rokaya, D.; et al. Review—Lab-in-a-Mouth and Advanced Point-of-Care Sensing Systems: Detecting Bioinformation from the Oral Cavity and Saliva. *ECS Sens. Plus* **2022**, *1*, No. 021603.
- (614) Chandran, M.; Veerapandian, M.; Dhanasekaran, B.; Govindaraju, S.; Yun, K. Advanced nanomaterials for health monitoring and diagnostics in next-generation wearable sensors. *Mater. Sci. Eng., R* **2025**, *165*, No. 101015.
- (615) Arakawa, T.; Kuroki, Y.; Nitta, H.; et al. Mouthguard biosensor with telemetry system for monitoring of saliva glucose: A novel cavitas sensor. *Biosens. Bioelectron.* **2016**, *84*, 106–111.

- (616) Kim, J.; Valdés-Ramírez, G.; Bandodkar, A. J.; et al. Non-invasive mouthguard biosensor for continuous salivary monitoring of metabolites. *Analyst* **2014**, *139*, 1632–1636.
- (617) Ichikawa, K.; Iitani, K.; Kawase, G.; et al. Real-time salivary turbidity measurement with a mouthguard-type wearable sensor. *Sens. Actuators, A* **2024**, *376*, No. 115660.
- (618) Jeon, S.; et al. A Wearable Electrochemical Biosensor for Salivary Detection of Periodontal Inflammation Biomarkers: Molecularly Imprinted Polymer Sensor with Deep Learning Integration. *Adv. Sci.* **2025**, *12* (40), No. e09658, DOI: 10.1002/advs.202509658.
- (619) Ichikawa, K.; Yoda, N.; Hijikata, W. Self-powered mouthguard-type bite force sensor with wide dynamic range. *Measurement* **2025**, *249*, No. 117064.
- (620) Rawassizadeh, R.; Price, B. A.; Petre, M. Wearables. *Commun. ACM* **2015**, *58*, 45–47.
- (621) Lu, T.-C.; Fu, C.-M.; Ma, M.; Fang, C.-C.; Turner, A. Healthcare Applications of Smart Watches. *Appl. Clin. Inform.* **2016**, *07*, 850–869.
- (622) Scheidt, D. T. et al. Exploring Novel Healthcare Monitoring Modalities: Wearable, Ingestible, and Implantable Biosensors. In *Nano-bioelectronics for Precision Health Monitoring*; Springer Nature Singapore: Singapore, 2025; pp 137–175 DOI: 10.1007/978-981-95-2117-3_5.
- (623) Henriksen, A.; Haugen Mikalsen, M.; Woldaregay, A. Z.; et al. Using Fitness Trackers and Smartwatches to Measure Physical Activity in Research: Analysis of Consumer Wrist-Worn Wearables. *J. Med. Internet Res.* **2018**, *20*, No. e110.
- (624) Isakadze, N.; Martin, S. S. How useful is the smartwatch ECG? *Trends Cardiovasc. Med.* **2020**, *30*, 442–448.
- (625) Chen, Y.; Wang, W.; Guo, Y.; et al. A Single-Center Validation of the Accuracy of a Photoplethysmography-Based Smartwatch for Screening Obstructive Sleep Apnea. *Nat. Sci. Sleep* **2021**, *Volume 13*, 1533–1544.
- (626) Kamišalić, A.; Fister, I.; Turkanović, M.; Karakatić, S. Sensors and Functionalities of Non-Invasive Wrist-Wearable Devices: A Review. *Sensors* **2018**, *18*, 1714.
- (627) Khan, Y.; Ostfeld, A. E.; Lochner, C. M.; Pierre, A.; Arias, A. C. Monitoring of Vital Signs with Flexible and Wearable Medical Devices. *Adv. Mater.* **2016**, *28*, 4373–4395.
- (628) Ma, X.; Wang, P.; Huang, L.; et al. A monolithically integrated in-textile wristband for wireless epidermal biosensing. *Sci. Adv.* **2023**, *9*, No. ead2763, DOI: 10.1126/sciadv.adj2763.
- (629) Cai, X.; Xia, R. Z.; Liu, Z. H.; et al. Fully Integrated Multiplexed Wristwatch for Real-Time Monitoring of Electrolyte Ions in Sweat. *ACS Nano* **2024**, *18*, 12808–12819.
- (630) Escobedo, P.; Ramos-Lorente, C. E.; Martínez-Olmos, A.; et al. Wireless wearable wristband for continuous sweat pH monitoring. *Sens. Actuators, B* **2021**, *327*, No. 128948.
- (631) Zhao, J. et al. 3D E-textile for Exercise Physiology and Clinical Maternal Health Monitoring. ArXiv arXiv:2407.07954v1, 2024.
- (632) Chatterjee, K.; Ghosh, T. K. 3D Printing of Textiles: Potential Roadmap to Printing with Fibers. *Adv. Mater.* (2019 DOI: 10.1002/adma.201902086.
- (633) Li, S.; Li, H.; Lu, Y.; et al. Advanced Textile-Based Wearable Biosensors for Healthcare Monitoring. *Biosensors* **2023**, *13* (10), 909 DOI: 10.3390/bios13100909.
- (634) Iqbal, S. M. A.; Mahgoub, I.; Du, E.; Leavitt, M. A.; Asghar, W. Advances in healthcare wearable devices. *npj Flexible Electron.* **2021**, *5*, No. 9, DOI: 10.1038/s41528-021-00107-x.
- (635) Meena, J. S.; Choi, S. Bin.; Jung, S.-B.; Kim, J.-W. Electronic textiles: New age of wearable technology for healthcare and fitness solutions. *Mater. Today Bio* **2023**, *19*, No. 100565, DOI: 10.1016/j.mtbio.2023.100565.
- (636) Luo, Y.; et al. Learning human–environment interactions using conformal tactile textiles. *Nat. Electron.* **2021**, *4* (3), 193–201, DOI: 10.1038/s41928-021-00558-0.
- (637) Wang, C.; Fu, L.; Ametefe, D. S.; Wang, S.; John, D. E-textiles in healthcare: a systematic literature review of wearable technologies for monitoring and enhancing human health. *Neural Comput. Appl.* **2024**, *37* (4), 2089–2111, DOI: 10.1007/s00521-024-10947-z.
- (638) Lu, W.; Wu, G.; Gan, L.; Zhang, Y.; Li, K. Functional fibers/textiles for smart sensing devices and applications in personal healthcare systems. *Anal. Methods* **2024**, *16* (31), 5372–5390, DOI: 10.1039/d4ay01127a.
- (639) Hernandez, N.; Castro, L.; Medina-Quero, J.; et al. Scoping Review of Healthcare Literature on Mobile, Wearable, and Textile Sensing Technology for Continuous Monitoring. *J. Healthcare Inform. Res.* **2021**, *5* (3), 270–299, DOI: 10.1007/s41666-020-00087-z.
- (640) Martinez, R. V. Can e-textiles make their way into mass production? *Nat. Rev. Electr. Eng.* **2024**, *1* (1), 4–5, DOI: 10.1038/s44287-023-00007-4.
- (641) Azeem, M.; Shahid, M.; Masin, I.; Petru, M. Design and development of textile-based wearable sensors for real-time biomedical monitoring; a review. *J. Text. Inst.* **2024**, *116* (1), 80–95, DOI: 10.1080/00405000.2024.2318500.
- (642) Azani, M.-R.; Hassanpour, A. Electronic textiles (E-Textiles): types, fabrication methods, and recent strategies to overcome durability challenges (washability & flexibility). *J. Mater. Sci.: Mater. Electron.* **2024**, *35* (29), No. 1897, DOI: 10.1007/s10854-024-13347-0.
- (643) Liao, M.; Wang, C.; Hong, Y.; et al. Industrial scale production of fibre batteries by a solution-extrusion method. *Nat. Nanotechnol.* **2022**, *17* (4), 372–377, DOI: 10.1038/s41565-021-01062-4.
- (644) Shi, H. H.; Pan, Y.; Xu, L.; et al. Sustainable electronic textiles towards scalable commercialization. *Nat. Mater.* **2023**, *22* (11), 1294–1303, DOI: 10.1038/s41563-023-01615-z.
- (645) Sala de Medeiros, M.; Goswami, D.; Chanci, D.; Moreno, C.; Martinez, R. V. Washable, breathable, and stretchable e-textiles wirelessly powered by omniphobic silk-based coils. *Nano Energy* **2021**, *87*, No. 106155, DOI: 10.1016/j.nanoen.2021.106155.
- (646) Du, K.; Lin, R.; Yin, L.; et al. Electronic textiles for energy, sensing, and communication. *iScience* **2022**, *25* (5), No. 104174, DOI: 10.1016/j.isci.2022.104174.
- (647) Lee, S. Analysis of electrical and comfort properties of conductive knitted fabrics based on blending ratio of silver-coated yarns for smart clothing. *J. Eng. Fiber. Fabr.* **2022**, *17*, No. 15589250221104474, DOI: 10.1177/15589250221104474.
- (648) Komolafe, A.; Zaghari, B.; Torah, R.; et al. E-Textile Technology Review—From Materials to Application. *IEEE Access* **2021**, *9*, 97152–97179, DOI: 10.1109/access.2021.3094303.
- (649) Zaman, S. uz.; Tao, X.; Cochrane, C.; Koncar, V. Smart E-Textile Systems: A Review for Healthcare Applications. *Electronics* **2022**, *11* (1), 99 DOI: 10.3390/electronics11010099.
- (650) Fernandes, S.; Ramos, A.; Vega-Barbas, M.; et al. Smart Textile Technology for the Monitoring of Mental Health. *Sensors* **2025**, *25* (4), 1148 DOI: 10.3390/s25041148.
- (651) Libanori, A.; Chen, G.; Zhao, X.; Zhou, Y.; Chen, J. Smart textiles for personalized healthcare. *Nat. Electron.* **2022**, *5* (3), 142–156, DOI: 10.1038/s41928-022-00723-z.
- (652) Latko-Duralek, P.; Rzempoluch, J.; Staniszevska, M.; et al. The Antifungal Fibers of Polyamide 12 Containing Silver and Metal Oxides. *Materials* **2023**, *16* (17), 5837 DOI: 10.3390/ma16175837.
- (653) Wang, J.; Lu, C.; Zhang, K. Textile-Based Strain Sensor for Human Motion Detection. *Energy Environ. Mater.* **2019**, *3* (1), 80–100, DOI: 10.1002/eeem2.12041.
- (654) Wicaksono, I.; Tucker, C. I.; Sun, T.; et al. A tailored, electronic textile conformable suit for large-scale spatiotemporal physiological sensing in vivo. *npj Flexible Electron.* **2020**, *4*, No. 5.
- (655) Islam, M. R.; et al. Advances in Printed Electronic Textiles. *Adv. Sci.* **2023**, *11* (6), No. 2304140, DOI: 10.1002/advs.202304140.
- (656) Muth, J. T.; Vogt, D. M.; Truby, R. L.; et al. Embedded 3D Printing of Strain Sensors within Highly Stretchable Elastomers. *Adv. Mater.* **2014**, *26* (36), 6307–6312, DOI: 10.1002/adma.201400334.
- (657) Shahariar, H.; Kim, I.; Soewardiman, H.; Jur, J. S. Inkjet Printing of Reactive Silver Ink on Textiles. *ACS Appl. Mater. Interfaces* **2019**, *11* (6), 6208–6216, DOI: 10.1021/acsami.8b18231.

- (658) Fromme, N. P.; Li, Y.; Camenzind, M.; Toncelli, C.; Rossi, R. Metal-Textile Laser Welding for Wearable Sensors Applications. *Adv. Electron. Mater.* **2021**, *7* (4), No. 2001238, DOI: 10.1002/aelm.202001238.
- (659) Khan, S.; Lorenzelli, L.; Dahiya, R. S. Technologies for Printing Sensors and Electronics Over Large Flexible Substrates: A Review. *IEEE Sens. J.* **2015**, *15* (6), 3164–3185, DOI: 10.1109/jSEN.2014.2375203.
- (660) La, T.; Qiu, S.; Scott, D. K.; et al. Two-Layered and Stretchable e-Textile Patches for Wearable Healthcare Electronics. *Adv. Healthcare Mater.* **2018**, *7* (22), No. 1801033, DOI: 10.1002/adhm.201801033.
- (661) Dulal, M.; Afroj, S.; Ahn, J.; et al. Toward Sustainable Wearable Electronic Textiles. *ACS Nano* **2022**, *16* (12), 19755–19788, DOI: 10.1021/acsnano.2c07723.
- (662) Liang, J.; Tong, K.; Pei, Q. A Water-Based Silver-Nanowire Screen-Print Ink for the Fabrication of Stretchable Conductors and Wearable Thin-Film Transistors. *Adv. Mater.* **2016**, *28* (28), 2828, DOI: 10.1002/adma.201600772.
- (663) Karim, N.; Afroj, S.; Tan, S.; Novoselov, K. S.; Yeates, S. G. All Inkjet-Printed Graphene-Silver Composite Ink on Textiles for Highly Conductive Wearable Electronics Applications. *Sci. Rep.* **2019**, *9* (1), No. 8035, DOI: 10.1038/s41598-019-44420-y.
- (664) Wang, C.; Xia, K.; Wang, H.; et al. Advanced Carbon for Flexible and Wearable Electronics. *Adv. Mater.* **2019**, *31* (9), No. 1801072, DOI: 10.1002/adma.201801072.
- (665) Wang, Z.; Liu, Y.; Zhou, Z.; Chen, P.; Peng, H. Towards integrated textile display systems. *Nat. Rev. Electr. Eng.* **2024**, *1* (7), 466–477, DOI: 10.1038/s44287-024-00063-4.
- (666) Fernández-Caramés, T.; Fraga-Lamas, P. Towards The Internet of Smart Clothing: A Review on IoT Wearables and Garments for Creating Intelligent Connected E-Textiles. *Electronics* **2018**, *7* (12), 405, DOI: 10.3390/electronics7120405.
- (667) Dahiya, A. S.; et al. Review—Energy Autonomous Wearable Sensors for Smart Healthcare: A Review. *J. Electrochem. Soc.* **2019**, *167* (3), No. 037516, DOI: 10.1149/2.0162003jes.
- (668) Lu, C.; Jiang, H.; Cheng, X.; et al. High-performance fibre battery with polymer gel electrolyte. *Nature* **2024**, *629* (8010), 86–91, DOI: 10.1038/s41586-024-07343-x.
- (669) Rishani, N.; Elayan, H.; Shubair, R.; Kiourti, A. Wearable, Epidermal and Implantable Sensors for Medical Applications. *ArXiv* arXiv:1810.00321, 2018.
- (670) Lin, R.; Kim, H. J.; Achavananthadith, S.; et al. Wireless battery-free body sensor networks using near-field-enabled clothing. *Nat. Commun.* **2020**, *11* (1), No. 444, DOI: 10.1038/s41467-020-14311-2.
- (671) Shi, X.; Zuo, Y.; Zhai, P.; et al. Large-area display textiles integrated with functional systems. *Nature* **2021**, *591* (7849), 240–245, DOI: 10.1038/s41586-021-03295-8.
- (672) Yan, W.; Noel, G.; Loke, G.; et al. Single fibre enables acoustic fabrics via nanometre-scale vibrations. *Nature* **2022**, *603* (7902), 616–623, DOI: 10.1038/s41586-022-04476-9.
- (673) Lin, R.; Kim, H. J.; Achavananthadith, S.; et al. Digitally-embroidered liquid metal electronic textiles for wearable wireless systems. *Nat. Commun.* **2022**, *13*, No. 2190.
- (674) Fathme, A. M. M. K.; Islam, M. M.; Liu, X.; Liu, L. Development and Characterization of High-Performance Conductive Coatings on Cotton Yarn for Advanced Machine-Sewn Electronic Textiles. *Adv. Mater. Technol.* **2024**, *10* (8), No. 2401722, DOI: 10.1002/admt.202401722.
- (675) Bi, L.; et al. MXene Functionalized Kevlar Yarn via Automated, Continuous Dip Coating. *Adv. Funct. Mater.* **2023**, *34* (14), No. 2312434, DOI: 10.1002/adfm.202312434.
- (676) Gaubert, V.; Yendigoyeva, A.; Muniraj, V. K. A.; et al. Enhancing Electrical and Mechanical Properties of Conductive Textile for Wearable Embedded Systems Through Copper Electroplating. *Adv. Eng. Mater.* **2024**, *26* (5), No. 2301436, DOI: 10.1002/adem.202301436.
- (677) Zhang, Z.; Kong, Y.; Gao, J.; et al. Engineering strong man-made cellulosic fibers: a review of the wet spinning process based on cellulose nanofibrils. *Nanoscale* **2024**, *16* (13), 6383–6401, DOI: 10.1039/d3nr06126d.
- (678) Martinez, A. P.; Ng, A.; Nah, S. H.; Yang, S. Active-Textile Yarns and Embroidery Enabled by Wet-Spun Liquid Crystalline Elastomer Filaments. *Adv. Funct. Mater.* **2024**, *34* (34), No. 2400742, DOI: 10.1002/adfm.202400742.
- (679) Zhang, J.; Seyedin, S.; Qin, S.; et al. Fast and scalable wet-spinning of highly conductive PEDOT:PSS fibers enables versatile applications. *J. Mater. Chem. A* **2019**, *7* (11), 6401–6410, DOI: 10.1039/c9ta00022d.
- (680) Sarabia-Riquelme, R.; Andrews, R.; Anthony, J. E.; Weisenberger, M. C. Highly conductive wet-spun PEDOT:PSS fibers for applications in electronic textiles. *J. Mater. Chem. C* **2020**, *8* (33), 11618–11630, DOI: 10.1039/d0tc02558e.
- (681) Martinez, A. P.; et al. High-Speed and Scalable Wet Spinning of Graphene/Liquid Crystalline Elastomer Composite Filaments. *Adv. Funct. Mater.* **2024**, *35* (24), No. 2422176, DOI: 10.1002/adfm.202422176.
- (682) Gao, C.; et al. Continuous dry-wet spinning of white, stretchable, and conductive fibers of poly(3-hydroxybutyrate-co-4-hydroxybutyrate) and ATO@TiO₂ nanoparticles for wearable e-textiles. *J. Mater. Chem. C* **2020**, *8* (25), 8362–8367, DOI: 10.1039/d0tc01310b.
- (683) Fu, X.; Wan, G.; Guo, H.; et al. Self-healing actuatable electroluminescent fibres. *Nat. Commun.* **2024**, *15* (1), No. 10498, DOI: 10.1038/s41467-024-53955-2.
- (684) Gupta, N.; Cheung, H.; Payra, S.; et al. A single-fibre computer enables textile networks and distributed inference. *Nature* **2025**, *639* (8053), 79–86, DOI: 10.1038/s41586-024-08568-6.
- (685) Sahasrabudhe, A.; et al. Multifunctional microelectronic fibers enable wireless modulation of gut and brain neural circuits. *Nat. Biotechnol.* **2023**, *42* (6), 892–904, DOI: 10.1038/s41587-023-01833-5.
- (686) Khudiyev, T.; Grena, B.; Loke, G.; et al. Thermally drawn rechargeable battery fiber enables pervasive power. *Mater. Today* **2022**, *52*, 80–89, DOI: 10.1016/j.mattod.2021.11.020.
- (687) Kang, M.; Kim, T.-W. Recent Advances in Fiber-Shaped Electronic Devices for Wearable Applications. *Appl. Sci* **2021**, *11* (13), 6131, DOI: 10.3390/app11136131.
- (688) Liu, H.; Shi, Y.; Pan, Y.; Wang, Z.; Wang, B. Sensory interactive fibers and textiles. *npj Flexible Electron.* **2025**, *9* (1), No. 23, DOI: 10.1038/s41528-025-00398-4.
- (689) Fakharuddin, A.; Li, H.; Di Giacomo, F.; et al. Fiber-Shaped Electronic Devices. *Adv. Energy Mater.* **2021**, *11* (34), No. 2101443, DOI: 10.1002/aenm.202101443.
- (690) Seyedin, S.; Carey, T.; Arbab, A.; et al. Fibre electronics: towards scaled-up manufacturing of integrated e-textile systems. *Nanoscale* **2021**, *13* (30), 12818–12847, DOI: 10.1039/d1nr02061g.
- (691) Yang, W.; Lin, S.; Gong, W.; et al. Single body-coupled fiber enables chipless textile electronics. *Science* **2024**, *384* (6691), 74–81, DOI: 10.1126/science.adk3755.
- (692) Wang, L.; Wang, L.; Zhang, Y.; et al. Weaving Sensing Fibers into Electrochemical Fabric for Real-Time Health Monitoring. *Adv. Funct. Mater.* **2018**, *28* (42), No. 1804456, DOI: 10.1002/adfm.201804456.
- (693) Tian, X.; Lee, P. M.; Tan, Y. J.; et al. Wireless body sensor networks based on metamaterial textiles. *Nat. Electron.* **2019**, *2*, 243–251.
- (694) Hajiaghajani, A.; Afandizadeh Zargari, A. H.; Dautta, M.; et al. Textile-integrated metamaterials for near-field multibody area networks. *Nat. Electron.* **2021**, *4* (11), 808–817, DOI: 10.1038/s41928-021-00663-0.
- (695) Takahashi, R.; Yukita, W.; Sasatani, T.; et al. Twin Meander Coil. *Proc. ACM Interact. Mob. Wearable Ubiquitous Technol.* **2021**, *5* (4), 1–21, DOI: 10.1145/3494996.

- (696) Nguyen, D. T.; Zeng, Q.; Tian, X.; et al. Ambient health sensing on passive surfaces using metamaterials. *Sci. Adv.* **2024**, *10*, No. eadj6613, DOI: 10.1126/sciadv.adj6613.
- (697) Lee, P.; Kim, H.; Kim, Y.; et al. Beyond Pathogen Filtration: Possibility of Smart Masks as Wearable Devices for Personal and Group Health and Safety Management. *JMIR mHealth uHealth* **2022**, *10*, No. e38614.
- (698) Kim, D.; Lee, J.; Park, M. K.; Ko, S. H. Recent developments in wearable breath sensors for healthcare monitoring. *Commun. Mater.* **2024**, *5*, No. 41, DOI: 10.1038/s43246-024-00480-w.
- (699) Suo, J.; Liu, Y.; Wu, C.; et al. Wide-Bandwidth Nano-composite-Sensor Integrated Smart Mask for Tracking Multiphase Respiratory Activities. *Adv. Sci.* **2022**, *9* (31), No. 2203565, DOI: 10.1002/advs.202203565.
- (700) Zhong, J.; Li, Z.; Takakuwa, M.; et al. Smart Face Mask Based on an Ultrathin Pressure Sensor for Wireless Monitoring of Breath Conditions. *Adv. Mater.* **2022**, *34* (6), No. 2107758, DOI: 10.1002/adma.202107758.
- (701) Cui, S.; et al. Laser direct writing of Ga₂O₃/liquid metal-based flexible humidity sensors. *Opto-Electron. Adv.* **2023**, *6*, 220172.
- (702) Fakir, M. H.; Kim, J. K. Prediction of individual thermal sensation from exhaled breath temperature using a smart face mask. *Build. Environ.* **2022**, *207*, No. 108507.
- (703) Alshabouna, F.; Lee, H. S.; Barandun, G.; et al. PEDOT:PSS-modified cotton conductive thread for mass manufacturing of textile-based electrical wearable sensors by computerized embroidery. *Mater. Today* **2022**, *59*, 56–67.
- (704) Xue, Y.; Zhao, X.; Wu, F.; et al. Near-Field Electrochemistry Enables a Wearable Sensor-Embedded Smart Facemask for Personalized Respiratory Assessment. *ACS Sens.* **2025**, *10*, 2378–2385.
- (705) Kim, J.-H.; Marcus, C.; Ono, R.; et al. A conformable sensory face mask for decoding biological and environmental signals. *Nat. Electron.* **2022**, *5*, 794–807.
- (706) Adeel, M.; Cotur, Y.; Naik, A.; Gonzalez-Macia, L.; Güder, F. Face masks as a platform for wearable sensors. *Nat. Electron.* **2022**, *5*, 719–720.
- (707) Gao, Z.; Lou, Z.; Chen, S.; et al. Fiber gas sensor-integrated smart face mask for room-temperature distinguishing of target gases. *Nano Res.* **2018**, *11*, 511–519.
- (708) Wu, G.; Du, H.; Cha, Y. L.; et al. A wearable mask sensor based on polyaniline/CNT nanocomposites for monitoring ammonia gas and human breathing. *Sens. Actuators, B* **2023**, *375*, No. 132858.
- (709) Moon, S. M.; Lee, S.; Min, H.; et al. Design and Integration of a Gas Sensor Module that Indicates the End of Service Life of a Gas Mask Canister. *Adv. Mater. Technol.* **2022**, *7* (3), No. 2100711, DOI: 10.1002/admt.202100711.
- (710) Wang, B.; Yang, D.; Chang, Z.; et al. Wearable bioelectronic masks for wireless detection of respiratory infectious diseases by gaseous media. *Matter* **2022**, *5*, 4347–4362.
- (711) Huang, S.; Bai, X.; Xu, Z.; et al. Charge transfer-enhanced OECT glucose sensor with superabsorbent polymer composites for non-invasive breath analysis. *Sens. Actuators, B* **2025**, *444*, No. 138516.
- (712) Ning, Z.; Long, Z.; Yang, G.; Xing, L.; Xue, X. Self-Powered Wearable Biosensor in a Baby Diaper for Monitoring Neonatal Jaundice through a Hydrovoltaic-Biosensing Coupling Effect of ZnO Nanoarray. *Biosensors* **2022**, *12*, (2022).
- (713) Seo, W.; Yu, W.; Tan, T.; Ziaie, B.; Jung, B. Diaper-Embedded Urinary Tract Infection Monitoring Sensor Module Powered by Urine-Activated Batteries. *IEEE Trans. Biomed. Circuits Syst.* **2017**, *11*, 681–691.
- (714) Paalanne, N.; Wikstedt, L.; Pokka, T.; et al. Diaper-embedded urine test device for the screening of urinary tract infections in children: A cohort study. *BMC Pediatr.* **2020**, *20*, No. 378.
- (715) Yu, W.; Seo, W.; Tan, T.; Jung, B.; Ziaie, B. A diaper-embedded disposable nitrite sensor with integrated on-board urine-activated battery for UTI screening. In *2016 38th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*; IEEE, 2016; pp 303–306.
- (716) Glinka, M.; Jazdzewska, K.; Vakh, C.; et al. Assessment of baby disposable diapers application for urine collection and determination of phthalate metabolites. *Ecotoxicol. Environ. Saf.* **2024**, *272*, No. 116033, DOI: 10.1016/j.ecoenv.2024.116033.
- (717) Su, H.; Sun, F.; Lu, Z.; et al. A wearable sensing system based on smartphone and diaper to detect urine in-situ for patients with urinary incontinence. *Sens. Actuators, B* **2022**, *357*, No. 131459, DOI: 10.1016/j.snb.2022.131459.
- (718) Li, X.; Zhan, C.; Huang, Q.; et al. Smart Diaper Based on Integrated Multiplex Carbon Nanotube-Coated Electrode Array Sensors for in Situ Urine Monitoring. *ACS Appl. Nano Mater.* **2022**, *5*, 4767–4778.
- (719) Zhang, J.; Liu, J.; Su, H.; et al. A wearable self-powered biosensor system integrated with diaper for detecting the urine glucose of diabetic patients. *Sens. Actuators, B* **2021**, *341*, No. 130046.
- (720) Liu, G.; An, Z.; Lu, Y.; et al. High-output moisture-enabled electricity generator for fully self-powered wearable physical and biochemical monitoring. *Nano Energy* **2024**, *119*, No. 109098, DOI: 10.1016/j.nanoen.2023.109098.
- (721) Shitanda, I.; Fujimura, Y.; Takarada, T.; et al. Self-Powered Diaper Sensor with Wireless Transmitter Powered by Paper-Based Biofuel Cell with Urine Glucose as Fuel. *ACS Sens.* **2021**, *6*, 3409–3415.
- (722) Apoorva, S.; Nguyen, N. T.; Sreejith, K. R. Recent developments and future perspectives of microfluidics and smart technologies in wearable devices. *Lab Chip* **2024**, *24*, 1833.
- (723) He, F.; You, X.; Gong, H.; et al. Stretchable, Biocompatible, and Multifunctional Silk Fibroin-Based Hydrogels toward Wearable Strain/Pressure Sensors and Triboelectric Nanogenerators. *ACS Appl. Mater. Interfaces* **2020**, *12*, 6442–6450.
- (724) Liu, Y.-J.; Cao, W.-T.; Ma, M.-G.; Wan, P. Ultrasensitive Wearable Soft Strain Sensors of Conductive, Self-healing, and Elastic Hydrogels with Synergistic “Soft and Hard” Hybrid Networks. *ACS Appl. Mater. Interfaces* **2017**, *9*, 25559–25570.
- (725) Jing, X.; Mi, H. Y.; Lin, Y. J.; et al. Highly Stretchable and Biocompatible Strain Sensors Based on Mussel-Inspired Super-Adhesive Self-Healing Hydrogels for Human Motion Monitoring. *ACS Appl. Mater. Interfaces* **2018**, *10*, 20897–20909.
- (726) Zhao, W.; Li, Y.; Tian, J.; et al. Highly Stretchable Sensitive Multiscale Hydrogel Inspired by Biological Muscles for Wearing Sensors. *ACS Appl. Mater. Interfaces* **2024**, *16*, 58313–58325.
- (727) Liu, H.; Wang, X.; Cao, Y.; et al. Freezing-Tolerant, Highly Sensitive Strain and Pressure Sensors Assembled from Ionic Conductive Hydrogels with Dynamic Cross-Links. *ACS Appl. Mater. Interfaces* **2020**, *12*, 25334–25344.
- (728) Kwon, D.; Lee, T. I.; Shim, J.; et al. Highly Sensitive, Flexible, and Wearable Pressure Sensor Based on a Giant Piezocapacitive Effect of Three-Dimensional Microporous Elastomeric Dielectric Layer. *ACS Appl. Mater. Interfaces* **2016**, *8*, 16922–16931.
- (729) Pang, Y.; Tian, H.; Tao, L.; et al. Flexible, Highly Sensitive, and Wearable Pressure and Strain Sensors with Graphene Porous Network Structure. *ACS Appl. Mater. Interfaces* **2016**, *8*, 26458–26462.
- (730) Yang, J. C.; Kim, J. O.; Oh, J.; et al. Microstructured Porous Pyramid-Based Ultrahigh Sensitive Pressure Sensor Insensitive to Strain and Temperature. *ACS Appl. Mater. Interfaces* **2019**, *11*, 19472–19480.
- (731) Luo, Y.; Shao, J.; Chen, S.; et al. Flexible Capacitive Pressure Sensor Enhanced by Tilted Micropillar Arrays. *ACS Appl. Mater. Interfaces* **2019**, *11*, 17796–17803.
- (732) Qin, Y.; Ma, X.; Ruan, Z.; et al. Improvement of Thermal Stability of Charges in Poly(lactic Acid) Electret Films for Biodegradable Electromechanical Sensors. *ACS Appl. Mater. Interfaces* **2024**, *16*, 62680–62692.
- (733) Jafarizadeh, B.; Chowdhury, A. H.; Islam Sozal, M. S.; et al. Wearable System Integrating Dual Piezoresistive and Photoplethys-

mography Sensors for Simultaneous Pulse Wave Monitoring. *ACS Appl. Mater. Interfaces* **2024**, *16*, 65402–65413.

(734) Park, J. J.; Hyun, W. J.; Mun, S. C.; Park, Y. T.; Park, O. O. Highly Stretchable and Wearable Graphene Strain Sensors with Controllable Sensitivity for Human Motion Monitoring. *ACS Appl. Mater. Interfaces* **2015**, *7*, 6317–6324.

(735) Deng, Z.; Hu, T.; Lei, Q.; et al. Stimuli-Responsive Conductive Nanocomposite Hydrogels with High Stretchability, Self-Healing, Adhesiveness, and 3D Printability for Human Motion Sensing. *ACS Appl. Mater. Interfaces* **2019**, *11*, 6796–6808.

(736) Sharma, S.; Chhetry, A.; Sharifuzzaman, M.; Yoon, H.; Park, J. Y. Wearable Capacitive Pressure Sensor Based on MXene Composite Nanofibrous Scaffolds for Reliable Human Physiological Signal Acquisition. *ACS Appl. Mater. Interfaces* **2020**, *12*, 22212–22224.

(737) Zhang, S.; Liu, H.; Yang, S.; et al. Ultrasensitive and Highly Compressible Piezoresistive Sensor Based on Polyurethane Sponge Coated with a Cracked Cellulose Nanofibril/Silver Nanowire Layer. *ACS Appl. Mater. Interfaces* **2019**, *11*, 10922–10932.

(738) Liu, S.; Wu, Y.; Jiang, L.; et al. Highly Stretchable, Tissue-like Ag Nanowire-Enhanced Ionogel Nanocomposites as an Ionogel-Based Wearable Sensor for Body Motion Monitoring. *ACS Appl. Mater. Interfaces* **2024**, *16*, 46538–46547.

(739) Chen, J.; Zhang, J.; Luo, Z.; et al. Superelastic, Sensitive, and Low Hysteresis Flexible Strain Sensor Based on Wave-Patterned Liquid Metal for Human Activity Monitoring. *ACS Appl. Mater. Interfaces* **2020**, *12*, 22200–22211.

(740) Zhao, Z.; Li, Q.; Dong, Y.; et al. Washable Patches with Gold Nanowires/Textiles in Wearable Sensors for Health Monitoring. *ACS Appl. Mater. Interfaces* **2022**, *14*, 18884–18900.

(741) Seyedin, S.; Razal, J. M.; Innis, P. C.; et al. Knitted Strain Sensor Textiles of Highly Conductive All-Polymeric Fibers. *ACS Appl. Mater. Interfaces* **2015**, *7*, 21150–21158.

(742) Lou, M.; Abdalla, I.; Zhu, M.; et al. Highly Wearable, Breathable, and Washable Sensing Textile for Human Motion and Pulse Monitoring. *ACS Appl. Mater. Interfaces* **2020**, *12*, 19965–19973.

(743) Min, S.; et al. Wearable blood pressure sensors for cardiovascular monitoring and machine learning algorithms for blood pressure estimation. *Nat. Rev. Cardiol.* **2025**, *22* (9), 629–648, DOI: 10.1038/s41569-025-01127-0.

(744) Lawes, C. M.; Hoorn, S.; Vander, Rodgers, A. Global burden of blood-pressure-related disease, 2001. *Lancet* **2008**, *371*, 1513–1518.

(745) McEvoy, J. W.; Chen, Y.; Rawlings, A.; et al. Diastolic Blood Pressure, Subclinical Myocardial Damage, and Cardiac Events. *J. Am. Coll. Cardiol.* **2016**, *68*, 1713–1722.

(746) de la Sierra, A. Early detection and management of the high-risk patient with elevated blood pressure. *Vasc. Health Risk Manage.* **2008**, *Volume 4*, 289–296.

(747) Domiano, K. L.; Hinck, S. M.; Savinske, D. L.; Hope, K. L. Comparison of Upper Arm and Forearm Blood Pressure. *Clin. Nurs. Res.* **2008**, *17*, 241–250.

(748) Prieto-Avalos, G.; Cruz-Ramos, N. A.; Alor-Hernández, G.; et al. Wearable Devices for Physical Monitoring of Heart: A Review. *Biosensors* **2022**, *12*, 292.

(749) Wu, S.; Moody, K.; Kollipara, A.; Zhu, Y. Highly Sensitive, Stretchable, and Robust Strain Sensor Based on Crack Propagation and Opening. *ACS Appl. Mater. Interfaces* **2023**, *15*, 1798–1807.

(750) Pang, C.; Lee, G. Y.; Kim, T. i.; et al. A flexible and highly sensitive strain-gauge sensor using reversible interlocking of nanofibres. *Nat. Mater.* **2012**, *11*, 795–801.

(751) Chen, M.; Luo, W.; Xu, Z.; et al. An ultrahigh resolution pressure sensor based on percolative metal nanoparticle arrays. *Nat. Commun.* **2019**, *10*, No. 4024.

(752) Wei, Y.; Chen, S.; Lin, Y.; Yuan, X.; Liu, L. Silver nanowires coated on cotton for flexible pressure sensors. *J. Mater. Chem. C* **2016**, *4*, 935–943.

(753) Eom, J.; Jaisutti, R.; Lee, H.; et al. Highly Sensitive Textile Strain Sensors and Wireless User-Interface Devices Using All-

Polymeric Conducting Fibers. *ACS Appl. Mater. Interfaces* **2017**, *9*, 10190–10197.

(754) Moghadam, B. H.; Hasanazadeh, M.; Simchi, A. Self-Powered Wearable Piezoelectric Sensors Based on Polymer Nanofiber–Metal–Organic Framework Nanoparticle Composites for Arterial Pulse Monitoring. *ACS Appl. Nano Mater.* **2020**, *3*, 8742–8752.

(755) Yang, L.; Wang, H.; Yuan, W.; et al. Wearable Pressure Sensors Based on MXene/Tissue Papers for Wireless Human Health Monitoring. *ACS Appl. Mater. Interfaces* **2021**, *13*, 60531–60543.

(756) Bijender; Kumar, A. Flexible and wearable capacitive pressure sensor for blood pressure monitoring. *Sens. Biosensing Res.* **2021**, *33*, No. 100434.

(757) Boutry, C. M.; Nguyen, A.; Lawal, Q. O.; et al. A Sensitive and Biodegradable Pressure Sensor Array for Cardiovascular Monitoring. *Adv. Mater.* **2015**, *27*, 6954–6961.

(758) Chen, L. Y.; Tee, B. K.; Chortos, A. L.; et al. Continuous wireless pressure monitoring and mapping with ultra-small passive sensors for health monitoring and critical care. *Nat. Commun.* **2014**, *5*, No. 5028.

(759) Li, J.; Jia, H.; Zhou, J.; et al. Thin, soft, wearable system for continuous wireless monitoring of artery blood pressure. *Nat. Commun.* **2023**, *14*, No. 5009.

(760) Cepeda, E.; Peluffo-Ordóñez, D. H.; Rosero-Montalvo, P.; et al. Heart Rate Detection using a Piezoelectric Ceramic Sensor: Preliminary results. *Bionatura* **2022**, *7*, 1–8.

(761) Chang, G.; Pan, X.; Hao, Y.; et al. PVDF/ZnO piezoelectric nanofibers designed for monitoring of internal micro-pressure. *RSC Adv.* **2024**, *14*, 11775–11783.

(762) Mandal, A.; Morali, A.; Skorobogatiy, M.; Bodkhe, S. 3D Printing of Polyvinylidene Fluoride-Based Piezoelectric Sensors for Noninvasive Continuous Blood Pressure Monitoring. *Adv. Eng. Mater.* **2024**, *26* (1), No. 2301292, DOI: 10.1002/adem.202301292.

(763) Zhang, T.; Yao, C.; Xu, X.; et al. Nanopores-templated CNT/PDMS Microcolumn Substrate for the Fabrication of Wearable Triboelectric Nanogenerator Sensors to Monitor Human Pulse and Blood Pressure. *Adv. Mater. Technol.* **2025**, *10* (2), No. 2400749, DOI: 10.1002/admt.202400749.

(764) Li, R.; Nie, B.; Digiglio, P.; Pan, T. Microfluidics: A Flexible, Transparent, Pressure-Sensitive Microfluidic Film. *Adv. Funct. Mater.* **2014**, *24*, 6195–6203.

(765) Zhou, S.; Park, G.; Longardner, K.; et al. Clinical validation of a wearable ultrasound sensor of blood pressure. *Nat. Biomed. Eng.* **2025**, *9* (6), 865–881, DOI: 10.1038/s41551-024-01279-3.

(766) Rachim, V. P.; Chung, W.-Y. Multimodal Wrist Biosensor for Wearable Cuff-less Blood Pressure Monitoring System. *Sci. Rep.* **2019**, *9*, No. 7947.

(767) Kachuee, M.; Kiani, M. M.; Mohammadzade, H.; Shabany, M. Cuffless Blood Pressure Estimation Algorithms for Continuous Health-Care Monitoring. *IEEE Trans. Biomed. Eng.* **2017**, *64*, 859–869.

(768) Peng, C.; Chen, M.; Sim, H. K.; Zhu, Y.; Jiang, X. Noninvasive and Nonocclusive Blood Pressure Monitoring via a Flexible Piezo-Composite Ultrasonic Sensor. *IEEE Sens. J.* **2021**, *21*, 2642–2650.

(769) Seok, D.; Lee, S.; Kim, M.; Cho, J.; Kim, C. Motion Artifact Removal Techniques for Wearable EEG and PPG Sensor Systems. *Front. Electron.* **2021**, *2*, No. 685513, DOI: 10.3389/felec.2021.685513.

(770) Kario, K. Management of Hypertension in the Digital Era. *Hypertension* **2020**, *76*, 640–650.

(771) Seçkin, A. a.; Ateş, B.; Seçkin, M. Review on Wearable Technology in Sports: Concepts, Challenges and Opportunities. *Appl. Sci.* **2023**, *13*, 10399.

(772) Li, R. T.; Kling, S. R.; Salata, M. J.; et al. Wearable Performance Devices in Sports Medicine. *Sports Health: A Multidisciplinary Approach* **2016**, *8*, 74–78.

(773) Moutquin, J. M.; Rainville, C.; Giroux, L.; et al. A prospective study of blood pressure in pregnancy: Prediction of preeclampsia. *Am. J. Obstet. Gynecol.* **1985**, *151*, 191–196.

- (774) Zhu, Z.; Liu, T.; Li, G.; Li, T.; Inoue, Y. Wearable Sensor Systems for Infants. *Sensors* **2015**, *15*, 3721–3749.
- (775) Shaffer, F.; Ginsberg, J. P. An Overview of Heart Rate Variability Metrics and Norms. *Front. Public Health* **2017**, *5*, No. 290215, DOI: 10.3389/fpubh.2017.00258.
- (776) Alugubelli, N.; Abuissa, H.; Roka, A. Wearable Devices for Remote Monitoring of Heart Rate and Heart Rate Variability—What We Know and What Is Coming. *Sensors* **2022**, *22*, 8903.
- (777) Galli, A.; Montree, R. J. H.; Que, S.; Peri, E.; Vullings, R. An Overview of the Sensors for Heart Rate Monitoring Used in Extramural Applications. *Sensors* **2022**, *22*, 4035.
- (778) Ito, S. High-intensity interval training for health benefits and care of cardiac diseases - The key to an efficient exercise protocol. *World J. Cardiol.* **2019**, *11*, 171–188.
- (779) Gill, S. K.; Barsky, A.; Guan, X.; et al. Consumer wearable devices for evaluation of heart rate control using digoxin versus beta-blockers: the RATE-AF randomized trial. *Nat. Med.* **2024**, *30*, 2030–2036.
- (780) Etiwy, M.; Akhrass, Z.; Gillinov, L.; et al. Accuracy of wearable heart rate monitors in cardiac rehabilitation. *Cardiovasc. Diagn. Ther.* **2019**, *9*, 262–271.
- (781) Wei, X.; Yohannan, S.; Richards, J. R. Physiology, Cardiac Repolarization Dispersion and Reserve. In *StatPearls*; StatPearls Publishing: Treasure Island (FL), 2025.
- (782) Ben Othman, G.; Kumar, A. A.; Hassine, F. B.; et al. Sustainability and predictive accuracy evaluation of gel and embroidered electrodes for ECG monitoring. *Biomed. Signal Process. Control* **2024**, *96*, No. 106632.
- (783) Zhang, J.; Guo, Y.; Dong, X.; et al. Opportunities and challenges of noise interference suppression algorithms for dynamic ECG signals in wearable devices: A review. *Measurement* **2025**, *250*, No. 117067.
- (784) Joutsen, A.; Cömert, A.; Kaappa, E.; et al. ECG signal quality in intermittent long-term dry electrode recordings with controlled motion artifacts. *Sci. Rep.* **2024**, *14*, No. 8882.
- (785) Qin, Q.; Li, J.; Yao, S.; et al. Electrocardiogram of a Silver Nanowire Based Dry Electrode: Quantitative Comparison With the Standard Ag/AgCl Gel Electrode. *IEEE Access* **2019**, *7*, 20789–20800.
- (786) Myers, A. C.; Huang, H.; Zhu, Y. Wearable silver nanowire dry electrodes for electrophysiological sensing. *RSC Adv.* **2015**, *5*, 11627–11632.
- (787) Yao, S.; Yang, J.; Poblete, F. R.; Hu, X.; Zhu, Y. Multifunctional Electronic Textiles Using Silver Nanowire Composites. *ACS Appl. Mater. Interfaces* **2019**, *11*, 31028–31037.
- (788) Cui, Z.; Han, Y.; Huang, Q.; Dong, J.; Zhu, Y. Electrohydrodynamic printing of silver nanowires for flexible and stretchable electronics. *Nanoscale* **2018**, *10*, 6806–6811.
- (789) Lee, S. P.; Ha, G.; Wright, D. E.; et al. Highly flexible, wearable, and disposable cardiac biosensors for remote and ambulatory monitoring. *npj Digit. Med.* **2018**, *1*, No. 2.
- (790) Badertscher, P.; Brasier, N. Wearable technology in clinical practice: The Basel Wearable Clinic. *Nat. Rev. Bioeng.* **2025**, *3*, 1002–1003.
- (791) Fine, J.; Branen, K. L.; Rodriguez, A. J.; et al. Sources of Inaccuracy in Photoplethysmography for Continuous Cardiovascular Monitoring. *Biosensors* **2021**, *11*, 126.
- (792) Elgendi, M. On the Analysis of Fingertip Photoplethysmogram Signals. *Curr. Cardiol. Rev.* **2012**, *8*, 14–25.
- (793) Weng, W.-H.; Baur, S.; Daswani, M.; et al. Predicting cardiovascular disease risk using photoplethysmography and deep learning. *PLoS Global Public Health* **2024**, *4*, No. e0003204.
- (794) Bent, B.; Goldstein, B. A.; Kibbe, W. A.; Dunn, J. P. Investigating sources of inaccuracy in wearable optical heart rate sensors. *npj Digit. Med.* **2020**, *3*, No. 18.
- (795) Koerber, D.; Khan, S.; Shamsheri, T.; Kirubarajan, A.; Mehta, S. Accuracy of Heart Rate Measurement with Wrist-Worn Wearable Devices in Various Skin Tones: a Systematic Review. *J. Racial Ethn. Health Disparities* **2023**, *10*, 2676–2684.
- (796) Meng, K.; Xiao, X.; Wei, W.; et al. Wearable Pressure Sensors for Pulse Wave Monitoring. *Adv. Mater.* **2022**, *34* (21), No. 2109357, DOI: 10.1002/adma.202109357.
- (797) Romano, C.; Schena, E.; Formica, D.; Massaroni, C. Comparison between Chest-Worn Accelerometer and Gyroscope Performance for Heart Rate and Respiratory Rate Monitoring. *Biosensors* **2022**, *12*, 834.
- (798) Cleveland Clinic. Heart rate monitors: how they work and accuracy. <https://my.clevelandclinic.org/health/diagnostics/23429-heart-rate-monitor>.
- (799) Kang, T. W.; Lee, J.; Kwon, Y.; Lee, Y. J.; Yeo, W. Recent Progress in the Development of Flexible Wearable Electrodes for Electrocardiogram Monitoring During Exercise. *Adv. Nanobiomed. Res.* **2024**, *4* (8), No. 2300169, DOI: 10.1002/anbr.202470081.
- (800) Kwak, Y. H.; Kim, W.; Park, K. B.; Kim, K.; Seo, S. Flexible heartbeat sensor for wearable device. *Biosens. Bioelectron.* **2017**, *94*, 250–255.
- (801) Wang, S.; Zhang, Z.; Chen, Z.; Mei, D.; Wang, Y. Development of Pressure Sensor Based Wearable Pulse Detection Device for Radial Pulse Monitoring. *Micromachines* **2022**, *13*, 1699.
- (802) Wang, L.; Tian, S.; Zhu, R. A new method of continuous blood pressure monitoring using multichannel sensing signals on the wrist. *Microsyst. Nanoeng.* **2023**, *9*, No. 117.
- (803) Tian, S.; Wang, L.; Zhu, R. A flexible multimodal pulse sensor for wearable continuous blood pressure monitoring. *Mater. Horiz.* **2024**, *11*, 2428–2437.
- (804) Park, J.-H.; Jang, D.-G.; Park, J.; Youm, S.-K. Wearable Sensing of In-Ear Pressure for Heart Rate Monitoring with a Piezoelectric Sensor. *Sensors* **2015**, *15*, 23402–23417.
- (805) Boukhayma, A.; Barison, A.; Haddad, S.; Caizzone, A. Earbud-Embedded Micro-Power mm-Sized Optical Sensor for Accurate Heart Beat Monitoring. *IEEE Sens. J.* **2021**, *21*, 19967–19977.
- (806) Venema, B.; Schiefer, J.; Blazek, V.; Blánik, N.; Leonhardt, S. Evaluating Innovative In-Ear Pulse Oximetry for Unobtrusive Cardiovascular and Pulmonary Monitoring During Sleep. *IEEE J. Transl. Eng. Health Med.* **2013**, *1*, 2700208.
- (807) Bubb, C. A.; Weber, M.; Kretsch, N.; et al. Wearable in-ear pulse oximetry validly measures oxygen saturation between 70% and 100%: A prospective agreement study. *Digit. Health* **2023**, *9*, No. 20552076231211169, DOI: 10.1177/20552076231211169.
- (808) Adams, T.; Wagner, S.; Baldinger, M.; et al. Accurate detection of heart rate using in-ear photoplethysmography in a clinical setting. *Front. Digit. Health* **2022**, *4*, No. 909519, DOI: 10.3389/fdgth.2022.909519.
- (809) Flowers, A.; Pillay, D. Sports audiology: Ear hygiene practices of gym users who wear earphones. *S. Afr. J. Commun. Disord.* **2021**, *68* (1), No. a794, DOI: 10.4102/sajcd.v68i1.794.
- (810) Chen, G.; Xiao, X.; Zhao, X.; et al. Electronic Textiles for Wearable Point-of-Care Systems. *Chem. Rev.* **2022**, *122*, 3259–3291.
- (811) Yin, J.; Wang, S.; Di Carlo, A.; et al. Smart textiles for self-powered biomonitors. *Med-X* **2023**, *1*, No. 3.
- (812) Lo Presti, D.; Santucci, F.; Massaroni, C.; et al. A multi-point heart rate monitoring using a soft wearable system based on fiber optic technology. *Sci. Rep.* **2021**, *11*, No. 21162.
- (813) Lin, J.; Fu, R.; Zhong, X.; et al. Wearable sensors and devices for real-time cardiovascular disease monitoring. *Cell Rep. Phys. Sci.* **2021**, *2*, No. 100541.
- (814) Nasiri, S.; Khosravani, M. R. Progress and challenges in fabrication of wearable sensors for health monitoring. *Sens. Actuators, A* **2020**, *312*, No. 112105.
- (815) Shetti, N. P.; Mishra, A.; Basu, S.; et al. Skin-Patchable Electrodes for Biosensor Applications: A Review. *ACS Biomater. Sci. Eng.* **2020**, *6*, 1823–1835.
- (816) Sun, X.; Guo, X.; Gao, J.; et al. E-Skin and Its Advanced Applications in Ubiquitous Health Monitoring. *Biomedicines* **2024**, *12*, 2307.
- (817) Liu, Y.; Tao, J.; Yang, W.; et al. Biodegradable, Breathable Leaf Vein-Based Tactile Sensors with Tunable Sensitivity and Sensing

Range. *Small* **2022**, *18* (8), No. 2106906, DOI: 10.1002/smll.202106906.

(818) Geneva, I. I.; Cuzzo, B.; Fazili, T.; Javid, W. Normal Body Temperature: A Systematic Review. *Open Forum Infect. Dis.* **2019**, *6* (4), No. ofz032, DOI: 10.1093/ofid/ofz032.

(819) Laupland, K. B. Fever in the critically ill medical patient. *Crit. Care Med.* **2009**, *37*, S273–S278.

(820) Evans, S. S.; Repasky, E. A.; Fisher, D. T. Fever and the thermal regulation of immunity: the immune system feels the heat. *Nat. Rev. Immunol.* **2015**, *15*, 335–349.

(821) Hymczak, H.; Gołab, A.; Mendrala, K.; et al. Core Temperature Measurement—Principles of Correct Measurement, Problems, and Complications. *Int. J. Environ. Res. Public Health* **2021**, *18*, 10606.

(822) Yu, L.; Delgado, J.; De Mezerille, R. Thermal mapping: Assessing the optimal sites for temperature measurement in the human body and emerging technologies. *Physiol. Rep.* **2024**, *12* (14), No. e16155, DOI: 10.14814/phy2.16155.

(823) LIU, G.; ZHANG, Y.; TANG, L. Research progress in core body temperature measurement during target temperature management. *J. Integr. Nursing* **2022**, *4*, 36–41.

(824) Liu, L.; Dou, Y.; Wang, J.; et al. Recent Advances in Flexible Temperature Sensors: Materials, Mechanism, Fabrication, and Applications. *Adv. Sci.* **2024**, *11* (36), No. 2405003, DOI: 10.1002/advs.202405003.

(825) Vaghasiya, J. V.; Mayorga-Martinez, C. C.; Pumera, M. Wearable sensors for telehealth based on emerging materials and nanoarchitectonics. *npj Flexible Electron.* **2023**, *7*, No. 26.

(826) Sonigara, K. K.; Vaghasiya, J. V.; Mayorga-Martinez, C. C.; Pumera, M. Flexible energy storage patch based on NiPS3/graphene zinc-ion hybrid supercapacitor for integrated biosensors. *Chem. Eng. J.* **2023**, *473*, No. 145204.

(827) Park, T. H.; Park, S.; Yu, S.; et al. Highly Sensitive On-Skin Temperature Sensors Based on Biocompatible Hydrogels with Thermoresponsive Transparency and Resistivity. *Adv. Healthcare Mater.* **2021**, *10* (14), No. 2100469, DOI: 10.1002/adhm.202100469.

(828) Yu, Y.; Peng, S.; Islam, M.; Wu, S.; Wang, C. H. Wearable Supercapacitive Temperature Sensors with High Accuracy Based on Ionically Conductive Organogel and Macro-Kirigami Electrode. *Adv. Mater. Technol.* **2023**, *8* (4), No. 2201020, DOI: 10.1002/admt.202201020.

(829) Shin, J.; Jeong, B.; Kim, J.; et al. Sensitive Wearable Temperature Sensor with Seamless Monolithic Integration. *Adv. Mater.* **2020**, *32* (2), No. 1905527, DOI: 10.1002/adma.201905527.

(830) Xin, Y.; Nesser, H.; Zhou, J.; Lubineau, G. Evolution of the Seebeck effect in nanoparticle-percolated networks under applied strain. *Appl. Mater. Today* **2022**, *28*, No. 101503.

(831) Ota, H.; Chao, M.; Gao, Y.; et al. 3D Printed “Earable” Smart Devices for Real-Time Detection of Core Body Temperature. *ACS Sens.* **2017**, *2*, 990–997.

(832) Davidson, A.; Buis, A.; Glesk, I. Toward Novel Wearable Pyroelectric Temperature Sensor for Medical Applications. *IEEE Sens. J.* **2017**, *17*, 6682–6689.

(833) Fan, W.; Liu, T.; Wu, F.; et al. An Antisweat Interference and Highly Sensitive Temperature Sensor Based on Poly(3,4-ethylenedioxythiophene)–Poly(styrenesulfonate) Fiber Coated with Polyurethane/Graphene for Real-Time Monitoring of Body Temperature. *ACS Nano* **2023**, *17*, 21073–21082.

(834) Zhu, Z.; Su, Y.; Chen, J.; et al. PEDOT:PSS-Based Wearable Flexible Temperature Sensor and Integrated Sensing Matrix for Human Body Monitoring. *ACS Appl. Mater. Interfaces* **2024**, *16* (41), 56082–56094, DOI: 10.1021/acsami.4c11251.

(835) Zhao, X.; Long, Y.; Yang, T.; Li, J.; Zhu, H. Simultaneous High Sensitivity Sensing of Temperature and Humidity with Graphene Woven Fabrics. *ACS Appl. Mater. Interfaces* **2017**, *9*, 30171–30176.

(836) Lim, S.; Suk, J. W. Flexible temperature sensors based on two-dimensional materials for wearable devices. *J. Phys. D: Appl. Phys.* **2023**, *56*, No. 063001.

(837) Choi, S.-G.; Kang, S.-H.; Lee, J.-Y.; Park, J.-H.; Kang, S.-K. Recent advances in wearable iontronic sensors for healthcare applications. *Front. Bioeng. Biotechnol.* **2023**, *11*, No. 1335188, DOI: 10.3389/fbioe.2023.1335188.

(838) Mo, F.; Zhou, P.; Lin, S.; Zhong, J.; Wang, Y. A Review of Conductive Hydrogel-Based Wearable Temperature Sensors. *Adv. Healthcare Mater.* **2024**, *13* (26), No. 2401503, DOI: 10.1002/adhm.202401503.

(839) Soomro, A. M.; Jawed, B.; Qayoom, A.; et al. Textile-Based Flexible Temperature Sensors for Wearable and Sports Applications. *Phys. Status Solidi A* **2024**, *221* (2), No. 2300523, DOI: 10.1002/pssa.202300523.

(840) Seifi, M.; Hamed, S.; Kordrostami, Z. Fabrication of a high-sensitive wearable temperature sensor with an improved response time based on PEDOT:PSS/rGO on a flexible kapton substrate. *J. Mater. Sci.: Mater. Electron.* **2022**, *33*, 6954–6968.

(841) Yu, Y.; Peng, S.; Blanloeil, P.; Wu, S.; Wang, C. H. Wearable Temperature Sensors with Enhanced Sensitivity by Engineering Microcrack Morphology in PEDOT:PSS–PDMS Sensors. *ACS Appl. Mater. Interfaces* **2020**, *12*, 36578–36588.

(842) Lee, J.-W.; Choi, Y.; Jang, J.; et al. High sensitivity flexible paper temperature sensor and body-attachable patch for thermometers. *Sens. Actuators, A* **2020**, *313*, No. 112205.

(843) Hu, R.; et al. Design, Fabrication, and Wearable Medical Application of a High-Resolution Flexible Capacitive Temperature Sensor Based on the Thermotropic Phase Transition Composites of PEO/PVDF-HFP/H₃PO₄. *Adv. Funct. Mater.* **2024**, *35* (11), No. 2417205, DOI: 10.1002/adfm.202417205.

(844) Zhang, H.; Zhang, Y.; Liu, Y.; et al. Self-Assembled and Multilayer-Overlapped ESM-PDA@rGO Nanofilm-Based Flexible Wearable Sensor for Real-Time Body Temperature Monitoring. *ACS Appl. Mater. Interfaces* **2023**, *15* (49), 57626–57635, DOI: 10.1021/acsami.3c13700.

(845) Wang, W.; Jin, L.; Shi, G.; et al. Flexible wearable optical microfiber sensor for identifying bending direction and body temperature. *Opt. Laser Technol.* **2024**, *179*, No. 111281.

(846) Jha, R.; Mishra, P.; Kumar, S. Advancements in optical fiber-based wearable sensors for smart health monitoring. *Biosens. Bioelectron.* **2024**, *254*, No. 116232.

(847) Wang, P.; Li, Y.; Cao, J.; et al. A direct written strain-insensitive temperature sensor based on a multi-polygonal structure for human temperature monitoring. *Nano Energy* **2024**, *131*, No. 110199.

(848) Yu, Y.; Peng, S.; Sha, Z.; et al. High-precision, stretchable kirigami-capacitive sensor with ultra-low cross-sensitivity for body temperature monitoring. *J. Mater. Chem. A* **2021**, *9*, 24874–24886.

(849) Kim, M.; Joe, D. J.; Doh, I.; Cho, Y. Piezoelectric Nanocomposite-Based Multifunctional Wearable Bioelectronics for Mental Stress Analysis Utilizing Physiological Signals. *Adv. Mater. Technol.* **2024**, *9* (5), No. 2301610, DOI: 10.1002/admt.202301610.

(850) Guo, X.; Niu, Y.; Yin, Z.; et al. Bionic Microstructure-Inspired Dual-Mode Flexible Sensor with Photothermal Effect for Ultra-sensitive Temperature and Strain Monitoring. *Adv. Mater. Technol.* **2024**, *9* (23), No. 2400701, DOI: 10.1002/admt.202400701.

(851) Khan, S.; Kim, J.; Kang, T. U.; et al. Compact Vital-Sensing Band with Uninterrupted Power Supply for Core Body Temperature and Pulse Rate Monitoring. *ACS Sens.* **2024**, *9*, 5885–5895.

(852) Ding, X.; Xu, J.; Xu, J.; et al. Standalone Stretchable Biophysical Sensing System Based on Laser Direct Write of Patterned Porous Graphene/Co₃O₄ Nanocomposites. *ACS Sens.* **2024**, *9*, 3730–3740.

(853) Zhang, X.; Chen, J.; Zheng, Z.; et al. Flexible Temperature Sensor with High Reproducibility and Wireless Closed-Loop System for Decoupled Multimodal Health Monitoring and Personalized Thermoregulation. *Adv. Mater.* **2024**, *36* (45), No. 2407859, DOI: 10.1002/adma.202407859.

(854) Ryu, H.; Song, J. W.; Luan, H.; et al. Materials and Device Designs for Wireless Monitoring of Temperature and Thermal

Transport Properties of Wound Beds during Healing. *Adv. Healthcare Mater.* **2024**, 13 (5), No. 2302797, DOI: 10.1002/adhm.202302797.

(855) Kang, M.; Jeong, H.; Park, S. W.; et al. Wireless graphene-based thermal patch for obtaining temperature distribution and performing thermography. *Sci. Adv.* **2022**, 8 (15), No. eabm6693, DOI: 10.1126/sciadv.abm6693.

(856) Ye, C.; Wang, M.; Min, J.; et al. A wearable aptamer nanobiosensor for non-invasive female hormone monitoring. *Nat. Nanotechnol.* **2024**, 19, 330–337.

(857) Maroli, G.; Rosati, G.; Suárez-García, S.; et al. Wearable, battery-free, wireless multiplexed printed sensors for heat stroke prevention with mussel-inspired bio-adhesive membranes. *Biosens. Bioelectron.* **2024**, 260, No. 116421.

(858) Scroccarello, A.; Álvarez-Diduk, R.; Della Pelle, F.; et al. One-Step Laser Nanostructuring of Reduced Graphene Oxide Films Embedding Metal Nanoparticles for Sensing Applications. *ACS Sens.* **2023**, 8, 598–609.

(859) Maroli, G.; Rosati, G.; Suárez-García, S.; et al. Wearable, battery-free, wireless multiplexed printed sensors for heat stroke prevention with mussel-inspired bio-adhesive membranes. *Biosens. Bioelectron.* **2024**, 260, No. 116421.

(860) Zhou, S.; Gao, X.; Park, G.; et al. Transcranial volumetric imaging using a conformal ultrasound patch. *Nature* **2024**, 629 (8013), 810–818, DOI: 10.1038/s41586-024-07381-5.

(861) Wu, J.; Deng, J.; Theocharidis, G.; et al. Adhesive anti-fibrotic interfaces on diverse organs. *Nature* **2024**, 630 (8016), 360–367, DOI: 10.1038/s41586-024-07426-9.

(862) Wang, C.; Chen, X.; Wang, L.; et al. Bioadhesive ultrasound for long-term continuous imaging of diverse organs. *Science* **2022**, 377 (6605), 517–523, DOI: 10.1126/science.abo2542.

(863) Wang, C.; Fan, K.; Shirzaei Sani, E.; et al. A microfluidic wearable device for wound exudate management and analysis in human chronic wounds. *Sci. Transl. Med.* **2025**, 17 (795), No. eadt0882, DOI: 10.1126/scitranslmed.adt0882.

(864) Chen, W.; Lin, J.; Ye, Z.; et al. Customized surface adhesive and wettability properties of conformal electronic devices. *Mater. Horiz.* **2024**, 11 (24), 6289–6225, DOI: 10.1039/d4mh00753k.

(865) Wu, S. J.; Zhao, X. Bioadhesive Technology Platforms. *Chem. Rev.* **2023**, 123 (24), 14084–14118, DOI: 10.1021/acs.chemrev.3c00380.

(866) Park, J.; Kim, H. W.; Lim, S.; et al. Conformal Fixation Strategies and Bioadhesives for Soft Bioelectronics. *Adv. Funct. Mater.* **2024**, 34 (22), No. 2313728, DOI: 10.1002/adfm.202313728.

(867) Cai, L.; Burton, A.; Gonzales, D. A.; et al. Osseosurface electronics—thin, wireless, battery-free and multimodal musculoskeletal biointerfaces. *Nat. Commun.* **2021**, 12, No. 6707, DOI: 10.1038/s41467-021-27003-2.

(868) Abramson, A.; Chan, C. T.; Khan, Y.; et al. A flexible electronic strain sensor for the real-time monitoring of tumor regression. *Sci. Adv.* **2022**, 8 (37), No. eabn6550, DOI: 10.1126/sciadv.abn6550.

(869) Wu, R.; Zhu, T.; Cheng, Y.; et al. Materials, structure design, performances of multifunctional flexible devices for healthcare. *Prog. Mater. Sci.* **2025**, 153, No. 101491, DOI: 10.1016/j.pmatsci.2025.101491.

(870) Shin, J. H.; Choi, J. Y.; June, K.; Choi, H.; Kim, T. Polymeric Conductive Adhesive-Based Ultrathin Epidermal Electrodes for Long-Term Monitoring of Electrophysiological Signals. *Adv. Mater.* **2024**, 36 (23), No. 2313157, DOI: 10.1002/adma.202313157.

(871) Deng, J.; Yuk, H.; Wu, J.; et al. Electrical bioadhesive interface for bioelectronics. *Nat. Mater.* **2021**, 20 (2), 229–236, DOI: 10.1038/s41563-020-00814-2.

(872) Kim, H.; et al. Skin preparation—free, stretchable microneedle adhesive patches for reliable electrophysiological sensing and exoskeleton robot control. *Sci. Adv.* **2024**, 10 (3), No. eadk5260, DOI: 10.1126/sciadv.adk5260.

(873) Choi, Y.; Kang, K.; Son, D.; Shin, M. Molecular Rationale for the Design of Instantaneous, Strain-Tolerant Polymeric Adhesive in a

Stretchable Underwater Human-Machine Interface. *ACS Nano* **2022**, 16 (1), 1368–1380, DOI: 10.1021/acsnano.1c09393.

(874) Zhang, Z.; Yang, J.; Wang, H.; et al. A 10-micrometer-thick nanomesh-reinforced gas-permeable hydrogel skin sensor for long-term electrophysiological monitoring. *Sci. Adv.* **2024**, 10 (2), No. eadj5389, DOI: 10.1126/sciadv.adj5389.

(875) Kim, D. W.; et al. Electrostatic–Mechanical Synergistic In Situ Multiscale Tissue Adhesion for Sustainable Residue-Free Bioelectronics Interfaces. *Adv. Mater.* **2022**, 34 (5), No. 2105338, DOI: 10.1002/adma.202105338.

(876) Zhao, C.; Park, J.; Root, S. E.; Bao, Z. Skin-inspired soft bioelectronic materials, devices and systems. *Nat. Rev. Bioeng.* **2024**, 2 (8), 671–690, DOI: 10.1038/s44222-024-00194-1.

(877) Zhang, B.; Li, J.; Zhou, J.; et al. A three-dimensional liquid diode for soft, integrated permeable electronics. *Nature* **2024**, 628 (8006), 84–92, DOI: 10.1038/s41586-024-07161-1.

(878) Wang, W.; Harimurti, S.; Inoue, D.; et al. Janus Membrane-Based Wearable pH Sensor with Sweat Absorption, Gas Permeability, and Self-Adhesiveness. *ACS Appl. Mater. Interfaces* **2024**, 16 (21), 27065–27074, DOI: 10.1021/acsami.4c02189.

(879) Lin, M.; Zhang, Z.; Gao, X.; et al. A fully integrated wearable ultrasound system to monitor deep tissues in moving subjects. *Nat. Biotechnol.* **2024**, 42 (3), 448–457, DOI: 10.1038/s41587-023-01800-0.

(880) Xiang, C.; Wen, C.; Wang, Z.; et al. Multifunctional Conductive Hydrogel for Sensing Underwater Applications and Wearable Electroencephalogram Recording. *ACS Appl. Mater. Interfaces* **2025**, 17 (5), 8327–8339, DOI: 10.1021/acsami.4c19660.

(881) Zhou, W.; Cheng, F.; Cai, C.; Fu, Y. Bioinspired dry-state polylactic acid adhesives-based wearable sensor with reversible adhesive performance in harsh environments via building hierarchical liquid metal bead structure. *Compos. Sci. Technol.* **2023**, 242, No. 110207, DOI: 10.1016/j.compscitech.2023.110207.

(882) Han, F.; Chen, S.; Wang, F.; et al. High-Conductivity, Self-Healing, and Adhesive Ionic Hydrogels for Health Monitoring and Human-Machine Interactions Under Extreme Cold Conditions. *Adv. Sci.* **2025**, 12 (16), No. 2412726, DOI: 10.1002/adv.202412726.

(883) Shang, Z.; Liu, G.; Sun, Y.; et al. Mussel-Inspired Wet-Adhesive Multifunctional Organohydrogel with Extreme Environmental Tolerance for Wearable Strain Sensor. *ACS Appl. Mater. Interfaces* **2023**, 15 (37), 44342–44353, DOI: 10.1021/acsami.3c10213.

(884) Zhou, H.; et al. From Glutinous-Rice-Inspired Adhesive Organohydrogels to Flexible Electronic Devices Toward Wearable Sensing, Power Supply, and Energy Storage. *Adv. Funct. Mater.* **2021**, 32 (1), No. 2108423, DOI: 10.1002/adfm.202108423.

(885) Yu, Z.; Wu, P. Water-Resistant Ionogel Electrode with Tailorable Mechanical Properties for Aquatic Ambulatory Physiological Signal Monitoring. *Adv. Funct. Mater.* **2021**, 31 (51), No. 2107226, DOI: 10.1002/adfm.202107226.

(886) Jeon, J.; Kim, J.; Park, S.; et al. Double-Sided Pressure-Sensitive Adhesive Materials under Human-Centric Extreme Environments. *ACS Appl. Mater. Interfaces* **2024**, 16 (36), 48257–48268, DOI: 10.1021/acsami.4c09327.

(887) Wang, D.; Tian, H.; Liu, H.; et al. Bioinspired Dry Adhesives for Highly Adaptable and Stable Manipulating Irregular Objects under Vibration. *Adv. Sci.* **2023**, 10 (21), No. 2302512, DOI: 10.1002/adv.202302512.

(888) Hoerni, J. A. Planar silicon diodes and transistors. In *1960 International Electron Devices Meeting*; IEEE, 1960 DOI: 10.1109/iedm.1960.187178.

(889) Sun, Y.; Rogers, J. A. Inorganic Semiconductors for Flexible Electronics. *Adv. Mater.* **2007**, 19 (15), 1897–1916, DOI: 10.1002/adma.200602223.

(890) Forrest, S. R.; Thompson, M. E. Introduction: Organic Electronics and Optoelectronics. *Chem. Rev.* **2007**, 107 (4), 923–925, DOI: 10.1021/cr0501590.

(891) Rogers, J. A.; Bao, Z.; Baldwin, K.; et al. Paper-like electronic displays: Large-area rubber-stamped plastic sheets of electronics and

- microencapsulated electrophoretic inks. *Proc. Natl. Acad. Sci. U.S.A.* **2001**, *98* (9), 4835–4840, DOI: 10.1073/pnas.091588098.
- (892) Hou, S.; et al. Stretchable Electronics with Strain-Resistive Performance. *Small* **2023**, *20* (11), No. 2306749, DOI: 10.1002/sml.202306749.
- (893) Ko, H. C.; Stoykovich, M. P.; Song, J.; et al. A hemispherical electronic eye camera based on compressible silicon optoelectronics. *Nature* **2008**, *454* (7205), 748–753, DOI: 10.1038/nature07113.
- (894) Kim, D.-H.; Lu, N.; Ma, R.; et al. Epidermal Electronics. *Science* **2011**, *333* (6044), 838–843, DOI: 10.1126/science.1206157.
- (895) Kim, E. H.; Han, H.; Yu, S.; et al. Interactive Skin Display with Epidermal Stimuli Electrode. *Adv. Sci.* **2019**, *6* (13), No. 1802351, DOI: 10.1002/advs.201802351.
- (896) Chung, W. G.; Kim, E.; Song, H.; et al. Recent Advances in Electrophysiological Recording Platforms for Brain and Heart Organoids. *Adv. Nanobiomed. Res.* **2022**, *2* (12), No. 2200081, DOI: 10.1002/anbr.202200081.
- (897) Kim, M.; Kim, S.; Kwon, Y. W.; et al. Emerging Bio-Interfacing Wearable Devices for Signal Monitoring: Overview of the Mechanisms and Diverse Sensor Designs to Target Distinct Physiological Bio-Parameters. *Adv. Sens. Res.* **2023**, *2* (4), No. 2200049, DOI: 10.1002/adsr.202200049.
- (898) Park, J.; Choi, S.; Janardhan, A. H.; et al. Electromechanical cardioplasty using a wrapped elasto-conductive epicardial mesh. *Sci. Transl. Med.* **2016**, *8* (344), No. 344ra86, DOI: 10.1126/scitranslmed.aad8568.
- (899) Lee, B.; Cho, H.; Moon, S.; et al. Omnidirectional printing of elastic conductors for three-dimensional stretchable electronics. *Nat. Electron.* **2023**, *6* (4), 307–318, DOI: 10.1038/s41928-023-00949-5.
- (900) Lim, C.; Lee, S.; Kang, H.; et al. Highly Conductive and Stretchable Hydrogel Nanocomposite Using Whiskered Gold Nanosheets for Soft Bioelectronics. *Adv. Mater.* **2024**, *36* (39), No. 2407931, DOI: 10.1002/adma.202407931.
- (901) Lim, C.; Park, C.; Sunwoo, S. H.; et al. Facile and Scalable Synthesis of Whiskered Gold Nanosheets for Stretchable, Conductive, and Biocompatible Nanocomposites. *ACS Nano* **2022**, *16* (7), 10431–10442, DOI: 10.1021/acsnano.2c00880.
- (902) Liu, S.; Shah, D. S.; Kramer-Bottiglio, R. Highly stretchable multilayer electronic circuits using biphasic gallium-indium. *Nat. Mater.* **2021**, *20* (6), 851–858, DOI: 10.1038/s41563-021-00921-8.
- (903) Nam, S.; Cha, G. D.; Sunwoo, S.; et al. Needle-Like Multifunctional Biphasic Microfiber for Minimally Invasive Implantable Bioelectronics. *Adv. Mater.* **2024**, *36* (36), No. 2404101, DOI: 10.1002/adma.202404101.
- (904) Huang, Y.; Peng, C.; Li, Y.; Yang, Y.; Feng, W. Elastomeric polymers for conductive layers of flexible sensors: Materials, fabrication, performance, and applications. *Aggregate* **2023**, *4* (4), No. e319, DOI: 10.1002/agt.2.319.
- (905) Shin, Y.; Lee, H. S.; Jeong, H.; Kim, D.-H. Recent advances in conductive hydrogels for soft biointegrated electronics: Materials, properties, and device applications. *Wearable Electron.* **2024**, *1*, 255–280, DOI: 10.1016/j.wees.2024.10.004.
- (906) Zhao, Y.; Ohm, Y.; Liao, J.; et al. A self-healing electrically conductive organogel composite. *Nat. Electron.* **2023**, *6* (3), 206–215, DOI: 10.1038/s41928-023-00932-0.
- (907) Shin, Y.; Lee, H. S.; Kim, J. U.; et al. Functional-hydrogel-based electronic-skin patch for accelerated healing and monitoring of skin wounds. *Biomaterials* **2025**, *314*, No. 122802, DOI: 10.1016/j.biomaterials.2024.122802.
- (908) Dutta, P.; Gao, Y.; Rathi, M.; et al. High mobility single-crystalline-like silicon thin films on inexpensive flexible metal foils by plasma enhanced chemical vapor deposition. *Acta Mater.* **2018**, *147* (51), 58 DOI: 10.1016/j.actamat.2018.01.008.
- (909) Yoon, J.; Jo, S.; Chun, I. S.; et al. GaAs photovoltaics and optoelectronics using releasable multilayer epitaxial assemblies. *Nature* **2010**, *465* (7296), 329–333, DOI: 10.1038/nature09054.
- (910) Sun, Y.; Kumar, V.; Adesida, I.; Rogers, J. A. Buckled and Wavy Ribbons of GaAs for High-Performance Electronics on Elastomeric Substrates. *Adv. Mater.* **2006**, *18* (21), 2857–2862, DOI: 10.1002/adma.200600646.
- (911) Choi, W. M.; Song, J.; Khang, D. Y.; et al. Biaxially Stretchable “Wavy” Silicon Nanomembranes. *Nano Lett.* **2007**, *7* (6), 1655–1663, DOI: 10.1021/nl0706244.
- (912) de Böck, R.; van de Ven, A. A. F.; Saes, L. H.; Maubach, J. M. L.; Koren, B. Moisture-induced buckling of paper sheets, analysis and simulation. *Int. J. Solids Struct.* **2017**, *128*, 296–308, DOI: 10.1016/j.ijsolstr.2017.08.038.
- (913) Jiang, H.; Sun, Y.; Rogers, J. A.; Huang, Y. Post-buckling analysis for the precisely controlled buckling of thin film encapsulated by elastomeric substrates. *Int. J. Solids Struct.* **2008**, *45* (7-8), 2014–2023, DOI: 10.1016/j.ijsolstr.2007.11.007.
- (914) Kaltenbrunner, M.; Sekitani, T.; Reeder, J.; et al. An ultra-lightweight design for imperceptible plastic electronics. *Nature* **2013**, *499* (7459), 458–463, DOI: 10.1038/nature12314.
- (915) Kim, D.-H.; Song, J.; Choi, W. M.; et al. Materials and noncoplanar mesh designs for integrated circuits with linear elastic responses to extreme mechanical deformations. *Proc. Natl. Acad. Sci. U.S.A.* **2008**, *105* (48), 18675–18680, DOI: 10.1073/pnas.0807476105.
- (916) Ahn, B. Y.; Duoss, E. B.; Motala, M. J.; et al. Omnidirectional Printing of Flexible, Stretchable, and Spanning Silver Microelectrodes. *Science* **2009**, *323* (5921), 1590–1593, DOI: 10.1126/science.1168375.
- (917) Hu, H.; Zhu, X.; Wang, C.; et al. Stretchable ultrasonic transducer arrays for three-dimensional imaging on complex surfaces. *Sci. Adv.* **2018**, *4* (3), No. eaar3979, DOI: 10.1126/sciadv.aar3979.
- (918) Dagdeviren, C.; Shi, Y.; Joe, P.; et al. Conformal piezoelectric systems for clinical and experimental characterization of soft tissue biomechanics. *Nat. Mater.* **2015**, *14* (7), 728–736, DOI: 10.1038/nmat4289.
- (919) Xu, S.; Zhang, Y.; Cho, J.; et al. Stretchable batteries with self-similar serpentine interconnects and integrated wireless recharging systems. *Nat. Commun.* **2013**, *4* (1), No. 1543, DOI: 10.1038/ncomms2553.
- (920) Kim, D.; Liu, Z.; Kim, Y.; et al. Optimized Structural Designs for Stretchable Silicon Integrated Circuits. *Small* **2009**, *5* (24), 2841–2847, DOI: 10.1002/sml.200900853.
- (921) Xu, S.; Yan, Z.; Jang, K. I.; et al. Assembly of micro/nanomaterials into complex, three-dimensional architectures by compressive buckling. *Science* **2015**, *347* (6218), 154–159, DOI: 10.1126/science.1260960.
- (922) Yan, Z.; Han, M.; Yang, Y.; et al. Deterministic assembly of 3D mesostructures in advanced materials via compressive buckling: A short review of recent progress. *Extreme Mech. Lett.* **2017**, *11*, 96–104, DOI: 10.1016/j.eml.2016.12.006.
- (923) Zhang, Y.; Fu, H.; Su, Y.; et al. Mechanics of ultra-stretchable self-similar serpentine interconnects. *Acta Mater.* **2013**, *61* (20), 7816–7827, DOI: 10.1016/j.actamat.2013.09.020.
- (924) Lee, Y. K.; Jang, K.; Ma, Y.; et al. Chemical Sensing Systems that Utilize Soft Electronics on Thin Elastomeric Substrates with Open Cellular Designs. *Adv. Funct. Mater.* **2017**, *27* (9), No. 1605476, DOI: 10.1002/adfm.201605476.
- (925) Zirbel, S. A.; Lang, R. J.; Thomson, M. W.; et al. Accommodating Thickness in Origami-Based Deployable Arrays. *J. Mech. Design* **2013**, *135* (11), No. 111005, DOI: 10.1115/1.4025372.
- (926) Bertoldi, K.; Vitelli, V.; Christensen, J.; van Hecke, M. Flexible mechanical metamaterials. *Nat. Rev. Mater.* **2017**, *2* (11), 1–11, DOI: 10.1038/natrevmats.2017.66.
- (927) Coles, L.; Ventrella, D.; Carnicer-Lombarte, A.; et al. Origami-inspired soft fluidic actuation for minimally invasive large-area electrocorticography. *Nat. Commun.* **2024**, *15* (1), No. 6290, DOI: 10.1038/s41467-024-50597-2.
- (928) Shyu, T. C.; Damasceno, P. F.; Dodd, P. M.; et al. A kirigami approach to engineering elasticity in nanocomposites through patterned defects. *Nat. Mater.* **2015**, *14* (8), 785–789, DOI: 10.1038/nmat4327.

- (929) Lamoureux, A.; Lee, K.; Shlian, M.; Forrest, S. R.; Shtein, M. Dynamic kirigami structures for integrated solar tracking. *Nat. Commun.* **2015**, *6* (1), No. 8092, DOI: [10.1038/ncomms9092](https://doi.org/10.1038/ncomms9092).
- (930) Lacour, S. P.; Wagner, S.; Huang, Z.; Suo, Z. Stretchable gold conductors on elastomeric substrates. *Appl. Phys. Lett.* **2003**, *82* (15), 2404–2406, DOI: [10.1063/1.1565683](https://doi.org/10.1063/1.1565683).
- (931) Mineev, I. R.; Musienko, P.; Hirsch, A.; et al. Electronic dura mater for long-term multimodal neural interfaces. *Science* **2015**, *347* (6218), 159–163, DOI: [10.1126/science.1260318](https://doi.org/10.1126/science.1260318).
- (932) Chae, S.; Choi, W. J.; Nebel, L. J.; et al. Kinetically controlled metal-elastomer nanophases for environmentally resilient stretchable electronics. *Nat. Commun.* **2024**, *15* (1), No. 3071, DOI: [10.1038/s41467-024-47223-6](https://doi.org/10.1038/s41467-024-47223-6).
- (933) Jiang, Z.; Chen, N.; Yi, Z.; et al. A 1.3-micrometre-thick elastic conductor for seamless on-skin and implantable sensors. *Nat. Electron.* **2022**, *5* (11), 784–793, DOI: [10.1038/s41928-022-00868-x](https://doi.org/10.1038/s41928-022-00868-x).
- (934) Kang, J.; Mun, J.; Zheng, Y.; et al. Tough-interface-enabled stretchable electronics using non-stretchable polymer semiconductors and conductors. *Nat. Nanotechnol.* **2022**, *17* (12), 1265–1271, DOI: [10.1038/s41565-022-01246-6](https://doi.org/10.1038/s41565-022-01246-6).
- (935) Zheng, Y.; Wang, G. N.; Kang, J.; et al. An Intrinsically Stretchable High-Performance Polymer Semiconductor with Low Crystallinity. *Adv. Funct. Mater.* **2019**, *29* (46), No. 1905340, DOI: [10.1002/adfm.201905340](https://doi.org/10.1002/adfm.201905340).
- (936) Liu, D.; Mun, J.; Chen, G.; et al. A Design Strategy for Intrinsically Stretchable High-Performance Polymer Semiconductors: Incorporating Conjugated Rigid Fused-Rings with Bulky Side Groups. *J. Am. Chem. Soc.* **2021**, *143*, 11679–11689.
- (937) Liu, W.; et al. High-efficiency stretchable light-emitting polymers from thermally activated delayed fluorescence. *Nat. Mater.* **2023**, *22* (6), 737–745, DOI: [10.1038/s41563-023-01529-w](https://doi.org/10.1038/s41563-023-01529-w).
- (938) Mun, J.; Wang, G. N.; Oh, J. Y.; et al. Effect of Nonconjugated Spacers on Mechanical Properties of Semiconducting Polymers for Stretchable Transistors. *Adv. Funct. Mater.* **2018**, *28* (43), No. 1804222, DOI: [10.1002/adfm.201804222](https://doi.org/10.1002/adfm.201804222).
- (939) Zhang, S.; Alesadi, A.; Mason, G. T.; et al. Molecular Origin of Strain-Induced Chain Alignment in PDPP-Based Semiconducting Polymeric Thin Films. *Adv. Funct. Mater.* **2021**, *31* (21), No. 2100161, DOI: [10.1002/adfm.202100161](https://doi.org/10.1002/adfm.202100161).
- (940) Chen, W.; et al. Elastic–Plastic Fully π -Conjugated Polymer with Excellent Energy Dissipation Capacity for Ultra-Deep-Blue Flexible Polymer Light-Emitting Diodes with CIE_y = 0.04. *Adv. Mater.* **2024**, *36* (30), No. 2402708, DOI: [10.1002/adma.202402708](https://doi.org/10.1002/adma.202402708).
- (941) Yue, H.; Wang, Y.; Luo, S.; et al. In situ continuous hydrogen-bonded engineering for intrinsically stretchable and healable high-mobility polymer semiconductors. *Sci. Adv.* **2024**, *10* (40), No. eadq0171, DOI: [10.1126/sciadv.adq0171](https://doi.org/10.1126/sciadv.adq0171).
- (942) Zhuo, Z.; et al. Intrinsically stretchable fully π -conjugated polymer film via fluid conjugated molecular external-plasticizing for flexible light-emitting diodes. *Nat. Commun.* **2024**, *15* (1), No. 7990, DOI: [10.1038/s41467-024-50358-1](https://doi.org/10.1038/s41467-024-50358-1).
- (943) Kim, J.-H.; Park, J.-W. Intrinsically stretchable organic light-emitting diodes. *Sci. Adv.* **2021**, *7* (9), No. eabd9715, DOI: [10.1126/sciadv.abd9715](https://doi.org/10.1126/sciadv.abd9715).
- (944) Wang, Y.; Zhu, C.; Pfattner, R.; et al. A highly stretchable, transparent, and conductive polymer. *Sci. Adv.* **2017**, *3* (3), No. e1602076, DOI: [10.1126/sciadv.1602076](https://doi.org/10.1126/sciadv.1602076).
- (945) Lingstedt, L. V.; Ghittorelli, M.; Lu, H.; et al. Effect of DMSO Solvent Treatments on the Performance of PEDOT:PSS Based Organic Electrochemical Transistors. *Adv. Electron. Mater.* **2019**, *5* (3), No. 1800804, DOI: [10.1002/aelm.201800804](https://doi.org/10.1002/aelm.201800804).
- (946) Chen, R.; He, H.; Hong, X. Z.; et al. PEDOT:PSS as Stretchable Conductors with Good Wettability on the Substrate through the Simultaneous Plasticization and Secondary Doping with a Cationic or Anionic Surfactant. *Macromolecules* **2022**, *55* (12), 4967–4978, DOI: [10.1021/acs.macromol.2c00592](https://doi.org/10.1021/acs.macromol.2c00592).
- (947) Mun, J.; Ochiai, Y.; Wang, W.; et al. A design strategy for high mobility stretchable polymer semiconductors. *Nat. Commun.* **2021**, *12* (1), No. 3572, DOI: [10.1038/s41467-021-23798-2](https://doi.org/10.1038/s41467-021-23798-2).
- (948) Zheng, Y.; Yu, Z.; Zhang, S.; et al. A molecular design approach towards elastic and multifunctional polymer electronics. *Nat. Commun.* **2021**, *12* (1), No. 5701, DOI: [10.1038/s41467-021-25719-9](https://doi.org/10.1038/s41467-021-25719-9).
- (949) Wang, Y.; Chen, K.; Prine, N.; et al. Stretchable and Self-Healable Semiconductive Composites Based on Hydrogen Bonding Cross-linked Elastomeric Matrix. *Adv. Funct. Mater.* **2023**, *33* (42), No. 2303031, DOI: [10.1002/adfm.202303031](https://doi.org/10.1002/adfm.202303031).
- (950) Lee, M. Y.; Dharmapurikar, S.; Lee, S. J.; et al. Regular H-Bonding-Containing Polymers with Stretchability up to 100% External Strain for Self-Healable Plastic Transistors. *Chem. Mater.* **2020**, *32* (5), 1914–1924, DOI: [10.1021/acs.chemmater.9b04574](https://doi.org/10.1021/acs.chemmater.9b04574).
- (951) Yin, Y.; Mohanty, S.; Cooper, C. B.; Bao, Z.; Cai, W. Network Evolution Controlling Strain-Induced Damage and Self-Healing of Elastomers with Dynamic Bonds. *Macromolecules* **2024**, *57* (13), 6410–6418, DOI: [10.1021/acs.macromol.4c00409](https://doi.org/10.1021/acs.macromol.4c00409).
- (952) Kim, H. J.; Choi, H.; Kim, D.-H.; Son, D. Stretchable Functional Nanocomposites for Soft Implantable Bioelectronics. *Nano Lett.* **2024**, *24* (28), 8453–8464, DOI: [10.1021/acs.nanolett.4c01163](https://doi.org/10.1021/acs.nanolett.4c01163).
- (953) Choi, S.; Han, S. I.; Jung, D.; et al. Highly conductive, stretchable and biocompatible Ag–Au core–sheath nanowire composite for wearable and implantable bioelectronics. *Nat. Nanotechnol.* **2018**, *13* (11), 1048–1056, DOI: [10.1038/s41565-018-0226-8](https://doi.org/10.1038/s41565-018-0226-8).
- (954) Lee, H.; Kim, H. J.; Shin, Y.; Kim, D.-H. Phase-separated stretchable conductive nanocomposite to reduce contact resistance of skin electronics. *Sci. Rep.* **2024**, *14* (1), No. 1393, DOI: [10.1038/s41598-024-51980-1](https://doi.org/10.1038/s41598-024-51980-1).
- (955) Jung, D.; Kim, Y.; Lee, H.; et al. Metal-Like Stretchable Nanocomposite Using Locally-Bundled Nanowires for Skin-Mountable Devices (Adv. Mater. 44/2023). *Adv. Mater.* **2023**, *35* (44), No. 2370320, DOI: [10.1002/adma.202370320](https://doi.org/10.1002/adma.202370320).
- (956) Jung, D.; Lim, C.; Shim, H. J.; et al. Highly conductive and elastic nanomembrane for skin electronics. *Science* **2021**, *373* (6558), 1022–1026, DOI: [10.1126/science.abh4357](https://doi.org/10.1126/science.abh4357).
- (957) Jung, D.; Lim, C.; Park, C.; et al. Adaptive Self-Organization of Nanomaterials Enables Strain-Insensitive Resistance of Stretchable Metallic Nanocomposites. *Adv. Mater.* **2022**, *34* (23), No. 2200980, DOI: [10.1002/adma.202200980](https://doi.org/10.1002/adma.202200980).
- (958) Matsuhisa, N.; Inoue, D.; Zalar, P.; et al. Printable elastic conductors by in situ formation of silver nanoparticles from silver flakes. *Nat. Mater.* **2017**, *16* (8), 834–840, DOI: [10.1038/nmat4904](https://doi.org/10.1038/nmat4904).
- (959) Sunwoo, S.-H.; Han, S. I.; Jung, D.; et al. Stretchable Low-Impedance Conductor with Ag–Au–Pt Core–Shell–Shell Nanowires and In Situ Formed Pt Nanoparticles for Wearable and Implantable Device. *ACS Nano* **2023**, *17* (8), 7550–7561, DOI: [10.1021/acsnano.2c12659](https://doi.org/10.1021/acsnano.2c12659).
- (960) Sunwoo, S.-H.; Cha, M. J.; Han, S. I.; et al. Ventricular tachyarrhythmia treatment and prevention by subthreshold stimulation with stretchable epicardial multichannel electrode array. *Sci. Adv.* **2023**, *9* (13), No. eadf6856, DOI: [10.1126/sciadv.adf6856](https://doi.org/10.1126/sciadv.adf6856).
- (961) Yun, G.; Tang, S. Y.; Zhao, Q.; et al. Liquid Metal Composites with Anisotropic and Unconventional Piezoconductivity. *Matter* **2020**, *3* (3), 824–841, DOI: [10.1016/j.matt.2020.05.022](https://doi.org/10.1016/j.matt.2020.05.022).
- (962) Yun, G.; Tang, S. Y.; Sun, S.; et al. Liquid metal-filled magnetorheological elastomer with positive piezoconductivity. *Nat. Commun.* **2019**, *10* (1), No. 1300, DOI: [10.1038/s41467-019-09325-4](https://doi.org/10.1038/s41467-019-09325-4).
- (963) Li, J.; Liu, Y.; Yuan, L.; et al. A tissue-like neurotransmitter sensor for the brain and gut. *Nature* **2022**, *606* (7912), 94–101, DOI: [10.1038/s41586-022-04615-2](https://doi.org/10.1038/s41586-022-04615-2).
- (964) Shin, Y.; Lee, H. S.; Hong, Y. J.; et al. Low-impedance tissue-device interface using homogeneously conductive hydrogels chemically bonded to stretchable bioelectronics. *Sci. Adv.* **2024**, *10* (12), No. eadi7724, DOI: [10.1126/sciadv.adi7724](https://doi.org/10.1126/sciadv.adi7724).
- (965) Zhou, T.; Yuk, H.; Hu, F.; et al. 3D printable high-performance conducting polymer hydrogel for all-hydrogel bioelectronic interfaces. *Nat. Mater.* **2023**, *22* (7), 895–902, DOI: [10.1038/s41563-023-01569-2](https://doi.org/10.1038/s41563-023-01569-2).

- (966) Arwani, R. T.; Tan, S. C. L.; Sundarapandi, A.; et al. Stretchable ionic–electronic bilayer hydrogel electronics enable in situ detection of solid-state epidermal biomarkers. *Nat. Mater.* **2024**, *23* (8), 1115–1122, DOI: 10.1038/s41563-024-01918-9.
- (967) Hur, J.; Im, K.; Kim, S. W.; et al. Polypyrrole/Agarose-Based Electronically Conductive and Reversibly Restorable Hydrogel. *ACS Nano* **2014**, *8* (10), 10066–10076, DOI: 10.1021/nn502704g.
- (968) Chae, S. H.; Lee, Y. H. Carbon nanotubes and graphene towards soft electronics. *Nano Converg.* **2014**, *1* (1), No. 15, DOI: 10.1186/s40580-014-0015-5.
- (969) Lee, D.; Kim, S. G.; Kim, J.; et al. Highly conductive and mechanically strong metal-free carbon nanotube composite fibers with self-doped polyaniline. *Carbon* **2023**, *213*, No. 118308, DOI: 10.1016/j.carbon.2023.118308.
- (970) Lei, T.; et al. Dispersion of High-Purity Semiconducting Arc-Discharged Carbon Nanotubes Using Backbone Engineered Diketopyrrolopyrrole (DPP)-Based Polymers. *Adv. Electron. Mater.* **2015**, *2* (1), No. 1500299, DOI: 10.1002/aelm.201500299.
- (971) Lei, T.; Shao, L. L.; Zheng, Y. Q.; et al. Low-voltage high-performance flexible digital and analog circuits based on ultrahigh-purity semiconducting carbon nanotubes. *Nat. Commun.* **2019**, *10* (1), No. 2161, DOI: 10.1038/s41467-019-10145-9.
- (972) Ismail, R. A.; Mohsin, M. H.; Ali, A. K.; Hassoon, K. I.; Erten-Ela, S. Preparation and characterization of carbon nanotubes by pulsed laser ablation in water for optoelectronic application. *Phys. E: Low. Dimens. Syst. Nanostruct.* **2020**, *119*, No. 113997, DOI: 10.1016/j.physe.2020.113997.
- (973) Lee, S.; Shin, S.; Lee, S.; et al. Ag Nanowire Reinforced Highly Stretchable Conductive Fibers for Wearable Electronics. *Adv. Funct. Mater.* **2015**, *25* (21), 3114–3121, DOI: 10.1002/adfm.201500628.
- (974) Sim, K.; Rao, Z.; Kim, H. J.; et al. Fully rubbery integrated electronics from high effective mobility intrinsically stretchable semiconductors. *Sci. Adv.* **2019**, *5* (2), No. eaav5749, DOI: 10.1126/sciadv.aav5749.
- (975) Ho, M. D.; Ling, Y.; Yap, L. W.; et al. Percolating Network of Ultrathin Gold Nanowires and Silver Nanowires toward “Invisible” Wearable Sensors for Detecting Emotional Expression and Apexcardiogram. *Adv. Funct. Mater.* **2017**, *27* (25), No. 1700845, DOI: 10.1002/adfm.201700845.
- (976) Gerlein, L. F.; Benavides-Guerrero, J. A.; Cloutier, S. G. High-performance silver nanowires transparent conductive electrodes fabricated using manufacturing-ready high-speed photonic sinterization solutions. *Sci. Rep.* **2021**, *11* (1), No. 24156, DOI: 10.1038/s41598-021-03528-w.
- (977) Guan, Y.-S.; Thukral, A.; Zhang, S.; et al. Air/water interfacial assembled rubbery semiconducting nanofilm for fully rubbery integrated electronics. *Sci. Adv.* **2020**, *6* (38), No. eabb3656, DOI: 10.1126/sciadv.abb3656.
- (978) Kim, H.-J.; Sim, K.; Thukral, A.; Yu, C. Rubbery electronics and sensors from intrinsically stretchable elastomeric composites of semiconductors and conductors. *Sci. Adv.* **2017**, *3* (9), No. e1701114, DOI: 10.1126/sciadv.1701114.
- (979) Xue, Z.; Song, H.; Rogers, J. A.; Zhang, Y.; Huang, Y. Mechanically-Guided Structural Designs in Stretchable Inorganic Electronics. *Adv. Mater.* (2019) DOI: 10.1002/adma.201902254.
- (980) Zhu, S.; et al. Ultrastretchable Fibers with Metallic Conductivity Using a Liquid Metal Alloy Core. *Adv. Funct. Mater.* **2012**, *23* (18), 2308–2314, DOI: 10.1002/adfm.201202405.
- (981) Li, G.; Wu, X.; Lee, D.-W. A galinstan-based inkjet printing system for highly stretchable electronics with self-healing capability. *Lab Chip* **2016**, *16* (8), 1366–1373, DOI: 10.1039/c6lc00046k.
- (982) Li, J.; Cao, J.; Lu, B.; Gu, G. 3D-printed PEDOT:PSS for soft robotics. *Nat. Rev. Mater.* **2023**, *8* (9), 604–622, DOI: 10.1038/s41578-023-00587-5.
- (983) Liu, Y.; Liu, J.; Chen, S.; et al. Soft and elastic hydrogel-based microelectronics for localized low-voltage neuromodulation. *Nat. Biomed. Eng.* **2019**, *3* (1), 58 DOI: 10.1038/s41551-018-0335-6.
- (984) Jiugao, Y.; Ning, W.; Xiaofei, M. The Effects of Citric Acid on the Properties of Thermoplastic Starch Plasticized by Glycerol. *Starch - Stärke* **2005**, *57* (10), 494–504, DOI: 10.1002/star.200500423.
- (985) Wang, Y.; et al. Three-Dimensional Stretchable Microelectronics by Projection Microstereolithography (PμSL). *ACS Appl. Mater. Interfaces* **2021**, *13* (7), 8901–8908, DOI: 10.1021/acsami.0c20162.
- (986) Jang, K.-I.; Li, K.; Chung, H. U.; et al. Self-assembled three dimensional network designs for soft electronics. *Nat. Commun.* **2017**, *8*, No. 15894, DOI: 10.1038/ncomms15894.
- (987) Fan, J. A.; Yeo, W. H.; Su, Y.; et al. Fractal design concepts for stretchable electronics. *Nat. Commun.* **2014**, *5* (1), No. 3266, DOI: 10.1038/ncomms4266.
- (988) Kim, J.; Lee, M.; Shim, H. J.; et al. Stretchable silicon nanoribbon electronics for skin prosthesis. *Nat. Commun.* **2014**, *5* (1), No. 5747, DOI: 10.1038/ncomms6747.
- (989) Jang, H.; Sel, K.; Kim, E.; et al. Graphene e-tattoos for unobstructive ambulatory electrodermal activity sensing on the palm enabled by heterogeneous serpentine ribbons. *Nat. Commun.* **2022**, *13* (1), No. 6604, DOI: 10.1038/s41467-022-34406-2.
- (990) Cao, J.; Liu, X.; Qiu, J.; et al. Anti-friction gold-based stretchable electronics enabled by interfacial diffusion-induced cohesion. *Nat. Commun.* **2024**, *15* (1), No. 1116, DOI: 10.1038/s41467-024-45393-x.
- (991) Han, W. B.; Ko, G. J.; Lee, K. G.; et al. Ultra-stretchable and biodegradable elastomers for soft, transient electronics. *Nat. Commun.* **2023**, *14* (1), No. 2263, DOI: 10.1038/s41467-023-38040-4.
- (992) Tutika, R.; Haque, A. B. M. T.; Bartlett, M. D. Self-healing liquid metal composite for reconfigurable and recyclable soft electronics. *Commun. Mater.* **2021**, *2* (1), No. 64, DOI: 10.1038/s43246-021-00169-4.
- (993) Ai, L.; Lin, W.; Cao, C.; et al. Tough soldering for stretchable electronics by small-molecule modulated interfacial assemblies. *Nat. Commun.* **2023**, *14* (1), No. 7723, DOI: 10.1038/s41467-023-43574-8.
- (994) Lee, W.; Kim, H.; Kang, I.; et al. Universal assembly of liquid metal particles in polymers enables elastic printed circuit board. *Science* **2022**, *378* (6620), 637–641, DOI: 10.1126/science.abo6631.
- (995) Kim, N.; Lienemann, S.; Petsagkourakis, I.; et al. Elastic conducting polymer composites in thermoelectric modules. *Nat. Commun.* **2020**, *11* (1), No. 1424, DOI: 10.1038/s41467-020-15135-w.
- (996) Yang, Y.; Zhao, G.; Cheng, X.; Deng, H.; Fu, Q. Stretchable and Healable Conductive Elastomer Based on PEDOT:PSS/Natural Rubber for Self-Powered Temperature and Strain Sensing. *ACS Appl. Mater. Interfaces* **2021**, *13* (12), 14599–14611, DOI: 10.1021/acsami.1c00879.
- (997) Kim, H. J.; Jung, D.; Sunwoo, S. H.; et al. Integration of Conductive Nanocomposites and Nanomembranes for High-Performance Stretchable Conductors. *Adv. Nanobiomed. Res.* **2023**, *3* (5), No. 2200153, DOI: 10.1002/anbr.202370051.
- (998) Jin, S.; Choi, H.; Seong, D.; et al. Injectable tissue prosthesis for instantaneous closed-loop rehabilitation. *Nature* **2023**, *623* (7985), 58–65, DOI: 10.1038/s41586-023-06628-x.
- (999) Ohm, Y.; et al. An electrically conductive silver–polyacrylamide–alginate hydrogel composite for soft electronics. *Nat. Electron.* **2021**, *4* (3), 185–192, DOI: 10.1038/s41928-021-00545-5.
- (1000) Park, J.; Kim, J. Y.; Heo, J. H.; et al. Intrinsically Nonswellable Multifunctional Hydrogel with Dynamic Nanoconfinement Networks for Robust Tissue-Adaptable Bioelectronics. *Adv. Sci.* **2023**, *10* (12), No. 2207237, DOI: 10.1002/advs.202207237.
- (1001) Wei, H.; Kong, D.; Li, T.; et al. Solution-Processable Conductive Composite Hydrogels with Multiple Synergetic Networks toward Wearable Pressure/Strain Sensors. *ACS Sens.* **2021**, *6*, 2938–2951.
- (1002) Feig, V. R.; Tran, H.; Lee, M.; Bao, Z. Mechanically tunable conductive interpenetrating network hydrogels that mimic the elastic

- moduli of biological tissue. *Nat. Commun.* **2018**, *9* (1), No. 2740, DOI: 10.1038/s41467-018-05222-4.
- (1003) He, H.; Li, H.; Pu, A.; et al. Hybrid assembly of polymeric nanofiber network for robust and electronically conductive hydrogels. *Nat. Commun.* **2023**, *14* (1), No. 759, DOI: 10.1038/s41467-023-36438-8.
- (1004) Wei, H.; Lei, M.; Zhang, P.; et al. Orthogonal photochemistry-assisted printing of 3D tough and stretchable conductive hydrogels. *Nat. Commun.* **2021**, *12* (1), No. 2082, DOI: 10.1038/s41467-021-21869-y.
- (1005) Li, G.; Huang, K.; Deng, J.; et al. Highly Conducting and Stretchable Double-Network Hydrogel for Soft Bioelectronics. *Adv. Mater.* **2022**, *34* (15), No. 2200261, DOI: 10.1002/adma.202200261.
- (1006) Zhao, Y.; Zhang, B.; Yao, B.; et al. Hierarchically Structured Stretchable Conductive Hydrogels for High-Performance Wearable Strain Sensors and Supercapacitors. *Matter* **2020**, *3* (4), 1196–1210, DOI: 10.1016/j.matt.2020.08.024.
- (1007) Song, P.; Qin, H.; Gao, H.-L.; Cong, H.-P.; Yu, S.-H. Self-healing and superstretchable conductors from hierarchical nanowire assemblies. *Nat. Commun.* **2018**, *9* (1), No. 2786, DOI: 10.1038/s41467-018-05238-w.
- (1008) Ohm, Y.; Liao, J.; Luo, Y.; Ford, M. J.; Majidi, C. Reconfigurable Electrical Networks within a Conductive Hydrogel Composite. *Adv. Mater.* **2023**, *35* (14), No. 2209408, DOI: 10.1002/adma.202209408.
- (1009) Zhou, Q.; Lyu, J.; Wang, G.; et al. Mechanically Strong and Multifunctional Hybrid Hydrogels with Ultrahigh Electrical Conductivity. *Adv. Funct. Mater.* **2021**, *31* (40), No. 2104536, DOI: 10.1002/adfm.202104536.
- (1010) Chong, J.; Sung, C.; Nam, K. S.; et al. Highly conductive tissue-like hydrogel interface through template-directed assembly. *Nat. Commun.* **2023**, *14* (1), No. 2206, DOI: 10.1038/s41467-023-37948-1.
- (1011) Sun, Y.; Cui, J.; Feng, S.; et al. Projection Stereolithography 3D Printing High-Conductive Hydrogel for Flexible Passive Wireless Sensing. *Adv. Mater.* **2024**, *36* (25), No. 2400103, DOI: 10.1002/adma.202400103.
- (1012) Jiang, Y.; Zhang, Z.; Wang, Y. X.; et al. Topological supramolecular network enabled high-conductivity, stretchable organic bioelectronics. *Science* **2022**, *375*, 1411–1417.
- (1013) Lu, B.; Yuk, H.; Lin, S.; et al. Pure PEDOT:PSS hydrogels. *Nat. Commun.* **2019**, *10* (1), No. 1043, DOI: 10.1038/s41467-019-09003-5.
- (1014) Yuk, H.; Lu, B.; Lin, S.; et al. 3D printing of conducting polymers. *Nat. Commun.* **2020**, *11* (1), No. 1604, DOI: 10.1038/s41467-020-15316-7.
- (1015) Wang, J.; Li, Q.; Li, K.; et al. Ultra-High Electrical Conductivity in Filler-Free Polymeric Hydrogels Toward Thermoelectrics and Electromagnetic Interference Shielding. *Adv. Mater.* **2022**, *34* (12), No. 2109904, DOI: 10.1002/adma.202109904.
- (1016) Wang, H.; Zhuang, T.; Wang, J.; et al. Multifunctional Filler-Free PEDOT:PSS Hydrogels with Ultrahigh Electrical Conductivity Induced by Lewis-Acid-Promoted Ion Exchange. *Adv. Mater.* **2023**, *35* (33), No. 2302919, DOI: 10.1002/adma.202302919.
- (1017) Won, D.; Kim, J.; Choi, J.; et al. Digital selective transformation and patterning of highly conductive hydrogel bioelectronics by laser-induced phase separation. *Sci. Adv.* **2022**, *8* (23), No. eabo3209, DOI: 10.1126/sciadv.abo3209.
- (1018) Zhang, T.; Wu, J.; Ran, F. Poly(3, 4-Ethylenedioxythiophene) as Promising Energy Storage Materials in Zinc-Ion Batteries. *Macromol. Rapid Commun.* **2024**, *45* (23), No. 2400476, DOI: 10.1002/marc.202400476.
- (1019) Dong, J.; Zhu, Y.; Liu, Z.; Wang, M. Liquid Metal-Based Devices: Material Properties, Fabrication and Functionalities. *Nanomaterials* **2021**, *11*, 3400.
- (1020) Wilcoxon, R.; Lower, N.; Dlouhy, D. A compliant thermal spreader with internal liquid metal cooling channels. In *2010 26th Annual IEEE Semiconductor Thermal Measurement and Management Symposium (SEMI-THERM)*; IEEE, 2010; pp 210–216 DOI: 10.1109/STHERM.2010.5444288.
- (1021) Hu, H.; Huang, H.; Li, M.; et al. A wearable cardiac ultrasound imager. *Nature* **2023**, *613*, 667–675.
- (1022) Li, G.; Zhang, M.; Liu, S.; et al. Three-dimensional flexible electronics using solidified liquid metal with regulated plasticity. *Nat. Electron.* **2023**, *6*, 154–163.
- (1023) Xu, K.; Fujita, Y.; Lu, Y.; et al. A Wearable Body Condition Sensor System with Wireless Feedback Alarm Functions. *Adv. Mater.* **2021**, *33* (18), No. 2008701, DOI: 10.1002/adma.202008701.
- (1024) Sempionatto, J. R.; Lasalde-Ramírez, J. A.; Mahato, K.; Wang, J.; Gao, W. Wearable chemical sensors for biomarker discovery in the omics era. *Nat. Rev. Chem.* **2022**, *6*, 899–915.
- (1025) Tang, L.; Yang, S.; Zhang, K.; Jiang, X. Skin Electronics from Biocompatible In Situ Welding Enabled By Intrinsically Sticky Conductors. *Adv. Sci.* **2022**, *9* (23), No. 2202043, DOI: 10.1002/advs.202202043.
- (1026) Chen, S.; Fan, S.; Qi, J.; et al. Ultrahigh Strain-Insensitive Integrated Hybrid Electronics Using Highly Stretchable Bilayer Liquid Metal Based Conductor. *Adv. Mater.* **2023**, *35* (5), No. 2208569, DOI: 10.1002/adma.202208569.
- (1027) Babatayn, W.; Buttner, U.; El-Atab, N.; Hussain, M. M. Graphene and Liquid Metal Integrated Multifunctional Wearable Platform for Monitoring Motion and Human–Machine Interfacing. *ACS Nano* **2022**, *16*, 20305–20317.
- (1028) Chen, H.; Furfaro, I.; Fernández Lavado, E.; Lacour, S. P. Liquid Metal-Based Sensor Skin Enabling Haptic Perception of Softness. *Adv. Funct. Mater.* **2024**, *34*, DOI: 10.1002/adfm.202308698.
- (1029) Gao, Y.; Ota, H.; Schaler, E. W.; et al. Wearable Microfluidic Diaphragm Pressure Sensor for Health and Tactile Touch Monitoring. *Adv. Mater.* **2017**, *29* (39), No. 1701985, DOI: 10.1002/adma.201701985.
- (1030) Yu, L.; Yeo, J. C.; Soon, R. H.; et al. Highly Stretchable, Weavable, and Washable Piezoresistive Microfiber Sensors. *ACS Appl. Mater. Interfaces* **2018**, *10*, 12773–12780.
- (1031) Qi, J.; Gao, F.; Sun, G.; Yeo, J. C.; Lim, C. T. HaptGlove—Untethered Pneumatic Glove for Multimode Haptic Feedback in Reality–Virtuality Continuum. *Adv. Sci.* **2023**, *10* (25), No. 2301044, DOI: 10.1002/advs.202301044.
- (1032) Li, Y.; Cui, Y.; Zhang, M.; et al. Ultrasensitive Pressure Sensor Sponge Using Liquid Metal Modulated Nitrogen-Doped Graphene Nanosheets. *Nano Lett.* **2022**, *22*, 2817–2825.
- (1033) Lee, S.; Byun, S.; Kim, C. Y.; et al. Beyond Human Touch Perception: An Adaptive Robotic Skin Based on Gallium Microgranules for Pressure Sensory Augmentation. *Adv. Mater.* **2022**, *34* (44), No. 2204805, DOI: 10.1002/adma.202204805.
- (1034) Li, X.; Zhu, P.; Zhang, S.; et al. A Self-Supporting, Conductor-Exposing, Stretchable, Ultrathin, and Recyclable Kirigami-Structured Liquid Metal Paper for Multifunctional E-Skin. *ACS Nano* **2022**, *16*, 5909–5919.
- (1035) Huang, Y.; Yang, F.; Liu, S.; et al. Liquid Metal-Based Epidermal Flexible Sensor for Wireless Breath Monitoring and Diagnosis Enabled by Highly Sensitive SnS₂ Nanosheets. *Research* **2021**, *2021*, No. 9847285, DOI: 10.34133/2021/9847285.
- (1036) Lee, G.; et al. A Personalized Electronic Tattoo for Healthcare Realized by On-the-Spot Assembly of an Intrinsically Conductive and Durable Liquid-Metal Composite. *Adv. Mater.* **2022**, *34* (32), No. 2204159, DOI: 10.1002/adma.202204159.
- (1037) Wang, S.; Nie, Y.; Zhu, H.; et al. Intrinsically stretchable electronics with ultrahigh deformability to monitor dynamically moving organs. *Sci. Adv.* **2022**, *8* (13), No. eabl5511, DOI: 10.1126/sciadv.abl5511.
- (1038) Jeong, Y. R.; Kim, J.; Xie, Z.; et al. A skin-attachable, stretchable integrated system based on liquid GaInSn for wireless human motion monitoring with multi-site sensing capabilities. *NPG Asia Mater.* **2017**, *9*, e443.
- (1039) Freepik. Freepik. <https://www.freepik.com/> (2026).

- (1040) Zheng, L.; Zhu, M.; Wu, B.; et al. Conductance-stable liquid metal sheath-core microfibers for stretchy smart fabrics and self-powered sensing. *Sci. Adv.* **2021**, *7* (22), No. eabg4041, DOI: 10.1126/sciadv.abg4041.
- (1041) Dong, C.; Leber, A.; Das Gupta, T.; et al. High-efficiency super-elastic liquid metal based triboelectric fibers and textiles. *Nat. Commun.* **2020**, *11*, No. 3537.
- (1042) Ren, W.; Sun, Y.; Zhao, D.; et al. High-performance wearable thermoelectric generator with self-healing, recycling, and Lego-like reconfiguring capabilities. *Sci. Adv.* **2021**, *7* (7), No. eabe0586, DOI: 10.1126/sciadv.abe0586.
- (1043) Ding, L.; Hang, C.; Yang, S.; et al. *In Situ* Deposition of Skin-Adhesive Liquid Metal Particles with Robust Wear Resistance for Epidermal Electronics. *Nano Lett.* **2022**, *22*, 4482–4490.
- (1044) Silva, C. A.; Iv, J.; Yin, L.; et al. Liquid Metal Based Island-Bridge Architectures for All Printed Stretchable Electrochemical Devices. *Adv. Funct. Mater.* **2020**, *30* (30), No. 2002041, DOI: 10.1002/adfm.202002041.
- (1045) Chen, S.; Fan, S.; Chan, H.; et al. Liquid Metal Functionalization Innovations in Wearables and Soft Robotics for Smart Healthcare Applications. *Adv. Funct. Mater.* **2024**, *34* (31), No. 2309989, DOI: 10.1002/adfm.202309989.
- (1046) Wu, Q.; Zhu, F.; Wu, Z.; et al. Suspension printing of liquid metal in yield-stress fluid for resilient 3D constructs with electromagnetic functions. *npj Flexible Electron.* **2022**, *6*, No. 50.
- (1047) He, Z.; Wang, Y.; Xiao, H.; et al. Highly stretchable, deformation-stable wireless powering antenna for wearable electronics. *Nano Energy* **2023**, *112*, No. 108461.
- (1048) Low, J.-H.; Chee, P.-S.; Lim, E.-H.; Ganesan, V. Design of a wireless smart insole using stretchable microfluidic sensor for gait monitoring. *Smart Mater. Struct.* **2020**, *29*, No. 065003.
- (1049) Alberto, J.; Leal, C.; Fernandes, C.; et al. Fully Untethered Battery-free Biomonitoring Electronic Tattoo with Wireless Energy Harvesting. *Sci. Rep.* **2020**, *10*, No. 5539.
- (1050) Ling, M. M.; Bao, Z. Thin Film Deposition, Patterning, and Printing in Organic Thin Film Transistors. *Chem. Mater.* **2004**, *16* (23), 4824–4840, DOI: 10.1021/cm0496117.
- (1051) Mei, J.; Diao, Y.; Appleton, A. L.; Fang, L.; Bao, Z. Integrated Materials Design of Organic Semiconductors for Field-Effect Transistors. *J. Am. Chem. Soc.* **2013**, *135* (18), 6724–6746, DOI: 10.1021/ja400881n.
- (1052) Jang, H.; Lee, W.; Won, S. M.; et al. Quantum Confinement Effects in Transferrable Silicon Nanomembranes and Their Applications on Unusual Substrates. *Nano Lett.* **2013**, *13* (11), 5600–5607, DOI: 10.1021/nl403251e.
- (1053) Dai, Y.; Dai, S.; Li, N.; et al. Stretchable Redox-Active Semiconducting Polymers for High-Performance Organic Electrochemical Transistors. *Adv. Mater.* **2022**, *34* (23), No. 2201178, DOI: 10.1002/adma.202201178.
- (1054) Rivnay, J.; Inal, S.; Salleo, A.; et al. Organic electrochemical transistors. *Nat. Rev. Mater.* **2018**, *3* (2), No. 17086, DOI: 10.1038/natrevmats.2017.86.
- (1055) Liu, K.; Ouyang, B.; Guo, X.; Guo, Y.; Liu, Y. Advances in flexible organic field-effect transistors and their applications for flexible electronics. *npj Flexible Electron.* **2022**, *6* (1), No. 1, DOI: 10.1038/s41528-022-00133-3.
- (1056) Liu, D.; Tian, X.; Bai, J.; et al. A wearable in-sensor computing platform based on stretchable organic electrochemical transistors. *Nat. Electron.* **2024**, *7* (12), 1176–1185, DOI: 10.1038/s41928-024-01250-9.
- (1057) Lago, N.; Cester, A. Flexible and Organic Neural Interfaces: A Review. *Appl. Sci.* **2017**, *7* (12), 1292 DOI: 10.3390/app7121292.
- (1058) Lin, P.; Yan, F. Organic Thin-Film Transistors for Chemical and Biological Sensing. *Adv. Mater.* **2011**, *24* (1), 34–51, DOI: 10.1002/adma.201103334.
- (1059) Wang, J.; Ochiai, Y.; Wu, N.; et al. Intrinsically stretchable organic photovoltaics by redistributing strain to PEDOT:PSS with enhanced stretchability and interfacial adhesion. *Nat. Commun.* **2024**, *15* (1), No. 4902, DOI: 10.1038/s41467-024-49352-4.
- (1060) Zhang, Y.; Liu, F.; Zhang, Y.; et al. Self-powered, light-controlled, bioresorbable platforms for programmed drug delivery. *Proc. Natl. Acad. Sci. U.S.A.* **2023**, *120* (11), No. e2217734120, DOI: 10.1073/pnas.2217734120.
- (1061) Kumar, R.; et al. All-Printed, Stretchable Zn-Ag₂O Rechargeable Battery via, Hyperelastic Binder for Self-Powering Wearable Electronics. *Adv. Energy Mater.* **2016**, 1602096.
- (1062) Liu, X.; Xu, F.; Li, Z.; et al. Design strategy for MXene and metal chalcogenides/oxides hybrids for supercapacitors, secondary batteries and electro/photocatalysis. *Coord. Chem. Rev.* **2022**, *464*, No. 214544, DOI: 10.1016/j.ccr.2022.214544.
- (1063) Tao, Y.; Guan, T.; Ma, Y.; Sang, G.; Liu, J. Stretchable soft batteries: From structures to materials. *Energy Storage Mater.* **2025**, *76*, No. 104085, DOI: 10.1016/j.ensm.2025.104085.
- (1064) Ali, A.; Shaukat, H.; Bibi, S.; et al. Recent progress in energy harvesting systems for wearable technology. *Energy Strategy Rev.* **2023**, *49*, No. 101124, DOI: 10.1016/j.esr.2023.101124.
- (1065) Zhang, Z.; Pang, H.; Georgiadis, A.; Cecati, C. Wireless Power Transfer—An Overview. *IEEE Trans. Ind. Electron.* **2019**, *66* (2), 1044–1058, DOI: 10.1109/tie.2018.2835378.
- (1066) Won, S. M.; Cai, L.; Gutruf, P.; Rogers, J. A. Wireless and battery-free technologies for neuroengineering. *Nat. Biomed. Eng.* **2023**, *7* (4), 405–423, DOI: 10.1038/s41551-021-00683-3.
- (1067) Falk, M.; Narváez Villarrubia, C. W.; Babanova, S.; Atanassov, P.; Shleev, S. Biofuel Cells for Biomedical Applications: Colonizing the Animal Kingdom. *ChemPhysChem* **2013**, *14* (10), 2045–2058, DOI: 10.1002/cphc.201300044.
- (1068) Bade, S. G. R.; Shan, X.; Hoang, P. T.; et al. Stretchable Light-Emitting Diodes with Organometal-Halide-Perovskite–Polymer Composite Emitters. *Adv. Mater.* **2017**, *29* (23), No. 1607053, DOI: 10.1002/adma.201607053.
- (1069) Zhang, Z.; Wang, W.; Jiang, Y.; et al. High-brightness all-polymer stretchable LED with charge-trapping dilution. *Nature* **2022**, *603* (7902), 624–630, DOI: 10.1038/s41586-022-04400-1.
- (1070) Sekitani, T.; Nakajima, H.; Maeda, H.; et al. Stretchable active-matrix organic light-emitting diode display using printable elastic conductors. *Nat. Mater.* **2009**, *8* (6), 494–499, DOI: 10.1038/nmat2459.
- (1071) Kim, J.; Salvatore, G. A.; Araki, H.; et al. Battery-free, stretchable optoelectronic systems for wireless optical characterization of the skin. *Sci. Adv.* **2016**, *2* (8), No. e1600418, DOI: 10.1126/sciadv.1600418.
- (1072) Bai, X.; et al. Advanced Stretchable Photodetectors: Strategies, Materials and Devices. *Chem. – Eur. J.* **2023**, *29* (13), No. e202203022, DOI: 10.1002/chem.202203022.
- (1073) Wang, Z.; Cheng, S.; Fukuda, K.; et al. Flexible near-infrared organic photodetectors for emergent wearable applications. *Wearable Electronics* **2024**, *1*, 53–77, DOI: 10.1016/j.wees.2024.04.001.
- (1074) Sanford, J. L.; Libsch, F. R. 4.2: TFT AMOLED Pixel Circuits and Driving Methods. *SID Symposium Digest of Technical Papers* **2003**, *34*, 10–13.
- (1075) Iwase, Y.; Tagawa, A.; Takeuchi, Y.; et al. A novel low-power gate driver architecture for large 8 K 120 Hz liquid crystal display employing IGZO technology. *J. Soc. Inf. Dispersion* **2018**, *26* (5), 304–313.
- (1076) Nathan, A.; Chaji, G. R.; Ashtiani, S. J. Driving Schemes for a-Si and LTPS AMOLED Displays. *J. Disp. Technol.* **2005**, *1*, 267–277.
- (1077) Lujan, R. A.; Street, R. A. Flexible X-Ray Detector Array Fabricated With Oxide Thin-Film Transistors. *IEEE Electron Device Lett.* **2012**, *33*, 688–690.
- (1078) van Breemen, A. J. J. M.; Simon, M.; Tousignant, O.; et al. Curved digital X-ray detectors. *npj Flexible Electron.* **2020**, *4*, No. 22.
- (1079) Takei, K.; Honda, W.; Harada, S.; Arie, T.; Akita, S. Toward Flexible and Wearable Human-Interactive Health-Monitoring Devices. *Adv. Healthcare Mater.* **2015**, *4*, 487–500.
- (1080) Zulqarnain, M.; Stanzione, S.; Rathinavel, G.; et al. A flexible ECG patch compatible with NFC RF communication. *npj Flexible Electron.* **2020**, *4*, No. 13.

- (1081) Lin, M.; Hu, H.; Zhou, S.; Xu, S. Soft wearable devices for deep-tissue sensing. *Nat. Rev. Mater.* **2022**, *7*, 850–869.
- (1082) Hadwen, B.; Broder, G. R.; Morganti, D.; et al. Programmable large area digital microfluidic array with integrated droplet sensing for bioassays. *Lab Chip* **2012**, *12*, 3305.
- (1083) Wang, D.; Liu, Z.; Li, J.; et al. Thin-film transistor arrays for biological sensing systems. *Flexible Printed Electron.* **2022**, *7*, No. 023004.
- (1084) Kurokawa, Y.; Ikeda, T.; Endo, M.; et al. UHF RFICs on Flexible and Glass Substrates for Secure RFID Systems. *IEEE J. Solid-State Circuits* **2008**, *43*, 292–299.
- (1085) Myny, K.; Smout, S.; Rockelé, M.; et al. A thin-film microprocessor with inkjet print-programmable memory. *Sci. Rep.* **2014**, *4*, No. 7398.
- (1086) Biggs, J.; Myers, J.; Kufel, J.; et al. A natively flexible 32-bit Arm microprocessor. *Nature* **2021**, *595*, 532–536.
- (1087) Han, X.; Seo, K. J.; Qiang, Y.; et al. Nanomeshed Si nanomembranes. *npj Flexible Electron.* **2019**, *3*, No. 9.
- (1088) Liu, J.; Wang, J.; Zhang, Z.; et al. Fully stretchable active-matrix organic light-emitting electrochemical cell array. *Nat. Commun.* **2020**, *11* (1), No. 3362, DOI: 10.1038/s41467-020-17084-w.
- (1089) Jang, H.; Dai, Z.; Ha, K.-H.; Ameri, S. K.; Lu, N. Stretchability of PMMA-supported CVD graphene and of its electrical contacts. *2d Mater.* **2020**, *7*, No. 014003.
- (1090) Basu, N.; Kumar, R.; Manikandan, D.; et al. Strain relaxation in monolayer MoS₂ over flexible substrate. *RSC Adv.* **2023**, *13*, 16241–16247.
- (1091) Yao, Y.; Dong, H.; Hu, W. Charge Transport in Organic and Polymeric Semiconductors for Flexible and Stretchable Devices. *Adv. Mater.* **2016**, *28*, 4513–4523.
- (1092) Uraide, A. R.; Lahiri, I.; Suresh, K. S. Graphene Properties, Synthesis and Applications: A Review. *JOM* **2023**, *75*, 614–630.
- (1093) Sze, S. M.; Ng, K. K. *Physics of Semiconductor Devices*; Wiley, 2006 DOI: 10.1002/0470068329.
- (1094) Feron, K.; Lim, R.; Sherwood, C.; et al. Organic Bioelectronics: Materials and Biocompatibility. *Int. J. Mol. Sci.* **2018**, *19*, 2382.
- (1095) Gaetani, R.; Derevyanchuk, Y.; Notargiacomo, A.; et al. Biocompatibility and Connectivity of Semiconductor Nanostructures for Cardiac Tissue Engineering Applications. *Bioengineering* **2022**, *9*, 621.
- (1096) Lemme, M. C.; Akinwande, D.; Huyghebaert, C.; Stampfer, C. 2D materials for future heterogeneous electronics. *Nat. Commun.* **2022**, *13*, No. 1392.
- (1097) Illarionov, Yu, Y. Y.; et al. Process implications on the stability and reliability of 300 mm FAB MoS₂ field-effect transistors. *npj 2D Mater. Appl.* **2024**, *8*, No. 8.
- (1098) Çeliker, H.; Dehaene, W.; Myny, K. Multi-project wafers for flexible thin-film electronics by independent foundries. *Nature* **2024**, *629*, 335–340.
- (1099) Zhang, Z.; Wang, Y.; Cai, T.; et al. Recent Progress on Flexible Silicon Nanomembranes for Advanced Electronics and Optoelectronics. *Adv. Funct. Mater.* **2025**, *35* (35), No. 2502191, DOI: 10.1002/adfm.202502191.
- (1100) Zhou, H.; Seo, J. H.; Paskiewicz, D. M.; et al. Fast flexible electronics with strained silicon nanomembranes. *Sci. Rep.* **2013**, *3*, No. 1291.
- (1101) Li, G.; Ma, Z.; You, C.; et al. Silicon nanomembrane phototransistor flipped with multifunctional sensors toward smart digital dust. *Sci. Adv.* **2020**, *6* (18), No. eaaz6511, DOI: 10.1126/sciadv.aaz6511.
- (1102) Kim, D.-H.; Ahn, J. H.; Choi, W. M.; et al. Stretchable and Foldable Silicon Integrated Circuits. *Science* **2008**, *320*, 507–511.
- (1103) Song, X.; Zhang, T.; Wu, L.; et al. Highly Stretchable High-Performance Silicon Nanowire Field Effect Transistors Integrated on Elastomer Substrates. *Adv. Sci.* **2022**, *9* (9), No. 2105623, DOI: 10.1002/advs.202105623.
- (1104) Zhang, T.; Sun, Y.; Hu, R.; Qian, W.; Yu, L. High-Performance Flexible Silicon Nanowire Field Effect Transistors on Plastic Substrates. *Adv. Electron. Mater.* **2025**, *11* (5), No. 2400615, DOI: 10.1002/aeml.202400615.
- (1105) Shim, G. W.; Hong, W.; Cha, J.; et al. TFT Channel Materials for Display Applications: From Amorphous Silicon to Transition Metal Dichalcogenides. *Adv. Mater.* **2020**, *32* (35), No. 1907166, DOI: 10.1002/adma.201907166.
- (1106) Indluru, A.; Alford, T. L. High-Temperature Stability and Enhanced Performance of a-Si:H TFT on Flexible Substrate Due to Improved Interface Quality. *IEEE Trans. Electron Devices* **2010**, *57*, 3006–3011.
- (1107) Chow, M. J.; Fomani, A. A.; Moradi, M.; et al. Effects of mechanical strain on amorphous silicon thin-film transistor electrical stability. *Appl. Phys. Lett.* **2013**, *102*, No. 233509.
- (1108) Pecora, A.; Maiolo, L.; Cuscunà, M.; et al. Low-temperature polysilicon thin film transistors on polyimide substrates for electronics on plastic. *Solid. State. Electron.* **2008**, *52*, 348–352.
- (1109) Fortunato, G.; Pecora, A.; Maiolo, L. Polysilicon thin-film transistors on polymer substrates. *Mater. Sci. Semicond. Process.* **2012**, *15*, 627–641.
- (1110) Sekitani, T.; Zschieschang, U.; Klauk, H.; Someya, T. Flexible organic transistors and circuits with extreme bending stability. *Nat. Mater.* **2010**, *9*, 1015–1022.
- (1111) Ling, H.; Liu, S.; Zheng, Z.; Yan, F. Organic Flexible Electronics. *Small Methods* **2018**, *2* (10), No. 1800070, DOI: 10.1002/smt.201800070.
- (1112) *Organic Flexible Electronics*; Elsevier, 2021 DOI: 10.1016/C2018-0-04244-9.
- (1113) Fukuda, K.; Takeda, Y.; Mizukami, M.; Kumaki, D.; Tokito, S. Fully Solution-Processed Flexible Organic Thin Film Transistor Arrays with High Mobility and Exceptional Uniformity. *Sci. Rep.* **2014**, *4*, No. 3947.
- (1114) Gelinck, G. H.; Huitema, H. E. A.; van Veenendaal, E.; et al. Flexible active-matrix displays and shift registers based on solution-processed organic transistors. *Nat. Mater.* **2004**, *3*, 106–110.
- (1115) Yi, H. T.; Payne, M. M.; Anthony, J. E.; Podzorov, V. Ultra-flexible solution-processed organic field-effect transistors. *Nat. Commun.* **2012**, *3*, No. 1259.
- (1116) Zschieschang, U.; Ante, F.; Yamamoto, T.; et al. Flexible Low-Voltage Organic Transistors and Circuits Based on a High-Mobility Organic Semiconductor with Good Air Stability. *Adv. Mater.* **2010**, *22*, 982–985.
- (1117) Oh, J.; Kim, J. H.; Lee, S. Y.; et al. Bending Performance of Flexible Organic Thin-Film Transistors With/Without Encapsulation Layer. *IEEE Trans. Device Mater. Reliab.* **2018**, *18*, 1–4.
- (1118) Guo, X.; Xu, Y.; Ogier, S.; et al. Current Status and Opportunities of Organic Thin-Film Transistor Technologies. *IEEE Trans. Electron. Devices* **2017**, *64*, 1906–1921.
- (1119) Mei, Z.; Deng, L.; Huang, B.; Tang, W.; Guo, X. Advances in Organic Thin-Film Transistor Technology for Flexible Analog Front Ends in Wearable Electronics. *IEEE J. Flexible Electron.* **2024**, *3*, 508–521.
- (1120) Nomura, K.; Takagi, A.; Kamiya, T.; et al. Amorphous Oxide Semiconductors for High-Performance Flexible Thin-Film Transistors. *Jpn. J. Appl. Phys.* **2006**, *45*, 4303.
- (1121) Nomura, K.; Ohta, H.; Takagi, A.; et al. Room-temperature fabrication of transparent flexible thin-film transistors using amorphous oxide semiconductors. *Nature* **2004**, *432*, 488–492.
- (1122) Petti, L.; Münzenrieder, N.; Vogt, C.; et al. Metal oxide semiconductor thin-film transistors for flexible electronics. *Appl. Phys. Rev.* **2016**, *3*, No. 021303.
- (1123) Park, J. W.; Kang, B. H.; Kim, H. J. A Review of Low-Temperature Solution-Processed Metal Oxide Thin-Film Transistors for Flexible Electronics. *Adv. Funct. Mater.* **2020**, *30* (20), No. 1904632, DOI: 10.1002/adfm.201904632.
- (1124) Zulqarnain, M.; Cantatore, E. Analog and Mixed Signal Circuit Design Techniques in Flexible Unipolar a-IGZO TFT Technology: Challenges and Recent Trends. *IEEE Open J. Circuits Syst.* **2021**, *2*, 743–756.

- (1125) Gao, L. Flexible Device Applications of 2D Semiconductors. *Small* **2017**, 13 (35), No. 1603994, DOI: 10.1002/sml.201603994.
- (1126) Akinwande, D.; Petrone, N.; Hone, J. Two-dimensional flexible nanoelectronics. *Nat. Commun.* **2014**, 5, No. 5678.
- (1127) Du, J.; Yu, H.; Liu, B.; et al. Strain Engineering in 2D Material-Based Flexible Optoelectronics. *Small Methods* **2021**, 5 (1), No. 2000919, DOI: 10.1002/smt.202000919.
- (1128) Cheng, R.; Jiang, S.; Chen, Y.; et al. Few-layer molybdenum disulfide transistors and circuits for high-speed flexible electronics. *Nat. Commun.* **2014**, 5, No. 5143.
- (1129) Chang, H.; Yogeesh, M. N.; Ghosh, R.; et al. Large-Area Monolayer MoS₂ for Flexible Low-Power RF Nanoelectronics in the GHz Regime. *Adv. Mater.* **2016**, 28, 1818–1823.
- (1130) Li, X.; Gao, T.; Wu, Y. Development of two-dimensional materials for electronic applications. *Sci. China Inf. Sci.* **2016**, 59, No. 061405.
- (1131) Kim, J.; Kim, J.; Ku, M.; et al. Intraocular Pressure Monitoring Following Islet Transplantation to the Anterior Chamber of the Eye. *Nano Lett.* **2020**, 20, 1517–1525.
- (1132) Kim, S.-Y.; Kim, J.; Cheong, W. H.; et al. Alcohol gas sensors capable of wireless detection using In₂O₃/Pt nanoparticles and Ag nanowires. *Sens. Actuators, B* **2018**, 259, 825–832.
- (1133) Lee, G.; et al. Implantable, Bioresorbable Radio Frequency Resonant Circuits for Magnetic Resonance Imaging. *Adv. Sci.* **2024**, 11 (27), No. 2301232, DOI: 10.1002/advs.202301232.
- (1134) Teengam, P.; Siangproh, W.; Tontisirin, S.; et al. NFC-enabling smartphone-based portable amperometric immunosensor for hepatitis B virus detection. *Sens. Actuators, B* **2021**, 326, No. 128825.
- (1135) Uzunçar, S.; Maroli, G.; Urban, M.; Merkoçi, A. Prussian-Blue Catalysis and NFC Synergy: a Battery-Free Laser-Induced Graphene-Based Platform for Urine Glucose Monitoring at Point-of-Care. *Adv. Sci.* **2025**, 12 (20), No. 2500365, DOI: 10.1002/advs.202500365.
- (1136) Yang, Q.; Rosati, G.; Abarintos, V.; et al. Wearable and fully printed microfluidic nanosensor for sweat rate, conductivity, and copper detection with healthcare applications. *Biosens. Bioelectron.* **2022**, 202, No. 114005.
- (1137) Yoo, J.-Y.; Oh, S.; Shalish, W.; et al. Wireless broadband acousto-mechanical sensing system for continuous physiological monitoring. *Nat. Med.* **2023**, 29, 3137–3148.
- (1138) Polley, C.; Jayarathna, T.; Gunawardana, U.; et al. Wearable Bluetooth Triage Healthcare Monitoring System. *Sensors* **2021**, 21, 7586.
- (1139) Martín, A.; Kim, J.; Kurniawan, J. F.; et al. Epidermal Microfluidic Electrochemical Detection System: Enhanced Sweat Sampling and Metabolite Detection. *ACS Sens.* **2017**, 2, 1860–1868.
- (1140) Qazi, R.; Parker, K. E.; Kim, C. Y.; et al. Scalable and modular wireless-network infrastructure for large-scale behavioural neuroscience. *Nat. Biomed. Eng.* **2022**, 6, 771–786.
- (1141) Affanni, A. Wireless Sensors System for Stress Detection by Means of ECG and EDA Acquisition. *Sensors* **2020**, 20, 2026.
- (1142) Gaetani, F.; Primiceri, P.; Antonio Zappatore, G.; Visconti, P. Hardware design and software development of a motion control and driving system for transradial prosthesis based on a wireless myoelectric armband. *IET Science, Meas. Technol.* **2019**, 13, 354–362.
- (1143) Fernández de Gorostiza, E.; Berzosa, J.; Mabe, J.; Cortiñas, R. A Method for Dynamically Selecting the Best Frequency Hopping Technique in Industrial Wireless Sensor Network Applications. *Sensors* **2018**, 18, 657.
- (1144) Thotahewa, K. M. S.; Redoute, J.-M.; Yuce, M. R. A Low-Power Wearable Dual-Band Wireless Body Area Network System: Development and Experimental Evaluation. *IEEE Trans. Microw. Theory Technol.* **2014**, 62, 2802–2811.
- (1145) Khan, M. A.; Quasim, M. T.; Alghamdi, N. S.; Khan, M. Y. A Secure Framework for Authentication and Encryption Using Improved ECC for IoT-Based Medical Sensor Data. *IEEE Access* **2020**, 8, 52018–52027.
- (1146) Fan, F.-R.; Tian, Z.-Q.; Lin Wang, Z. Flexible triboelectric generator. *Nano Energy* **2012**, 1, 328–334.
- (1147) Wu, C.; Wang, A. C.; Ding, W.; Guo, H.; Wang, Z. L. Triboelectric Nanogenerator: A Foundation of the Energy for the New Era. *Energy Mater.* **2019**, 9 (1), No. 1802906, DOI: 10.1002/aenm.201802906.
- (1148) Qin, Z.; Yin, Y.; Zhang, W.; Li, C.; Pan, K. Wearable and Stretchable Triboelectric Nanogenerator Based on Crumpled Nanofibrous Membranes. *ACS Appl. Mater. Interfaces* **2019**, 11, 12452–12459.
- (1149) Zhou, S.; Jia, C.; Shu, G.; et al. Recent advances in TENGs collecting acoustic energy: From low-frequency sound to ultrasound. *Nano Energy* **2024**, 129, No. 109951.
- (1150) Zhao, G.; Zhang, Y.; Shi, N.; et al. Transparent and stretchable triboelectric nanogenerator for self-powered tactile sensing. *Nano Energy* **2019**, 59, 302–310.
- (1151) Wang, S.; Ding, L.; Wang, Y.; Gong, X. Multifunctional triboelectric nanogenerator towards impact energy harvesting and safeguards. *Nano Energy* **2019**, 59, 434–442.
- (1152) Ren, Z.; Nie, J.; Shao, J.; et al. Fully Elastic and Metal-Free Tactile Sensors for Detecting both Normal and Tangential Forces Based on Triboelectric Nanogenerators. *Adv. Funct. Mater.* **2018**, 28 (31), No. 1802989, DOI: 10.1002/adfm.201802989.
- (1153) Xia, K.; Zhu, Z.; Fu, J.; et al. A triboelectric nanogenerator based on waste tea leaves and packaging bags for powering electronic office supplies and behavior monitoring. *Nano Energy* **2019**, 60, 61–71.
- (1154) Bairagi, S.; Shahid-ul-Islam; Shahadat, M.; Mulvihill, D. M.; Ali, W. Mechanical energy harvesting and self-powered electronic applications of textile-based piezoelectric nanogenerators: A systematic review. *Nano Energy* **2023**, 111, No. 108414.
- (1155) Vidal, J. V.; Slabov, V.; Kholkin, A. L.; dos Santos, M. P. S. Hybrid Triboelectric-Electromagnetic Nanogenerators for Mechanical Energy Harvesting: A Review. *Nano-Micro Lett* **2021**, 13, No. 199.
- (1156) Korkmaz, S.; Kariper, I. A. Pyroelectric nanogenerators (PyNGs) in converting thermal energy into electrical energy: Fundamentals and current status. *Nano Energy* **2021**, 84, No. 105888.
- (1157) Zhang, Y.; Phuong, P. T. T.; Roake, E.; et al. Thermal Energy Harvesting Using Pyroelectric-Electrochemical Coupling in Ferroelectric Materials. *Joule* **2020**, 4, 301–309.
- (1158) Wang, D.; Ding, J.; Ma, Y.; et al. Multi-heterojunctioned plastics with high thermoelectric figure of merit. *Nature* **2024**, 632, 528–535.
- (1159) Qin, B.; Kanatzidis, M. G.; Zhao, L.-D. The development and impact of tin selenide on thermoelectrics. *Science* **2024**, 386 (6719), No. eadp2444, DOI: 10.1126/science.adp2444.
- (1160) Slaughter, G.; Kulkarni, T. Enzymatic Glucose Biofuel Cell and its Application. *J. Biochips Tissue Chips* **2015**, 05 (1), 111 DOI: 10.4172/2153-0777.1000110.
- (1161) Garland, N. T.; Kaveti, R.; Bandodkar, A. J. Biofluid-Activated Biofuel Cells, Batteries, and Supercapacitors: A Comprehensive Review. *Adv. Mater.* **2023**, 35 (52), No. 2303197, DOI: 10.1002/adma.202303197.
- (1162) Xue, G.; Xu, Y.; Ding, T.; et al. Water-evaporation-induced electricity with nanostructured carbon materials. *Nat. Nanotechnol.* **2017**, 12, 317–321.
- (1163) Liu, X.; Gao, H.; Ward, J. E.; et al. Power generation from ambient humidity using protein nanowires. *Nature* **2020**, 578, 550–554.
- (1164) Sun, Z.; Wen, X.; Wang, L.; et al. Emerging design principles, materials, and applications for moisture-enabled electric generation. *eScience* **2022**, 2, 32–46.
- (1165) Gai, Y.; Wang, E.; Liu, M.; et al. A Self-Powered Wearable Sensor for Continuous Wireless Sweat Monitoring. *Small Methods* **2022**, 6 (10), No. 2200653, DOI: 10.1002/smt.202200653.
- (1166) Ding, S.; Saha, T.; Yin, L.; et al. A fingertip-wearable microgrid system for autonomous energy management and metabolic monitoring. *Nat. Electron.* **2024**, 7, 788–799.
- (1167) Park, D. Y.; Joe, D. J.; Kim, D. H.; et al. Self-Powered Real-Time Arterial Pulse Monitoring Using Ultrathin Epidermal Piezo-

- electric Sensors. *Adv. Mater.* **2017**, 29 (37), No. 1702308, DOI: 10.1002/adma.201702308.
- (1168) Wang, C.; Qu, X.; Zheng, Q.; et al. Stretchable, Self-Healing, and Skin-Mounted Active Sensor for Multipoint Muscle Function Assessment. *ACS Nano* **2021**, 15, 10130–10140.
- (1169) Luo, J.; Li, Y.; He, M.; et al. Rehabilitation of Total Knee Arthroplasty by Integrating Conjoint Isometric Myodynamia and Real-Time Rotation Sensing System. *Adv. Sci.* **2022**, 9 (8), No. 2105219, DOI: 10.1002/advs.202105219.
- (1170) Zou, Y.; Gai, Y.; Tan, P.; et al. Stretchable graded multichannel self-powered respiratory sensor inspired by shark gill. *Fundamental Research* **2022**, 2, 619–628.
- (1171) Ouyang, H.; Tian, J.; Sun, G.; et al. Self-Powered Pulse Sensor for Antidiastole of Cardiovascular Disease. *Adv. Mater.* **2017**, 29 (40), No. 1703456, DOI: 10.1002/adma.201703456.
- (1172) Xia, Y.; Sun, C.; Liu, W.; et al. Ultra-Broadband Flexible Thin-Film Sensor for Sound Monitoring and Ultrasonic Diagnosis. *Small* **2024**, 20 (10), No. 2305678, DOI: 10.1002/smll.202305678.
- (1173) Liu, Y.; Zhu, Y.; Liu, J.; et al. Design of Bionic Cochlear Basilar Membrane Acoustic Sensor for Frequency Selectivity Based on Film Triboelectric Nanogenerator. *Nanoscale Res. Lett.* **2018**, 13, No. 191.
- (1174) Jang, J.; Lee, J.; Jang, J. H.; Choi, H. A Triboelectric-Based Artificial Basilar Membrane to Mimic Cochlear Tonotopy. *Adv. Healthcare Mater.* **2016**, 5, 2481–2487.
- (1175) Chen, C.; Wen, Z.; Shi, J.; et al. Micro triboelectric ultrasonic device for acoustic energy transfer and signal communication. *Nat. Commun.* **2020**, 11, No. 4143.
- (1176) Sun, C.; Si, S.; Liu, J.; et al. Flexible, ultra-wideband acoustic device for ultrasound energy harvesting and passive wireless sensing. *Nano Energy* **2023**, 112, No. 108430.
- (1177) Riul, A.; de Sousa, H. C.; Malmegrim, R. R.; et al. Wine classification by taste sensors made from ultra-thin films and using neural networks. *Sens. Actuators, B* **2004**, 98, 77–82.
- (1178) Paulovich, F. V.; De Oliveira, M. C. F.; Oliveira, O. N. A Future with Ubiquitous Sensing and Intelligent Systems. *ACS Sens.* **2018**, 3, 1433–1438.
- (1179) Rodrigues, J. F.; Florea, L.; de Oliveira, M. C. F.; Diamond, D.; Oliveira, O. N. Big data and machine learning for materials science. *Discovery Mater.* **2021**, 1, No. 12.
- (1180) Oliveira, O. N.; Oliveira, M. C. F. Materials Discovery With Machine Learning and Knowledge Discovery. *Front. Chem.* **2022**, 10, No. 930369, DOI: 10.3389/fchem.2022.930369.
- (1181) Oliveira, O. N.; Oliveira, M. C. F. Sensing and Biosensing in the World of Autonomous Machines and Intelligent Systems. *Frontiers in Sensors* **2021**, 2, No. 752754, DOI: 10.3389/fsens.2021.752754.
- (1182) Barton, D. *Data Analytics: A Comprehensive Beginner's Guide - from a to Z. Simple Steps*; CreateSpace Independent Publishing Platform, 2016.
- (1183) Mitchell, T. M. *Mach. Learn.*; McGraw-Hill, 1997.
- (1184) Card, S. K.; Mackinlay, J.; Shneiderman, B. *Readings in Information Visualization: Using Vision to Think*; Elsevier Science, 1999.
- (1185) Keim, D. A. Information visualization and visual data mining. *IEEE Trans. Visual. Comput. Graph.* **2002**, 8, 1–8.
- (1186) de Oliveira, M. C. F.; Levkowitz, H. From visual data exploration to visual data mining: A survey. *IEEE Trans. Visual. Comput. Graph.* **2003**, 9, 378–394.
- (1187) Keim, D. et al. Visual Analytics: Definition, Process, and Challenges. In *Information Visualization*; Springer Berlin Heidelberg: Berlin, Heidelberg pp 154–175 DOI: 10.1007/978-3-540-70956-5_7.
- (1188) Hastie, T.; Tibshirani, R.; Friedman, J. H. *The Elements of Statistical Learning: Data Mining, Inference, and Prediction*; Springer, 2001.
- (1189) *Visual Data Mining*; Springer Berlin Heidelberg: Berlin, Heidelberg, 2008; Vol. 4404.
- (1190) Cortes, C.; Vapnik, V. Support-vector networks. *Mach. Learn.* **1995**, 20, 273–297.
- (1191) LeCun, Y.; Bengio, Y.; Hinton, G. Deep learning. *Nature* **2015**, 521, 436–444.
- (1192) Krizhevsky, A.; Sutskever, I.; Hinton, G. E. ImageNet Classification with Deep Convolutional Neural Networks. In *Adv. Neural Inf. Process. Syst.*; Pereira, F.; Burges, C. J.; Bottou, L.; Weinberger, K. Q., Eds.; Curran Associates, Inc., 2012; Vol. 25.
- (1193) Lecun, Y.; Bottou, L.; Bengio, Y.; Haffner, P. Gradient-based learning applied to document recognition. *Proc. IEEE* **1998**, 86, 2278–2324.
- (1194) Ribeiro, M. T.; Singh, S.; Guestrin, C. 'Why Should I Trust You?': Explaining the Predictions of Any Classifier. ArXiv arXiv:1602.04938.
- (1195) Molnar, C. *Interpretable Machine Learning*, 2025; Vol. 1.
- (1196) Miorandi, D.; Sicari, S.; De Pellegrini, F.; Chlamtac, I. Internet of things: Vision, applications and research challenges. *Ad Hoc Networks* **2012**, 10, 1497–1516.
- (1197) Ashton, K. That 'Internet of Things' Thing. *RFID Journal* <https://www.rfidjournal.com/expert-views/that-internet-of-things-thing/73881/> (2009).
- (1198) Mahato, K.; Saha, T.; Ding, S.; et al. Hybrid multimodal wearable sensors for comprehensive health monitoring. *Nat. Electron.* **2024**, 7, 735–750.
- (1199) Xiang, K.; Liu, M.; Chen, J.; et al. AI-Assisted Insole Sensing System for Multifunctional Plantar-Healthcare Applications. *ACS Appl. Mater. Interfaces* **2024**, 16, 32662–32678.
- (1200) Zhu, G.; Wang, X.; Zhang, G.; et al. Multifunctional Flexible Sensor with the Hybrid Staggered-Rib Conductive Network for Intelligent Recognition of Human Biomechanical and Electrophysiological Signals. *ACS Appl. Nano Mater.* **2024**, 7, 24061–24070.
- (1201) Lee, S. H.; Kim, Y. S.; Yeo, M. K.; et al. Fully portable continuous real-time auscultation with a soft wearable stethoscope designed for automated disease diagnosis. *Sci. Adv.* **2022**, 8 (21), No. eabo5867, DOI: 10.1126/sciadv.abo5867.
- (1202) Oiticica, P. R. A.; Angelim, M. K. S. C.; Soares, J. C.; et al. Using Machine Learning and Optical Microscopy Image Analysis of Immunosensors Made on Plasmonic Substrates: Application to Detect the SARS-CoV-2 Virus. *ACS Sens.* **2025**, 10, 1407–1418.
- (1203) Feng, H.; et al. Bimodal Visual Sensors Based on Mechanoluminescence and Biosensing for Artificial Intelligence-Assisted Orthodontics. *Adv. Funct. Mater.* **2024**, 35 (10), No. 2416437, DOI: 10.1002/adfm.202416437.
- (1204) He, K.; Zhang, X.; Ren, S.; Sun, J. Deep Residual Learning for Image Recognition. ArXiv arXiv:1512.03385 (2015).
- (1205) Li, Y.; Lei, X.; Guo, D.; et al. Laser-Induced Skin-like Flexible Pressure Sensor for Artificial Intelligence Speech Recognition. *ACS Appl. Mater. Interfaces* **2024**, 16, 10380–10388.
- (1206) Neto, M. P.; Soares, A. C.; Oliveira, O. N.; Paulovich, F. V. Machine Learning Used to Create a Multidimensional Calibration Space for Sensing and Biosensing Data. *Bull. Chem. Soc. Jpn.* **2021**, 94, 1553–1562.
- (1207) Kimura, N.; Aota, T.; Aso, Y.; et al. Predicting positron emission tomography brain amyloid positivity using interpretable machine learning models with wearable sensor data and lifestyle factors. *Alzheimers Res. Ther.* **2023**, 15, 212.
- (1208) Gough, D. A.; Kumosa, L. S.; Routh, T. L.; Lin, J. T.; Lucisano, J. Y. Function of an Implanted Tissue Glucose Sensor for More than 1 Year in Animals. *Sci. Transl. Med.* **2010**, 2 (42), No. 42ra53, DOI: 10.1126/scitranslmed.3001148.
- (1209) Woods, J. E. et al. Millimeter-sized battery-free epidural cortical stimulators DOI: 10.1101/2023.09.13.23295460 (2023).
- (1210) Nuffield Department of Surgical Sciences. First-in-human implant of a 'closed-loop' bioelectronic research system for investigating treatments for brain disorders. <https://www.nds.ox.ac.uk/news/bioelectronic-research-system-for-investigating-treatments-for-brain-disorders> (2022).
- (1211) Neuralink. PRIME Study Progress Update - Second Participant. <https://neuralink.com/blog/prime-study-progress-update-second-participant/> (2024).

- (1212) Neuralink. Neuralink Blog. <https://neuralink.com/blog/> (2025).
- (1213) Tesla. We, robot. https://www.tesla.com/we-robot?utm_source=chatgpt.com (2025).
- (1214) Karim, M. E. Biosensors: Ethical, Regulatory, and Legal Issues. In *Handbook of Cell Biosensors*; Springer International Publishing: Cham, 2020; pp 1–27 DOI: 10.1007/978-3-319-47405-2_23-1.
- (1215) Vilariño, N.; Fonfria, E. S.; Louzao, M. C.; Botana, L. M. Use of Biosensors as Alternatives to Current Regulatory Methods for Marine Biotoxins. *Sensors* **2009**, *9*, 9414–9443.
- (1216) Brasier, N.; Gao, W.; Wang, J.; et al. The values driving development and investigation of innovative healthcare technology: Stakeholder voices. *Cell Biomater.* **2025**, *1*, No. 100014.
- (1217) Warren, J. L.; Warren, J. S. The Case for Understanding Interdisciplinary Relationships in Health Care. *Ochsner J.* **2023**, *23*, 94–97.
- (1218) Lara, B.; Gallo-Salazar, C.; Puente, C.; et al. Interindividual variability in sweat electrolyte concentration in marathoners. *J. Int. Soc. Sports Nutr.* **2016**, *13* (1), No. 31, DOI: 10.1186/s12970-016-0141-z.
- (1219) Harshman, S. W.; Strayer, K. E.; Davidson, C. N.; et al. Rate normalization for sweat metabolomics biomarker discovery. *Talanta* **2021**, *223*, No. 121797.
- (1220) Basu, P.; Banerjee, A.; Okoro, P. D.; et al. Integration of Functional Polymers and Biosensors to Enhance Wound Healing. *Adv. Healthcare Mater.* **2024**, *13*, No. 2401461.
- (1221) Liu, Y.; Wang, H.; Zhu, Y. Recycling of Nanowire Percolation Network for Sustainable Soft Electronics. *Adv. Electron. Mater.* **2021**, *7* (9), No. 2100588, DOI: 10.1002/aelm.202100588.
- (1222) Ahmad, M.; Shukla, D.; Zhu, Y.; Velev, O. D. Biodegradable Chitosan-Based Stretchable Electronics with Recyclable Silver Nanowires. *ACS Appl. Mater. Interfaces* **2025**, *17*, 17316–17329.
- (1223) Meza, L.; Shukla, D.; Sadeghifar, H.; et al. Sustainable Soft Electronics with Biodegradable Regenerated Cellulose Films and Printed Recyclable Silver Nanowires. *Adv. Sustainable Syst.* **2025**, *9* (2), No. 2400713, DOI: 10.1002/adsu.202400713.
- (1224) Liu, Y.; Ahmad, M.; Venditti, R. A.; Velev, O. D.; Zhu, Y. Sustainable Soft Electronics Combining Recyclable Metal Nanowire Circuits and Biodegradable Gel Film Substrates. *Adv. Electron. Mater.* **2024**, *10* (6), No. 2300792, DOI: 10.1002/aelm.202300792.
- (1225) Huang, Q.; Zhu, Y. Gravure Printing of Water-based Silver Nanowire ink on Plastic Substrate for Flexible Electronics. *Sci. Rep.* **2018**, *8*, No. 15167.
- (1226) Shukla, D.; Liu, Y.; Zhu, Y. Eco-friendly screen printing of silver nanowires for flexible and stretchable electronics. *Nanoscale* **2023**, *15*, 2767–2778.
- (1227) Liu, Y.; Zheng, M.; O'Connor, B.; Dong, J.; Zhu, Y. Curvilinear soft electronics by micromolding of metal nanowires in capillaries. *Sci. Adv.* **2022**, *8* (46), No. eadd6996, DOI: 10.1126/sciadv.add6996.
- (1228) Huang, Q.; Zhu, Y. Printing Conductive Nanomaterials for Flexible and Stretchable Electronics: A Review of Materials, Processes, and Applications. *Adv. Mater. Technol.* **2019**, *4* (5), No. 1800546, DOI: 10.1002/admt.201800546.
- (1229) Huang, Q.; Zhu, Y. Patterning of Metal Nanowire Networks: Methods and Applications. *ACS Appl. Mater. Interfaces* **2021**, *13*, 60736–60762.
- (1230) Nathan, V.; Jafari, R. Particle Filtering and Sensor Fusion for Robust Heart Rate Monitoring Using Wearable Sensors. *IEEE J. Biomed. Health Inform.* **2018**, *22*, 1834–1846.
- (1231) Yao, S.; Ren, P.; Song, R.; et al. Nanomaterial-Enabled Flexible and Stretchable Sensing Systems: Processing, Integration, and Applications. *Adv. Mater.* **2020**, *32* (15), No. 1902343, DOI: 10.1002/adma.201902343.
- (1232) Yao, S.; Swetha, P.; Zhu, Y. Nanomaterial-Enabled Wearable Sensors for Healthcare. *Adv. Healthcare Mater.* **2018**, *7* (1), No. 1700889, DOI: 10.1002/adhm.201700889.
- (1233) Fioravanti, V. B. O.; Freitas, P. G.; Rodrigues, P. G.; et al. Machine learning framework for Inter-Beat Interval estimation using wearable Photoplethysmography sensors. *Biomed. Signal Process. Control* **2024**, *88*, No. 105689.
- (1234) Ishaque, S.; Khan, N.; Krishnan, S. Trends in Heart-Rate Variability Signal Analysis. *Front. Digit. Health* **2021**, *3*, No. 639444, DOI: 10.3389/fdgth.2021.639444.
- (1235) Fang, Y.; Zhao, X.; Tat, T.; et al. All-in-one conformal epidermal patch for multimodal biosensing. *Matter* **2021**, *4*, 1102–1105.
- (1236) Alsheikhy, A.; Said, Y. F.; Shawly, T.; Lahza, H. A Model to Predict Heartbeat Rate Using Deep Learning Algorithms. *Healthcare* **2023**, *11*, 330.
- (1237) Naeini, E. K.; Sarhaddi, F.; Azimi, I.; et al. A Deep Learning–based PPG Quality Assessment Approach for Heart Rate and Heart Rate Variability. *ACM Trans. Comput. Healthcare* **2023**, *4*, 1–22.
- (1238) Luo, Y.; Abidian, M. R.; Ahn, J. H.; et al. Technology Roadmap for Flexible Sensors. *ACS Nano* **2023**, *17*, 5211–5295.
- (1239) Jacob, C.; Lindeque, J.; Müller, R.; et al. A sociotechnical framework to assess patient-facing eHealth tools: results of a modified Delphi process. *npj Digit. Med.* **2023**, *6*, No. 232.
- (1240) Canali, S.; Schiaffonati, V.; Aliverti, A. Challenges and recommendations for wearable devices in digital health: Data quality, interoperability, health equity, fairness. *PLoS Digit. Health* **2022**, *1*, No. e0000104.
- (1241) Brasier, N.; Geissmann, L.; Käch, M.; et al. Device- and Analytics-Agnostic Infrastructure for Continuous Inpatient Monitoring: A Technical Note. *Digit. Biomark.* **2020**, *4*, 62–68.
- (1242) Brasier, N.; Domenghino, A.; Meier, C. A.; et al. Implementing clinical needs into the development of wearable health-monitoring technology. *Nat. Nanotechnol.* **2025**, *20*, 1350–1353.
- (1243) Brazaca, L.; Bonacin, J.; Muñoz, R.; Janegitz, B.; Carrilho, E. Electrochemical Biosensors for the Detection of Viruses: Must-Have Products or Just Science for Publication? *Braz. J. Anal. Chem.* (2024) DOI: 10.30744/brjac.2179-3425.letter.brazacaLC.
- (1244) Farivar, S.; Abouzahra, M.; Ghasemaghahi, M. Wearable device adoption among older adults: A mixed-methods study. *Int. J. Inf. Manage.* **2020**, *55*, No. 102209.
- (1245) Javdan, M.; Ghasemaghahi, M.; Abouzahra, M. Psychological barriers of using wearable devices by seniors: A mixed-methods study. *Comput. Human Behav.* **2023**, *141*, No. 107615.
- (1246) Mashamba-Thompson, T. P.; Pfavayi, L. T.; Mutapi, F. Blind spots in the implementation of point-of-care diagnostics for underserved communities. *Nat. Rev. Bioeng.* **2023**, *1*, 876–878.
- (1247) Sen-Gupta, E.; Wright, D.; Caccese, J.; et al. A Pivotal Study to Validate the Performance of a Novel Wearable Sensor and System for Biometric Monitoring in Clinical and Remote Environments. *Digit. Biomark.* **2019**, *3* (1), 1–13, DOI: 10.1159/000493642.
- (1248) Shah, K.; Wang, A.; Chen, Y.; et al. Automated loss of pulse detection on a consumer smartwatch. *Nature* **2025**, *642* (8066), 174–181, DOI: 10.1038/s41586-025-08810-9.
- (1249) Doherty, A.; Bucci, S.; Kenny, A.; et al. Passive sensing at scale to transform understanding of poor mental health. *Lancet Digit. Health* **2025**, *7*, e172–e174.